

Data Path : Z:\HPCHEM1\BNA M\DATA\BM100517\
 Data File : BM011930.D
 Acq On : 05 Oct 2017 18:20
 Operator : SJ/JU
 Sample : PB102799BL
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SBLK99

Quant Time: Oct 06 16:04:12 2017
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM-EPA-BM100317.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Oct 06 04:40:59 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.87	152	266189	20.00	ng/ul	0.00
18) Naphthalene-d8	10.67	136	1096195	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.52	164	634544	20.00	ng/ul	0.00
61) Phenanthrene-d10	17.26	188	1299647	20.00	ng/ul	0.00
75) Chrysene-d12	21.43	240	1117391	20.00	ng/ul	0.00
83) Perylene-d12	23.76	264	1206269	20.00	ng/ul	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
3) 1,4-Dioxane-d8	3.29	96	36416	6.47	ng/uL	0.00
5) Phenol-d5	7.02	99	634485	27.70	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.20	67	398149	29.37	ng/ul	0.00
9) 2-Chlorophenol-d4	7.40	132	559887	30.47	ng/ul	0.00
13) 4-Methylphenol-d8	8.57	113	552455	27.91	ng/ul	0.00
19) Nitrobenzene-d5	9.03	128	267657	34.29	ng/ul	0.00
22) 2-Nitrophenol-d4	9.76	143	308786	35.12	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.29	165	543812	30.60	ng/ul	0.00
29) 4-Chloroaniline-d4	10.81	131	690502	35.61	ng/ul	0.00
43) Dimethylphthalate-d6	13.92	166	1820482	33.10	ng/ul	0.00
46) Acenaphthylene-d8	14.20	160	2173254	33.63	ng/ul	0.00
51) 4-Nitrophenol-d4	14.70	143	210324	21.89	ng/ul	0.00
57) Fluorene-d10	15.50	176	1456944	30.02	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.63	200	239802	27.42	ng/ul	0.00
70) Anthracene-d10	17.36	188	2210188	34.51	ng/ul	0.00
76) Pyrene-d10	19.64	212	2285216	45.34	ng/ul	0.00
87) Benzo(a)pyrene-d12	23.60	264	2051404	35.35	ng/ul	0.00

Target Compounds	Ovalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA M\DATA\BM100517\
Data File : BM011930.D
Acq On : 05 Oct 2017 18:20
Operator : SJ/JU
Sample : PB102799BL
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK99

Quant Time: Oct 06 16:04:12 2017
Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM-EPA-BM100317.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Oct 06 04:40:59 2017
Response via : Initial Calibration

