

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM100520\
 Data File : BM027612.D
 Acq On : 05 Oct 2020 19:12
 Operator : JU/CG
 Sample : PB132047BL
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SBLK47

Quant Time: Oct 06 01:23:48 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-SIM-BM092320.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Oct 05 10:43:07 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.798	152	1053	0.40	ng/ul	0.00
2) Naphthalene-d8	10.579	136	3013	0.40	ng/ul	0.00
6) Acenaphthene-d10	14.423	164	2013	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.168	188	5062	0.40	ng/ul	0.00
16) Chrysene-d12	21.343	240	5317	0.40	ng/ul	0.00
20) Perylene-d12	23.600	264	5674	0.40	ng/ul	# 0.00

System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.166	152	1961	0.36	ng/ul	0.00
14) Fluoranthene-d10	19.190	212	5950	0.43	ng/ul	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM100520\
Data File : BM027612.D
Acq On : 05 Oct 2020 19:12
Operator : JU/CG
Sample : PB132047BL
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK47

Quant Time: Oct 06 01:23:48 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-SIM-BM092320.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Oct 05 10:43:07 2020
Response via : Initial Calibration

