

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM100522\
 Data File : BM036839.D
 Acq On : 05 Oct 2022 17:22
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC040

Quant Time: Oct 06 02:35:33 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM100322.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Oct 04 03:17:03 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	97	0.00
2	1,4-Dioxane	0.486	0.466	4.1	96	0.00
3	Pyridine	1.499	1.437	4.1	93	0.00
4	n-Nitrosodimethylamine	0.651	0.585	10.1	87	0.00
5 S	2-Fluorophenol	1.127	1.158	-2.8	98	0.00
6	Aniline	2.010	1.983	1.3	91	0.00
7 S	Phenol-d6	1.583	1.620	-2.3	95	0.00
8	2-Chlorophenol	1.345	1.387	-3.1	97	0.00
9	Benzaldehyde	1.014	0.960	5.3	93	0.00
10 C	Phenol	1.813	1.820	-0.4	94	0.00
11	bis(2-Chloroethyl)ether	1.481	1.415	4.5	90	0.00
12	1,3-Dichlorobenzene	1.473	1.490	-1.2	97	0.00
13 C	1,4-Dichlorobenzene	1.510	1.540	-2.0	98	0.00
14	1,2-Dichlorobenzene	1.446	1.464	-1.2	98	0.00
15	Benzyl Alcohol	1.233	1.169	5.2	87	0.00
16	2,2'-oxybis(1-Chloropropane	2.163	1.864	13.8	82	0.00
17	2-Methylphenol	1.184	1.182	0.2	94	0.00
18	Hexachloroethane	0.557	0.555	0.4	94	0.00
19 P	n-Nitroso-di-n-propylamine	1.202	1.107	7.9	85	0.00
20	3+4-Methylphenols	1.627	1.637	-0.6	93	0.00
21 I	Naphthalene-d8	1.000	1.000	0.0	91	0.00
22	Acetophenone	0.502	0.513	-2.2	90	0.00
23 S	Nitrobenzene-d5	0.374	0.391	-4.5	90	0.00
24	Nitrobenzene	0.395	0.404	-2.3	89	0.00
25	Isophorone	0.691	0.688	0.4	85	0.00
26 C	2-Nitrophenol	0.132	0.175	-32.6#	102	0.00
27	2,4-Dimethylphenol	0.256	0.273	-6.6	92	0.00
28	bis(2-Chloroethoxy)methane	0.418	0.408	2.4	86	0.00
29 C	2,4-Dichlorophenol	0.270	0.298	-10.4	94	0.00
30	1,2,4-Trichlorobenzene	0.291	0.312	-7.2	95	0.00
31	Naphthalene	1.080	1.108	-2.6	92	0.00
32	Benzoic acid	0.195	0.228	-16.9	100	-0.01
33	4-Chloroaniline	0.434	0.443	-2.1	89	0.00
34 C	Hexachlorobutadiene	0.159	0.174	-9.4	98	0.00
35	Caprolactam	0.093	0.094	-1.1	85	0.00
36 C	4-Chloro-3-methylphenol	0.318	0.325	-2.2	87	0.00
37	2-Methylnaphthalene	0.726	0.741	-2.1	89	0.00
38	1-Methylnaphthalene	0.710	0.721	-1.5	89	0.00
39 I	Acenaphthene-d10	1.000	1.000	0.0	87	0.00
40	1,2,4,5-Tetrachlorobenzene	0.491	0.550	-12.0	94	0.00
41 P	Hexachlorocyclopentadiene	0.240	0.277	-15.4	96	0.00
42 S	2,4,6-Tribromophenol	0.157	0.180	-14.6	91	0.00
43 C	2,4,6-Trichlorophenol	0.337	0.390	-15.7	93	0.00
44	2,4,5-Trichlorophenol	0.375	0.426	-13.6	92	0.00
45 S	2-Fluorobiphenyl	1.297	1.422	-9.6	91	0.00
46	1,1'-Biphenyl	1.474	1.565	-6.2	88	0.00
47	2-Chloronaphthalene	1.132	1.207	-6.6	89	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM100522\
 Data File : BM036839.D
 Acq On : 05 Oct 2022 17:22
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC040

Quant Time: Oct 06 02:35:33 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM100322.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Oct 04 03:17:03 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
48	2-Nitroaniline	0.360	0.382	-6.1	82	0.00
49	Acenaphthylene	1.834	1.960	-6.9	87	0.00
50	Dimethylphthalate	1.410	1.448	-2.7	84	0.00
51	2,6-Dinitrotoluene	0.282	0.309	-9.6	86	0.00
52 C	Acenaphthene	1.140	1.189	-4.3	87	0.00
53	3-Nitroaniline	0.338	0.366	-8.3	84	0.00
54 P	2,4-Dinitrophenol	0.141	0.160	-13.5	92	0.00
55	Dibenzofuran	1.725	1.792	-3.9	87	0.00
56 P	4-Nitrophenol	0.270	0.277	-2.6	82	0.00
57	2,4-Dinitrotoluene	0.366	0.404	-10.4	85	0.00
58	Fluorene	1.413	1.478	-4.6	86	0.00
59	2,3,4,6-Tetrachlorophenol	0.310	0.346	-11.6	90	0.00
60	Diethylphthalate	1.463	1.452	0.8	80	0.00
61	4-Chlorophenyl-phenylether	0.618	0.656	-6.1	88	0.00
62	4-Nitroaniline	0.347	0.373	-7.5	83	0.00
63	Azobenzene	1.627	1.545	5.0	76	0.00
64 I	Phenanthrene-d10	1.000	1.000	0.0	82	0.00
65	4,6-Dinitro-2-methylphenol	0.099	0.116	-17.2	91	0.00
66 c	n-Nitrosodiphenylamine	0.618	0.661	-7.0	83	0.00
67	4-Bromophenyl-phenylether	0.184	0.201	-9.2	86	0.00
68	Hexachlorobenzene	0.204	0.226	-10.8	88	0.00
69	Atrazine	0.170	0.189	-11.2	81	0.00
70 C	Pentachlorophenol	0.125	0.142	-13.6	87	0.00
71	Phenanthrene	1.120	1.179	-5.3	83	0.00
72	Anthracene	1.068	1.139	-6.6	82	0.00
73	Carbazole	0.997	1.053	-5.6	81	0.00
74	Di-n-butylphthalate	1.248	1.272	-1.9	75	0.00
75 C	Fluoranthene	1.152	1.212	-5.2	81	0.00
76 I	Chrysene-d12	1.000	1.000	0.0	84	0.00
77	Benzydine	0.420	0.364	13.3	65	0.00
78	Pyrene	1.506	1.560	-3.6	81	0.00
79 S	Terphenyl-d14	1.004	1.072	-6.8	83	0.00
80	Butylbenzylphthalate	0.621	0.625	-0.6	74	0.00
81	Benzo(a)anthracene	1.299	1.355	-4.3	84	0.00
82	3,3'-Dichlorobenzidine	0.353	0.407	-15.3	89	0.00
83	Chrysene	1.251	1.301	-4.0	84	0.00
84	Bis(2-ethylhexyl)phthalate	0.911	0.874	4.1	70	0.00
85 c	Di-n-octyl phthalate	1.468	1.334	9.1	71	0.00
86 I	Perylene-d12	1.000	1.000	0.0	94	0.00
87	Indeno(1,2,3-cd)pyrene	1.321	1.508	-14.2	102	0.00
88	Benzo(b)fluoranthene	1.267	1.288	-1.7	88	0.00
89	Benzo(k)fluoranthene	1.304	1.338	-2.6	92	0.00
90 C	Benzo(a)pyrene	1.019	1.086	-6.6	93	0.00
91	Dibenzo(a,h)anthracene	1.100	1.244	-13.1	100	0.00
92	Benzo(g,h,i)perylene	1.068	1.233	-15.4	103	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM100522\
Data File : BM036839.D
Acq On : 05 Oct 2022 17:22
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTDCCC040

Quant Time: Oct 06 02:35:33 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM100322.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Oct 04 03:17:03 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
----------	-------	------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 1