




Data Path : Z:\HPCHEM1\BNA M\DATA\BM100617\
Data File : BM011959.D
Acq On : 07 Oct 2017 02:42
Operator : SJ/JU
Sample : I5462-05
Misc :
ALS Vial : 22 Sample Multiplier: 1

Quant Time: Oct 07 07:01:16 2017
Quant Method : Z:\HPCHEM1\BNA M\METHODS\SOM-EPA-BM100317.M
Quant Title : SVOA CALIBRATION
QLast Update : Sat Oct 07 05:58:48 2017
Response via : Initial Calibration



TIC: BM011959.D

(85) Benzo(b)fluoranthene

23.039min (+0.000) 3.15ng/ul m

response 332007

Ion	Exp%	Act%
252.00	100	100
253.00	22.40	24.24
125.00	4.50	11.52#
0.00	0.00	0.00