

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM100621\
 Data File : BM032339.D
 Acq On : 06 Oct 2021 12:45
 Operator : CG/JU
 Sample : PB139008BL
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SBLK008

Quant Time: Oct 06 13:15:53 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-BM100521.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Oct 06 10:06:22 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.619	152	249239	20.00	ng/ul	0.00
20) Naphthalene-d8	10.413	136	1043212	20.00	ng/ul	0.00
38) Acenaphthene-d10	14.266	164	678234	20.00	ng/ul	0.00
64) Phenanthrene-d10	16.995	188	1451383	20.00	ng/ul	0.00
79) Chrysene-d12	21.160	240	1423416	20.00	ng/ul	0.00
88) Perylene-d12	23.306	264	1433171	20.00	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	2.961	96	35659	5.74	ng/uL	0.00
4) Pyridine-d5	3.384	84	500962	28.53	ng/ul	0.00
7) Phenol-d5	6.808	99	595295	28.09	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	6.949	67	355326	29.20	ng/ul	0.00
11) 2-Chlorophenol-d4	7.160	132	475018	29.35	ng/ul	0.00
15) 4-Methylphenol-d8	8.349	113	468543	27.37	ng/ul	0.00
21) Nitrobenzene-d5	8.772	128	219601	29.84	ng/ul	0.00
24) 2-Nitrophenol-d4	9.502	143	237347	29.92	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.043	165	439805	27.85	ng/ul	0.00
31) 4-Chloroaniline-d4	10.543	131	662684	28.45	ng/ul	0.00
46) Dimethylphthalate-d6	13.672	166	1509866	32.17	ng/ul	0.00
49) Acenaphthylene-d8	13.960	160	1731604	30.37	ng/ul	0.00
54) 4-Nitrophenol-d4	14.472	143	236473	25.65	ng/ul	0.00
60) Fluorene-d10	15.260	176	1266304	31.81	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.372	200	189731	22.94	ng/ul	0.00
73) Anthracene-d10	17.095	188	2043054	32.34	ng/ul	0.00
81) Pyrene-d10	19.377	212	2359440	32.95	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.171	264	2225022	31.00	ng/ul	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM100621\
Data File : BM032339.D
Acq On : 06 Oct 2021 12:45
Operator : CG/JU
Sample : PB139008BL
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK008

Quant Time: Oct 06 13:15:53 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-BM100521.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Oct 06 10:06:22 2021
Response via : Initial Calibration

