

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM100717\
 Data File : BM011969.D
 Acq On : 07 Oct 2017 13:04
 Operator : SJ/JU
 Sample : I5735-02
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 B0AE0

Quant Time: Oct 07 22:32:54 2017
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM100317.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Oct 07 05:58:48 2017
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.86	152	251002	20.00	ng/ul	0.00
18) Naphthalene-d8	10.66	136	1014425	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.50	164	592688	20.00	ng/ul	0.00
61) Phenanthrene-d10	17.24	188	1337710	20.00	ng/ul	0.00
75) Chrysene-d12	21.42	240	1459213	20.00	ng/ul	0.00
83) Perylene-d12	23.74	264	1540251	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.27	96	23484	4.43	ng/uL	0.00
5) Phenol-d5	7.02	99	405741	18.79	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.18	67	253442	19.83	ng/ul	0.00
9) 2-Chlorophenol-d4	7.39	132	358262	20.68	ng/ul	0.00
13) 4-Methylphenol-d8	8.56	113	346139	18.54	ng/ul	0.00
19) Nitrobenzene-d5	9.02	128	161687	22.38	ng/ul	0.00
22) 2-Nitrophenol-d4	9.75	143	185038	22.74	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.28	165	342438	20.83	ng/ul	0.00
29) 4-Chloroaniline-d4	10.80	131	433311	24.15	ng/ul	0.00
43) Dimethylphthalate-d6	13.90	166	1090055	21.22	ng/ul	0.00
46) Acenaphthylene-d8	14.19	160	1340340	22.21	ng/ul	0.00
51) 4-Nitrophenol-d4	14.70	143	105715	11.78	ng/ul	0.00
57) Fluorene-d10	15.49	176	920120	20.30	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.62	200	141318	15.70	ng/ul	0.00
70) Anthracene-d10	17.34	188	1445630	21.93	ng/ul	0.00
76) Pyrene-d10	19.63	212	1607750	24.43	ng/ul	0.00
87) Benzo(a)pyrene-d12	23.59	264	1607476	21.69	ng/ul	0.00

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_M\DATA\BM100717\
Data File : BM011969.D
Acq On : 07 Oct 2017 13:04
Operator : SJ/JU
Sample : I5735-02
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
B0AE0

Quant Time: Oct 07 22:32:54 2017
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM100317.M
Quant Title : SVOA CALIBRATION
QLast Update : Sat Oct 07 05:58:48 2017
Response via : Initial Calibration

