

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM100722\
 Data File : BM036892.D
 Acq On : 08 Oct 2022 12:43
 Operator : CG/JU
 Sample : N4873-10
 Misc :
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 EXZ56

Quant Time: Oct 10 01:25:43 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM100622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Oct 10 00:54:15 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.959	152	7349	0.400	ng/ul	0.00
4) Naphthalene-d8	10.765	136	25743	0.400	ng/ul	# 0.00
9) Acenaphthene-d10	14.584	164	14244	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.322	188	29042	0.400	ng/ul	0.00
17) Chrysene-d12	21.488	240	21896	0.400	ng/ul	0.00
23) Perylene-d12	23.891	264	17672	0.400	ng/ul	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.352	96	37729	3.640	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.349	152	11284	0.290	ng/ul	0.00
18) Fluoranthene-d10	19.340	212	26928	0.411	ng/ul	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.386	88	1376	0.136	ng/ul#	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM100722\
Data File : BM036892.D
Acq On : 08 Oct 2022 12:43
Operator : CG/JU
Sample : N4873-10
Misc :
ALS Vial : 47 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
EXZ56

Quant Time: Oct 10 01:25:43 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM100622.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Oct 10 00:54:15 2022
Response via : Initial Calibration

