

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM100722\
 Data File : BM036884.D
 Acq On : 08 Oct 2022 07:57
 Operator : CG/JU
 Sample : PB147999BL
 Misc :
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SBLK999

Quant Time: Oct 10 01:24:49 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM100622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Oct 10 00:54:15 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.959	152	8307	0.400	ng/u1	0.00
4) Naphthalene-d8	10.765	136	29112	0.400	ng/u1 #	0.00
9) Acenaphthene-d10	14.585	164	16211	0.400	ng/u1	0.00
13) Phenanthrene-d10	17.323	188	33692	0.400	ng/u1	0.00
17) Chrysene-d12	21.490	240	27536	0.400	ng/u1	0.00
23) Perylene-d12	23.890	264	24749	0.400	ng/u1 #	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.352	96	53735	4.587	ng/u1	0.00
6) 2-Methylnaphthalene-d10	12.349	152	12843	0.292	ng/u1	0.00
18) Fluoranthene-d10	19.341	212	25646	0.311	ng/u1	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM100722\
Data File : BM036884.D
Acq On : 08 Oct 2022 07:57
Operator : CG/JU
Sample : PB147999BL
Misc :
ALS Vial : 39 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK999

Quant Time: Oct 10 01:24:49 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM100622.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Oct 10 00:54:15 2022
Response via : Initial Calibration

