

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM100722\
 Data File : BM036896.D
 Acq On : 08 Oct 2022 15:07
 Operator : CG/JU
 Sample : N4873-17
 Misc :
 ALS Vial : 51 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 EXZ62

Quant Time: Oct 10 01:26:11 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM100622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Oct 10 00:54:15 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.964	152	7059	0.400	ng/ul	0.00
4) Naphthalene-d8	10.765	136	24807	0.400	ng/ul	# 0.00
9) Acenaphthene-d10	14.585	164	16323	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.322	188	30635	0.400	ng/ul	0.00
17) Chrysene-d12	21.487	240	24286	0.400	ng/ul	0.00
23) Perylene-d12	23.890	264	19520	0.400	ng/ul	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.361	96	37597	3.777	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.349	152	14249	0.380	ng/ul	0.00
18) Fluoranthene-d10	19.341	212	34630	0.477	ng/ul	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.403	88	1316	0.135	ng/ul#	78


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM100722\
Data File : BM036896.D
Acq On : 08 Oct 2022 15:07
Operator : CG/JU
Sample : N4873-17
Misc :
ALS Vial : 51 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
EXZ62


Quant Time: Oct 10 01:26:11 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM100622.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Oct 10 00:54:15 2022
Response via : Initial Calibration





#2
 1,4-Dioxane
 Concen: 0.135 ng/ul
 RT: 3.403 min Scan# 70
 Delta R.T. 0.013 min
 Lab File: BM036896.D
 Acq: 08 Oct 2022 15:07

Instrument :
 BNA_M
 ClientSampleId :
 EXZ62



Tgt Ion: 88 Resp: 1316
 Ion Ratio Lower Upper
 88 100
 43 60.2 37.8 56.6#
 58 35.3 42.6 63.8#

