


Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM100722\
 Data File : BM036848.D
 Acq On : 07 Oct 2022 09:07
 Operator : CG/JU
 Sample : DFTPP073
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP073

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM100622.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Thu Oct 06 17:02:30 2022



AutoFind: Scans 2494, 2495, 2496; Background Corrected with Scan 2487

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	41.4	168021	PASS
68	69	0.00	2	1.5	2734	PASS
69	198	0.00	100	43.5	176435	PASS
70	69	0.00	2	0.5	932	PASS
127	198	10	80	53.6	217232	PASS
197	198	0.00	2	0.3	1122	PASS
198	198	100	100	100.0	405483	PASS
199	198	5	9	6.8	27543	PASS
275	198	10	60	22.4	90739	PASS
365	198	1	100	2.6	10730	PASS
441	198	0.01	100	9.0	36605	PASS
442	198	50	100	55.7	225920	PASS
443	442	15	24	19.1	43040	PASS

m/z	Abundance
39.00	197.0
39.90	1846.0
40.90	231.0
41.90	151.0
44.00	3774.0
44.90	285.0
46.80	160.0
48.90	699.0
50.90	371.0
58.90	222.0
61.00	173.0
73.00	1438.0
73.90	179.0
75.00	383.0
77.00	772.0
77.90	386.0
78.90	259.0
84.00	667.0
85.80	359.0
87.00	231.0
88.00	150.0
91.00	222.0
95.90	1137.0
96.70	198.0
102.80	290.0
114.90	218.0
116.90	169.0
118.90	232.0
132.90	1150.0
133.90	162.0
134.90	288.0
147.00	443.0
160.80	156.0
163.00	263.0
164.90	185.0
176.90	286.0
178.80	227.0
191.00	1081.0
191.90	353.0
192.90	710.0
207.00	10383.0
207.90	2337.0
208.90	1217.0
209.80	237.0
221.00	281.0
264.90	247.0
267.00	216.0
280.90	1373.0
281.90	295.0
283.00	216.0

Instrument :
BNA_M
ClientSampleId :
DFTPP073

m/z	Abundance
36.00	404.0
37.00	1049.0
38.00	2779.0
39.00	16752.0
40.00	2752.0
41.00	720.0
41.90	222.0
43.00	261.0
44.00	4513.0
44.90	884.0
47.00	336.0
48.00	167.0
49.00	1945.0
50.00	56496.0
51.00	225280.0
52.00	11735.0
53.00	553.0
55.00	1149.0
55.90	6237.0
57.00	15597.0
58.00	597.0
58.90	378.0
60.00	306.0
60.90	3434.0
62.00	3327.0
63.00	9467.0
64.00	1278.0
65.00	4437.0
65.90	371.0
67.00	507.0
68.00	3664.0
68.90	230656.0
70.00	1254.0
70.80	192.0
73.00	3269.0
74.00	24144.0
75.00	37232.0
76.00	13234.0
77.00	276992.0
78.00	20992.0
79.00	16528.0
80.00	13392.0
81.00	17616.0
81.90	4974.0
83.00	4837.0
83.90	873.0
85.00	3135.0
86.00	5742.0
87.00	2924.0
88.00	1204.0
88.90	487.0
90.10	159.0
90.90	4032.0
92.00	5101.0
93.00	29808.0
93.90	2194.0
94.90	534.0
95.90	2559.0
96.90	524.0
97.90	22176.0
98.90	23424.0
99.90	1843.0
100.90	10828.0
101.90	646.0
102.90	3526.0
103.90	5813.0
105.00	5987.0
106.00	1796.0
107.00	79800.0
108.00	11738.0
108.90	1987.0
109.90	138944.0
111.00	22416.0
111.90	2769.0
112.90	985.0
113.70	188.0
114.90	466.0
115.90	4156.0
116.90	67936.0
118.00	5540.0
118.90	1154.0
120.00	1180.0
120.90	266.0
122.00	4779.0
122.90	7088.0
123.90	3285.0
125.00	3122.0
126.00	742.0
127.00	275264.0
127.90	20856.0
128.90	103888.0
129.90	8739.0

Instrument :
BNA_M
ClientSampleId :
DFTPP073

130.90	1395.0
132.00	847.0
132.90	1781.0
134.00	3175.0
135.00	8020.0
135.90	3718.0
136.90	5868.0
137.80	1255.0
138.90	412.0
140.00	1125.0
140.90	13001.0
141.90	4144.0
142.90	2982.0
143.90	644.0
144.80	619.0
145.90	2447.0
146.90	7545.0
147.90	17552.0
148.90	4081.0
149.90	1344.0
151.00	1958.0
152.00	706.0
152.90	3669.0
153.90	2994.0
155.00	6849.0
156.00	9963.0
156.90	2071.0
157.90	2045.0
158.90	1873.0
159.90	3439.0
160.90	5728.0
161.90	1790.0
162.90	989.0
164.00	721.0
164.90	4404.0
165.90	3346.0
166.90	25440.0
167.90	23624.0
168.90	3210.0
170.00	871.0
170.80	734.0
171.90	2448.0
172.90	3385.0
173.90	5427.0
175.00	9782.0
175.90	2702.0
176.90	5167.0
177.90	1583.0
178.90	18744.0
179.90	15966.0
181.00	5992.0
182.00	994.0
182.70	523.0
184.00	1443.0
184.90	8397.0
185.90	65064.0
186.90	19184.0
187.90	2448.0
188.90	4361.0
189.80	676.0
190.90	3035.0
191.90	6630.0
192.90	6914.0
193.90	1658.0
194.90	760.0
195.90	12509.0
197.00	1586.0
197.90	476096.0
198.90	33536.0
199.90	2674.0
201.40	1914.0
202.90	3429.0
203.90	16792.0
204.90	29608.0
206.00	121792.0
207.00	28312.0
207.90	6415.0
209.00	3229.0
210.00	2415.0
210.90	4421.0
212.00	1185.0
212.80	479.0
213.90	203.0
214.90	1192.0
215.90	2617.0
216.90	31528.0
217.90	4074.0
219.00	452.0
219.90	491.0
221.00	17232.0
221.70	4559.0
222.90	5892.0
224.00	63384.0
225.00	17112.0
226.00	2581.0
226.90	25584.0

Instrument :
BNA_M
ClientSampleId :
DFTPP073

227.90	3788.0
228.90	5802.0
229.90	828.0
230.90	2053.0
231.90	396.0
233.00	418.0
233.90	1900.0
234.90	1450.0
235.90	1650.0
236.90	4092.0
237.90	495.0
238.90	1136.0
239.90	943.0
240.90	1452.0
241.90	3335.0
242.90	3857.0
244.00	50808.0
245.00	6611.0
245.90	9354.0
246.90	1969.0
247.80	304.0
248.90	1966.0
249.80	273.0
250.90	415.0
251.80	560.0
253.00	1314.0
253.90	1038.0
254.90	240320.0
255.90	35120.0
256.90	3667.0
257.90	13150.0
258.90	2297.0
259.90	604.0
260.80	388.0
263.80	365.0
264.90	6332.0
265.90	813.0
267.00	295.0
268.00	157.0
269.90	174.0
270.90	868.0
272.00	616.0
272.90	6690.0
273.90	18224.0
274.90	96888.0
275.90	14017.0
276.90	8918.0
277.90	2080.0
278.90	441.0
281.00	1580.0
282.00	652.0
282.90	1236.0
283.90	755.0
284.90	1774.0
285.90	322.0
287.90	185.0
288.90	322.0
289.90	412.0
290.90	373.0
291.90	319.0
292.90	2291.0
293.90	671.0
294.90	509.0
295.90	28552.0
296.90	4060.0
297.90	181.0
301.00	276.0
302.00	458.0
302.90	3366.0
303.90	959.0
304.90	285.0
307.90	518.0
308.90	307.0
309.90	398.0
311.90	200.0
313.00	185.0
314.00	1484.0
314.90	3150.0
315.90	1734.0
317.00	231.0
320.90	912.0
322.00	395.0
323.00	9398.0
323.90	1496.0
324.90	270.0
325.80	179.0
326.90	1802.0
327.90	1072.0
328.80	222.0
331.80	440.0
332.90	673.0
333.90	5429.0
334.90	1134.0
335.80	273.0
339.00	173.0
341.00	1038.0

Instrument :
BNA_M
ClientSampleId :
DFTPP073

341.90	335.0
343.00	153.0
345.90	1664.0
346.80	311.0
350.70	160.0
352.00	2105.0
353.00	1618.0
354.00	2783.0
355.00	594.0
358.90	217.0
364.90	10309.0
365.90	1445.0
369.90	194.0
370.90	635.0
372.00	3631.0
372.90	1091.0
382.90	1335.0
384.00	365.0
384.80	151.0
390.00	354.0
391.00	318.0
391.90	268.0
400.90	158.0
401.90	1547.0
402.90	2072.0
403.90	1159.0
404.90	458.0
405.90	237.0
414.90	156.0
420.90	1938.0
422.00	1881.0
422.90	11224.0
424.00	6094.0
424.90	1009.0
441.00	33048.0
442.00	205248.0
443.00	39376.0
444.00	3525.0
445.10	218.0

Instrument :
BNA_M
ClientSampleId :
DFTPP073

m/z	Abundance
35.90	261.0
37.00	870.0
38.00	2610.0
39.00	14759.0
40.00	2464.0
41.00	653.0
42.90	276.0
44.00	3793.0
45.00	651.0
46.90	303.0
48.00	249.0
48.90	1939.0
50.00	50208.0
51.00	193728.0
52.00	9795.0
52.90	481.0
55.00	972.0
55.90	5442.0
57.00	13446.0
57.90	573.0
58.90	301.0
60.10	261.0
60.90	2629.0
61.90	3234.0
63.00	8147.0
64.00	1022.0
65.00	4075.0
66.00	350.0
67.00	461.0
67.90	2991.0
68.90	208960.0
70.00	1040.0
71.80	151.0
73.00	3244.0
74.00	21432.0
75.00	33960.0
76.00	11932.0
77.00	250048.0
78.00	17928.0
79.00	14667.0
79.90	11429.0
80.90	16816.0
81.90	4429.0
83.00	4521.0
84.00	1146.0
85.00	3102.0
86.00	5416.0
87.00	2520.0
88.00	981.0
88.90	751.0
91.00	3867.0
91.90	4850.0
93.00	27160.0
93.90	2121.0
94.90	607.0
96.00	2245.0
96.90	570.0
97.90	20760.0
99.00	20240.0
100.00	2059.0
100.90	9610.0
101.90	622.0
102.90	3322.0
103.90	5491.0
105.00	5934.0
105.90	1558.0
107.00	72288.0
108.00	10609.0
109.00	1363.0
109.90	128984.0
110.90	20080.0
111.90	2200.0
112.90	888.0
113.90	250.0
114.90	568.0
116.00	3627.0
116.90	64192.0
117.90	4769.0
118.90	1091.0
119.90	1071.0
121.00	345.0
121.90	4328.0
123.00	7707.0
123.90	3534.0
125.00	3052.0
126.00	613.0
127.00	258368.0
127.90	19552.0
128.90	98776.0
129.90	7666.0
130.90	1599.0
131.90	877.0

Instrument :
BNA_M
ClientSampleId :
DFTPP073

133.00	1461.0
133.90	2759.0
134.90	8409.0
136.00	3289.0
136.90	5618.0
137.70	842.0
138.90	551.0
139.90	1121.0
140.90	12037.0
141.90	4154.0
142.90	2512.0
143.90	995.0
144.90	687.0
145.90	2699.0
146.90	7221.0
147.90	16100.0
148.90	3805.0
149.90	1362.0
151.00	1435.0
151.90	866.0
152.90	3575.0
153.90	3098.0
154.90	6876.0
156.00	9890.0
157.00	1887.0
157.90	2197.0
158.90	1605.0
159.90	4274.0
160.90	6099.0
161.90	1715.0
162.90	1210.0
164.00	739.0
164.90	4157.0
165.90	3616.0
166.90	26864.0
167.90	22024.0
168.90	2860.0
169.90	744.0
170.80	1060.0
171.90	2031.0
172.90	2817.0
173.90	5266.0
175.00	10467.0
175.90	2421.0
176.90	4610.0
177.90	1887.0
178.90	18680.0
179.90	14054.0
180.90	5742.0
181.90	933.0
183.00	620.0
183.90	1414.0
184.90	8481.0
186.00	65768.0
187.00	18648.0
187.90	2316.0
188.90	4461.0
189.80	778.0
190.90	3548.0
191.90	6204.0
193.00	7858.0
193.90	1328.0
194.90	1046.0
195.90	12017.0
196.90	1781.0
197.90	494144.0
198.90	33088.0
200.00	2349.0
201.40	1982.0
202.90	3116.0
203.90	18032.0
205.00	30680.0
206.00	124496.0
207.00	28448.0
207.90	6560.0
208.90	3040.0
210.00	2045.0
210.90	5247.0
211.80	931.0
212.90	470.0
213.90	154.0
214.90	1163.0
215.90	2551.0
216.90	32600.0
218.00	4176.0
218.90	425.0
221.00	18752.0
221.70	4322.0
222.90	7155.0
224.00	71920.0
225.00	17208.0
225.90	2438.0
226.90	29080.0
227.90	3990.0
228.90	5528.0
229.80	627.0

Instrument :
BNA_M
ClientSampleId :
DFTPP073

231.00	2524.0
231.80	489.0
232.80	485.0
233.90	1741.0
234.90	1811.0
235.90	1494.0
236.90	3639.0
238.00	631.0
238.90	1104.0
239.90	1088.0
240.90	1532.0
242.00	3447.0
242.90	3471.0
244.00	56032.0
245.00	7156.0
245.90	10047.0
246.90	1805.0
247.90	406.0
248.90	2016.0
249.90	296.0
250.90	680.0
251.80	342.0
252.90	1354.0
254.00	1341.0
254.90	267200.0
255.90	37384.0
256.90	3133.0
257.90	15553.0
258.90	2520.0
259.90	450.0
260.90	383.0
262.90	174.0
263.90	562.0
264.90	6472.0
265.90	1055.0
266.80	316.0
269.90	209.0
270.90	673.0
272.00	602.0
272.90	8016.0
273.90	20248.0
274.90	112768.0
276.00	15570.0
276.90	9428.0
277.90	2360.0
278.90	455.0
280.90	1691.0
281.90	711.0
283.00	1291.0
283.90	762.0
285.00	1549.0
285.80	432.0
288.80	234.0
289.90	435.0
290.80	326.0
291.90	568.0
292.90	2235.0
294.00	906.0
294.90	521.0
295.90	33968.0
296.90	4323.0
297.80	371.0
300.90	432.0
301.90	548.0
303.00	3820.0
303.90	824.0
307.90	549.0
308.90	352.0
309.90	441.0
312.80	246.0
314.00	1249.0
314.90	3528.0
315.90	1783.0
317.00	310.0
320.90	934.0
321.80	532.0
323.00	10023.0
324.00	1862.0
325.00	205.0
325.80	237.0
326.90	2073.0
327.90	1217.0
331.90	715.0
332.90	1070.0
334.00	7013.0
335.00	1535.0
335.90	186.0
338.90	197.0
339.90	223.0
341.00	996.0
342.00	515.0
345.90	2314.0
346.90	455.0
351.90	2963.0
352.90	1947.0
353.90	3267.0

Instrument :
BNA_M
ClientSampleId :
DFTPP073

355.00	689.0
359.00	229.0
364.90	13898.0
365.90	1839.0
366.90	230.0
369.90	243.0
370.90	881.0
372.00	4715.0
373.00	1527.0
382.90	1553.0
384.00	312.0
385.00	183.0
389.90	799.0
391.00	374.0
391.80	272.0
400.70	307.0
402.00	2358.0
403.00	2773.0
403.90	1249.0
405.10	339.0
415.10	177.0
420.90	2493.0
422.00	2559.0
423.00	16159.0
424.00	6715.0
424.90	1036.0
441.00	45512.0
442.00	277056.0
443.00	54456.0
444.00	4722.0
445.00	177.0

Instrument :
BNA_M
ClientSampleId :
DFTPP073

m/z	Abundance
35.90	298.0
36.90	646.0
38.10	1066.0
39.00	6745.0
39.90	2334.0
41.00	501.0
42.90	366.0
44.00	4346.0
44.90	491.0
46.90	269.0
48.90	1350.0
50.00	22544.0
51.00	86168.0
52.00	4201.0
52.80	310.0
55.00	441.0
56.00	2641.0
57.00	5724.0
57.90	215.0
58.90	281.0
60.00	181.0
61.00	1480.0
61.90	1365.0
63.00	3616.0
63.90	604.0
64.90	1776.0
67.90	1547.0
68.90	89688.0
70.00	503.0
73.00	2223.0
74.00	9252.0
75.00	15512.0
76.00	5006.0
77.00	112040.0
78.00	9083.0
79.00	6091.0
79.90	5132.0
81.00	8150.0
81.90	1950.0
83.00	1758.0
83.90	584.0
85.00	1434.0
85.90	2941.0
86.90	1333.0
87.90	539.0
88.90	454.0
90.90	2060.0
91.90	2110.0
92.90	11543.0
93.90	959.0
94.90	263.0
95.90	2145.0
96.80	476.0
97.90	9217.0
99.00	8954.0
99.90	565.0
100.90	3893.0
101.90	232.0
103.00	1223.0
103.90	3114.0
104.90	2140.0
105.90	1135.0
107.00	32736.0
107.90	4845.0
109.00	815.0
109.90	58496.0
110.90	9003.0
112.00	990.0
112.90	335.0
113.80	162.0
115.00	422.0
115.90	1580.0
116.90	28984.0
117.90	2382.0
118.90	713.0
119.90	428.0
120.90	192.0
121.90	2031.0
122.90	3685.0
123.90	1698.0
124.90	1394.0
125.90	368.0
127.00	118064.0
127.90	8844.0
128.90	45560.0
129.90	3745.0
131.00	662.0
132.00	539.0
133.00	1419.0
133.90	1480.0
134.90	4361.0
136.00	1500.0

Instrument :
BNA_M
ClientSampleId :
DFTPP073

136.90	2557.0
138.00	486.0
139.00	306.0
139.80	423.0
140.90	5876.0
141.90	1568.0
142.90	1394.0
144.00	426.0
144.90	317.0
145.90	823.0
146.90	3780.0
147.90	8132.0
148.90	1942.0
149.90	606.0
151.20	534.0
151.70	443.0
152.90	1783.0
153.80	1656.0
154.90	3405.0
155.90	5000.0
157.00	826.0
157.90	1254.0
158.90	951.0
159.90	2002.0
161.00	2644.0
161.90	871.0
162.90	636.0
163.90	323.0
164.90	1960.0
165.90	1989.0
166.90	13475.0
167.90	11328.0
168.90	1500.0
169.70	297.0
170.80	305.0
171.90	1261.0
172.90	1387.0
173.90	2604.0
175.00	4276.0
175.80	1393.0
176.90	2278.0
177.80	818.0
178.90	8996.0
179.90	7040.0
181.00	2855.0
182.10	515.0
182.80	525.0
183.80	586.0
184.90	3807.0
186.00	32464.0
186.90	9244.0
188.00	972.0
188.80	1728.0
189.90	282.0
190.90	2482.0
191.90	2608.0
192.90	3821.0
193.90	784.0
194.80	612.0
195.90	7849.0
196.70	920.0
197.90	246208.0
198.90	16004.0
199.90	1403.0
201.30	1101.0
202.00	253.0
202.90	1612.0
204.00	8919.0
204.90	15185.0
206.00	63664.0
207.00	20952.0
207.90	4830.0
208.90	2498.0
210.00	1041.0
210.90	2186.0
211.70	471.0
212.80	316.0
214.90	666.0
215.90	1160.0
216.90	15974.0
217.90	2179.0
218.90	165.0
220.90	9196.0
221.80	2190.0
222.90	3638.0
224.00	34312.0
224.90	8520.0
226.00	1112.0
226.90	14705.0
227.90	2004.0
228.90	3591.0
229.90	321.0
230.90	1170.0
233.00	222.0
233.90	1049.0
234.90	1070.0

Instrument :
BNA_M
ClientSampleId :
DFTPP073

235.90	790.0
236.90	1911.0
237.90	321.0
238.90	603.0
239.90	674.0
241.00	656.0
242.00	2033.0
242.90	2081.0
244.00	29496.0
245.00	3361.0
245.90	5223.0
247.00	1111.0
247.80	216.0
248.90	1014.0
249.90	220.0
250.90	430.0
252.00	294.0
252.90	342.0
253.90	787.0
254.90	144576.0
255.90	20600.0
256.90	1776.0
257.90	8771.0
258.90	1318.0
260.80	308.0
264.90	3438.0
265.80	511.0
266.80	260.0
270.80	344.0
271.90	390.0
272.90	4083.0
273.90	11654.0
274.90	62560.0
275.90	7948.0
276.90	5115.0
277.90	1147.0
278.90	193.0
280.90	1736.0
282.00	425.0
283.00	989.0
284.00	407.0
284.90	1072.0
285.90	224.0
288.90	265.0
289.90	223.0
291.00	184.0
291.90	255.0
292.80	1221.0
293.90	493.0
295.00	431.0
295.90	19432.0
296.90	2197.0
297.80	267.0
300.90	160.0
301.90	181.0
302.90	2139.0
303.90	594.0
304.90	172.0
307.90	177.0
308.90	241.0
312.90	237.0
313.90	736.0
314.90	1910.0
315.90	1163.0
320.80	665.0
321.80	221.0
323.00	5844.0
324.00	1191.0
326.90	1212.0
327.90	561.0
331.90	367.0
332.80	513.0
334.00	4264.0
334.90	906.0
338.90	276.0
340.90	828.0
345.90	1595.0
346.90	478.0
350.90	169.0
351.90	1799.0
352.90	1498.0
353.90	1976.0
355.00	357.0
358.80	188.0
364.90	7984.0
365.90	1771.0
369.90	188.0
370.90	406.0
371.90	3023.0
372.90	967.0
382.90	949.0
383.90	336.0
389.90	384.0
391.00	354.0
392.00	267.0
400.80	162.0

Instrument :
BNA_M
ClientSampleId :
DFTPP073

402.00	1212.0
402.90	1874.0
403.90	1042.0
414.90	248.0
420.90	1645.0
421.90	1778.0
423.00	10986.0
424.00	3891.0
424.90	543.0
440.00	194.0
441.00	31256.0
442.00	195456.0
443.00	35288.0
444.00	3182.0
445.00	202.0

Instrument :
BNA_M
ClientSampleId :
DFTPP073