


Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM100724\
Data File : BM047878.D
Acq On : 07 Oct 2024 09:21
Operator : RC/JU
Sample : DFTPP046
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP046

Integration File: LSCINT.P

Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM092724.MA.M
Title : SVOA CALIBRATION
Last Update : Fri Oct 04 08:32:31 2024



AutoFind: Scans 2487, 2488, 2489; Background Corrected with Scan 2481

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	30.8	303405	PASS
68	69	0.00	2	1.5	4620	PASS
69	198	0.00	100	31.8	313655	PASS
70	69	0.00	2	0.4	1378	PASS
127	198	10	80	43.6	429576	PASS
197	198	0.00	2	0.5	5120	PASS
198	198	100	100	100.0	985453	PASS
199	198	5	9	6.7	66317	PASS
275	198	10	60	25.7	253696	PASS
365	198	1	100	3.6	35467	PASS
441	198	0.01	100	14.6	143973	PASS
442	198	50	100	91.8	904939	PASS
443	442	15	24	19.5	176448	PASS

m/z	Abundance
35.90	156.0
36.90	165.0
38.00	340.0
39.10	624.0
40.00	3936.0
41.10	1566.0
42.00	339.0
43.10	1996.0
44.00	11388.0
45.00	458.0
47.00	279.0
48.90	633.0
51.20	467.0
52.90	306.0
55.00	1528.0
55.90	498.0
57.10	1501.0
58.90	331.0
62.10	188.0
63.20	210.0
65.00	388.0
67.00	549.0
68.00	275.0
69.10	1054.0
70.10	534.0
71.10	792.0
72.00	151.0
73.10	5302.0
74.10	711.0
75.00	703.0
77.20	793.0
78.00	347.0
79.00	560.0
80.00	270.0
81.00	553.0
82.10	479.0
83.10	757.0
84.10	956.0
85.00	528.0
85.80	387.0
87.20	165.0
88.10	154.0
89.00	209.0
91.10	1047.0
92.10	361.0
93.10	489.0
94.10	204.0
95.20	590.0
96.10	519.0
97.10	635.0
98.10	314.0
98.90	294.0
101.90	151.0
103.00	288.0
104.00	223.0
105.10	631.0
106.00	284.0
107.10	331.0
109.10	375.0
110.10	200.0
111.10	270.0
114.90	500.0
117.00	321.0
119.20	501.0
120.20	179.0
121.00	344.0
122.10	218.0
123.00	306.0
125.10	398.0
127.00	291.0
128.20	432.0
129.10	426.0
131.10	412.0
132.00	197.0
133.10	689.0
133.90	209.0
135.10	302.0
137.00	242.0
139.10	373.0
140.00	268.0
141.10	261.0
143.10	169.0
145.10	202.0
146.00	180.0
147.10	2144.0
148.10	667.0
149.00	521.0
150.00	158.0
150.90	161.0
152.00	267.0
152.90	206.0
155.00	152.0

Instrument :
BNA_M
ClientSampleId :
DFTPP046

156.00	155.0
157.20	289.0
159.00	193.0
161.00	288.0
162.00	237.0
163.00	356.0
165.00	298.0
166.20	362.0
167.10	1829.0
168.10	351.0
169.00	270.0
173.10	173.0
175.10	273.0
176.10	187.0
177.00	318.0
178.10	277.0
179.00	222.0
181.00	191.0
181.90	150.0
183.20	162.0
187.00	377.0
189.00	171.0
191.00	986.0
192.10	330.0
192.90	761.0
194.00	235.0
195.00	161.0
197.00	219.0
198.10	232.0
202.90	197.0
204.00	293.0
205.00	339.0
207.10	5535.0
208.00	1384.0
209.00	732.0
210.10	184.0
211.10	164.0
219.10	167.0
220.10	170.0
221.10	1679.0
222.10	514.0
223.10	503.0
225.10	210.0
249.00	151.0
253.00	197.0
265.10	182.0
267.10	681.0
267.90	195.0
281.10	2426.0
282.00	624.0
282.90	419.0
295.10	256.0
325.00	309.0
327.10	185.0
341.00	779.0
342.10	346.0
355.10	1209.0
356.10	537.0
357.00	168.0
358.10	169.0
401.00	210.0
403.00	156.0
429.10	369.0

Instrument :
BNA_M
ClientSampleId :
DFTPP046

m/z	Abundance
36.00	514.0
37.00	1146.0
38.00	3325.0
39.10	24912.0
40.00	3964.0
41.10	2160.0
42.10	467.0
43.10	2195.0
44.00	9405.0
45.00	1000.0
47.00	342.0
47.90	287.0
48.20	250.0
49.10	2981.0
50.10	88232.0
51.10	361024.0
52.10	17632.0
53.10	1106.0
54.00	293.0
55.10	2744.0
56.10	10675.0
57.10	25624.0
58.00	1444.0
59.00	642.0
60.00	491.0
61.10	4334.0
62.10	5502.0
63.10	14678.0
64.00	1937.0
65.10	7189.0
66.00	641.0
67.10	1045.0
68.10	5794.0
69.00	363008.0
70.10	2071.0
71.00	1235.0
72.00	389.0
73.10	4070.0
74.10	35872.0
75.10	57256.0
76.10	18928.0
77.10	438464.0
78.10	30024.0
79.10	24384.0
80.00	19936.0
81.10	29048.0
82.10	7541.0
83.10	7454.0
84.10	1493.0
85.10	6121.0
86.00	8431.0
87.00	3766.0
88.00	1899.0
89.00	825.0
89.90	241.0
91.00	7218.0
92.10	7424.0
93.00	46032.0
94.10	3000.0
95.10	1602.0
96.10	2684.0
97.10	1455.0
98.10	34024.0
99.00	28520.0
100.00	2569.0
101.00	16928.0
102.00	945.0
103.10	5055.0
104.10	9537.0
105.10	10666.0
106.10	3514.0
107.10	133312.0
108.10	20136.0
109.10	3368.0
110.10	242240.0
111.10	36776.0
112.10	4779.0
113.10	1716.0
114.00	336.0
115.00	832.0
116.10	6656.0
117.00	105968.0
118.10	8524.0
119.10	1407.0
120.10	1721.0
121.00	765.0
122.10	7824.0
123.00	13099.0
124.00	5417.0
125.00	5860.0
126.20	2275.0
127.10	487808.0

Instrument :
BNA_M
ClientSampleId :
DFTPP046

128.10	37032.0
129.00	184640.0
130.10	14592.0
131.00	3515.0
132.10	1503.0
133.00	1221.0
134.00	4327.0
135.10	14158.0
136.00	6170.0
137.10	7807.0
138.00	2148.0
139.10	1131.0
140.10	2554.0
141.10	23088.0
142.10	7089.0
143.00	4865.0
144.10	1555.0
145.00	2004.0
146.10	4298.0
147.10	11033.0
148.00	28448.0
149.10	6112.0
150.00	1473.0
151.20	2810.0
152.00	2479.0
153.10	6814.0
154.00	5710.0
155.10	12891.0
156.10	20264.0
157.10	3542.0
158.00	4210.0
159.10	3181.0
160.00	6828.0
161.10	11427.0
162.00	3054.0
162.90	873.0
164.00	1168.0
165.00	8428.0
166.10	6791.0
167.10	49472.0
168.10	27376.0
169.10	4563.0
170.00	1464.0
171.10	2118.0
172.00	4757.0
173.10	5904.0
174.10	10878.0
175.10	19896.0
176.10	5801.0
177.00	8568.0
178.00	3172.0
179.00	36096.0
180.10	24512.0
181.00	13064.0
182.20	1863.0
182.90	1017.0
184.00	2574.0
185.10	17160.0
186.10	138112.0
187.10	38832.0
188.10	3803.0
189.10	8376.0
190.10	1847.0
191.10	4204.0
192.10	12219.0
193.10	13641.0
194.10	3068.0
195.00	1495.0
196.10	26176.0
197.10	5843.0
198.00	1068032.0
199.00	75320.0
200.00	5833.0
201.50	3945.0
203.00	6552.0
204.10	37288.0
205.10	63568.0
206.10	268096.0
207.10	38904.0
208.10	9160.0
209.00	3029.0
210.20	4042.0
211.10	9532.0
213.10	729.0
214.10	334.0
215.10	2906.0
216.10	5097.0
217.00	70800.0
218.10	9153.0
219.10	1135.0
221.10	41920.0
221.90	9297.0
223.10	15065.0
224.10	149760.0
225.10	40520.0
226.10	4331.0

Instrument :
BNA_M
ClientSampleId :
DFTPP046

227.10	60296.0
228.10	8998.0
229.10	12523.0
230.10	1945.0
231.10	4755.0
232.10	1172.0
233.00	1299.0
234.00	3675.0
235.10	4778.0
236.10	3686.0
237.00	4748.0
238.00	971.0
239.00	2292.0
240.10	1897.0
241.00	4138.0
242.10	7596.0
243.10	8662.0
244.10	116768.0
245.10	15819.0
246.10	22176.0
247.10	4205.0
248.10	781.0
249.00	3679.0
250.10	787.0
251.00	985.0
252.10	1147.0
253.10	2621.0
254.10	3574.0
255.10	588800.0
256.10	88344.0
257.10	7020.0
258.10	33552.0
259.10	5720.0
260.00	964.0
261.00	1366.0
262.10	266.0
263.00	350.0
264.00	957.0
265.10	13005.0
266.10	1584.0
267.10	427.0
267.80	225.0
270.00	769.0
271.10	1822.0
272.10	1937.0
273.10	18200.0
274.10	46232.0
275.10	262208.0
276.10	36352.0
277.10	20672.0
278.10	3378.0
279.10	872.0
281.00	1448.0
282.10	841.0
283.10	2694.0
284.10	1845.0
285.10	3465.0
286.10	792.0
288.20	281.0
289.10	976.0
290.10	1002.0
290.90	618.0
292.00	1070.0
293.10	4916.0
294.00	1411.0
295.00	1244.0
296.10	75264.0
297.10	10802.0
298.10	913.0
301.10	969.0
302.10	1250.0
303.20	9230.0
304.10	2204.0
305.00	335.0
307.00	152.0
308.10	997.0
309.00	671.0
310.10	1006.0
311.00	252.0
312.00	228.0
313.00	744.0
314.10	3586.0
315.10	8417.0
316.10	4219.0
317.10	757.0
319.00	161.0
320.00	343.0
321.00	2306.0
322.10	1085.0
323.10	23664.0
324.10	4299.0
325.10	360.0
326.00	504.0
327.10	5111.0
328.10	2189.0
329.20	641.0

Instrument :
BNA_M
ClientSampleId :
DFTPP046

332.10	1722.0
333.20	2144.0
334.10	15041.0
335.10	4400.0
336.10	395.0
339.10	455.0
340.10	269.0
341.20	2856.0
342.00	781.0
343.10	224.0
346.10	5731.0
347.10	1216.0
348.10	188.0
350.30	163.0
351.10	410.0
352.10	7209.0
353.10	5120.0
354.20	7434.0
355.20	2424.0
356.00	287.0
359.20	555.0
360.00	255.0
364.10	203.0
365.10	33792.0
366.10	5041.0
367.20	309.0
370.20	624.0
371.10	2079.0
372.10	12829.0
373.10	3033.0
374.20	472.0
376.90	265.0
383.10	3504.0
384.00	951.0
385.10	297.0
390.10	1507.0
391.10	1127.0
392.10	829.0
393.10	269.0
395.10	220.0
401.10	1129.0
402.10	5141.0
403.10	7676.0
404.10	2557.0
405.10	548.0
415.00	203.0
421.10	6850.0
422.10	6670.0
423.10	43632.0
424.10	9930.0
425.00	1273.0
438.00	185.0
440.10	161.0
441.10	128624.0
442.10	795840.0
443.10	159488.0
444.20	15068.0
445.10	688.0

Instrument :
BNA_M
ClientSampleId :
DFTPP046

m/z	Abundance
36.00	445.0
37.00	1391.0
38.10	3400.0
39.10	23736.0
40.00	4061.0
41.10	2223.0
42.00	515.0
43.10	2367.0
44.00	9720.0
45.00	706.0
46.90	257.0
48.10	315.0
49.10	2585.0
50.10	87424.0
51.10	341184.0
52.10	17848.0
53.00	896.0
54.20	290.0
55.00	2808.0
56.10	9864.0
57.00	23656.0
58.10	1412.0
59.00	410.0
60.00	421.0
61.10	4127.0
62.10	4886.0
63.10	14248.0
64.00	1684.0
65.10	7097.0
66.00	757.0
67.10	1058.0
68.00	5392.0
69.00	357184.0
70.10	2330.0
71.20	1278.0
72.10	214.0
73.00	3641.0
74.10	34568.0
75.10	55648.0
76.10	19456.0
77.10	428992.0
78.10	30528.0
79.00	24192.0
80.00	19848.0
81.00	27392.0
82.10	8261.0
83.00	7653.0
84.00	1514.0
85.10	6559.0
86.10	8054.0
87.10	3563.0
88.10	1506.0
89.10	879.0
90.10	243.0
91.10	6685.0
92.00	6790.0
93.00	43984.0
94.10	3129.0
95.00	1525.0
96.10	3013.0
97.10	1557.0
98.10	34224.0
99.00	29176.0
100.00	2300.0
101.00	16736.0
102.00	1022.0
103.00	5363.0
104.00	9436.0
105.00	10558.0
106.10	3760.0
107.10	134400.0
108.10	20464.0
109.10	3893.0
110.00	242176.0
111.00	35104.0
112.10	4698.0
113.10	1393.0
114.10	408.0
115.00	909.0
116.10	6957.0
117.00	110264.0
118.10	8655.0
119.10	1498.0
120.00	1881.0
120.90	740.0
122.10	8754.0
123.00	13872.0
124.00	6206.0
125.10	6143.0
126.10	2322.0
127.10	491392.0
128.10	35768.0

Instrument :
BNA_M
ClientSampleId :
DFTPP046

129.00	185984.0
130.10	14518.0
131.00	3173.0
132.10	1625.0
133.00	1328.0
134.10	4625.0
135.10	15247.0
136.00	6344.0
137.10	8348.0
137.90	1847.0
139.00	1149.0
140.00	2429.0
141.10	24040.0
142.00	7045.0
143.00	6143.0
144.00	1288.0
145.00	1900.0
146.00	4290.0
147.10	12830.0
148.00	29200.0
149.00	6222.0
150.10	1916.0
151.10	3153.0
151.90	1903.0
153.10	7909.0
154.00	5884.0
155.10	14711.0
156.10	21568.0
157.10	4016.0
158.00	4165.0
159.10	2936.0
160.10	7118.0
161.10	11823.0
162.00	3372.0
163.10	1625.0
164.00	1372.0
165.10	8814.0
166.10	7824.0
167.10	50856.0
168.10	27544.0
169.10	4792.0
170.00	1561.0
171.00	2285.0
172.10	4357.0
173.10	5762.0
174.10	10178.0
175.10	20968.0
176.10	5584.0
177.00	8417.0
178.10	3320.0
179.00	36680.0
180.10	25608.0
181.10	12539.0
182.10	1860.0
183.20	1216.0
184.10	3143.0
185.10	18016.0
186.10	146560.0
187.10	40472.0
188.20	4771.0
189.00	8652.0
190.00	1457.0
191.10	5193.0
192.10	11949.0
193.10	13528.0
194.10	2931.0
195.10	1581.0
196.10	28768.0
197.10	6497.0
198.00	1137152.0
199.00	75864.0
200.00	6316.0
201.60	3657.0
203.10	6868.0
204.10	39952.0
205.10	69920.0
206.10	286720.0
207.10	41752.0
208.10	10327.0
209.10	3555.0
210.10	4772.0
211.10	10673.0
213.10	912.0
214.10	540.0
215.00	3211.0
216.10	6580.0
217.00	75264.0
218.10	9811.0
219.10	973.0
221.10	42256.0
221.90	9975.0
223.10	17928.0
224.10	167616.0
225.10	39144.0
226.20	4364.0
227.10	66648.0

Instrument :
BNA_M
ClientSampleId :
DFTPP046

228.10	9096.0
229.10	14258.0
230.00	1773.0
231.10	5717.0
232.10	1236.0
233.10	1350.0
234.00	3842.0
235.00	4688.0
236.10	3099.0
237.10	4787.0
238.00	705.0
239.00	2602.0
240.10	1793.0
241.00	3198.0
242.10	9723.0
243.10	8839.0
244.10	130808.0
245.10	16912.0
246.10	23104.0
247.10	4983.0
248.00	1203.0
249.00	4557.0
250.00	1167.0
251.00	1480.0
252.10	1352.0
253.00	2738.0
254.10	4098.0
255.10	666880.0
256.10	97064.0
257.10	7062.0
258.10	37648.0
259.10	5447.0
260.00	1045.0
261.00	1205.0
261.90	290.0
263.00	544.0
264.10	980.0
265.10	15209.0
266.10	2515.0
267.00	526.0
268.10	280.0
269.00	220.0
270.00	761.0
271.00	2302.0
272.10	2021.0
273.10	19576.0
274.10	53936.0
275.10	294528.0
276.10	39504.0
277.10	24600.0
278.10	4132.0
279.00	1270.0
281.10	1678.0
282.10	900.0
283.10	3049.0
284.10	2282.0
285.10	4588.0
286.00	833.0
287.80	213.0
289.10	925.0
290.10	762.0
291.00	676.0
292.10	1095.0
293.10	5866.0
294.10	1250.0
295.00	1446.0
296.10	86696.0
297.10	11361.0
298.10	1014.0
299.10	249.0
301.10	1338.0
302.10	1442.0
303.10	9807.0
304.10	2585.0
305.10	396.0
307.00	202.0
308.00	1107.0
309.00	754.0
310.10	920.0
311.00	283.0
312.00	205.0
313.00	994.0
314.20	3974.0
315.10	9813.0
316.10	5128.0
317.10	722.0
320.00	368.0
321.10	2280.0
322.10	1327.0
323.10	28872.0
324.10	5461.0
325.00	518.0
326.10	597.0
327.10	5157.0
328.10	2616.0
329.10	599.0

Instrument :
BNA_M
ClientSampleId :
DFTPP046

330.00	160.0
331.00	187.0
332.10	2030.0
333.10	3030.0
334.10	18904.0
335.10	4823.0
336.20	537.0
339.10	345.0
340.00	372.0
341.10	3878.0
342.10	1042.0
343.10	197.0
346.10	6972.0
347.10	1333.0
349.90	253.0
351.00	372.0
352.10	8722.0
353.20	6215.0
354.10	8673.0
355.10	1838.0
356.10	411.0
357.20	279.0
359.10	748.0
360.00	153.0
364.00	302.0
365.10	40800.0
366.10	6408.0
367.10	499.0
370.10	754.0
371.10	2452.0
372.20	15845.0
373.10	3921.0
374.20	442.0
376.90	294.0
383.10	4642.0
384.20	1248.0
385.10	421.0
390.10	1774.0
391.10	1375.0
392.10	1022.0
393.10	203.0
395.00	155.0
397.20	177.0
401.10	996.0
402.10	6034.0
403.10	9810.0
404.10	3517.0
405.10	466.0
410.10	387.0
415.10	530.0
416.00	200.0
420.00	178.0
421.10	8657.0
422.10	7568.0
423.10	55296.0
424.10	12431.0
425.20	1182.0
438.00	374.0
439.00	201.0
440.10	240.0
441.10	164992.0
442.10	1033152.0
443.20	205312.0
444.10	18376.0
445.20	1063.0

Instrument :
BNA_M
ClientSampleId :
DFTPP046

m/z	Abundance
35.90	319.0
37.10	775.0
38.10	2626.0
39.10	15505.0
40.00	4068.0
41.10	1690.0
42.10	645.0
43.10	2199.0
44.00	12176.0
45.00	686.0
46.10	164.0
47.00	371.0
48.00	190.0
49.10	2234.0
50.10	51672.0
51.10	209408.0
52.10	10485.0
53.10	791.0
54.10	311.0
55.00	2760.0
56.00	6763.0
57.00	15880.0
58.00	1023.0
59.00	554.0
59.90	202.0
61.10	2383.0
62.00	2695.0
63.10	8214.0
64.10	1278.0
65.10	4556.0
66.00	734.0
67.00	1101.0
68.00	3500.0
69.00	223936.0
70.10	1336.0
71.20	1230.0
72.10	271.0
73.10	2763.0
74.10	20616.0
75.10	35416.0
76.10	12908.0
77.10	264448.0
78.10	16760.0
79.00	15825.0
80.00	12207.0
81.10	18784.0
82.00	5450.0
83.10	5388.0
84.00	1008.0
85.10	4086.0
86.10	4633.0
87.10	2244.0
88.00	897.0
89.10	672.0
91.00	4652.0
92.10	4509.0
93.00	28224.0
94.00	1932.0
95.10	1139.0
96.10	2466.0
97.10	1254.0
98.00	21280.0
99.00	18456.0
100.00	1380.0
101.00	9857.0
102.00	593.0
103.00	3232.0
104.00	6565.0
105.10	6630.0
106.00	2091.0
107.00	82240.0
108.10	12809.0
109.00	2210.0
110.00	155008.0
111.00	23032.0
112.10	2813.0
113.00	1160.0
114.00	187.0
115.00	636.0
116.00	4186.0
117.00	67816.0
118.00	4575.0
119.10	920.0
120.10	1382.0
121.00	579.0
122.00	5290.0
123.00	8634.0
124.00	3678.0
125.00	3840.0
126.20	1432.0
127.10	310400.0
128.10	22544.0

Instrument :
BNA_M
ClientSampleId :
DFTPP046

129.00	116992.0
130.00	9509.0
131.00	2183.0
132.10	1083.0
133.00	1106.0
134.00	3496.0
135.00	9276.0
136.00	3711.0
137.10	4883.0
138.00	1058.0
139.00	751.0
140.10	1731.0
141.00	14496.0
142.10	5176.0
143.10	3253.0
144.10	1122.0
145.00	1129.0
146.10	2782.0
147.10	8882.0
148.00	19008.0
149.10	3963.0
150.10	1001.0
151.10	1869.0
152.00	1222.0
153.10	4591.0
154.00	4183.0
155.10	8641.0
156.10	12639.0
157.10	2620.0
158.00	2799.0
159.00	2308.0
160.00	5092.0
161.10	7353.0
162.00	1993.0
163.00	824.0
164.00	752.0
165.00	6133.0
166.10	4541.0
167.10	33768.0
168.10	17896.0
169.10	2879.0
170.00	1216.0
171.00	1377.0
172.00	3044.0
173.10	3675.0
174.10	6609.0
175.10	13047.0
176.10	3415.0
177.10	5628.0
178.00	2290.0
179.00	25856.0
180.10	17464.0
181.00	8058.0
182.10	1616.0
183.00	823.0
184.00	1884.0
185.10	11909.0
186.10	94976.0
187.10	25984.0
188.10	3091.0
189.00	6169.0
190.00	1160.0
191.00	3565.0
192.00	8137.0
193.10	9758.0
194.10	2103.0
195.10	1295.0
196.10	18792.0
197.10	3677.0
198.00	751872.0
199.00	47768.0
200.00	3622.0
201.60	3339.0
203.10	4957.0
204.10	26096.0
205.10	46360.0
206.10	189312.0
207.10	29176.0
208.10	6881.0
209.10	2546.0
210.10	2888.0
211.10	6455.0
211.80	1526.0
213.10	712.0
214.00	208.0
215.00	2126.0
216.00	3653.0
217.00	50736.0
218.00	6665.0
219.10	761.0
221.10	30104.0
222.00	6504.0
223.10	11106.0
224.10	110864.0
225.10	27392.0
226.10	2698.0

Instrument :
BNA_M
ClientSampleId :
DFTPP046

227.10	44864.0
228.10	6214.0
229.00	9341.0
230.00	1005.0
231.10	4069.0
232.10	589.0
233.00	751.0
234.10	2622.0
235.00	2995.0
236.10	2320.0
237.10	3728.0
238.00	779.0
239.10	1630.0
240.10	1293.0
241.00	2691.0
242.10	5736.0
243.10	6242.0
244.10	88904.0
245.10	11401.0
246.10	15831.0
247.00	3394.0
247.90	615.0
249.10	3404.0
250.00	766.0
251.00	838.0
252.10	959.0
253.00	2210.0
254.10	2506.0
255.10	457792.0
256.10	63080.0
257.10	4639.0
258.10	25064.0
259.10	4124.0
260.00	772.0
261.10	932.0
262.00	222.0
264.00	772.0
265.10	10459.0
266.10	1413.0
267.00	318.0
267.90	173.0
270.10	619.0
271.00	1431.0
272.10	1408.0
273.10	13823.0
274.10	36264.0
275.10	204352.0
276.10	28392.0
277.10	16760.0
278.10	2933.0
279.10	535.0
281.10	1280.0
282.10	850.0
283.10	2797.0
284.10	1375.0
285.10	3400.0
285.90	613.0
287.90	169.0
289.00	712.0
290.00	527.0
291.00	495.0
292.10	1100.0
293.00	3986.0
294.10	1200.0
295.10	1216.0
296.10	59640.0
297.10	7853.0
298.10	424.0
299.10	180.0
301.20	754.0
302.10	899.0
303.10	7290.0
304.10	1888.0
307.10	211.0
308.10	937.0
309.10	686.0
310.10	880.0
312.00	227.0
313.10	421.0
314.10	2467.0
315.00	6895.0
316.10	3788.0
316.90	594.0
320.10	213.0
321.10	2202.0
322.10	1010.0
323.10	22088.0
324.10	3765.0
325.00	388.0
326.10	369.0
327.10	4110.0
328.10	2333.0
329.10	350.0
332.10	1378.0
333.10	1954.0
334.10	13800.0

Instrument :
BNA_M
ClientSampleId :
DFTPP046

335.10	3218.0
336.10	550.0
339.10	230.0
340.10	285.0
341.00	2543.0
342.00	532.0
346.00	5123.0
347.00	901.0
349.90	264.0
350.90	393.0
352.10	6986.0
353.10	4961.0
354.10	6616.0
355.10	1353.0
356.10	297.0
358.90	528.0
364.00	229.0
365.10	31808.0
366.10	5002.0
367.10	341.0
370.10	875.0
371.00	1356.0
372.10	13044.0
373.10	3211.0
374.20	306.0
377.10	336.0
383.10	3359.0
384.10	1133.0
385.00	312.0
390.10	1649.0
391.10	1378.0
392.20	961.0
393.10	164.0
397.10	171.0
401.20	833.0
402.10	5702.0
403.10	7994.0
404.10	2699.0
405.10	501.0
409.90	311.0
415.10	301.0
415.90	163.0
421.10	7103.0
422.20	5823.0
423.10	47096.0
424.10	9371.0
425.10	1186.0
438.10	230.0
439.10	282.0
440.10	371.0
441.10	138304.0
442.10	885824.0
443.10	164544.0
444.20	14612.0
445.10	674.0

Instrument :
BNA_M
ClientSampleId :
DFTPP046