

Data Path : Z:\HPCHEM1\BNA M\DATA\BM100815\
 Data File : BM002876.D
 Acq On : 08 Oct 2015 22:47
 Operator : UM/IZ
 Sample : PB85741BL
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 PB85741BL

Quant Time: Oct 09 07:05:43 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\8270-SIM-BM100815.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 09 06:47:52 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.68	152	42683	5.00	ng	0.00
7) Naphthalene-d8	11.52	136	170973	5.00	ng	0.00
13) Acenaphthene-d10	15.25	164	94908	5.00	ng	0.00
19) Phenanthrene-d10	17.97	188	195511	5.00	ng	0.02
26) Chrysene-d12	22.05	240	261082	5.00	ng	0.00
35) Perylene-d12	24.63	264	260306	5.00	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	6.13	112	65413	7.17	ng	-0.02
5) Phenol-d6	7.80	99	84734	7.51	ng	0.00
8) Nitrobenzene-d5	9.88	82	33523	3.90	ng	0.00
14) 2,4,6-Tribromophenol	16.74	330	24274	7.72	ng	0.00
15) 2-Fluorobiphenyl	13.89	172	112504	3.91	ng	0.00
29) Terphenyl-d14	20.49	244	131681	3.78	ng	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA M\DATA\BM100815\
Data File : BM002876.D
Acq On : 08 Oct 2015 22:47
Operator : UM/IZ
Sample : PB85741BL
Misc :
ALS Vial : 12 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
PB85741BL

Quant Time: Oct 09 07:05:43 2015
Quant Method : Z:\HPCHEM1\BNA M\METHODS\8270-SIM-BM100815.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 09 06:47:52 2015
Response via : Initial Calibration

