

Data Path : Z:\svoasrv\HPCHEM1\BNA M\Data\BM100820\
 Data File : BM027646.D
 Acq On : 08 Oct 2020 13:06
 Operator : JU/CG
 Sample : SSTDO.404
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTDO.4104

Quant Time: Oct 08 13:41:41 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\SFAM-EPA-SIM-BM100820.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 08 13:40:34 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.79	152	882	0.40	ng/ul	0.00
2) Naphthalene-d8	10.57	136	2461	0.40	ng/ul	0.00
7) Acenaphthene-d10	14.42	164	1742	0.40	ng/ul	0.00
11) Phenanthrene-d10	17.16	188	4522	0.40	ng/ul	0.00
17) Chrysene-d12	21.34	240	4925	0.40	ng/ul	0.00
21) Perylene-d12	23.59	264	5020	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.16	152	1904	0.38	ng/ul	0.00
15) Fluoranthene-d10	19.18	212	4473	0.29	ng/ul	0.00
Target Compounds					Ovalue	
3) Naphthalene	10.62	128	2860	0.419	ng/ul	100
5) 2-Methylnaphthalene	12.23	142	2119	0.402	ng/ul	100
6) 1-Methylnaphthalene	12.45	142	2388	0.396	ng/ul	100
8) Acenaphthylene	14.14	152	3077	0.392	ng/ul	100
9) Acenaphthene	14.48	153	2437	0.416	ng/ul	100
10) Fluorene	15.47	166	3139	0.407	ng/ul	100
12) Pentachlorophenol	16.81	266	750	0.424	ng/ul	100
13) Phenanthrene	17.20	178	5628	0.418	ng/ul	100
14) Anthracene	17.29	178	5239	0.405	ng/ul	100
16) Fluoranthene	19.21	202	5439	0.287	ng/ul	100
18) Pyrene	19.58	202	5477	0.293	ng/ul	100
19) Benzo(a)anthracene	21.32	228	7808	0.403	ng/ul	100
20) Chrysene	21.37	228	8237	0.414	ng/ul	100
22) Benzo(b)fluoranthene	22.91	252	8534	0.413	ng/ul	100
23) Benzo(k)fluoranthene	22.96	252	8536	0.419	ng/ul	100
24) Benzo(a)pyrene	23.49	252	7721	0.418	ng/ul	100
25) Indeno(1,2,3-cd)pyrene	25.86	276	10272	0.418	ng/ul#	100
26) Dibenzo(a,h)anthracene	25.87	278	8512	0.417	ng/ul	100
27) Benzo(a,h,i)perylene	26.55	276	8955	0.425	ng/ul	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA M\Data\BM100820\
Data File : BM027646.D
Acq On : 08 Oct 2020 13:06
Operator : JU/CG
Sample : SST0.404
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTD0.4104

Quant Time: Oct 08 13:41:41 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\SFAM-EPA-SIM-BM100820.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 08 13:40:34 2020
Response via : Initial Calibration

