

Data Path : Z:\svoasrv\HPCHEM1\BNA M\Data\BM100820\
 Data File : BM027648.D
 Acq On : 08 Oct 2020 14:18
 Operator : JU/CG
 Sample : SSTD1.606
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD1.6106

Quant Time: Oct 08 14:50:16 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\SFAM-EPA-SIM-BM100820.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 08 13:40:34 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.79	152	849	0.40	ng/ul	0.00
2) Naphthalene-d8	10.57	136	2348	0.40	ng/ul	0.00
7) Acenaphthene-d10	14.42	164	1572	0.40	ng/ul	0.00
11) Phenanthrene-d10	17.16	188	3754	0.40	ng/ul	0.00
17) Chrysene-d12	21.33	240	3942	0.40	ng/ul	0.00
21) Perylene-d12	23.59	264	3904	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.16	152	6829	1.45	ng/ul	0.00
15) Fluoranthene-d10	19.18	212	18769	1.46	ng/ul	0.00
Target Compounds					Ovalue	
3) Naphthalene	10.62	128	10313	1.583	ng/ul	95
5) 2-Methylnaphthalene	12.23	142	7491	1.488	ng/ul	98
6) 1-Methylnaphthalene	12.45	142	8598	1.496	ng/ul	100
8) Acenaphthylene	14.13	152	10626	1.500	ng/ul	98
9) Acenaphthene	14.48	153	8315	1.573	ng/ul	97
10) Fluorene	15.46	166	10560	1.515	ng/ul	99
12) Pentachlorophenol	16.81	266	2376	1.619	ng/ul	98
13) Phenanthrene	17.20	178	17680	1.582	ng/ul	99
14) Anthracene	17.29	178	16807	1.564	ng/ul	99
16) Fluoranthene	19.21	202	23826	1.512	ng/ul	100
18) Pyrene	19.57	202	23809	1.591	ng/ul	99
19) Benzo(a)anthracene	21.32	228	23722	1.529	ng/ul	99
20) Chrysene	21.37	228	25086	1.577	ng/ul	99
22) Benzo(b)fluoranthene	22.91	252	25616	1.593	ng/ul	93
23) Benzo(k)fluoranthene	22.95	252	25654	1.619	ng/ul#	93
24) Benzo(a)pyrene	23.49	252	22842	1.590	ng/ul#	91
25) Indeno(1,2,3-cd)pyrene	25.86	276	30755	1.610	ng/ul#	100
26) Dibenzo(a,h)anthracene	25.87	278	25569	1.609	ng/ul	94
27) Benzo(a,h,i)perylene	26.55	276	26795	1.635	ng/ul	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA M\Data\BM100820\
Data File : BM027648.D
Acq On : 08 Oct 2020 14:18
Operator : JU/CG
Sample : SSTD1.606
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTD1.6106

Quant Time: Oct 08 14:50:16 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA M\METHODS\SFAM-EPA-SIM-BM100820.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 08 13:40:34 2020
Response via : Initial Calibration

