

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM100820\
 Data File : BM027646.D
 Acq On : 08 Oct 2020 13:06
 Operator : JU/CG
 Sample : SST0.404
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SST0.4104

Quant Time: Oct 08 13:40:59 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM100820.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 08 13:40:34 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.788	152	882	0.40	ng/ul	0.00
2) Naphthalene-d8	10.570	136	2461	0.40	ng/ul	0.00
7) Acenaphthene-d10	14.416	164	1742	0.40	ng/ul	0.00
11) Phenanthrene-d10	17.157	188	4522	0.40	ng/ul	0.00
17) Chrysene-d12	21.336	240	4925	0.40	ng/ul	0.00
21) Perylene-d12	23.590	264	5020	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.158	152	1904	0.38	ng/ul	0.00
15) Fluoranthene-d10	19.182	212	4473	0.29	ng/ul	0.00
Target Compounds						Qvalue
3) Naphthalene	10.619	128	2860	0.42	ng/ul	100
5) 2-Methylnaphthalene	12.234	142	2119	0.40	ng/ul	100
6) 1-Methylnaphthalene	12.454	142	2388	0.40	ng/ul	100
8) Acenaphthylene	14.135	152	3077	0.39	ng/ul	100
9) Acenaphthene	14.476	153	2437	0.42	ng/ul	100
10) Fluorene	15.466	166	3139	0.41	ng/ul	100
12) Pentachlorophenol	16.807	266	750	0.42	ng/ul	100
13) Phenanthrene	17.199	178	5628	0.42	ng/ul	100
14) Anthracene	17.289	178	5239	0.40	ng/ul	100
16) Fluoranthene	19.213	202	5439	0.29	ng/ul	100
18) Pyrene	19.575	202	5477	0.29	ng/ul	100
19) Benzo(a)anthracene	21.321	228	7808	0.40	ng/ul	100
20) Chrysene	21.369	228	8237	0.41	ng/ul	100
22) Benzo(b)fluoranthene	22.912	252	8534	0.41	ng/ul	100
23) Benzo(k)fluoranthene	22.958	252	8536	0.42	ng/ul	100
24) Benzo(a)pyrene	23.491	252	7721	0.42	ng/ul	100
25) Indeno(1,2,3-cd)pyrene	25.858	276	10272	0.42	ng/ul#	100
26) Dibenzo(a,h)anthracene	25.870	278	8512	0.42	ng/ul	100
27) Benzo(g,h,i)perylene	26.551	276	8955	0.43	ng/ul	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM100820\
Data File : BM027646.D
Acq On : 08 Oct 2020 13:06
Operator : JU/CG
Sample : SST0.404
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTD0.4104

Quant Time: Oct 08 13:40:59 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM100820.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 08 13:40:34 2020
Response via : Initial Calibration

