

Data Path : Z:\HPCHEM1\BNA_M\Data\BM100915\
 Data File : BM002880.D
 Acq On : 09 Oct 2015 10:03
 Operator : UM/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC001

Quant Time: Oct 10 04:51:04 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\8270-SIM-BM100815.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 09 06:47:52 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	127	-0.02
2	1,4-Dioxane	0.499	0.442	11.4	121	0.00
3	n-Nitrosodimethylamine	0.389	0.368	5.4	116	0.00
4 S	2-Fluorophenol	1.069	0.981	8.2	118	0.00
5 S	Phenol-d6	1.321	1.191	9.8	118	0.00
6	bis(2-Chloroethyl)ether	1.303	1.145	12.1	115	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	124	-0.02
8 S	Nitrobenzene-d5	0.251	0.237	5.6	119	0.00
9	Nitrobenzene	0.266	0.251	5.6	120	-0.01
10	Naphthalene	1.110	1.081	2.6	124	-0.02
11	Hexachlorobutadiene	0.176	0.175	0.6	127	0.00
12	2-Methylnaphthalene	0.738	0.717	2.8	123	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	119	0.00
14 S	2,4,6-Tribromophenol	0.166	0.138	16.9	107	0.00
15 S	2-Fluorobiphenyl	1.514	1.565	-3.4	124	-0.01
16	Acenaphthylene	2.122	2.034	4.1	115	0.00
17	Acenaphthene	1.370	1.343	2.0	122	-0.02
18	Fluorene	1.678	1.581	5.8	115	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	129	0.00
20	4-Bromophenyl-phenylether	0.237	0.234	1.3	124	0.00
21	Hexachlorobenzene	0.292	0.252	13.7	113	0.00
22	Pentachlorophenol	0.027	0.018	33.3#	101	0.00
23	Phenanthrene	1.463	1.286	12.1	116	0.00
24	Anthracene	1.237	1.025	17.1	111	0.00
25	Fluoranthene	1.621	1.362	16.0	115	0.00
26 I	Chrysene-d12	1.000	1.000	0.0	109	0.00
27	Benzidine	0.688	0.546	20.6	98	0.00
28	Pyrene	1.349	1.454	-7.8	115	0.00
29 S	Terphenyl-d14	0.667	0.711	-6.6	118	-0.02
30	Benzo(a)anthracene	1.435	1.377	4.0	112	0.00
31	3,3'-Dichlorobenzidine	0.551	0.471	14.5	97	0.00
32	Chrysene	1.344	1.369	-1.9	111	-0.02
33	Bis(2-ethylhexyl)phthalate	0.874	0.827	5.4	87	-0.02
34	Indeno(1,2,3-cd)pyrene	1.611	1.361	15.5	95	0.00
35 I	Perylene-d12	1.000	1.000	0.0	103	0.00
36	Benzo(b)fluoranthene	1.431	1.392	2.7	98	0.00
37	Benzo(k)fluoranthene	1.382	1.422	-2.9	113	0.00
38 C	Benzo(a)pyrene	1.316	1.250	5.0	101	0.00
39	Dibenzo(a,h)anthracene	1.282	1.141	11.0	95	0.00
40	Benzo(g,h,i)perylene	1.324	1.180	10.9	94	0.00

Data Path : Z:\HPCHEM1\BNA_M\Data\BM100915\
Data File : BM002880.D
Acq On : 09 Oct 2015 10:03
Operator : UM/IZ
Sample : SSTDCCC001
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTDCCC001

Quant Time: Oct 10 04:51:04 2015
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\8270-SIM-BM100815.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 09 06:47:52 2015
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range				
SPCC's out = 0 CCC's out = 0				