

Data Path : Z:\HPCHEM1\BNA_M\Data\BM100915\
 Data File : BM002889.D
 Acq On : 09 Oct 2015 17:47
 Operator : UM/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC001

Quant Time: Oct 12 05:15:35 2015
 Quant Method : Z:\HPCHEM1\BNA_M\METHODS\8270-SIM-BM100815.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 09 06:47:52 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	152	-0.02
2	1,4-Dioxane	1.000	0.956	4.4	141	0.00
3	n-Nitrosodimethylamine	1.000	0.949	5.1	140	0.00
4 S	2-Fluorophenol	1.000	0.870	13.0	134	0.00
5 S	Phenol-d6	1.000	0.910	9.0	143	0.00
6	bis(2-Chloroethyl)ether	1.000	0.955	4.5	150	0.00
7 I	Naphthalene-d8	5.000	5.000	0.0	150	-0.01
8 S	Nitrobenzene-d5	1.000	0.953	4.7	145	0.00
9	Nitrobenzene	1.000	0.953	4.7	146	0.00
10	Naphthalene	1.000	0.961	3.9	148	-0.01
11	Hexachlorobutadiene	1.000	0.976	2.4	151	0.00
12	2-Methylnaphthalene	1.000	0.959	4.1	147	0.00
13 I	Acenaphthene-d10	5.000	5.000	0.0	149	0.00
14 S	2,4,6-Tribromophenol	1.000	0.827	17.3	133	0.00
15 S	2-Fluorobiphenyl	1.000	0.986	1.4	147	0.00
16	Acenaphthylene	1.000	0.964	3.6	145	0.00
17	Acenaphthene	1.000	0.938	6.2	146	0.00
18	Fluorene	1.000	0.949	5.1	145	0.00
19 I	Phenanthrene-d10	5.000	5.000	0.0	131	0.02
20	4-Bromophenyl-phenylether	1.000	0.913	8.7	117	0.00
21	Hexachlorobenzene	1.000	1.117	-11.7	149	0.00
22	Pentachlorophenol	1.000	0.771	22.9	89	0.02
23	Phenanthrene	1.000	0.998	0.2	134	0.00
24	Anthracene	1.000	1.086	-8.6	149	0.00
25	Fluoranthene	1.000	0.925	7.5	130	0.01
26 I	Chrysene-d12	5.000	5.000	0.0	126	0.00
27	Benzydine	1.000	0.763	23.7	101	0.00
28	Pyrene	1.000	1.032	-3.2	128	0.01
29 S	Terphenyl-d14	1.000	1.056	-5.6	136	0.00
30	Benzo(a)anthracene	1.000	0.901	9.9	122	0.00
31	3,3'-Dichlorobenzidine	1.000	0.814	18.6	107	0.00
32	Chrysene	1.000	0.965	3.5	123	0.00
33	Bis(2-ethylhexyl)phthalate	1.000	0.873	12.7	93	0.00
34	Indeno(1,2,3-cd)pyrene	1.000	0.681	31.9#	88	0.00
35 I	Perylene-d12	5.000	5.000	0.0	101	0.00
36	Benzo(b)fluoranthene	1.000	0.987	1.3	98	0.00
37	Benzo(k)fluoranthene	1.000	1.030	-3.0	111	0.00
38 C	Benzo(a)pyrene	1.000	0.944	5.6	98	0.00
39	Dibenzo(a,h)anthracene	1.000	0.849	15.1	88	0.00
40	Benzo(g,h,i)perylene	1.000	0.848	15.2	88	0.00

Data Path : Z:\HPCHEM1\BNA_M\Data\BM100915\
Data File : BM002889.D
Acq On : 09 Oct 2015 17:47
Operator : UM/IZ
Sample : SSTDCCC001
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTDCCC001

Quant Time: Oct 12 05:15:35 2015
Quant Method : Z:\HPCHEM1\BNA_M\METHODS\8270-SIM-BM100815.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 09 06:47:52 2015
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev Area	% Dev(min)

(#) = Out of Range				
SPCC's out = 0 CCC's out = 0				