

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM100921\
 Data File : BM032406.D
 Acq On : 08 Oct 2021 22:58
 Operator : CG/JU
 Sample : M4136-03MSD
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 PSC124027MSD

Manual Integrations
 APPROVED

mohammad
 10/11/2021 2:13:34 PM

Quant Time: Oct 09 02:17:40 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM100221.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 08 17:11:04 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.613	152	361321	20.00	ng	0.00
21) Naphthalene-d8	10.407	136	1440698	20.00	ng	# 0.00
39) Acenaphthene-d10	14.265	164	817484	20.00	ng	0.00
64) Phenanthrene-d10	16.989	188	1457315	20.00	ng	# 0.00
76) Chrysene-d12	21.153	240	952815	20.00	ng	# 0.00
86) Perylene-d12	23.300	264	991354	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.190	112	2524238	117.71	ng	0.00
7) Phenol-d6	6.819	99	3192102	108.85	ng	0.00
23) Nitrobenzene-d5	8.772	82	1930505	74.53	ng	0.00
42) 2,4,6-Tribromophenol	15.754	330	894303	97.72	ng	0.00
45) 2-Fluorobiphenyl	12.883	172	3015880	54.85	ng	0.00
79) Terphenyl-d14	19.600	244	2700803	59.70	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	2.996	88	335360	36.75	ng	90
3) Pyridine	3.401	79	809840	32.79	ng	94
4) n-Nitrosodimethylamine	3.319	42	428283	43.24	ng	# 58
6) Aniline	6.942	93	1327871	40.98	ng	90
8) 2-Chlorophenol	7.189	128	1121599	47.99	ng	91
9) Benzaldehyde	6.742	77	639409	37.36	ng	83
10) Phenol	6.842	94	1347250	47.71	ng	82
11) bis(2-Chloroethyl)ether	7.036	93	1024205	45.61	ng	92
12) 1,3-Dichlorobenzene	7.507	146	1184742	44.97	ng	# 90
13) 1,4-Dichlorobenzene	7.648	146	1208133	45.06	ng	98
14) 1,2-Dichlorobenzene	7.972	146	1161497	45.13	ng	97
15) Benzyl Alcohol	7.860	79	949529	47.28	ng	87
16) 2,2'-oxybis(1-Chloropr...	8.148	45	1253445	44.06	ng	98
17) 2-Methylphenol	8.083	107	909205	47.65	ng	# 89
18) Hexachloroethane	8.701	117	435240	47.68	ng	91
19) n-Nitroso-di-n-propyla...	8.425	70	716524	44.94	ng	# 82
20) 3+4-Methylphenols	8.413	107	1189413	46.20	ng	95
22) Acetophenone	8.436	105	1514784	43.63	ng	# 87
24) Nitrobenzene	8.813	77	1137484	45.53	ng	# 87
25) Isophorone	9.330	82	1980185	46.01	ng	96
26) 2-Nitrophenol	9.525	139	582411	50.10	ng	# 73
27) 2,4-Dimethylphenol	9.601	122	1003730	51.18	ng	95
28) bis(2-Chloroethoxy)met...	9.819	93	1292686	41.77	ng	100
29) 2,4-Dichlorophenol	10.066	162	1004665	46.00	ng	96
30) 1,2,4-Trichlorobenzene	10.277	180	1057601	42.69	ng	97
31) Naphthalene	10.454	128	3220136	43.56	ng	100
32) Benzoic acid	9.783	122	617567	42.32	ng	# 83
33) 4-Chloroaniline	10.560	127	923211	30.25	ng	94
34) Hexachlorobutadiene	10.760	225	633517	42.79	ng	98
35) Caprolactam	11.336	113	302392	45.18	ng	# 76
36) 4-Chloro-3-methylphenol	11.713	107	1033363	43.94	ng	96
37) 2-Methylnaphthalene	12.071	142	2227410m	44.33	ng	
38) 1-Methylnaphthalene	12.295	142	2055112	41.86	ng	# 97
40) 1,2,4,5-Tetrachloroben...	12.454	216	1048676	44.53	ng	# 97
41) Hexachlorocyclopentadiene	12.442	237	858440	76.47	ng	97
43) 2,4,6-Trichlorophenol	12.695	196	747716	45.89	ng	97

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM100921\
 Data File : BM032406.D
 Acq On : 08 Oct 2021 22:58
 Operator : CG/JU
 Sample : M4136-03MSD
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 PSC124027MSD


Manual Integrations
 APPROVED

mohammad
 10/11/2021 2:13:34 PM

Quant Time: Oct 09 02:17:40 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM100221.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 08 17:11:04 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)	
44) 2,4,5-Trichlorophenol	12.771	196	797435	45.53	ng	#	91
46) 1,1'-Biphenyl	13.095	154	2625646	43.52	ng		99
47) 2-Chloronaphthalene	13.130	162	2077783	44.91	ng		99
48) 2-Nitroaniline	13.336	65	622630	49.50	ng	#	78
49) Acenaphthylene	13.983	152	3293904	45.86	ng		99
50) Dimethylphthalate	13.724	163	2504639	43.39	ng		99
51) 2,6-Dinitrotoluene	13.830	165	554349	47.41	ng	#	65
52) Acenaphthene	14.330	154	2120406m	44.75	ng		
53) 3-Nitroaniline	14.165	138	442248	33.66	ng		86
54) 2,4-Dinitrophenol	14.365	184	326584	50.33	ng	#	64
55) Dibenzofuran	14.660	168	3003573	41.64	ng		93
56) 4-Nitrophenol	14.495	139	906231	83.63	ng	#	67
57) 2,4-Dinitrotoluene	14.624	165	731016	47.42	ng	#	58
58) Fluorene	15.307	166	2204307	40.18	ng		99
59) 2,3,4,6-Tetrachlorophenol	14.895	232	632594m	41.93	ng		
60) Diethylphthalate	15.083	149	2444741	43.14	ng		96
61) 4-Chlorophenyl-phenyle...	15.301	204	1104689	38.88	ng		93
62) 4-Nitroaniline	15.330	138	527134	40.20	ng	#	68
63) Azobenzene	15.595	77	2289316	42.43	ng		83
65) 4,6-Dinitro-2-methylph...	15.389	198	259321	31.61	ng	#	72
66) n-Nitrosodiphenylamine	15.518	169	2053775	47.73	ng		98
67) 4-Bromophenyl-phenylether	16.189	248	696696	45.93	ng		90
68) Hexachlorobenzene	16.324	284	754625	45.20	ng	#	84
69) Atrazine	16.459	200	700411	47.80	ng		96
70) Pentachlorophenol	16.659	266	861216	83.57	ng		98
71) Phenanthrene	17.036	178	3352784	43.83	ng		100
72) Anthracene	17.124	178	3384237	45.28	ng		99
73) Carbazole	17.389	167	2981417	40.17	ng		98
74) Di-n-butylphthalate	17.948	149	3745677	46.98	ng	#	97
75) Fluoranthene	19.042	202	3324961	39.20	ng		97
77) Benzidine	19.224	184	2288064	96.67	ng		98
78) Pyrene	19.400	202	3283407	50.82	ng		99
80) Butylbenzylphthalate	20.294	149	1349333	54.98	ng		87
81) Benzo(a)anthracene	21.141	228	2623294	44.05	ng		98
82) 3,3'-Dichlorobenzidine	21.077	252	948637	50.56	ng		99
83) Chrysene	21.194	228	2591366	44.98	ng		99
84) Bis(2-ethylhexyl)phtha...	21.065	149	1881064	56.85	ng	#	93
85) Di-n-octyl phthalate	21.906	149	2924480	52.71	ng	#	89
87) Indeno(1,2,3-cd)pyrene	25.429	276	3059053	50.08	ng		96
88) Benzo(b)fluoranthene	22.665	252	2678221	44.41	ng		99
89) Benzo(k)fluoranthene	22.706	252	2565087	43.54	ng		99
90) Benzo(a)pyrene	23.206	252	2630910	53.80	ng		98
91) Dibenzo(a,h)anthracene	25.429	278	2574925	50.00	ng	#	96
92) Benzo(g,h,i)perylene	26.082	276	2705980	53.15	ng	#	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1
1,4-Dichlorobenzene-d4
Concen: 20.00 ng
RT: 7.613 min Scan# 812
Delta R.T. -0.006 min
Lab File: BM032406.D
Acq: 08 Oct 2021 22:58





Tgt Ion: 152 Resp: 361321

Ion	Ratio	Lower	Upper
152	100		
150	156.2	122.4	183.6
115	59.4	53.8	80.8

Instrument :
BNA_M
ClientSampleId :
PSC124027MSD

Manual Integrations
APPROVED




mohammad
10/11/2021 2:13:34 PM




#2
1,4-Dioxane
Concen: 36.75 ng
RT: 2.996 min Scan# 27
Delta R.T. -0.000 min
Lab File: BM032406.D
Acq: 08 Oct 2021 22:58

Tgt Ion: 88 Resp: 335360

Ion	Ratio	Lower	Upper
88	100		
58	72.4	50.4	75.6
43	25.7	22.2	33.2









#3
Pyridine
Concen: 32.79 ng
RT: 3.401 min Scan# 96
Delta R.T. -0.000 min
Lab File: BM032406.D
Acq: 08 Oct 2021 22:58

Tgt Ion: 79 Resp: 809840
Ion Ratio Lower Upper
79 100
52 66.8 48.6 73.0
51 33.4 27.4 41.2

Instrument :
BNA_M
ClientSampleId :
PSC124027MSD




Manual Integrations
APPROVED


mohammad
10/11/2021 2:13:34 PM



#4
n-Nitrosodimethylamine
Concen: 43.24 ng
RT: 3.319 min Scan# 82
Delta R.T. -0.000 min
Lab File: BM032406.D
Acq: 08 Oct 2021 22:58

Tgt Ion: 42 Resp: 428283
Ion Ratio Lower Upper
42 100
74 147.8 99.5 149.3
44 6.8 57.3 85.9









#5
2-Fluorophenol
Concen: 117.71 ng
RT: 5.190 min Scan# 400
Delta R.T. -0.000 min
Lab File: BM032406.D
Acq: 08 Oct 2021 22:58

Tgt Ion: 112 Resp: 2524238
Ion Ratio Lower Upper
112 100
64 56.1 53.7 80.5
63 29.4 33.9 50.9#

Instrument :
BNA_M
ClientSampleId :
PSC124027MSD




Manual Integrations
APPROVED


mohammad
10/11/2021 2:13:34 PM



#6
Aniline
Concen: 40.98 ng
RT: 6.942 min Scan# 698
Delta R.T. -0.000 min
Lab File: BM032406.D
Acq: 08 Oct 2021 22:58

Tgt Ion: 93 Resp: 1327871
Ion Ratio Lower Upper
93 100
66 37.2 35.3 52.9
65 19.0 18.7 28.1





#7
Phenol-d6
Concen: 108.85 ng
RT: 6.819 min Scan# 677
Delta R.T. 0.006 min
Lab File: BM032406.D
Acq: 08 Oct 2021 22:58





Instrument :
BNA_M
ClientSampleId :
PSC124027MSD

Tgt Ion: 99 Resp: 3192102

Ion	Ratio	Lower	Upper
99	100		
42	15.7	23.0	34.6#
71	33.6	37.8	56.8#

Manual Integrations
APPROVED




mohammad
10/11/2021 2:13:34 PM




#8
2-Chlorophenol
Concen: 47.99 ng
RT: 7.189 min Scan# 740
Delta R.T. -0.000 min
Lab File: BM032406.D
Acq: 08 Oct 2021 22:58

Tgt Ion: 128 Resp: 1121599


Ion	Ratio	Lower	Upper
128	100		
130	32.2	13.0	53.0
64	43.5	33.4	73.4





#9
Benzaldehyde
Concen: 37.36 ng
RT: 6.742 min Scan# 664
Delta R.T. -0.000 min
Lab File: BM032406.D
Acq: 08 Oct 2021 22:58



Instrument :
BNA_M
ClientSampleId :
PSC124027MSD




Tgt Ion: 77 Resp: 639409
Ion Ratio Lower Upper
77 100
105 102.5 68.9 108.9
106 98.0 61.2 101.2

Manual Integrations
APPROVED



mohammad
10/11/2021 2:13:34 PM




#10
Phenol
Concen: 47.71 ng
RT: 6.842 min Scan# 681
Delta R.T. 0.006 min
Lab File: BM032406.D
Acq: 08 Oct 2021 22:58



Tgt Ion: 94 Resp: 1347250
Ion Ratio Lower Upper
94 100
65 27.4 16.9 56.9
66 37.0 30.0 70.0





#11
bis(2-Chloroethyl)ether
Concen: 45.61 ng
RT: 7.036 min Scan# 714
Delta R.T. -0.000 min
Lab File: BM032406.D
Acq: 08 Oct 2021 22:58





Tgt Ion: 93 Resp: 1024205

Ion	Ratio	Lower	Upper
93	100		
63	74.0	63.7	103.7
95	32.3	12.4	52.4

Instrument :
BNA_M
ClientSampleId :
PSC124027MSD

Manual Integrations
APPROVED




mohammad
10/11/2021 2:13:34 PM




#12
1,3-Dichlorobenzene
Concen: 44.97 ng
RT: 7.507 min Scan# 794
Delta R.T. -0.006 min
Lab File: BM032406.D
Acq: 08 Oct 2021 22:58

Tgt Ion: 146 Resp: 1184742

Ion	Ratio	Lower	Upper
146	100		
148	63.6	52.3	78.5
75	29.2	34.2	51.2









#13
1,4-Dichlorobenzene
Concen: 45.06 ng
RT: 7.648 min Scan# 818
Delta R.T. -0.006 min
Lab File: BM032406.D
Acq: 08 Oct 2021 22:58

Tgt Ion:146 Resp: 1208133
Ion Ratio Lower Upper
146 100
148 64.4 51.1 76.7
111 40.7 35.3 52.9

Instrument :
BNA_M
ClientSampleId :
PSC124027MSD




Manual Integrations
APPROVED


mohammad
10/11/2021 2:13:34 PM



#14
1,2-Dichlorobenzene
Concen: 45.13 ng
RT: 7.972 min Scan# 873
Delta R.T. -0.006 min
Lab File: BM032406.D
Acq: 08 Oct 2021 22:58


Tgt Ion:146 Resp: 1161497
Ion Ratio Lower Upper
146 100
148 64.2 51.7 77.5
111 43.6 38.2 57.2





#15
Benzyl Alcohol
Concen: 47.28 ng
RT: 7.860 min Scan# 854
Delta R.T. -0.000 min
Lab File: BM032406.D
Acq: 08 Oct 2021 22:58




Instrument :
BNA_M
ClientSampleId :
PSC124027MSD




Tgt Ion: 79 Resp: 949529
Ion Ratio Lower Upper
79 100
108 78.9 53.6 80.4
77 62.4 57.7 86.5

Manual Integrations
APPROVED



mohammad
10/11/2021 2:13:34 PM




#16
2,2'-oxybis(1-Chloropropane)
Concen: 44.06 ng
RT: 8.148 min Scan# 903
Delta R.T. -0.006 min
Lab File: BM032406.D
Acq: 08 Oct 2021 22:58




Tgt Ion: 45 Resp: 1253445
Ion Ratio Lower Upper
45 100
77 16.0 0.0 36.1
79 12.4 0.0 30.5





#17
2-Methylphenol
Concen: 47.65 ng
RT: 8.083 min Scan# 892
Delta R.T. -0.000 min
Lab File: BM032406.D
Acq: 08 Oct 2021 22:58




Instrument :
BNA_M
ClientSampleId :
PSC124027MSD




Tgt Ion:	107	Resp:	909205
Ion	Ratio	Lower	Upper
107	100		
108	112.7	87.9	131.9
77	43.8	48.2	72.2#
79	43.5	44.7	67.1#

Manual Integrations
APPROVED



mohammad
10/11/2021 2:13:34 PM




#18
Hexachloroethane
Concen: 47.68 ng
RT: 8.701 min Scan# 997
Delta R.T. -0.000 min
Lab File: BM032406.D
Acq: 08 Oct 2021 22:58



Tgt Ion:	117	Resp:	435240
Ion	Ratio	Lower	Upper
117	100		
119	97.8	77.7	116.5
201	104.5	70.2	105.2









#19
n-Nitroso-di-n-propylamine
Concen: 44.94 ng
RT: 8.425 min Scan# 950
Delta R.T. -0.000 min
Lab File: BM032406.D
Acq: 08 Oct 2021 22:58

Tgt Ion:	70	Resp:	716524
Ion	Ratio	Lower	Upper
70	100		
42	52.2	56.3	84.5#
101	10.8	7.8	11.6
130	22.4	14.2	21.4#

Instrument :
BNA_M
ClientSampleId :
PSC124027MSD




Manual Integrations
APPROVED


mohammad
10/11/2021 2:13:34 PM



#20
3+4-Methylphenols
Concen: 46.20 ng
RT: 8.413 min Scan# 948
Delta R.T. -0.000 min
Lab File: BM032406.D
Acq: 08 Oct 2021 22:58


Tgt Ion:	107	Resp:	1189413
Ion	Ratio	Lower	Upper
107	100		
108	90.2	67.6	107.6
77	44.1	21.9	61.9
79	28.9	15.3	55.3





#21
Naphthalene-d8
Concen: 20.00 ng
RT: 10.407 min Scan# 1287
Delta R.T. -0.000 min
Lab File: BM032406.D
Acq: 08 Oct 2021 22:58




Instrument :
BNA_M
ClientSampleId :
PSC124027MSD




Tgt Ion:136 Resp: 1440698
Ion Ratio Lower Upper
136 100
137 10.9 8.7 13.1
54 7.4 7.4 11.0
68 5.8 3.8 5.8#

Manual Integrations
APPROVED



mohammad
10/11/2021 2:13:34 PM




#22
Acetophenone
Concen: 43.63 ng
RT: 8.436 min Scan# 952
Delta R.T. -0.000 min
Lab File: BM032406.D
Acq: 08 Oct 2021 22:58



Tgt Ion:105 Resp: 1514784
Ion Ratio Lower Upper
105 100
71 4.4 3.8 5.6
51 25.3 29.5 44.3#
120 22.6 16.3 24.5









#23
Nitrobenzene-d5
Concen: 74.53 ng
RT: 8.772 min Scan# 1009
Delta R.T. -0.000 min
Lab File: BM032406.D
Acq: 08 Oct 2021 22:58

Tgt Ion: 82 Resp: 1930505
Ion Ratio Lower Upper
82 100
128 45.6 29.0 43.6#
54 48.7 45.8 68.8

Instrument :
BNA_M
ClientSampleId :
PSC124027MSD




Manual Integrations
APPROVED


mohammad
10/11/2021 2:13:34 PM



#24
Nitrobenzene
Concen: 45.53 ng
RT: 8.813 min Scan# 1016
Delta R.T. -0.000 min
Lab File: BM032406.D
Acq: 08 Oct 2021 22:58

Tgt Ion: 77 Resp: 1137484
Ion Ratio Lower Upper
77 100
123 47.7 30.1 45.1#
65 13.5 11.7 17.5









#25
Isophorone
Concen: 46.01 ng
RT: 9.330 min Scan# 1104
Delta R.T. -0.006 min
Lab File: BM032406.D
Acq: 08 Oct 2021 22:58

Tgt Ion: 82 Resp: 1980185
Ion Ratio Lower Upper
82 100
95 7.5 6.4 9.6
138 18.0 12.4 18.6

Instrument :
BNA_M
ClientSampleId :
PSC124027MSD




Manual Integrations
APPROVED


mohammad
10/11/2021 2:13:34 PM



#26
2-Nitrophenol
Concen: 50.10 ng
RT: 9.525 min Scan# 1137
Delta R.T. -0.000 min
Lab File: BM032406.D
Acq: 08 Oct 2021 22:58


Tgt Ion: 139 Resp: 582411
Ion Ratio Lower Upper
139 100
109 31.1 34.8 52.2#
65 47.8 59.9 89.9#





#27
2,4-Dimethylphenol
Concen: 51.18 ng
RT: 9.601 min Scan# 1150
Delta R.T. -0.000 min
Lab File: BM032406.D
Acq: 08 Oct 2021 22:58




Instrument :
BNA_M
ClientSampleId :
PSC124027MSD




Tgt Ion:122 Resp: 1003730
Ion Ratio Lower Upper
122 100
107 117.2 100.6 150.8
121 58.6 47.3 70.9

Manual Integrations
APPROVED



mohammad
10/11/2021 2:13:34 PM




#28
bis(2-Chloroethoxy)methane
Concen: 41.77 ng
RT: 9.819 min Scan# 1187
Delta R.T. -0.000 min
Lab File: BM032406.D
Acq: 08 Oct 2021 22:58




Tgt Ion: 93 Resp: 1292686
Ion Ratio Lower Upper
93 100
95 32.2 25.8 38.8
123 11.3 8.8 13.2





#29
2,4-Dichlorophenol
Concen: 46.00 ng
RT: 10.066 min Scan# 1229
Delta R.T. -0.000 min
Lab File: BM032406.D
Acq: 08 Oct 2021 22:58




Instrument :
BNA_M
ClientSampleId :
PSC124027MSD




Tgt Ion:162 Resp: 1004665
Ion Ratio Lower Upper
162 100
164 64.0 47.4 87.4
98 36.2 18.9 58.9

Manual Integrations
APPROVED



mohammad
10/11/2021 2:13:34 PM




#30
1,2,4-Trichlorobenzene
Concen: 42.69 ng
RT: 10.277 min Scan# 1265
Delta R.T. -0.000 min
Lab File: BM032406.D
Acq: 08 Oct 2021 22:58



Tgt Ion:180 Resp: 1057601
Ion Ratio Lower Upper
180 100
182 96.6 75.5 113.3
145 30.5 26.2 39.2









#31
Naphthalene
Concen: 43.56 ng
RT: 10.454 min Scan# 1295
Delta R.T. -0.006 min
Lab File: BM032406.D
Acq: 08 Oct 2021 22:58

Tgt Ion:128 Resp: 3220136
Ion Ratio Lower Upper
128 100
129 11.0 8.6 13.0
127 13.2 10.4 15.6

Instrument :
BNA_M
ClientSampleId :
PSC124027MSD




Manual Integrations
APPROVED


mohammad
10/11/2021 2:13:34 PM



#32
Benzoic acid
Concen: 42.32 ng
RT: 9.783 min Scan# 1181
Delta R.T. 0.012 min
Lab File: BM032406.D
Acq: 08 Oct 2021 22:58

Tgt Ion:122 Resp: 617567
Ion Ratio Lower Upper
122 100
105 128.1 120.3 160.3
77 92.9 100.0 140.0#









#33
4-Chloroaniline
Concen: 30.25 ng
RT: 10.560 min Scan# 1313
Delta R.T. -0.006 min
Lab File: BM032406.D
Acq: 08 Oct 2021 22:58

Tgt Ion:	127	Resp:	923211
Ion	Ratio	Lower	Upper
127	100		
129	32.4	25.6	38.4
65	30.4	30.4	45.6
92	20.1	16.2	24.4

Instrument :
BNA_M
ClientSampleId :
PSC124027MSD




Manual Integrations
APPROVED


mohammad
10/11/2021 2:13:34 PM



#34
Hexachlorobutadiene
Concen: 42.79 ng
RT: 10.760 min Scan# 1347
Delta R.T. -0.006 min
Lab File: BM032406.D
Acq: 08 Oct 2021 22:58


Tgt Ion:	225	Resp:	633517
Ion	Ratio	Lower	Upper
225	100		
223	63.3	52.6	78.8
227	65.3	51.2	76.8








#35
Caprolactam
Concen: 45.18 ng
RT: 11.336 min Scan# 1445
Delta R.T. 0.012 min
Lab File: BM032406.D
Acq: 08 Oct 2021 22:58


Instrument :
BNA_M
ClientSampleId :
PSC124027MSD





Tgt Ion:113 Resp: 302392
Ion Ratio Lower Upper
113 100
55 152.9 170.5 210.5#
56 116.6 125.2 165.2#




#36
4-Chloro-3-methylphenol
Concen: 43.94 ng
RT: 11.713 min Scan# 1509
Delta R.T. 0.006 min
Lab File: BM032406.D
Acq: 08 Oct 2021 22:58



Tgt Ion:107 Resp: 1033363
Ion Ratio Lower Upper
107 100
144 25.6 19.3 28.9
142 79.5 60.5 90.7









#37
2-Methylnaphthalene
Concen: 44.33 ng m
RT: 12.071 min Scan# 1570
Delta R.T. -0.000 min
Lab File: BM032406.D
Acq: 08 Oct 2021 22:58

Instrument :
BNA_M
ClientSampleId :
PSC124027MSD

Tgt Ion:142 Resp: 2227410
Ion Ratio Lower Upper
142 100
141 89.0 71.8 107.6
115 35.5 36.3 54.5#




Manual Integrations
APPROVED


mohammad
10/11/2021 2:13:34 PM



#38
1-Methylnaphthalene
Concen: 41.86 ng
RT: 12.295 min Scan# 1608
Delta R.T. -0.000 min
Lab File: BM032406.D
Acq: 08 Oct 2021 22:58


Tgt Ion:142 Resp: 2055112
Ion Ratio Lower Upper
142 100
141 92.0 75.6 113.4
116 4.0 4.0 6.0#





#39
Acenaphthene-d10
Concen: 20.00 ng
RT: 14.265 min Scan# 1943
Delta R.T. -0.000 min
Lab File: BM032406.D
Acq: 08 Oct 2021 22:58




Instrument :
BNA_M
ClientSampleId :
PSC124027MSD




Tgt Ion:164 Resp: 817484
Ion Ratio Lower Upper
164 100
162 100.4 80.6 121.0
160 44.0 35.1 52.7

Manual Integrations
APPROVED



mohammad
10/11/2021 2:13:34 PM




#40
1,2,4,5-Tetrachlorobenzene
Concen: 44.53 ng
RT: 12.454 min Scan# 1635
Delta R.T. -0.000 min
Lab File: BM032406.D
Acq: 08 Oct 2021 22:58



Tgt Ion:216 Resp: 1048676
Ion Ratio Lower Upper
216 100
214 78.3 62.5 93.7
179 21.2 18.7 28.1
108 20.3 21.0 31.4#









#41
Hexachlorocyclopentadiene
Concen: 76.47 ng
RT: 12.442 min Scan# 1633
Delta R.T. -0.006 min
Lab File: BM032406.D
Acq: 08 Oct 2021 22:58

Tgt Ion:237 Resp: 858440
Ion Ratio Lower Upper
237 100
235 62.1 44.4 84.4
272 12.1 0.0 31.4

Instrument :
BNA_M
ClientSampleId :
PSC124027MSD




Manual Integrations
APPROVED


mohammad
10/11/2021 2:13:34 PM




#42
2,4,6-Tribromophenol
Concen: 97.72 ng
RT: 15.754 min Scan# 2196
Delta R.T. -0.000 min
Lab File: BM032406.D
Acq: 08 Oct 2021 22:58

Tgt Ion:330 Resp: 894303
Ion Ratio Lower Upper
330 100
332 96.3 75.7 113.5
141 37.8 50.7 76.1#









#43
2,4,6-Trichlorophenol
Concen: 45.89 ng
RT: 12.695 min Scan# 1676
Delta R.T. -0.000 min
Lab File: BM032406.D
Acq: 08 Oct 2021 22:58





Tgt Ion:196 Resp: 747716
Ion Ratio Lower Upper
196 100
198 94.3 73.6 110.4
200 30.3 22.1 33.1



#44
2,4,5-Trichlorophenol
Concen: 45.53 ng
RT: 12.771 min Scan# 1689
Delta R.T. -0.000 min
Lab File: BM032406.D
Acq: 08 Oct 2021 22:58




Tgt Ion:196 Resp: 797435
Ion Ratio Lower Upper
196 100
198 96.5 77.0 115.4
97 47.9 53.9 80.9#
132 27.2 23.8 35.8



Instrument :
BNA_M
ClientSampleId :
PSC124027MSD


Manual Integrations
APPROVED

mohammad
10/11/2021 2:13:34 PM



#45
2-Fluorobiphenyl
Concen: 54.85 ng
RT: 12.883 min Scan# 1708
Delta R.T. -0.000 min
Lab File: BM032406.D
Acq: 08 Oct 2021 22:58




Instrument :
BNA_M
ClientSampleId :
PSC124027MSD



Tgt Ion:172 Resp: 3015880
Ion Ratio Lower Upper
172 100
171 35.6 28.1 42.1
170 23.7 18.6 27.8




Manual Integrations
APPROVED


mohammad
10/11/2021 2:13:34 PM



#46
1,1'-Biphenyl
Concen: 43.52 ng
RT: 13.095 min Scan# 1744
Delta R.T. -0.000 min
Lab File: BM032406.D
Acq: 08 Oct 2021 22:58

Tgt Ion:154 Resp: 2625646
Ion Ratio Lower Upper
154 100
153 40.1 20.3 60.3
76 12.3 0.0 33.5





#47
2-Chloronaphthalene
Concen: 44.91 ng
RT: 13.130 min Scan# 1750
Delta R.T. -0.006 min
Lab File: BM032406.D
Acq: 08 Oct 2021 22:58





Instrument :
BNA_M
ClientSampleId :
PSC124027MSD

Tgt Ion:162 Resp: 2077783

Ion	Ratio	Lower	Upper
162	100		
127	40.6	33.2	49.8
164	32.8	25.7	38.5

Manual Integrations
APPROVED




mohammad
10/11/2021 2:13:34 PM




#48
2-Nitroaniline
Concen: 49.50 ng
RT: 13.336 min Scan# 1785
Delta R.T. -0.000 min
Lab File: BM032406.D
Acq: 08 Oct 2021 22:58

Tgt Ion: 65 Resp: 622630

Ion	Ratio	Lower	Upper
65	100		
92	71.7	51.2	76.8
138	112.4	66.8	100.2









#49
Acenaphthylene
Concen: 45.86 ng
RT: 13.983 min Scan# 1895
Delta R.T. -0.000 min
Lab File: BM032406.D
Acq: 08 Oct 2021 22:58

Tgt Ion:152 Resp: 3293904
Ion Ratio Lower Upper
152 100
151 20.5 15.7 23.5
153 13.3 10.6 15.8

Instrument :
BNA_M
ClientSampleId :
PSC124027MSD




Manual Integrations
APPROVED


mohammad
10/11/2021 2:13:34 PM



#50
Dimethylphthalate
Concen: 43.39 ng
RT: 13.724 min Scan# 1851
Delta R.T. -0.000 min
Lab File: BM032406.D
Acq: 08 Oct 2021 22:58

Tgt Ion:163 Resp: 2504639
Ion Ratio Lower Upper
163 100
194 4.1 2.9 4.3
164 9.9 8.0 12.0









#51
2,6-Dinitrotoluene
Concen: 47.41 ng
RT: 13.830 min Scan# 1869
Delta R.T. -0.000 min
Lab File: BM032406.D
Acq: 08 Oct 2021 22:58

Tgt Ion:165 Resp: 554349
Ion Ratio Lower Upper
165 100
63 54.1 72.4 108.6#
89 56.8 69.3 103.9#

Instrument :
BNA_M
ClientSampleId :
PSC124027MSD




Manual Integrations
APPROVED


mohammad
10/11/2021 2:13:34 PM



#52
Acenaphthene
Concen: 44.75 ng m
RT: 14.330 min Scan# 1954
Delta R.T. -0.000 min
Lab File: BM032406.D
Acq: 08 Oct 2021 22:58


Tgt Ion:154 Resp: 2120406
Ion Ratio Lower Upper
154 100
153 112.0 92.4 138.6
152 53.8 43.7 65.5





#53
3-Nitroaniline
Concen: 33.66 ng
RT: 14.165 min Scan# 1926
Delta R.T. -0.000 min
Lab File: BM032406.D
Acq: 08 Oct 2021 22:58




Instrument :
BNA_M
ClientSampleId :
PSC124027MSD




Tgt Ion:138 Resp: 442248
Ion Ratio Lower Upper
138 100
108 11.1 8.3 12.5
92 114.5 105.2 157.8

Manual Integrations
APPROVED



mohammad
10/11/2021 2:13:34 PM




#54
2,4-Dinitrophenol
Concen: 50.33 ng
RT: 14.365 min Scan# 1960
Delta R.T. -0.000 min
Lab File: BM032406.D
Acq: 08 Oct 2021 22:58



Tgt Ion:184 Resp: 326584
Ion Ratio Lower Upper
184 100
63 62.2 93.8 140.6#
154 66.8 64.0 96.0









#55
Dibenzenofuran
Concen: 41.64 ng
RT: 14.660 min Scan# 2010
Delta R.T. -0.006 min
Lab File: BM032406.D
Acq: 08 Oct 2021 22:58

Tgt Ion:168 Resp: 3003573
Ion Ratio Lower Upper
168 100
139 38.3 35.0 52.4
169 13.7 10.8 16.2

Instrument :
BNA_M
ClientSampleId :
PSC124027MSD




Manual Integrations
APPROVED


mohammad
10/11/2021 2:13:34 PM



#56
4-Nitrophenol
Concen: 83.63 ng
RT: 14.495 min Scan# 1982
Delta R.T. 0.006 min
Lab File: BM032406.D
Acq: 08 Oct 2021 22:58


Tgt Ion:139 Resp: 906231
Ion Ratio Lower Upper
139 100
109 75.2 92.3 132.3#
65 93.0 110.5 150.5#





#57
2,4-Dinitrotoluene
Concen: 47.42 ng
RT: 14.624 min Scan# 2004
Delta R.T. 0.006 min
Lab File: BM032406.D
Acq: 08 Oct 2021 22:58




Instrument :
BNA_M
ClientSampleId :
PSC124027MSD




Tgt Ion:165 Resp: 731016
Ion Ratio Lower Upper
165 100
63 40.0 70.3 105.5#
89 68.4 83.1 124.7#
182 2.8 2.2 3.4

Manual Integrations
APPROVED



mohammad
10/11/2021 2:13:34 PM




#58
Fluorene
Concen: 40.18 ng
RT: 15.307 min Scan# 2120
Delta R.T. -0.006 min
Lab File: BM032406.D
Acq: 08 Oct 2021 22:58



Tgt Ion:166 Resp: 2204307
Ion Ratio Lower Upper
166 100
165 97.9 79.3 118.9
167 13.5 10.6 16.0









#59
2,3,4,6-Tetrachlorophenol
Concen: 41.93 ng m
RT: 14.895 min Scan# 2050
Delta R.T. -0.000 min
Lab File: BM032406.D
Acq: 08 Oct 2021 22:58

Tgt Ion:232 Resp: 632594
Ion Ratio Lower Upper
232 100
131 45.6 50.6 75.8#
130 2.4 3.0 4.4#
166 31.5 28.8 43.2

Instrument :
BNA_M
ClientSampleId :
PSC124027MSD




Manual Integrations
APPROVED


mohammad
10/11/2021 2:13:34 PM



#60
Diethylphthalate
Concen: 43.14 ng
RT: 15.083 min Scan# 2082
Delta R.T. -0.006 min
Lab File: BM032406.D
Acq: 08 Oct 2021 22:58

Tgt Ion:149 Resp: 2444741
Ion Ratio Lower Upper
149 100
177 22.1 15.7 23.5
150 12.8 9.6 14.4









#61
4-Chlorophenyl-phenylether
Concen: 38.88 ng
RT: 15.301 min Scan# 2119
Delta R.T. -0.000 min
Lab File: BM032406.D
Acq: 08 Oct 2021 22:58

Tgt Ion:204 Resp: 1104689
Ion Ratio Lower Upper
204 100
206 32.6 26.7 40.1
141 61.8 55.8 83.8

Instrument :
BNA_M
ClientSampleId :
PSC124027MSD




Manual Integrations
APPROVED


mohammad
10/11/2021 2:13:34 PM



#62
4-Nitroaniline
Concen: 40.20 ng
RT: 15.330 min Scan# 2124
Delta R.T. -0.000 min
Lab File: BM032406.D
Acq: 08 Oct 2021 22:58


Tgt Ion:138 Resp: 527134
Ion Ratio Lower Upper
138 100
92 49.1 40.6 80.6
108 78.4 103.8 143.8#





#63
Azobenzene
Concen: 42.43 ng
RT: 15.595 min Scan# 2169
Delta R.T. -0.000 min
Lab File: BM032406.D
Acq: 08 Oct 2021 22:58




Instrument :
BNA_M
ClientSampleId :
PSC124027MSD




Tgt Ion: 77 Resp: 2289316
Ion Ratio Lower Upper
77 100
182 27.4 0.0 39.3
105 22.8 0.0 36.1
51 28.4 19.1 59.1

Manual Integrations
APPROVED



mohammad
10/11/2021 2:13:34 PM




#64
Phenanthrene-d10
Concen: 20.00 ng
RT: 16.989 min Scan# 2406
Delta R.T. -0.006 min
Lab File: BM032406.D
Acq: 08 Oct 2021 22:58



Tgt Ion: 188 Resp: 1457315
Ion Ratio Lower Upper
188 100
94 9.9 6.3 9.5#
80 10.3 7.3 10.9









#65
4,6-Dinitro-2-methylphenol
Concen: 31.61 ng
RT: 15.389 min Scan# 2134
Delta R.T. -0.000 min
Lab File: BM032406.D
Acq: 08 Oct 2021 22:58

Tgt Ion:198 Resp: 259321
Ion Ratio Lower Upper
198 100
51 40.2 49.7 89.7#
105 42.2 34.9 74.9

Instrument :
BNA_M
ClientSampleId :
PSC124027MSD




Manual Integrations
APPROVED


mohammad
10/11/2021 2:13:34 PM




#66
n-Nitrosodiphenylamine
Concen: 47.73 ng
RT: 15.518 min Scan# 2156
Delta R.T. -0.000 min
Lab File: BM032406.D
Acq: 08 Oct 2021 22:58

Tgt Ion:169 Resp: 2053775
Ion Ratio Lower Upper
169 100
168 66.3 55.0 82.4
167 36.4 29.8 44.8









#67
4-Bromophenyl-phenylether
Concen: 45.93 ng
RT: 16.189 min Scan# 2270
Delta R.T. -0.000 min
Lab File: BM032406.D
Acq: 08 Oct 2021 22:58





Tgt Ion:248 Resp: 696696
Ion Ratio Lower Upper
248 100
250 98.8 77.4 116.2
141 75.9 74.1 111.1



#68
Hexachlorobenzene
Concen: 45.20 ng
RT: 16.324 min Scan# 2293
Delta R.T. -0.000 min
Lab File: BM032406.D
Acq: 08 Oct 2021 22:58




Tgt Ion:284 Resp: 754625
Ion Ratio Lower Upper
284 100
142 36.2 43.0 64.6#
249 32.5 27.6 41.4



Instrument :
BNA_M
ClientSampleId :
PSC124027MSD

Manual Integrations
APPROVED

mohammad
10/11/2021 2:13:34 PM







#69
Atrazine
Concen: 47.80 ng
RT: 16.459 min Scan# 2316
Delta R.T. -0.000 min
Lab File: BM032406.D
Acq: 08 Oct 2021 22:58

Tgt Ion:200 Resp: 700411
Ion Ratio Lower Upper
200 100
173 26.4 5.8 45.8
215 47.5 23.6 63.6

Instrument :
BNA_M
ClientSampleId :
PSC124027MSD




Manual Integrations
APPROVED


mohammad
10/11/2021 2:13:34 PM



#70
Pentachlorophenol
Concen: 83.57 ng
RT: 16.659 min Scan# 2350
Delta R.T. -0.000 min
Lab File: BM032406.D
Acq: 08 Oct 2021 22:58


Tgt Ion:266 Resp: 861216
Ion Ratio Lower Upper
266 100
268 62.7 50.3 75.5
264 62.5 52.1 78.1





#71
Phenanthrene
Concen: 43.83 ng
RT: 17.036 min Scan# 2414
Delta R.T. -0.000 min
Lab File: BM032406.D
Acq: 08 Oct 2021 22:58




Instrument :
BNA_M
ClientSampleId :
PSC124027MSD




Tgt Ion:178 Resp: 3352784
Ion Ratio Lower Upper
178 100
176 19.6 15.6 23.4
179 15.3 12.4 18.6

Manual Integrations
APPROVED



mohammad
10/11/2021 2:13:34 PM




#72
Anthracene
Concen: 45.28 ng
RT: 17.124 min Scan# 2429
Delta R.T. -0.000 min
Lab File: BM032406.D
Acq: 08 Oct 2021 22:58




Tgt Ion:178 Resp: 3384237
Ion Ratio Lower Upper
178 100
176 18.8 14.8 22.2
179 15.6 12.3 18.5





#73
 Carbazole
 Concen: 40.17 ng
 RT: 17.389 min Scan# 2474
 Delta R.T. -0.000 min
 Lab File: BM032406.D
 Acq: 08 Oct 2021 22:58




Instrument :
 BNA_M
 ClientSampleId :
 PSC124027MSD




Tgt Ion:167 Resp: 2981417
 Ion Ratio Lower Upper
 167 100
 166 21.6 17.4 26.0
 139 13.4 11.9 17.9

Manual Integrations
 APPROVED



mohammad
 10/11/2021 2:13:34 PM




#74
 Di-n-butylphthalate
 Concen: 46.98 ng
 RT: 17.948 min Scan# 2569
 Delta R.T. -0.000 min
 Lab File: BM032406.D
 Acq: 08 Oct 2021 22:58



Tgt Ion:149 Resp: 3745677
 Ion Ratio Lower Upper
 149 100
 150 9.2 7.3 10.9
 104 5.0 5.6 8.4#









#75
Fluoranthene
Concen: 39.20 ng
RT: 19.042 min Scan# 2755
Delta R.T. -0.000 min
Lab File: BM032406.D
Acq: 08 Oct 2021 22:58

Tgt Ion:202 Resp: 3324961
Ion Ratio Lower Upper
202 100
101 11.5 0.0 29.9
203 17.8 0.0 36.8

Instrument :
BNA_M
ClientSampleId :
PSC124027MSD




Manual Integrations
APPROVED


mohammad
10/11/2021 2:13:34 PM



#76
Chrysene-d12
Concen: 20.00 ng
RT: 21.153 min Scan# 3114
Delta R.T. -0.006 min
Lab File: BM032406.D
Acq: 08 Oct 2021 22:58

Tgt Ion:240 Resp: 952815
Ion Ratio Lower Upper
240 100
120 12.0 7.8 11.6#
236 25.9 20.9 31.3









#77
Benzidine
Concen: 96.67 ng
RT: 19.224 min Scan# 2786
Delta R.T. -0.000 min
Lab File: BM032406.D
Acq: 08 Oct 2021 22:58

Tgt Ion:184 Resp: 2288064
Ion Ratio Lower Upper
184 100
185 14.1 10.6 16.0
183 11.8 9.0 13.4

Instrument :
BNA_M
ClientSampleId :
PSC124027MSD




Manual Integrations
APPROVED


mohammad
10/11/2021 2:13:34 PM



#78
Pyrene
Concen: 50.82 ng
RT: 19.400 min Scan# 2816
Delta R.T. -0.006 min
Lab File: BM032406.D
Acq: 08 Oct 2021 22:58


Tgt Ion:202 Resp: 3283407
Ion Ratio Lower Upper
202 100
200 21.3 16.6 25.0
203 18.0 13.7 20.5





#79
Terphenyl-d14
Concen: 59.70 ng
RT: 19.600 min Scan# 2850
Delta R.T. -0.006 min
Lab File: BM032406.D
Acq: 08 Oct 2021 22:58




Instrument :
BNA_M
ClientSampleId :
PSC124027MSD




Tgt Ion:244 Resp: 2700803
Ion Ratio Lower Upper
244 100
212 7.5 6.5 9.7
122 12.1 8.7 13.1

Manual Integrations
APPROVED



mohammad
10/11/2021 2:13:34 PM




#80
Butylbenzylphthalate
Concen: 54.98 ng
RT: 20.294 min Scan# 2968
Delta R.T. -0.000 min
Lab File: BM032406.D
Acq: 08 Oct 2021 22:58




Tgt Ion:149 Resp: 1349333
Ion Ratio Lower Upper
149 100
91 67.9 64.0 96.0
206 21.6 14.4 21.6





#81
Benzo(a)anthracene
Concen: 44.05 ng
RT: 21.141 min Scan# 3112
Delta R.T. -0.000 min
Lab File: BM032406.D
Acq: 08 Oct 2021 22:58




Instrument :
BNA_M
ClientSampleId :
PSC124027MSD




Tgt Ion:228 Resp: 2623294
Ion Ratio Lower Upper
228 100
226 27.9 21.4 32.2
229 20.0 15.3 22.9

Manual Integrations
APPROVED



mohammad
10/11/2021 2:13:34 PM




#82
3,3'-Dichlorobenzidine
Concen: 50.56 ng
RT: 21.077 min Scan# 3101
Delta R.T. -0.000 min
Lab File: BM032406.D
Acq: 08 Oct 2021 22:58




Tgt Ion:252 Resp: 948637
Ion Ratio Lower Upper
252 100
254 63.9 51.6 77.4
126 10.9 9.7 14.5





#83
Chrysene
Concen: 44.98 ng
RT: 21.194 min Scan# 3121
Delta R.T. -0.000 min
Lab File: BM032406.D
Acq: 08 Oct 2021 22:58




Instrument :
BNA_M
ClientSampleId :
PSC124027MSD




Tgt Ion:228 Resp: 2591366
Ion Ratio Lower Upper
228 100
226 30.8 24.0 36.0
229 20.0 16.0 24.0

Manual Integrations
APPROVED



mohammad
10/11/2021 2:13:34 PM




#84
Bis(2-ethylhexyl)phthalate
Concen: 56.85 ng
RT: 21.065 min Scan# 3099
Delta R.T. -0.000 min
Lab File: BM032406.D
Acq: 08 Oct 2021 22:58




Tgt Ion:149 Resp: 1881064
Ion Ratio Lower Upper
149 100
167 28.2 19.7 29.5
279 4.3 2.6 4.0#





#85
Di-n-octyl phthalate
Concen: 52.71 ng
RT: 21.906 min Scan# 3242
Delta R.T. -0.000 min
Lab File: BM032406.D
Acq: 08 Oct 2021 22:58




Instrument :
BNA_M
ClientSampleId :
PSC124027MSD




Tgt Ion:149 Resp: 2924480
Ion Ratio Lower Upper
149 100
167 1.6 1.2 1.8
43 8.2 10.4 15.6#

Manual Integrations
APPROVED



mohammad
10/11/2021 2:13:34 PM




#86
Perylene-d12
Concen: 20.00 ng
RT: 23.300 min Scan# 3479
Delta R.T. -0.000 min
Lab File: BM032406.D
Acq: 08 Oct 2021 22:58




Tgt Ion:264 Resp: 991354
Ion Ratio Lower Upper
264 100
260 23.9 19.3 28.9
265 22.7 18.3 27.5





#87
Indeno(1,2,3-cd)pyrene
Concen: 50.08 ng
RT: 25.429 min Scan# 3841
Delta R.T. -0.000 min
Lab File: BM032406.D
Acq: 08 Oct 2021 22:58




Instrument :
BNA_M
ClientSampleId :
PSC124027MSD




Tgt Ion:276 Resp: 3059053
Ion Ratio Lower Upper
276 100
138 26.5 18.4 27.6
277 25.6 20.1 30.1

Manual Integrations
APPROVED



mohammad
10/11/2021 2:13:34 PM




#88
Benzo(b)fluoranthene
Concen: 44.41 ng
RT: 22.665 min Scan# 3371
Delta R.T. -0.000 min
Lab File: BM032406.D
Acq: 08 Oct 2021 22:58




Tgt Ion:252 Resp: 2678221
Ion Ratio Lower Upper
252 100
253 22.2 18.0 27.0
125 10.5 7.5 11.3





#89
Benzo(k)fluoranthene
Concen: 43.54 ng
RT: 22.706 min Scan# 3378
Delta R.T. -0.000 min
Lab File: BM032406.D
Acq: 08 Oct 2021 22:58




Instrument :
BNA_M
ClientSampleId :
PSC124027MSD




Tgt Ion:252 Resp: 2565087
Ion Ratio Lower Upper
252 100
253 22.0 17.3 25.9
125 10.3 7.5 11.3

Manual Integrations
APPROVED



mohammad
10/11/2021 2:13:34 PM




#90
Benzo(a)pyrene
Concen: 53.80 ng
RT: 23.206 min Scan# 3463
Delta R.T. -0.006 min
Lab File: BM032406.D
Acq: 08 Oct 2021 22:58




Tgt Ion:252 Resp: 2630910
Ion Ratio Lower Upper
252 100
253 22.0 17.4 26.2
125 11.8 8.1 12.1





#91
Dibenzo(a,h)anthracene
Concen: 50.00 ng
RT: 25.429 min Scan# 3841
Delta R.T. -0.006 min
Lab File: BM032406.D
Acq: 08 Oct 2021 22:58




Instrument :
BNA_M
ClientSampleId :
PSC124027MSD




Tgt Ion:278 Resp: 2574925
Ion Ratio Lower Upper
278 100
139 16.9 10.9 16.3#
279 23.7 18.2 27.2

Manual Integrations
APPROVED

mohammad
10/11/2021 2:13:34 PM



#92
Benzo(g,h,i)perylene
Concen: 53.15 ng
RT: 26.082 min Scan# 3952
Delta R.T. -0.000 min
Lab File: BM032406.D
Acq: 08 Oct 2021 22:58



Tgt Ion:276 Resp: 2705980
Ion Ratio Lower Upper
276 100
277 24.2 19.4 29.0
138 23.1 14.6 22.0#

