

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM101022\
 Data File : BM036901.D
 Acq On : 10 Oct 2022 11:04
 Operator : CG/JU
 Sample : PB148169BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SBLK169

Quant Time: Oct 10 12:09:52 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM100622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Oct 10 00:54:15 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.955	152	7167	0.400	ng/ul	0.00
4) Naphthalene-d8	10.759	136	24961	0.400	ng/ul #	0.00
9) Acenaphthene-d10	14.584	164	13541	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.317	188	27233	0.400	ng/ul	0.00
17) Chrysene-d12	21.488	240	19298	0.400	ng/ul	0.00
23) Perylene-d12	23.885	264	16350	0.400	ng/ul #	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.348	96	56230	5.563	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.349	152	11336	0.301	ng/ul	0.00
18) Fluoranthene-d10	19.340	212	21448	0.372	ng/ul	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM101022\
Data File : BM036901.D
Acq On : 10 Oct 2022 11:04
Operator : CG/JU
Sample : PB148169BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK169

Quant Time: Oct 10 12:09:52 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM100622.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Oct 10 00:54:15 2022
Response via : Initial Calibration

