

Data Path : Z:\svoasrv\HPCHEM1\BNA\_M\Data\BM101022\  
 Data File : BM036949.D  
 Acq On : 11 Oct 2022 17:41  
 Operator : CG/JU  
 Sample : PB148212BL  
 Misc :  
 ALS Vial : 52 Sample Multiplier: 1

Instrument :  
 BNA\_M  
 ClientSampleId :  
 SBLK212

Quant Time: Oct 12 01:27:15 2022  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA\_M\METHODS\SFAM-EPA-SIM-BM100622.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Tue Oct 11 09:03:52 2022  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon | Response | Conc  | Units | Dev(Min) |
|-----------------------------|--------|------|----------|-------|-------|----------|
| -----                       |        |      |          |       |       |          |
| Internal Standards          |        |      |          |       |       |          |
| 1) 1,4-Dichlorobenzene-d4   | 7.955  | 152  | 10615    | 0.400 | ng/u1 | 0.00     |
| 4) Naphthalene-d8           | 10.759 | 136  | 37811    | 0.400 | ng/u1 | # 0.00   |
| 9) Acenaphthene-d10         | 14.580 | 164  | 21340    | 0.400 | ng/u1 | 0.00     |
| 13) Phenanthrene-d10        | 17.317 | 188  | 43264    | 0.400 | ng/u1 | 0.00     |
| 17) Chrysene-d12            | 21.485 | 240  | 29070    | 0.400 | ng/u1 | 0.00     |
| 23) Perylene-d12            | 23.885 | 264  | 22236    | 0.400 | ng/u1 | # 0.00   |
| System Monitoring Compounds |        |      |          |       |       |          |
| 3) 1,4-Dioxane-d8           | 3.348  | 96   | 63124    | 4.217 | ng/u1 | 0.00     |
| 6) 2-Methylnaphthalene-d10  | 12.343 | 152  | 15791    | 0.277 | ng/u1 | 0.00     |
| 18) Fluoranthene-d10        | 19.336 | 212  | 29377    | 0.338 | ng/u1 | 0.00     |

| Target Compounds | Qvalue |
|------------------|--------|
| -----            |        |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA\_M\Data\BM101022\  
Data File : BM036949.D  
Acq On : 11 Oct 2022 17:41  
Operator : CG/JU  
Sample : PB148212BL  
Misc :  
ALS Vial : 52 Sample Multiplier: 1

Instrument :  
BNA\_M  
ClientSampleId :  
SBLK212

Quant Time: Oct 12 01:27:15 2022  
Quant Method : Z:\SVOASRV\HPCHEM1\BNA\_M\METHODS\SFAM-EPA-SIM-BM100622.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Tue Oct 11 09:03:52 2022  
Response via : Initial Calibration

