

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM101022\
 Data File : BM036975.D
 Acq On : 12 Oct 2022 09:47
 Operator : CG/JU
 Sample : PB148216BL
 Misc :
 ALS Vial : 79 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SBLK216

Quant Time: Oct 12 23:08:35 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM100622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Oct 11 09:03:52 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.947	152	7928	0.400	ng/ul	0.00
4) Naphthalene-d8	10.749	136	28003	0.400	ng/ul	# 0.00
9) Acenaphthene-d10	14.575	164	15528	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.309	188	30480	0.400	ng/ul	0.00
17) Chrysene-d12	21.482	240	17981	0.400	ng/ul	0.00
23) Perylene-d12	23.879	264	15445	0.400	ng/ul	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.344	96	55317	4.948	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.338	152	14244	0.337	ng/ul	0.00
18) Fluoranthene-d10	19.331	212	24538	0.456	ng/ul	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM101022\
Data File : BM036975.D
Acq On : 12 Oct 2022 09:47
Operator : CG/JU
Sample : PB148216BL
Misc :
ALS Vial : 79 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK216

Quant Time: Oct 12 23:08:35 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM100622.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Oct 11 09:03:52 2022
Response via : Initial Calibration

