


Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM101024\
Data File : BM047980.D
Acq On : 10 Oct 2024 16:20
Operator : RC/JU
Sample : DFTPP057
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP057

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-SIM-BM092524.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Thu Oct 10 18:01:13 2024



AutoFind: Scans 2484, 2485, 2486; Background Corrected with Scan 2476

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	29.8	352426	PASS
68	69	0.00	2	1.4	5402	PASS
69	198	0.00	100	31.5	372629	PASS
70	69	0.00	2	0.4	1639	PASS
127	198	10	80	43.4	512427	PASS
197	198	0.00	2	0.5	6468	PASS
198	198	100	100	100.0	1181867	PASS
199	198	5	9	6.9	81187	PASS
275	198	10	60	25.7	303168	PASS
365	198	1	100	3.6	42504	PASS
441	198	0.01	100	13.9	164160	PASS
442	198	50	100	88.1	1041771	PASS
443	442	15	24	19.5	203200	PASS

m/z	Abundance
36.00	507.0
37.90	379.0
39.10	870.0
40.00	3461.0
41.10	1554.0
42.00	753.0
43.10	2241.0
44.00	8654.0
45.00	583.0
45.90	249.0
46.90	222.0
48.00	157.0
49.00	768.0
50.10	305.0
51.10	513.0
52.00	264.0
52.90	285.0
55.00	1558.0
56.00	798.0
57.00	880.0
58.00	194.0
60.10	464.0
65.00	194.0
67.10	348.0
67.90	351.0
69.00	790.0
70.00	566.0
71.00	416.0
73.10	691.0
74.00	215.0
75.00	165.0
75.80	176.0
77.00	541.0
78.00	333.0
79.10	176.0
80.00	154.0
81.10	388.0
82.10	356.0
83.20	565.0
83.90	890.0
85.00	430.0
85.90	480.0
89.10	177.0
90.80	320.0
91.20	322.0
95.00	239.0
96.10	441.0
97.00	550.0
98.00	361.0
103.10	156.0
107.00	195.0
110.00	181.0
111.10	308.0
113.10	151.0
115.00	240.0
118.90	208.0
125.10	214.0
126.00	236.0
127.90	182.0
132.90	365.0
135.00	427.0
139.10	243.0
147.20	249.0
149.00	336.0
150.10	175.0
160.00	165.0
162.90	230.0
165.90	155.0
167.10	1115.0
168.00	176.0
177.00	285.0
191.00	398.0
193.00	459.0
207.00	2619.0
208.00	549.0
209.00	457.0
221.20	190.0
267.00	218.0
281.00	863.0
283.00	178.0

Instrument :
BNA_M
ClientSampleId :
DFTPP057

m/z	Abundance
35.90	469.0
37.10	1373.0
38.10	3556.0
39.10	26608.0
40.00	3829.0
41.10	1450.0
42.10	601.0
43.10	1487.0
44.00	6667.0
45.10	980.0
46.90	219.0
47.90	407.0
49.00	2577.0
50.10	87984.0
51.10	353024.0
52.10	18544.0
53.10	775.0
54.20	269.0
55.00	2309.0
56.10	10727.0
57.00	24904.0
58.00	1264.0
59.10	302.0
60.10	514.0
61.00	4376.0
62.10	4398.0
63.10	14607.0
64.10	2217.0
65.10	7633.0
66.00	709.0
67.20	756.0
68.00	5508.0
69.00	374336.0
70.10	2556.0
71.00	538.0
73.10	3095.0
74.10	36232.0
75.10	59816.0
76.10	20368.0
77.10	443264.0
78.10	30992.0
79.10	26312.0
80.10	21056.0
81.10	28528.0
82.00	7610.0
83.10	7461.0
84.10	1365.0
85.00	4777.0
86.10	8587.0
87.10	3929.0
88.00	1459.0
89.00	863.0
90.10	289.0
91.10	6428.0
92.10	6784.0
93.10	46512.0
94.00	3237.0
95.10	925.0
96.10	2358.0
97.10	715.0
98.10	35064.0
99.10	28576.0
100.10	2739.0
101.10	17856.0
102.10	890.0
103.00	5983.0
104.10	9761.0
105.10	9647.0
106.10	3379.0
107.10	135680.0
108.10	19848.0
109.10	3785.0
110.10	246912.0
111.10	37576.0
112.10	4400.0
113.00	1480.0
114.00	460.0
115.00	636.0
116.00	6964.0
117.10	106872.0
118.10	8579.0
119.00	762.0
120.00	1804.0
121.10	765.0
122.00	7908.0
123.00	13496.0
124.00	5962.0
125.10	5406.0
126.20	2233.0
127.10	486592.0
128.10	38328.0
129.10	181952.0

Instrument :
BNA_M
ClientSampleId :
DFTPP057

130.00	16206.0
131.00	3329.0
132.10	1700.0
133.10	653.0
134.10	5264.0
135.10	14568.0
136.10	6116.0
137.10	7867.0
138.00	1572.0
139.00	993.0
140.10	1891.0
141.00	21872.0
142.10	7568.0
143.00	5232.0
144.00	1491.0
145.00	1017.0
146.00	4109.0
147.00	11968.0
148.00	28456.0
149.10	6188.0
150.10	1978.0
151.20	3559.0
152.00	2011.0
153.00	8165.0
154.00	5592.0
155.10	13639.0
156.10	20352.0
157.10	3736.0
158.00	4238.0
159.00	3163.0
160.00	8042.0
161.10	11303.0
162.00	3169.0
163.10	988.0
164.00	1408.0
165.10	8806.0
166.10	7179.0
167.10	46800.0
168.10	27448.0
169.10	4586.0
170.00	1642.0
171.10	1911.0
172.00	3954.0
173.00	5633.0
174.10	9950.0
175.10	19560.0
176.10	5519.0
177.10	8203.0
178.00	2887.0
179.00	35984.0
180.10	26320.0
181.10	11858.0
182.00	1899.0
182.90	739.0
183.20	842.0
184.00	2869.0
185.10	17120.0
186.10	133184.0
187.10	39752.0
188.10	4720.0
189.10	8819.0
190.10	1349.0
191.10	3828.0
192.00	11208.0
193.10	12845.0
194.10	2889.0
195.20	1691.0
196.10	26976.0
197.00	5511.0
198.00	1069568.0
199.00	76984.0
200.00	4936.0
201.60	3806.0
203.10	6956.0
204.10	36800.0
205.10	61976.0
206.10	263808.0
207.10	37400.0
208.10	8555.0
209.10	3216.0
210.20	4082.0
211.10	9142.0
212.00	1620.0
213.10	787.0
214.00	352.0
215.10	2913.0
216.10	5579.0
217.00	69696.0
218.00	8890.0
219.10	855.0
221.10	39360.0
221.90	9413.0
223.10	16003.0
224.10	147072.0
225.10	39552.0
226.10	4355.0

Instrument :
BNA_M
ClientSampleId :
DFTPP057

227.10	59344.0
228.10	9361.0
229.10	13556.0
230.00	1559.0
231.10	5484.0
232.10	1010.0
233.10	1189.0
234.00	3537.0
235.10	4342.0
236.00	2595.0
237.00	4400.0
238.10	860.0
239.00	2136.0
240.10	1625.0
241.00	3487.0
242.10	7594.0
243.10	8580.0
244.10	118128.0
245.10	15783.0
246.10	20112.0
247.10	4081.0
248.00	1096.0
249.10	3831.0
250.10	781.0
251.00	1176.0
252.10	1435.0
253.10	2302.0
254.10	3235.0
255.10	580608.0
256.10	88704.0
257.10	6667.0
258.00	32808.0
259.00	4991.0
260.00	738.0
261.00	797.0
262.20	161.0
263.10	344.0
264.00	926.0
265.00	12657.0
266.10	2053.0
267.10	387.0
268.10	207.0
270.10	616.0
271.10	2001.0
272.10	1768.0
273.10	17600.0
274.10	47992.0
275.10	253440.0
276.10	34872.0
277.10	21176.0
278.10	3947.0
279.00	601.0
281.10	991.0
282.10	548.0
283.10	2445.0
284.10	1717.0
285.10	3709.0
286.10	726.0
288.00	273.0
289.00	881.0
290.10	757.0
291.00	894.0
292.10	1048.0
293.10	4800.0
294.10	1411.0
295.00	1623.0
296.10	72376.0
297.10	9446.0
298.00	622.0
299.00	201.0
301.10	1108.0
302.10	1201.0
303.10	8149.0
304.10	1951.0
305.20	263.0
308.10	915.0
309.10	860.0
310.10	847.0
311.20	219.0
312.00	323.0
313.10	611.0
314.10	3534.0
315.10	8175.0
316.10	4834.0
317.00	1093.0
319.00	290.0
320.00	245.0
321.10	2096.0
322.10	1137.0
323.10	22616.0
324.20	4663.0
325.10	381.0
325.90	435.0
327.00	4389.0
328.10	2337.0
329.00	494.0

Instrument :
BNA_M
ClientSampleId :
DFTPP057

330.10	192.0
332.10	2053.0
333.10	2105.0
334.10	14280.0
335.10	3705.0
336.20	527.0
339.10	398.0
340.00	347.0
341.10	2637.0
342.10	758.0
346.10	5492.0
347.00	843.0
348.10	156.0
350.10	277.0
350.90	326.0
352.10	7509.0
353.10	4915.0
354.10	7613.0
355.10	1202.0
357.10	194.0
359.10	596.0
361.20	229.0
365.10	31736.0
366.10	5253.0
367.00	316.0
370.10	782.0
371.10	1627.0
372.10	12239.0
373.20	3259.0
374.10	388.0
377.10	218.0
383.10	3487.0
384.10	1034.0
385.00	333.0
390.10	1650.0
391.10	1270.0
392.10	654.0
393.00	228.0
401.20	673.0
402.10	4947.0
403.10	7037.0
404.10	2885.0
405.10	368.0
408.00	158.0
410.00	368.0
415.00	257.0
416.10	156.0
421.10	5687.0
422.10	5643.0
423.10	42408.0
424.10	10009.0
425.10	1070.0
438.00	273.0
438.90	241.0
440.00	301.0
441.10	116864.0
442.10	728640.0
443.10	150592.0
444.20	14678.0
445.10	787.0

Instrument :
BNA_M
ClientSampleId :
DFTPP057

m/z	Abundance
36.00	549.0
37.00	1549.0
38.10	3788.0
39.10	27448.0
40.00	3230.0
41.10	1752.0
42.00	654.0
43.10	1259.0
44.00	6832.0
45.10	979.0
47.00	345.0
48.00	275.0
49.00	2720.0
50.10	98848.0
51.10	399616.0
52.10	19944.0
53.10	1344.0
54.00	228.0
55.00	2592.0
56.10	11146.0
57.00	26952.0
58.00	1327.0
59.00	419.0
60.10	654.0
61.00	4955.0
62.10	5649.0
63.10	16792.0
64.00	2042.0
65.10	8310.0
66.10	695.0
67.00	1010.0
68.10	6356.0
69.00	417792.0
70.10	2037.0
71.10	650.0
71.90	318.0
73.10	3810.0
74.10	39928.0
75.10	69376.0
76.10	22304.0
77.10	500736.0
78.10	33776.0
79.10	29768.0
80.10	24000.0
81.10	32784.0
82.10	9272.0
83.10	7968.0
84.00	1311.0
85.10	6499.0
86.10	9121.0
87.10	4159.0
88.00	1722.0
89.10	963.0
90.00	249.0
91.10	7596.0
92.00	8554.0
93.10	53112.0
94.00	4027.0
95.00	1005.0
96.10	3204.0
97.10	1091.0
98.10	41568.0
99.10	34568.0
100.00	2994.0
101.10	20480.0
102.00	1356.0
103.10	6337.0
104.00	11971.0
105.10	10995.0
106.10	4301.0
107.10	160960.0
108.10	25616.0
109.10	4167.0
110.10	293056.0
111.10	44568.0
112.00	5546.0
113.00	1571.0
114.10	381.0
115.10	774.0
116.00	7634.0
117.10	135424.0
118.10	9421.0
119.00	1205.0
120.00	2052.0
121.00	1027.0
122.10	9374.0
123.00	16262.0
124.00	7517.0
125.10	7273.0
126.10	2725.0
127.10	585408.0
128.10	43120.0

Instrument :
BNA_M
ClientSampleId :
DFTPP057

129.10	217664.0
130.00	18640.0
131.00	3808.0
132.20	1779.0
133.00	1001.0
134.10	6050.0
135.10	16856.0
136.10	6863.0
137.00	9523.0
138.00	1709.0
139.00	1326.0
140.00	2880.0
141.10	27944.0
142.10	9753.0
143.00	6496.0
144.10	1941.0
145.10	1724.0
146.10	5105.0
147.00	14085.0
148.00	32136.0
149.10	7306.0
150.10	2135.0
151.10	3636.0
152.00	2330.0
153.10	8013.0
154.10	6087.0
155.10	16896.0
156.10	25360.0
157.10	4660.0
158.10	5104.0
159.10	4377.0
160.00	8586.0
161.10	13952.0
162.10	3696.0
163.00	958.0
164.00	1520.0
165.10	10408.0
166.10	8523.0
167.10	60736.0
168.10	34632.0
169.10	5558.0
170.10	2038.0
171.00	2576.0
172.10	5020.0
173.10	7638.0
174.10	13265.0
175.10	23640.0
176.10	6800.0
177.00	11051.0
178.10	3761.0
179.00	46712.0
180.10	32504.0
181.10	15674.0
182.10	2659.0
183.10	1236.0
184.10	4141.0
185.10	22704.0
186.10	170944.0
187.10	50336.0
188.10	4924.0
189.10	11123.0
190.00	1499.0
191.00	5084.0
192.10	15751.0
193.10	17192.0
194.10	3722.0
195.10	2514.0
196.10	36048.0
197.00	7930.0
198.00	1356288.0
199.00	91992.0
200.10	7344.0
201.60	5226.0
203.10	8881.0
204.10	47728.0
205.10	82184.0
206.10	353664.0
207.10	45448.0
208.10	10883.0
209.10	3923.0
210.20	5072.0
211.10	12325.0
213.00	1064.0
214.10	408.0
215.10	3712.0
216.10	7586.0
217.00	93064.0
218.10	11658.0
219.00	1138.0
220.30	2036.0
221.10	52360.0
223.10	20896.0
224.10	194752.0
225.10	50816.0
226.10	5499.0
227.10	80784.0

Instrument :
BNA_M
ClientSampleId :
DFTPP057

228.10	10583.0
229.10	17192.0
230.00	2048.0
231.10	7156.0
232.10	1140.0
233.10	1295.0
234.10	5425.0
235.10	6080.0
236.10	3932.0
237.10	6845.0
238.10	879.0
239.10	2735.0
240.10	2779.0
241.00	5232.0
242.10	10281.0
243.10	10863.0
244.10	157120.0
245.10	22256.0
246.10	27976.0
247.10	5875.0
248.10	1437.0
249.10	5364.0
250.00	1070.0
250.90	1301.0
252.10	1287.0
253.10	3268.0
254.10	4841.0
255.10	796928.0
256.10	114656.0
257.10	9455.0
258.10	43720.0
259.10	6882.0
260.00	1128.0
261.00	1595.0
262.10	228.0
262.90	250.0
264.00	1066.0
265.10	16856.0
266.00	2528.0
267.00	379.0
268.00	213.0
269.10	231.0
270.10	805.0
271.10	2564.0
272.10	2335.0
273.10	22712.0
274.10	63608.0
275.10	346048.0
276.10	46960.0
277.10	28152.0
278.10	5138.0
279.10	1042.0
281.10	897.0
282.10	1131.0
283.10	3833.0
284.10	2132.0
285.10	5682.0
286.10	932.0
288.00	466.0
289.00	1046.0
290.00	1156.0
291.10	1212.0
292.10	1533.0
293.10	6817.0
294.20	1674.0
295.10	2030.0
296.10	101640.0
297.10	14044.0
298.10	893.0
299.10	254.0
301.10	1147.0
302.10	1749.0
303.10	11811.0
304.20	2797.0
305.10	392.0
306.10	180.0
307.10	191.0
308.10	1521.0
309.10	872.0
310.10	1480.0
311.00	238.0
312.00	346.0
313.10	929.0
314.20	4710.0
315.10	10944.0
316.10	6887.0
317.00	1047.0
319.10	330.0
320.00	393.0
321.10	2946.0
322.10	1707.0
323.10	33432.0
324.10	6553.0
325.10	570.0
326.10	670.0
327.10	6368.0

Instrument :
BNA_M
ClientSampleId :
DFTPP057

328.10	3012.0
329.00	717.0
330.00	206.0
331.10	167.0
332.10	2196.0
333.10	2792.0
334.10	21792.0
335.10	4989.0
336.10	788.0
339.00	355.0
340.00	397.0
341.10	3817.0
342.10	1149.0
346.10	7786.0
347.10	1664.0
348.00	268.0
350.20	336.0
351.10	577.0
352.10	10739.0
353.10	8383.0
354.10	10366.0
355.10	2082.0
356.10	332.0
357.10	313.0
359.10	703.0
360.10	199.0
361.20	288.0
364.00	296.0
365.10	48272.0
366.10	7840.0
367.10	442.0
370.10	1190.0
371.10	2906.0
372.10	18064.0
373.10	4415.0
374.10	638.0
377.10	422.0
383.10	5020.0
384.10	1554.0
385.20	544.0
390.10	2363.0
391.20	1875.0
392.10	1040.0
395.10	271.0
397.00	155.0
401.10	1190.0
402.10	7450.0
403.10	11686.0
404.10	4350.0
405.10	444.0
410.00	266.0
414.90	432.0
415.20	432.0
416.10	238.0
421.10	9656.0
422.10	9205.0
423.10	63720.0
424.10	13088.0
425.00	1269.0
438.00	482.0
439.10	308.0
440.00	347.0
441.10	184640.0
442.10	1187840.0
443.20	229504.0
444.10	22992.0
445.10	1430.0

Instrument :
BNA_M
ClientSampleId :
DFTPP057

m/z	Abundance
36.00	477.0
37.00	1025.0
38.10	3471.0
39.10	21648.0
40.10	4351.0
41.00	1638.0
42.00	707.0
43.10	1710.0
44.10	8149.0
45.00	1269.0
47.00	450.0
48.00	336.0
49.10	1985.0
50.10	76424.0
51.10	306176.0
52.10	16448.0
53.00	866.0
54.00	232.0
55.10	2296.0
56.10	8830.0
57.10	22168.0
58.00	1447.0
59.10	276.0
60.00	637.0
61.00	4147.0
62.10	4439.0
63.10	11979.0
64.10	2360.0
65.10	6667.0
66.10	782.0
67.00	708.0
68.10	5395.0
69.00	328128.0
70.00	2023.0
71.00	422.0
73.10	3340.0
74.10	32592.0
75.00	51992.0
76.10	16880.0
77.10	397824.0
78.10	26352.0
79.10	23112.0
80.00	17440.0
81.10	26624.0
82.10	7355.0
83.00	7045.0
83.90	1251.0
85.00	5213.0
86.00	7554.0
87.10	3460.0
88.10	1413.0
89.00	1065.0
90.10	154.0
91.10	6710.0
92.00	6451.0
93.10	40272.0
94.00	2951.0
95.10	763.0
96.10	2394.0
97.00	839.0
98.10	32248.0
99.10	27152.0
100.10	2482.0
101.00	15162.0
102.10	986.0
103.10	4707.0
104.00	9438.0
105.00	8737.0
106.10	3475.0
107.10	124944.0
108.10	18984.0
109.10	2996.0
110.00	231616.0
111.00	34432.0
112.00	3953.0
113.10	1540.0
114.00	365.0
115.20	462.0
116.10	6645.0
117.10	99816.0
118.10	7174.0
119.10	774.0
120.10	1661.0
120.90	610.0
121.30	525.0
122.00	7309.0
123.10	11421.0
124.10	6054.0
125.00	5524.0
126.10	2056.0
127.10	465280.0
128.10	33016.0

Instrument :
BNA_M
ClientSampleId :
DFTPP057

129.10	170432.0
130.00	14343.0
131.10	3336.0
132.00	1808.0
133.10	968.0
134.00	5070.0
135.10	13493.0
136.00	5894.0
137.10	7728.0
138.00	1749.0
139.10	1148.0
140.10	2128.0
141.00	21896.0
142.10	6632.0
143.10	5330.0
144.10	1271.0
145.10	1337.0
146.10	4125.0
147.00	10787.0
148.00	26904.0
149.00	5550.0
150.10	1739.0
151.10	3210.0
151.90	1685.0
153.10	6493.0
154.00	6100.0
155.10	13314.0
156.10	20176.0
157.10	3471.0
158.00	4726.0
159.10	3261.0
160.00	7350.0
161.10	11447.0
162.00	3103.0
163.00	938.0
164.10	1288.0
165.00	8403.0
166.00	7277.0
167.10	51048.0
168.10	26880.0
169.10	4276.0
170.10	1653.0
171.10	1814.0
172.00	4254.0
173.00	5719.0
174.10	10524.0
175.10	19544.0
176.10	5724.0
177.00	8512.0
178.10	2895.0
179.00	36456.0
180.10	25952.0
181.10	12257.0
182.10	2264.0
183.00	1247.0
184.10	3065.0
185.10	17328.0
186.10	140224.0
187.10	39672.0
188.10	3921.0
189.00	8703.0
190.00	1523.0
191.10	4432.0
192.00	12653.0
193.10	13520.0
194.10	3370.0
195.20	1514.0
196.10	28648.0
196.90	5963.0
198.00	1119744.0
199.00	74584.0
200.00	5318.0
201.60	5197.0
203.00	7233.0
204.10	39776.0
205.10	67088.0
206.10	284416.0
207.10	38464.0
208.00	8975.0
209.00	3291.0
210.10	4441.0
211.10	10258.0
213.00	798.0
213.90	385.0
215.00	3225.0
216.10	5657.0
217.00	77760.0
218.10	9896.0
219.10	941.0
221.10	43152.0
221.90	9991.0
223.10	15715.0
224.10	159616.0
225.10	41480.0
226.10	3549.0
227.10	66632.0

Instrument :
BNA_M
ClientSampleId :
DFTPP057

228.10	9218.0
229.00	13950.0
230.00	1856.0
231.10	5171.0
232.00	968.0
233.00	1238.0
234.10	4253.0
235.10	5155.0
236.10	3447.0
237.10	5833.0
238.00	895.0
239.10	2691.0
240.10	1814.0
241.00	3662.0
242.10	8723.0
243.10	8746.0
244.10	130760.0
245.10	16382.0
246.10	25512.0
247.10	4600.0
248.10	1283.0
249.10	4687.0
250.10	1036.0
251.00	1125.0
252.10	1515.0
253.00	3102.0
254.10	4060.0
255.10	676928.0
256.10	98120.0
257.10	7023.0
258.10	40320.0
259.10	5435.0
260.10	1064.0
261.10	1232.0
262.00	251.0
263.10	448.0
264.10	906.0
265.10	15611.0
266.10	2374.0
267.10	419.0
267.90	338.0
270.00	588.0
271.10	2037.0
272.10	1777.0
273.10	21360.0
274.10	54296.0
275.10	310016.0
276.10	41040.0
277.10	25232.0
278.10	4404.0
279.10	1067.0
281.10	1017.0
282.10	1025.0
283.20	2876.0
284.10	2148.0
285.10	4996.0
286.00	769.0
288.00	284.0
289.10	1251.0
290.00	710.0
291.10	752.0
292.00	1324.0
293.10	5889.0
294.10	1673.0
295.00	1677.0
296.10	91064.0
297.10	11646.0
298.00	752.0
299.00	375.0
301.10	1394.0
302.00	1595.0
303.10	10035.0
304.10	2905.0
305.20	382.0
307.20	434.0
308.10	1022.0
309.00	774.0
310.10	1305.0
311.00	231.0
312.00	361.0
313.10	873.0
314.10	4315.0
315.10	10148.0
316.10	5814.0
317.10	865.0
320.00	388.0
321.10	2778.0
322.10	1207.0
323.10	31104.0
324.10	6077.0
325.00	386.0
326.00	545.0
327.00	5738.0
328.10	2870.0
329.00	709.0
330.10	190.0

Instrument :
BNA_M
ClientSampleId :
DFTPP057

331.00	153.0
332.10	2248.0
333.10	2581.0
334.10	19776.0
335.10	5304.0
336.20	646.0
337.00	157.0
339.10	490.0
340.00	520.0
341.20	3418.0
342.10	1225.0
343.00	189.0
346.10	7855.0
347.00	1262.0
348.00	206.0
350.00	200.0
351.20	415.0
352.10	9603.0
353.10	6480.0
354.10	10261.0
355.10	1958.0
356.10	299.0
359.10	1056.0
361.10	222.0
364.20	236.0
365.10	47504.0
366.10	6541.0
367.10	496.0
370.10	1000.0
371.10	2534.0
372.10	18304.0
373.10	3866.0
374.10	645.0
377.00	254.0
383.10	5396.0
384.10	1281.0
385.10	516.0
390.10	2282.0
391.20	1961.0
392.20	1117.0
393.20	305.0
397.10	236.0
401.10	1304.0
402.10	8756.0
403.10	11124.0
404.10	3877.0
405.20	636.0
410.10	277.0
415.10	604.0
420.10	186.0
421.10	9843.0
422.10	8964.0
423.10	61168.0
424.10	13064.0
425.10	1380.0
437.30	192.0
438.20	352.0
439.00	344.0
440.00	344.0
441.10	190976.0
442.20	1208832.0
443.20	229504.0
444.10	20496.0
445.10	943.0

Instrument :
BNA_M
ClientSampleId :
DFTPP057