

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM101121\
 Data File : BM032417.D
 Acq On : 11 Oct 2021 12:08
 Operator : CG/JU
 Sample : M4125-03
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 BGB16

Quant Time: Oct 11 12:38:36 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-BM100521.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Oct 11 11:06:33 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.613	152	260142	20.000	ng/u1	0.00
20) Naphthalene-d8	10.407	136	1143891	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.266	164	765513	20.000	ng/u1	0.00
64) Phenanthrene-d10	16.995	188	1591124	20.000	ng/u1	0.00
79) Chrysene-d12	21.159	240	1565010	20.000	ng/u1	0.00
88) Perylene-d12	23.306	264	1592727	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	2.960	96	14357	2.213	ng/uL	0.00
4) Pyridine-d5	3.390	84	161057	8.787	ng/u1	0.00
7) Phenol-d5	6.807	99	163237	7.379	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	6.943	67	347574	27.365	ng/u1	0.00
11) 2-Chlorophenol-d4	7.154	132	399769	23.669	ng/u1	0.00
15) 4-Methylphenol-d8	8.342	113	297283	16.639	ng/u1	0.00
21) Nitrobenzene-d5	8.766	128	246482	30.540	ng/u1	0.00
24) 2-Nitrophenol-d4	9.495	143	273610	31.451	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.036	165	485961	28.061	ng/u1	0.00
31) 4-Chloroaniline-d4	10.542	131	542181	21.225	ng/u1	0.00
46) Dimethylphthalate-d6	13.666	166	1751874	33.068	ng/u1	0.00
49) Acenaphthylene-d8	13.954	160	2029353	31.532	ng/u1	0.00
54) 4-Nitrophenol-d4	14.471	143	73793	7.093	ng/u1	0.00
60) Fluorene-d10	15.254	176	1493091	33.231	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.371	200	276731	30.517	ng/u1	0.00
73) Anthracene-d10	17.089	188	2397435	34.615	ng/u1	0.00
81) Pyrene-d10	19.377	212	2701811	34.316	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.171	264	2575537	32.291	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM101121\
Data File : BM032417.D
Acq On : 11 Oct 2021 12:08
Operator : CG/JU
Sample : M4125-03
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
BGB16

Quant Time: Oct 11 12:38:36 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-BM100521.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Oct 11 11:06:33 2021
Response via : Initial Calibration

