

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM101121\
 Data File : BM032453.D
 Acq On : 12 Oct 2021 11:40
 Operator : CG/JU
 Sample : M4098-10
 Misc :
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 YAT21

Quant Time: Oct 12 12:17:10 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-BM100521.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Oct 12 09:54:27 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.607	152	269872	20.000	ng/u1	0.00
20) Naphthalene-d8	10.401	136	1235842	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.260	164	846919	20.000	ng/u1	0.00
64) Phenanthrene-d10	16.983	188	1794562	20.000	ng/u1	0.00
79) Chrysene-d12	21.148	240	1757067	20.000	ng/u1	0.00
88) Perylene-d12	23.295	264	1784541	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	2.955	96	31391	4.664	ng/uL	0.00
4) Pyridine-d5	3.396	84	83583	4.396	ng/u1	0.01
7) Phenol-d5	6.801	99	144777	6.309	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	6.937	67	372107	28.240	ng/u1	0.00
11) 2-Chlorophenol-d4	7.149	132	399593	22.806	ng/u1	0.00
15) 4-Methylphenol-d8	8.337	113	286346	15.449	ng/u1	0.00
21) Nitrobenzene-d5	8.760	128	264590	30.344	ng/u1	0.00
24) 2-Nitrophenol-d4	9.490	143	283531	30.166	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.031	165	476991	25.494	ng/u1	0.00
31) 4-Chloroaniline-d4	10.531	131	592310	21.462	ng/u1	0.00
46) Dimethylphthalate-d6	13.660	166	1843430	31.452	ng/u1	0.00
49) Acenaphthylene-d8	13.948	160	2148380	30.173	ng/u1	0.00
54) 4-Nitrophenol-d4	14.466	143	64147	5.573	ng/u1	0.00
60) Fluorene-d10	15.248	176	1568326	31.551	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.366	200	288207	28.180	ng/u1	0.00
73) Anthracene-d10	17.083	188	2620598	33.548	ng/u1	0.00
81) Pyrene-d10	19.371	212	3052490	34.533	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.165	264	2921980	32.696	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM101121\
Data File : BM032453.D
Acq On : 12 Oct 2021 11:40
Operator : CG/JU
Sample : M4098-10
Misc :
ALS Vial : 39 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
YAT21

Quant Time: Oct 12 12:17:10 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-BM100521.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Oct 12 09:54:27 2021
Response via : Initial Calibration

