

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM101121\
 Data File : BM032459.D
 Acq On : 12 Oct 2021 15:55
 Operator : CG/JU
 Sample : PB139788BL
 Misc :
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SBLK788

Quant Time: Oct 12 16:26:02 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-BM100521.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Oct 12 09:54:27 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.607	152	222593	20.00	ng/ul	0.00
20) Naphthalene-d8	10.395	136	993378	20.00	ng/ul	0.00
38) Acenaphthene-d10	14.260	164	691149	20.00	ng/ul	0.00
64) Phenanthrene-d10	16.983	188	1474280	20.00	ng/ul	0.00
79) Chrysene-d12	21.148	240	1424364	20.00	ng/ul	0.00
88) Perylene-d12	23.294	264	1373467	20.00	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	2.955	96	34575	6.23	ng/uL	0.00
4) Pyridine-d5	3.384	84	510530	32.55	ng/ul	0.00
7) Phenol-d5	6.801	99	635162	33.56	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	6.937	67	383875	35.32	ng/ul	0.00
11) 2-Chlorophenol-d4	7.148	132	509478	35.25	ng/ul	0.00
15) 4-Methylphenol-d8	8.337	113	535393	35.02	ng/ul	0.00
21) Nitrobenzene-d5	8.760	128	265819	37.93	ng/ul	0.00
24) 2-Nitrophenol-d4	9.489	143	297394	39.36	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.031	165	511794	34.03	ng/ul	0.00
31) 4-Chloroaniline-d4	10.531	131	791478	35.68	ng/ul	0.00
46) Dimethylphthalate-d6	13.660	166	1846620	38.61	ng/ul	0.00
49) Acenaphthylene-d8	13.948	160	2171274	37.37	ng/ul	0.00
54) 4-Nitrophenol-d4	14.466	143	308003	32.79	ng/ul	0.00
60) Fluorene-d10	15.248	176	1555533	38.35	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.366	200	249149	29.65	ng/ul	0.00
73) Anthracene-d10	17.083	188	2417735	37.67	ng/ul	0.00
81) Pyrene-d10	19.371	212	2693844	37.59	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.165	264	2454033	35.68	ng/ul	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM101121\
Data File : BM032459.D
Acq On : 12 Oct 2021 15:55
Operator : CG/JU
Sample : PB139788BL
Misc :
ALS Vial : 45 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK788

Quant Time: Oct 12 16:26:02 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-BM100521.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Oct 12 09:54:27 2021
Response via : Initial Calibration

