

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM101121\
 Data File : BM032473.D
 Acq On : 13 Oct 2021 00:59
 Operator : CG/JU
 Sample : PB139820BL
 Misc :
 ALS Vial : 59 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SBLK820

Quant Time: Oct 13 03:26:10 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-BM100521.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Oct 12 09:54:27 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.607	152	229085	20.00	ng/ul	0.00
20) Naphthalene-d8	10.395	136	1023010	20.00	ng/ul	0.00
38) Acenaphthene-d10	14.254	164	712483	20.00	ng/ul	0.00
64) Phenanthrene-d10	16.983	188	1497873	20.00	ng/ul	0.00
79) Chrysene-d12	21.148	240	1442839	20.00	ng/ul	0.00
88) Perylene-d12	23.289	264	1386679	20.00	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	2.955	96	32822	5.74	ng/uL	0.00
4) Pyridine-d5	3.384	84	464363	28.77	ng/ul	0.00
7) Phenol-d5	6.795	99	551342	28.30	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	6.931	67	328744	29.39	ng/ul	0.00
11) 2-Chlorophenol-d4	7.148	132	444659	29.90	ng/ul	0.00
15) 4-Methylphenol-d8	8.337	113	464533	29.52	ng/ul	0.00
21) Nitrobenzene-d5	8.760	128	229924	31.85	ng/ul	0.00
24) 2-Nitrophenol-d4	9.484	143	255783	32.88	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.025	165	448137	28.93	ng/ul	0.00
31) 4-Chloroaniline-d4	10.531	131	682817	29.89	ng/ul	0.00
46) Dimethylphthalate-d6	13.660	166	1567315	31.79	ng/ul	0.00
49) Acenaphthylene-d8	13.942	160	1887087	31.50	ng/ul	0.00
54) 4-Nitrophenol-d4	14.466	143	249693	25.79	ng/ul	0.00
60) Fluorene-d10	15.248	176	1346803	32.21	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.360	200	206617	24.20	ng/ul	0.00
73) Anthracene-d10	17.077	188	2121998	32.55	ng/ul	0.00
81) Pyrene-d10	19.365	212	2346903	32.33	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.153	264	2119596	30.52	ng/ul	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM101121\
Data File : BM032473.D
Acq On : 13 Oct 2021 00:59
Operator : CG/JU
Sample : PB139820BL
Misc :
ALS Vial : 59 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SBLK820

Quant Time: Oct 13 03:26:10 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-BM100521.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Oct 12 09:54:27 2021
Response via : Initial Calibration

