

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM101121\
 Data File : BM032444.D
 Acq On : 12 Oct 2021 06:15
 Operator : CG/JU
 Sample : PB139788BS
 Misc :
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SLCS788

Quant Time: Oct 12 06:46:45 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-BM100521.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Oct 11 11:06:33 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.613	152	270774	20.000	ng/u1	0.00
20) Naphthalene-d8	10.401	136	1264787	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.260	164	849222	20.000	ng/u1	0.00
64) Phenanthrene-d10	16.989	188	1731310	20.000	ng/u1	0.00
79) Chrysene-d12	21.153	240	1604813	20.000	ng/u1	0.00
88) Perylene-d12	23.294	264	1669774	20.000	ng/u1	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	2.954	96	36882	5.461	ng/uL	0.00
4) Pyridine-d5	3.384	84	517484	27.124	ng/u1	0.00
7) Phenol-d5	6.801	99	724376	31.460	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	6.937	67	405977	30.708	ng/u1	0.00
11) 2-Chlorophenol-d4	7.148	132	568543	32.340	ng/u1	0.00
15) 4-Methylphenol-d8	8.348	113	604102	32.483	ng/u1	0.00
21) Nitrobenzene-d5	8.766	128	293579	32.898	ng/u1	0.00
24) 2-Nitrophenol-d4	9.489	143	327285	34.025	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.036	165	607832	31.743	ng/u1	0.00
31) 4-Chloroaniline-d4	10.536	131	805010	28.502	ng/u1	0.00
46) Dimethylphthalate-d6	13.666	166	1831129	31.157	ng/u1	0.00
49) Acenaphthylene-d8	13.948	160	2234110	31.291	ng/u1	0.00
54) 4-Nitrophenol-d4	14.471	143	354600	30.723	ng/u1	0.00
60) Fluorene-d10	15.248	176	1539127	30.879	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.371	200	315440	31.969	ng/u1	0.00
73) Anthracene-d10	17.083	188	2358009	31.289	ng/u1	0.00
81) Pyrene-d10	19.371	212	2587081	32.044	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.165	264	2490473	29.783	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	2.990	88	78402	10.665	ng/uL	98
5) Pyridine	3.401	79	550169	29.159	ng/u1	96
6) Benzaldehyde	6.737	77	467954	43.174	ng/u1	98
8) Phenol	6.831	94	784250	33.623	ng/u1	98
10) Bis(2-Chloroethyl)ether	7.031	93	601755	32.787	ng/u1	99
12) 2-Chlorophenol	7.184	128	613410	33.775	ng/u1	98
13) 2-Methylphenol	8.078	108	609363	34.067	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.148	45	758396	33.367	ng/u1	100
16) Acetophenone	8.431	105	958101	35.166	ng/u1	99
17) N-Nitroso-di-n-propyla...	8.425	70	490453	36.459	ng/u1	97
18) 4-Methylphenol	8.407	108	661755	35.297	ng/u1	99
19) Hexachloroethane	8.695	117	243420	33.692	ng/u1	99
22) Nitrobenzene	8.807	77	699907	33.035	ng/u1	97
23) Isophorone	9.331	82	1417716	34.684	ng/u1	99
25) 2-Nitrophenol	9.519	139	365057	34.732	ng/u1	99
26) 2,4-Dimethylphenol	9.595	107	731870	32.567	ng/u1	99
27) Bis(2-Chloroethoxy)met...	9.813	93	876561	32.336	ng/u1	100
29) 2,4-Dichlorophenol	10.060	162	630985	33.474	ng/u1	99
30) Naphthalene	10.448	128	2046452	31.570	ng/u1	100
32) 4-Chloroaniline	10.560	127	857169	30.228	ng/u1	98
33) Hexachlorobutadiene	10.754	225	387993	31.905	ng/u1	98
34) Caprolactam	11.313	113	218302	32.126	ng/u1	97
35) 4-Chloro-3-methylphenol	11.707	107	710996	34.278	ng/u1	99
36) 2-Methylnaphthalene	12.066	142	1433484	32.514	ng/u1	99


Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM101121\
 Data File : BM032444.D
 Acq On : 12 Oct 2021 06:15
 Operator : CG/JU
 Sample : PB139788BS
 Misc :
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SLCS788

Quant Time: Oct 12 06:46:45 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-BM100521.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Oct 11 11:06:33 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.289	142	1448476	32.264	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.454	216	745754	31.847	ng/ul	98
40) Hexachlorocyclopentadiene	12.442	237	426595	31.264	ng/ul	100
41) 2,4,6-Trichlorophenol	12.689	196	517962	34.539	ng/ul	99
42) 2,4,5-Trichlorophenol	12.766	196	562397	34.175	ng/ul	98
43) 1,1'-Biphenyl	13.089	154	1936276	31.801	ng/ul	99
44) 2-Chloronaphthalene	13.130	162	1470209	31.682	ng/ul	99
45) 2-Nitroaniline	13.330	65	442504	35.707	ng/ul	95
47) Dimethylphthalate	13.713	163	1879744	31.935	ng/ul	100
48) 2,6-Dinitrotoluene	13.830	165	396880	36.182	ng/ul	93
50) Acenaphthylene	13.977	152	2456738	32.521	ng/ul	99
51) 3-Nitroaniline	14.160	138	414678	36.167	ng/ul	100
52) Acenaphthene	14.324	153	1592088	31.916	ng/ul	99
53) 2,4-Dinitrophenol	14.365	184	233856	33.146	ng/ul	97
55) 4-Nitrophenol	14.489	109	292388	31.390	ng/ul	96
56) Dibenzofuran	14.660	168	2279712	31.535	ng/ul	100
57) 2,4-Dinitrotoluene	14.618	165	566994	34.971	ng/ul	98
58) 2,3,4,6-Tetrachlorophenol	14.895	232	469527	35.132	ng/ul	98
59) Diethylphthalate	15.083	149	1904193	32.428	ng/ul	100
61) Fluorene	15.307	166	1743565	31.390	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.295	204	882286	31.447	ng/ul	99
63) 4-Nitroaniline	15.324	138	443864	41.876	ng/ul	99
66) 4,6-Dinitro-2-methylph...	15.383	198	330338	34.063	ng/ul	98
67) N-Nitrosodiphenylamine	15.512	169	1584557	32.804	ng/ul	99
68) 4-Bromophenyl-phenylether	16.183	248	543726	32.586	ng/ul	99
69) Hexachlorobenzene	16.318	284	620238	32.271	ng/ul	99
70) Atrazine	16.459	200	585400	32.066	ng/ul	98
71) Pentachlorophenol	16.654	266	401465	34.651	ng/ul	99
72) Phenanthrene	17.030	178	2864712	32.059	ng/ul	99
74) Anthracene	17.118	178	2877523	32.056	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.060	216	763562	31.412	ng/ul	100
76) Pentachlorobenzene	14.595	250	740144	30.659	ng/ul	99
77) Carbazole	17.389	167	2662972	33.263	ng/ul	100
78) Di-n-butylphthalate	17.942	149	3207050	34.674	ng/ul	99
80) Fluoranthene	19.042	202	3299429	33.123	ng/ul	100
82) Pyrene	19.400	202	3345286	32.691	ng/ul	97
83) Butylbenzylphthalate	20.289	149	1445401	37.223	ng/ul	99
84) 3,3'-Dichlorobenzidine	21.071	252	1040255	35.749	ng/ul	100
85) Benzo(a)anthracene	21.136	228	3131760	32.361	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.059	149	2014694	37.322	ng/ul	100
87) Chrysene	21.189	228	3066815	32.184	ng/ul	100
89) Di-n-octyl phthalate	21.900	149	3642331	38.301	ng/ul	100
90) Benzo(b)fluoranthene	22.659	252	3401317	33.481	ng/ul	99
91) Benzo(k)fluoranthene	22.700	252	3009166	32.658	ng/ul	99
93) Benzo(a)pyrene	23.206	252	3176777	33.053	ng/ul	100
94) Indeno(1,2,3-cd)pyrene	25.424	276	3253141	30.621	ng/ul	98
95) Dibenzo(a,h)anthracene	25.424	278	2834502	31.094	ng/ul	99
96) Benzo(g,h,i)perylene	26.076	276	2758693	29.673	ng/ul	99





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 10.665 ng/uL
RT: 2.990 min Scan# 26
Delta R.T. -0.000 min
Lab File: BM032444.D
Acq: 12 Oct 2021 06:15




Instrument :
BNA_M
ClientSampleId :
SLCS788


Tgt Ion: 88 Resp: 78402
Ion Ratio Lower Upper
88 100
43 26.2 22.2 33.2
58 72.6 59.5 89.3



#5
Pyridine
Concen: 29.159 ng/uL
RT: 3.401 min Scan# 96
Delta R.T. -0.006 min
Lab File: BM032444.D
Acq: 12 Oct 2021 06:15

Tgt Ion: 79 Resp: 550169
Ion Ratio Lower Upper
79 100
52 70.8 59.4 89.0
51 33.5 28.1 42.1









#6
Benzaldehyde
Concen: 43.174 ng/ul
RT: 6.737 min Scan# 663
Delta R.T. -0.006 min
Lab File: BM032444.D
Acq: 12 Oct 2021 06:15




Instrument :
BNA_M
ClientSampled :
SLCS788


Tgt Ion: 77 Resp: 467954
Ion Ratio Lower Upper
77 100
105 100.2 79.5 119.3
106 97.0 75.7 113.5



#8
Phenol
Concen: 33.623 ng/ul
RT: 6.831 min Scan# 679
Delta R.T. -0.000 min
Lab File: BM032444.D
Acq: 12 Oct 2021 06:15

Tgt Ion: 94 Resp: 784250
Ion Ratio Lower Upper
94 100
65 26.6 22.3 33.5
66 36.1 29.5 44.3









#10

Bis(2-Chloroethyl)ether
 Concen: 32.787 ng/ul
 RT: 7.031 min Scan# 713
 Delta R.T. -0.006 min
 Lab File: BM032444.D
 Acq: 12 Oct 2021 06:15

Instrument :
 BNA_M
 ClientSampleId :
 SLCS788




Tgt Ion: 93	Resp: 601755
Ion Ratio	Lower Upper
93 100	
63 75.2	59.1 88.7
95 32.4	26.0 39.0




#12

2-Chlorophenol
 Concen: 33.775 ng/ul
 RT: 7.184 min Scan# 739
 Delta R.T. -0.006 min
 Lab File: BM032444.D
 Acq: 12 Oct 2021 06:15

Tgt Ion:128	Resp: 613410
Ion Ratio	Lower Upper
128 100	
64 43.6	36.2 54.4
130 32.4	26.6 39.8









#13
2-Methylphenol
Concen: 34.067 ng/ul
RT: 8.078 min Scan# 891
Delta R.T. -0.000 min
Lab File: BM032444.D
Acq: 12 Oct 2021 06:15




Instrument :
BNA_M
ClientSampleId :
SLCS788


Tgt Ion:108 Resp: 609363
Ion Ratio Lower Upper
108 100
107 90.0 71.0 106.6



#14
2,2'-oxybis(1-Chloropropane)
Concen: 33.367 ng/ul
RT: 8.148 min Scan# 903
Delta R.T. -0.000 min
Lab File: BM032444.D
Acq: 12 Oct 2021 06:15

Tgt Ion: 45 Resp: 758396
Ion Ratio Lower Upper
45 100
77 15.9 12.8 19.2
79 12.2 9.6 14.4









#16
Acetophenone
Concen: 35.166 ng/ul
RT: 8.431 min Scan# 951
Delta R.T. -0.006 min
Lab File: BM032444.D
Acq: 12 Oct 2021 06:15




Instrument :
BNA_M
ClientSampleId :
SLCS788


Tgt Ion:105 Resp: 958101
Ion Ratio Lower Upper
105 100
77 77.7 60.9 91.3
51 25.8 20.5 30.7



#17
N-Nitroso-di-n-propylamine
Concen: 36.459 ng/ul
RT: 8.425 min Scan# 950
Delta R.T. -0.000 min
Lab File: BM032444.D
Acq: 12 Oct 2021 06:15

Tgt Ion: 70 Resp: 490453
Ion Ratio Lower Upper
70 100
42 51.5 44.2 66.2
101 10.1 8.2 12.2
130 21.6 17.4 26.2









#18
4-Methylphenol
Concen: 35.297 ng/ul
RT: 8.407 min Scan# 947
Delta R.T. -0.000 min
Lab File: BM032444.D
Acq: 12 Oct 2021 06:15




Instrument :
BNA_M
ClientSampleId :
SLCS788


Tgt Ion:108 Resp: 661755
Ion Ratio Lower Upper
108 100
107 111.4 87.9 131.9



#19
Hexachloroethane
Concen: 33.692 ng/ul
RT: 8.695 min Scan# 996
Delta R.T. -0.006 min
Lab File: BM032444.D
Acq: 12 Oct 2021 06:15

Tgt Ion:117 Resp: 243420
Ion Ratio Lower Upper
117 100
201 98.3 78.9 118.3
199 60.5 49.0 73.6









#22
Nitrobenzene
Concen: 33.035 ng/ul
RT: 8.807 min Scan# 1015
Delta R.T. -0.006 min
Lab File: BM032444.D
Acq: 12 Oct 2021 06:15




Instrument :
BNA_M
ClientSampled :
SLCS788


Tgt Ion: 77 Resp: 699907
Ion Ratio Lower Upper
77 100
123 47.0 35.9 53.9
65 13.3 10.4 15.6



#23
Isophorone
Concen: 34.684 ng/ul
RT: 9.331 min Scan# 1104
Delta R.T. -0.000 min
Lab File: BM032444.D
Acq: 12 Oct 2021 06:15

Tgt Ion: 82 Resp: 1417716
Ion Ratio Lower Upper
82 100
95 7.0 5.8 8.6
138 17.8 13.9 20.9









#25
2-Nitrophenol
Concen: 34.732 ng/ul
RT: 9.519 min Scan# 1136
Delta R.T. -0.006 min
Lab File: BM032444.D
Acq: 12 Oct 2021 06:15




Instrument :
BNA_M
ClientSampled :
SLCS788


Tgt Ion:139 Resp: 365057
Ion Ratio Lower Upper
139 100
65 51.3 40.2 60.4
109 31.6 24.7 37.1



#26
2,4-Dimethylphenol
Concen: 32.567 ng/ul
RT: 9.595 min Scan# 1149
Delta R.T. -0.006 min
Lab File: BM032444.D
Acq: 12 Oct 2021 06:15

Tgt Ion:107 Resp: 731870
Ion Ratio Lower Upper
107 100
121 50.8 40.9 61.3
122 86.7 69.8 104.6









#27
 Bis(2-Chloroethoxy)methane
 Concen: 32.336 ng/ul
 RT: 9.813 min Scan# 1186
 Delta R.T. -0.000 min
 Lab File: BM032444.D
 Acq: 12 Oct 2021 06:15




Instrument :
 BNA_M
 ClientSampleId :
 SLCS788


Tgt Ion: 93 Resp: 876561
 Ion Ratio Lower Upper
 93 100
 95 32.8 26.2 39.2
 123 11.0 8.6 12.8



#29
 2,4-Dichlorophenol
 Concen: 33.474 ng/ul
 RT: 10.060 min Scan# 1228
 Delta R.T. -0.006 min
 Lab File: BM032444.D
 Acq: 12 Oct 2021 06:15

Tgt Ion: 162 Resp: 630985
 Ion Ratio Lower Upper
 162 100
 164 64.6 51.7 77.5
 98 36.2 29.8 44.6









#30
Naphthalene
Concen: 31.570 ng/ul
RT: 10.448 min Scan# 1294
Delta R.T. -0.006 min
Lab File: BM032444.D
Acq: 12 Oct 2021 06:15




Instrument :
BNA_M
ClientSampled :
SLCS788


Tgt Ion:128 Resp: 2046452
Ion Ratio Lower Upper
128 100
129 11.0 8.9 13.3
127 13.3 10.5 15.7



#32
4-Chloroaniline
Concen: 30.228 ng/ul
RT: 10.560 min Scan# 1313
Delta R.T. -0.006 min
Lab File: BM032444.D
Acq: 12 Oct 2021 06:15

Tgt Ion:127 Resp: 857169
Ion Ratio Lower Upper
127 100
129 32.7 25.1 37.7









#33
Hexachlorobutadiene
Concen: 31.905 ng/ul
RT: 10.754 min Scan# 1346
Delta R.T. -0.006 min
Lab File: BM032444.D
Acq: 12 Oct 2021 06:15




Instrument :
BNA_M
ClientSampled :
SLCS788


Tgt Ion:225 Resp: 387993
Ion Ratio Lower Upper
225 100
223 63.0 51.1 76.7
227 63.3 52.0 78.0



#34
Caprolactam
Concen: 32.126 ng/ul
RT: 11.313 min Scan# 1441
Delta R.T. 0.012 min
Lab File: BM032444.D
Acq: 12 Oct 2021 06:15

Tgt Ion:113 Resp: 218302
Ion Ratio Lower Upper
113 100
55 159.8 123.8 185.8
56 121.8 98.6 147.8









#35
4-Chloro-3-methylphenol
Concen: 34.278 ng/ul
RT: 11.707 min Scan# 1508
Delta R.T. -0.000 min
Lab File: BM032444.D
Acq: 12 Oct 2021 06:15




Instrument :
BNA_M
ClientSampleId :
SLCS788


Tgt Ion:107 Resp: 710996
Ion Ratio Lower Upper
107 100
144 25.1 20.4 30.6
142 78.6 62.4 93.6



#36
2-Methylnaphthalene
Concen: 32.514 ng/ul
RT: 12.066 min Scan# 1569
Delta R.T. -0.006 min
Lab File: BM032444.D
Acq: 12 Oct 2021 06:15

Tgt Ion:142 Resp: 1433484
Ion Ratio Lower Upper
142 100
141 89.9 71.2 106.8









#37
1-Methylnaphthalene
Concen: 32.264 ng/ul
RT: 12.289 min Scan# 1607
Delta R.T. -0.006 min
Lab File: BM032444.D
Acq: 12 Oct 2021 06:15




Instrument :
BNA_M
ClientSampleId :
SLCS788


Tgt Ion:142 Resp: 1448476
Ion Ratio Lower Upper
142 100
141 91.6 73.0 109.6
116 3.9 3.0 4.6



#39
1,2,4,5-Tetrachlorobenzene
Concen: 31.847 ng/ul
RT: 12.454 min Scan# 1635
Delta R.T. -0.000 min
Lab File: BM032444.D
Acq: 12 Oct 2021 06:15

Tgt Ion:216 Resp: 745754
Ion Ratio Lower Upper
216 100
214 77.6 61.8 92.6
179 20.6 17.2 25.8
108 17.2 15.7 23.5









#40
Hexachlorocyclopentadiene
Concen: 31.264 ng/ul
RT: 12.442 min Scan# 1633
Delta R.T. -0.006 min
Lab File: BM032444.D
Acq: 12 Oct 2021 06:15




Instrument :
BNA_M
ClientSampled :
SLCS788


Tgt Ion:237	Resp: 426595
Ion Ratio	Lower Upper
237 100	
235 61.7	49.4 74.0
272 12.3	9.8 14.8



#41
2,4,6-Trichlorophenol
Concen: 34.539 ng/ul
RT: 12.689 min Scan# 1675
Delta R.T. -0.006 min
Lab File: BM032444.D
Acq: 12 Oct 2021 06:15

Tgt Ion:196	Resp: 517962
Ion Ratio	Lower Upper
196 100	
198 96.7	75.9 113.9
200 31.1	24.7 37.1









#42
2,4,5-Trichlorophenol
Concen: 34.175 ng/ul
RT: 12.766 min Scan# 1688
Delta R.T. -0.006 min
Lab File: BM032444.D
Acq: 12 Oct 2021 06:15




Instrument :
BNA_M
ClientSampled :
SLCS788


Tgt Ion:196 Resp: 562397
Ion Ratio Lower Upper
196 100
198 97.3 76.4 114.6
200 31.6 25.0 37.4



#43
1,1'-Biphenyl
Concen: 31.801 ng/ul
RT: 13.089 min Scan# 1743
Delta R.T. -0.006 min
Lab File: BM032444.D
Acq: 12 Oct 2021 06:15

Tgt Ion:154 Resp: 1936276
Ion Ratio Lower Upper
154 100
153 40.5 32.1 48.1
76 12.4 10.2 15.2









#44
2-Chloronaphthalene
Concen: 31.682 ng/ul
RT: 13.130 min Scan# 1750
Delta R.T. -0.000 min
Lab File: BM032444.D
Acq: 12 Oct 2021 06:15




Instrument :
BNA_M
ClientSampled :
SLCS788


Tgt Ion:162 Resp: 1470209
Ion Ratio Lower Upper
162 100
164 33.6 26.6 39.8
127 39.9 31.7 47.5



#45
2-Nitroaniline
Concen: 35.707 ng/ul
RT: 13.330 min Scan# 1784
Delta R.T. -0.006 min
Lab File: BM032444.D
Acq: 12 Oct 2021 06:15

Tgt Ion: 65 Resp: 442504
Ion Ratio Lower Upper
65 100
92 71.4 54.6 81.8
138 111.0 84.3 126.5









#47
Dimethylphthalate
Concen: 31.935 ng/ul
RT: 13.713 min Scan# 1849
Delta R.T. -0.006 min
Lab File: BM032444.D
Acq: 12 Oct 2021 06:15




Instrument :
BNA_M
ClientSampleId :
SLCS788


Tgt Ion:163 Resp: 1879744
Ion Ratio Lower Upper
163 100
194 4.0 3.1 4.7
164 10.3 8.1 12.1



#48
2,6-Dinitrotoluene
Concen: 36.182 ng/ul
RT: 13.830 min Scan# 1869
Delta R.T. -0.000 min
Lab File: BM032444.D
Acq: 12 Oct 2021 06:15

Tgt Ion:165 Resp: 396880
Ion Ratio Lower Upper
165 100
89 54.6 48.6 73.0
121 20.2 18.3 27.5









#50
Acenaphthylene
Concen: 32.521 ng/ul
RT: 13.977 min Scan# 1894
Delta R.T. -0.006 min
Lab File: BM032444.D
Acq: 12 Oct 2021 06:15




Instrument :
BNA_M
ClientSampleId :
SLCS788


Tgt Ion:152 Resp: 2456738
Ion Ratio Lower Upper
152 100
151 20.0 15.7 23.5
153 13.4 10.6 16.0



#51
3-Nitroaniline
Concen: 36.167 ng/ul
RT: 14.160 min Scan# 1925
Delta R.T. -0.000 min
Lab File: BM032444.D
Acq: 12 Oct 2021 06:15

Tgt Ion:138 Resp: 414678
Ion Ratio Lower Upper
138 100
108 10.8 8.9 13.3
92 116.8 93.3 139.9









#52
Acenaphthene
Concen: 31.916 ng/ul
RT: 14.324 min Scan# 1953
Delta R.T. -0.006 min
Lab File: BM032444.D
Acq: 12 Oct 2021 06:15




Instrument :
BNA_M
ClientSampleId :
SLCS788


Tgt Ion:153 Resp: 1592088
Ion Ratio Lower Upper
153 100
152 47.1 38.2 57.2
154 87.6 70.8 106.2



#53
2,4-Dinitrophenol
Concen: 33.146 ng/ul
RT: 14.365 min Scan# 1960
Delta R.T. -0.000 min
Lab File: BM032444.D
Acq: 12 Oct 2021 06:15

Tgt Ion:184 Resp: 233856
Ion Ratio Lower Upper
184 100
63 59.2 50.6 75.8
154 63.0 50.7 76.1









#55
4-Nitrophenol
Concen: 31.390 ng/ul
RT: 14.489 min Scan# 1981
Delta R.T. -0.000 min
Lab File: BM032444.D
Acq: 12 Oct 2021 06:15




Instrument :
BNA_M
ClientSampled :
SLCS788


Tgt Ion:109 Resp: 292388
Ion Ratio Lower Upper
109 100
139 135.0 101.0 151.6
65 127.3 102.6 154.0



#56
Dibenzofuran
Concen: 31.535 ng/ul
RT: 14.660 min Scan# 2010
Delta R.T. -0.000 min
Lab File: BM032444.D
Acq: 12 Oct 2021 06:15

Tgt Ion:168 Resp: 2279712
Ion Ratio Lower Upper
168 100
139 37.4 30.0 45.0









#57
2,4-Dinitrotoluene
Concen: 34.971 ng/ul
RT: 14.618 min Scan# 2003
Delta R.T. -0.000 min
Lab File: BM032444.D
Acq: 12 Oct 2021 06:15




Instrument :
BNA_M
ClientSampleId :
SLCS788


Tgt Ion:165 Resp: 566994
Ion Ratio Lower Upper
165 100
63 41.6 34.4 51.6
182 2.7 2.2 3.4



#58
2,3,4,6-Tetrachlorophenol
Concen: 35.132 ng/ul
RT: 14.895 min Scan# 2050
Delta R.T. -0.006 min
Lab File: BM032444.D
Acq: 12 Oct 2021 06:15

Tgt Ion:232 Resp: 469527
Ion Ratio Lower Upper
232 100
131 45.2 37.4 56.2
130 2.5 1.8 2.8
166 31.2 25.5 38.3









#59
Diethylphthalate
Concen: 32.428 ng/ul
RT: 15.083 min Scan# 2082
Delta R.T. -0.000 min
Lab File: BM032444.D
Acq: 12 Oct 2021 06:15




Instrument :
BNA_M
ClientSampleId :
SLCS788


Tgt Ion:149 Resp: 1904193
Ion Ratio Lower Upper
149 100
177 22.6 17.9 26.9
150 12.5 10.0 15.0



#61
Fluorene
Concen: 31.390 ng/ul
RT: 15.307 min Scan# 2120
Delta R.T. -0.000 min
Lab File: BM032444.D
Acq: 12 Oct 2021 06:15

Tgt Ion:166 Resp: 1743565
Ion Ratio Lower Upper
166 100
165 97.9 77.0 115.6
167 13.5 10.9 16.3









#62
4-Chlorophenyl-phenylether
Concen: 31.447 ng/ul
RT: 15.295 min Scan# 2118
Delta R.T. -0.006 min
Lab File: BM032444.D
Acq: 12 Oct 2021 06:15




Instrument :
BNA_M
ClientSampled :
SLCS788


Tgt Ion:204 Resp: 882286
Ion Ratio Lower Upper
204 100
206 32.4 26.2 39.4
141 61.1 48.6 73.0



#63
4-Nitroaniline
Concen: 41.876 ng/ul
RT: 15.324 min Scan# 2123
Delta R.T. -0.000 min
Lab File: BM032444.D
Acq: 12 Oct 2021 06:15

Tgt Ion:138 Resp: 443864
Ion Ratio Lower Upper
138 100
92 51.8 41.3 61.9
108 80.1 64.6 97.0









#66
4,6-Dinitro-2-methylphenol
Concen: 34.063 ng/ul
RT: 15.383 min Scan# 2133
Delta R.T. -0.006 min
Lab File: BM032444.D
Acq: 12 Oct 2021 06:15




Instrument :
BNA_M
ClientSampled :
SLCS788


Tgt Ion:198 Resp: 330338
Ion Ratio Lower Upper
198 100
182 4.3 4.1 6.1
77 25.9 21.4 32.0



#67
N-Nitrosodiphenylamine
Concen: 32.804 ng/ul
RT: 15.512 min Scan# 2155
Delta R.T. -0.000 min
Lab File: BM032444.D
Acq: 12 Oct 2021 06:15

Tgt Ion:169 Resp: 1584557
Ion Ratio Lower Upper
169 100
168 66.1 53.3 79.9
167 35.7 29.3 43.9









#68
4-Bromophenyl-phenylether
Concen: 32.586 ng/ul
RT: 16.183 min Scan# 2269
Delta R.T. -0.006 min
Lab File: BM032444.D
Acq: 12 Oct 2021 06:15




Instrument :
BNA_M
ClientSampleId :
SLCS788


Tgt Ion:248	Resp: 543726
Ion Ratio	Lower Upper
248	100
250	97.5 77.0 115.4
141	78.9 63.7 95.5



#69
Hexachlorobenzene
Concen: 32.271 ng/ul
RT: 16.318 min Scan# 2292
Delta R.T. -0.006 min
Lab File: BM032444.D
Acq: 12 Oct 2021 06:15

Tgt Ion:284	Resp: 620238
Ion Ratio	Lower Upper
284	100
142	37.4 29.4 44.0
249	32.6 25.5 38.3









#70
Atrazine
Concen: 32.066 ng/ul
RT: 16.459 min Scan# 2316
Delta R.T. -0.000 min
Lab File: BM032444.D
Acq: 12 Oct 2021 06:15




Instrument :
BNA_M
ClientSampled :
SLCS788


Tgt Ion:200 Resp: 585400
Ion Ratio Lower Upper
200 100
173 26.4 22.2 33.4
215 48.0 37.5 56.3



#71
Pentachlorophenol
Concen: 34.651 ng/ul
RT: 16.654 min Scan# 2349
Delta R.T. -0.006 min
Lab File: BM032444.D
Acq: 12 Oct 2021 06:15

Tgt Ion:266 Resp: 401465
Ion Ratio Lower Upper
266 100
264 63.5 49.9 74.9
268 63.8 50.5 75.7









#72
Phenanthrene
Concen: 32.059 ng/u1
RT: 17.030 min Scan# 2413
Delta R.T. -0.006 min
Lab File: BM032444.D
Acq: 12 Oct 2021 06:15




Instrument :
BNA_M
ClientSampleId :
SLCS788


Tgt Ion:178 Resp: 2864712
Ion Ratio Lower Upper
178 100
179 15.5 12.2 18.2
176 19.4 15.4 23.0



#74
Anthracene
Concen: 32.056 ng/u1
RT: 17.118 min Scan# 2428
Delta R.T. -0.006 min
Lab File: BM032444.D
Acq: 12 Oct 2021 06:15

Tgt Ion:178 Resp: 2877523
Ion Ratio Lower Upper
178 100
179 15.7 12.3 18.5
176 18.7 14.9 22.3









#75
1,2,3,4-Tetrachlorobenzene
Concen: 31.412 ng/uL
RT: 13.060 min Scan# 1738
Delta R.T. -0.006 min
Lab File: BM032444.D
Acq: 12 Oct 2021 06:15




Instrument :
BNA_M
ClientSampleId :
SLCS788


Tgt Ion:216 Resp: 763562
Ion Ratio Lower Upper
216 100
214 79.0 62.8 94.2
218 48.0 38.4 57.6



#76
Pentachlorobenzene
Concen: 30.659 ng/uL
RT: 14.595 min Scan# 1999
Delta R.T. -0.000 min
Lab File: BM032444.D
Acq: 12 Oct 2021 06:15

Tgt Ion:250 Resp: 740144
Ion Ratio Lower Upper
250 100
248 62.1 50.7 76.1
252 64.0 51.0 76.6









#77
 Carbazole
 Concen: 33.263 ng/ul
 RT: 17.389 min Scan# 2474
 Delta R.T. -0.000 min
 Lab File: BM032444.D
 Acq: 12 Oct 2021 06:15




Instrument :
 BNA_M
 ClientSampleId :
 SLCS788


Tgt Ion:167 Resp: 2662972
 Ion Ratio Lower Upper
 167 100
 166 21.8 17.4 26.2
 139 13.1 10.6 15.8



#78
 Di-n-butylphthalate
 Concen: 34.674 ng/ul
 RT: 17.942 min Scan# 2568
 Delta R.T. -0.006 min
 Lab File: BM032444.D
 Acq: 12 Oct 2021 06:15

Tgt Ion:149 Resp: 3207050
 Ion Ratio Lower Upper
 149 100
 150 9.5 7.4 11.0
 104 5.2 4.2 6.2









#80
Fluoranthene
Concen: 33.123 ng/ul
RT: 19.042 min Scan# 2755
Delta R.T. -0.000 min
Lab File: BM032444.D
Acq: 12 Oct 2021 06:15




Instrument :
BNA_M
ClientSampleId :
SLCS788


Tgt Ion:202 Resp: 3299429
Ion Ratio Lower Upper
202 100
101 11.2 9.1 13.7
100 8.6 7.0 10.4



#82
Pyrene
Concen: 32.691 ng/ul
RT: 19.400 min Scan# 2816
Delta R.T. -0.006 min
Lab File: BM032444.D
Acq: 12 Oct 2021 06:15

Tgt Ion:202 Resp: 3345286
Ion Ratio Lower Upper
202 100
101 14.0 10.3 15.5
100 11.4 8.1 12.1









#83
Butylbenzylphthalate
Concen: 37.223 ng/ul
RT: 20.289 min Scan# 2967
Delta R.T. -0.006 min
Lab File: BM032444.D
Acq: 12 Oct 2021 06:15




Instrument :
BNA_M
ClientSampled :
SLCS788


Tgt Ion:149 Resp: 1445401
Ion Ratio Lower Upper
149 100
91 71.6 56.3 84.5
206 20.5 16.8 25.2



#84
3,3'-Dichlorobenzidine
Concen: 35.749 ng/ul
RT: 21.071 min Scan# 3100
Delta R.T. -0.006 min
Lab File: BM032444.D
Acq: 12 Oct 2021 06:15

Tgt Ion:252 Resp: 1040255
Ion Ratio Lower Upper
252 100
254 63.8 50.9 76.3
126 11.9 9.0 13.4









#85
Benzo(a)anthracene
Concen: 32.361 ng/ul
RT: 21.136 min Scan# 3111
Delta R.T. -0.006 min
Lab File: BM032444.D
Acq: 12 Oct 2021 06:15




Instrument :
BNA_M
ClientSampled :
SLCS788


Tgt Ion:228 Resp: 3131760
Ion Ratio Lower Upper
228 100
229 20.2 15.6 23.4
226 27.8 21.8 32.6



#86
Bis(2-ethylhexyl)phthalate
Concen: 37.322 ng/ul
RT: 21.059 min Scan# 3098
Delta R.T. -0.006 min
Lab File: BM032444.D
Acq: 12 Oct 2021 06:15

Tgt Ion:149 Resp: 2014694
Ion Ratio Lower Upper
149 100
167 27.5 22.0 33.0
279 3.8 3.3 4.9









#87
Chrysene
Concen: 32.184 ng/ul
RT: 21.189 min Scan# 3120
Delta R.T. -0.006 min
Lab File: BM032444.D
Acq: 12 Oct 2021 06:15




Instrument :
BNA_M
ClientSampleId :
SLCS788


Tgt Ion:228 Resp: 3066815
Ion Ratio Lower Upper
228 100
226 30.1 24.1 36.1
229 19.9 15.5 23.3



#89
Di-n-octyl phthalate
Concen: 38.301 ng/ul
RT: 21.900 min Scan# 3241
Delta R.T. -0.006 min
Lab File: BM032444.D
Acq: 12 Oct 2021 06:15

Tgt Ion:149 Resp: 3642331









#90
Benzo(b)fluoranthene
Concen: 33.481 ng/ul
RT: 22.659 min Scan# 3370
Delta R.T. -0.006 min
Lab File: BM032444.D
Acq: 12 Oct 2021 06:15




Instrument :
BNA_M
ClientSampled :
SLCS788


Tgt Ion:252 Resp: 3401317
Ion Ratio Lower Upper
252 100
253 22.3 17.6 26.4
125 10.9 8.2 12.4



#91
Benzo(k)fluoranthene
Concen: 32.658 ng/ul
RT: 22.700 min Scan# 3377
Delta R.T. -0.006 min
Lab File: BM032444.D
Acq: 12 Oct 2021 06:15

Tgt Ion:252 Resp: 3009166
Ion Ratio Lower Upper
252 100
253 22.3 17.5 26.3
125 11.0 8.4 12.6









#93
Benzo(a)pyrene
Concen: 33.053 ng/ul
RT: 23.206 min Scan# 3463
Delta R.T. -0.006 min
Lab File: BM032444.D
Acq: 12 Oct 2021 06:15




Instrument :
BNA_M
ClientSampled :
SLCS788


Tgt Ion:252 Resp: 3176777
Ion Ratio Lower Upper
252 100
253 22.0 17.8 26.6
125 11.5 9.2 13.8



#94
Indeno(1,2,3-cd)pyrene
Concen: 30.621 ng/ul
RT: 25.424 min Scan# 3840
Delta R.T. -0.006 min
Lab File: BM032444.D
Acq: 12 Oct 2021 06:15

Tgt Ion:276 Resp: 3253141
Ion Ratio Lower Upper
276 100
138 28.9 21.8 32.6
277 26.0 20.4 30.6









#95
Dibenzo(a,h)anthracene
Concen: 31.094 ng/ul
RT: 25.424 min Scan# 3840
Delta R.T. -0.012 min
Lab File: BM032444.D
Acq: 12 Oct 2021 06:15

Instrument :
BNA_M
ClientSampled :
SLCS788

Tgt Ion:278 Resp: 2834502
Ion Ratio Lower Upper
278 100
139 17.6 13.4 20.2
279 23.6 19.0 28.4



#96
Benzo(g,h,i)perylene
Concen: 29.673 ng/ul
RT: 26.076 min Scan# 3951
Delta R.T. -0.006 min
Lab File: BM032444.D
Acq: 12 Oct 2021 06:15

Tgt Ion:276 Resp: 2758693
Ion Ratio Lower Upper
276 100
138 23.7 18.6 27.8
277 23.6 19.0 28.4

