

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM101121\
 Data File : BM032445.D
 Acq On : 12 Oct 2021 06:51
 Operator : CG/JU
 Sample : PB139789BS
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SLCS789

Quant Time: Oct 12 07:23:25 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-BM100521.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Oct 11 11:06:33 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.613	152	251535	20.000	ng/u1	0.00
20) Naphthalene-d8	10.401	136	1164722	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.260	164	780342	20.000	ng/u1	0.00
64) Phenanthrene-d10	16.989	188	1593951	20.000	ng/u1	0.00
79) Chrysene-d12	21.153	240	1498485	20.000	ng/u1	0.00
88) Perylene-d12	23.294	264	1547130	20.000	ng/u1	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	2.955	96	34119	5.438	ng/uL	0.00
4) Pyridine-d5	3.384	84	479321	27.045	ng/u1	0.00
7) Phenol-d5	6.801	99	676092	31.609	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	6.937	67	380020	30.943	ng/u1	0.00
11) 2-Chlorophenol-d4	7.148	132	529646	32.432	ng/u1	0.00
15) 4-Methylphenol-d8	8.342	113	566941	32.817	ng/u1	0.00
21) Nitrobenzene-d5	8.766	128	275026	33.467	ng/u1	0.00
24) 2-Nitrophenol-d4	9.489	143	309970	34.993	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.036	165	570971	32.380	ng/u1	0.00
31) 4-Chloroaniline-d4	10.536	131	722107	27.763	ng/u1	0.00
46) Dimethylphthalate-d6	13.666	166	1718220	31.816	ng/u1	0.00
49) Acenaphthylene-d8	13.948	160	2112751	32.204	ng/u1	0.00
54) 4-Nitrophenol-d4	14.471	143	330845	31.195	ng/u1	0.00
60) Fluorene-d10	15.248	176	1440388	31.449	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.371	200	296041	32.589	ng/u1	0.00
73) Anthracene-d10	17.083	188	2209616	31.846	ng/u1	0.00
81) Pyrene-d10	19.371	212	2435855	32.312	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.165	264	2381513	30.738	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	2.990	88	72597	10.631	ng/uL	98
5) Pyridine	3.402	79	510761	29.140	ng/u1	96
6) Benzaldehyde	6.737	77	436622	43.365	ng/u1	97
8) Phenol	6.831	94	723103	33.372	ng/u1	98
10) Bis(2-Chloroethyl)ether	7.031	93	561419	32.929	ng/u1	100
12) 2-Chlorophenol	7.184	128	571971	33.902	ng/u1	100
13) 2-Methylphenol	8.078	108	566812	34.112	ng/u1	100
14) 2,2'-oxybis(1-Chloropr...	8.148	45	697477	33.034	ng/u1	99
16) Acetophenone	8.431	105	898056	35.483	ng/u1	99
17) N-Nitroso-di-n-propyla...	8.425	70	452950	36.247	ng/u1	98
18) 4-Methylphenol	8.407	108	621233	35.670	ng/u1	99
19) Hexachloroethane	8.695	117	228025	33.975	ng/u1	99
22) Nitrobenzene	8.807	77	644441	33.030	ng/u1	98
23) Isophorone	9.331	82	1328004	35.281	ng/u1	99
25) 2-Nitrophenol	9.519	139	348172	35.971	ng/u1	98
26) 2,4-Dimethylphenol	9.595	107	698243	33.740	ng/u1	98
27) Bis(2-Chloroethoxy)met...	9.813	93	820497	32.868	ng/u1	99
29) 2,4-Dichlorophenol	10.060	162	589183	33.942	ng/u1	99
30) Naphthalene	10.448	128	1924817	32.245	ng/u1	100
32) 4-Chloroaniline	10.560	127	765815	29.326	ng/u1	99
33) Hexachlorobutadiene	10.754	225	362019	32.326	ng/u1	98
34) Caprolactam	11.313	113	205924	32.908	ng/u1	97
35) 4-Chloro-3-methylphenol	11.707	107	671363	35.148	ng/u1	100
36) 2-Methylnaphthalene	12.066	142	1343372	33.088	ng/u1	100


Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM101121\
 Data File : BM032445.D
 Acq On : 12 Oct 2021 06:51
 Operator : CG/JU
 Sample : PB139789BS
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SLCS789

Quant Time: Oct 12 07:23:25 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-BM100521.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Oct 11 11:06:33 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.289	142	1365884	33.038	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.448	216	699956	32.530	ng/ul	99
40) Hexachlorocyclopentadiene	12.442	237	406331	32.407	ng/ul	99
41) 2,4,6-Trichlorophenol	12.689	196	490732	35.612	ng/ul	99
42) 2,4,5-Trichlorophenol	12.766	196	532883	35.240	ng/ul	99
43) 1,1'-Biphenyl	13.089	154	1813350	32.411	ng/ul	99
44) 2-Chloronaphthalene	13.130	162	1385264	32.487	ng/ul	100
45) 2-Nitroaniline	13.330	65	420171	36.898	ng/ul	95
47) Dimethylphthalate	13.713	163	1784906	33.001	ng/ul	99
48) 2,6-Dinitrotoluene	13.830	165	374463	37.151	ng/ul	91
50) Acenaphthylene	13.977	152	2304403	33.197	ng/ul	99
51) 3-Nitroaniline	14.160	138	373408	35.442	ng/ul	99
52) Acenaphthene	14.324	153	1492418	32.559	ng/ul	99
53) 2,4-Dinitrophenol	14.360	184	220333	33.986	ng/ul	97
55) 4-Nitrophenol	14.489	109	276772	32.337	ng/ul	94
56) Dibenzofuran	14.660	168	2127403	32.026	ng/ul	100
57) 2,4-Dinitrotoluene	14.618	165	538208	36.126	ng/ul	97
58) 2,3,4,6-Tetrachlorophenol	14.895	232	444069	36.161	ng/ul	97
59) Diethylphthalate	15.083	149	1799542	33.351	ng/ul	100
61) Fluorene	15.307	166	1640728	32.145	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.295	204	825197	32.008	ng/ul	99
63) 4-Nitroaniline	15.324	138	414445	42.552	ng/ul	99
66) 4,6-Dinitro-2-methylph...	15.383	198	309481	34.663	ng/ul	99
67) N-Nitrosodiphenylamine	15.513	169	1492461	33.560	ng/ul	99
68) 4-Bromophenyl-phenylether	16.183	248	514115	33.466	ng/ul	99
69) Hexachlorobenzene	16.318	284	580687	32.817	ng/ul	98
70) Atrazine	16.454	200	554767	33.006	ng/ul	99
71) Pentachlorophenol	16.654	266	379198	35.550	ng/ul	99
72) Phenanthrene	17.030	178	2681942	32.600	ng/ul	100
74) Anthracene	17.118	178	2715543	32.859	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.060	216	715829	31.986	ng/ul	100
76) Pentachlorobenzene	14.589	250	701319	31.554	ng/ul	100
77) Carbazole	17.383	167	2487762	33.752	ng/ul	99
78) Di-n-butylphthalate	17.942	149	3049715	35.815	ng/ul	100
80) Fluoranthene	19.036	202	3117004	33.512	ng/ul	99
82) Pyrene	19.401	202	3152465	32.992	ng/ul	97
83) Butylbenzylphthalate	20.289	149	1352144	37.292	ng/ul	100
84) 3,3'-Dichlorobenzidine	21.071	252	949417	34.942	ng/ul	99
85) Benzo(a)anthracene	21.136	228	2969493	32.861	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.059	149	1906429	37.822	ng/ul	100
87) Chrysene	21.189	228	2894828	32.535	ng/ul	100
89) Di-n-octyl phthalate	21.900	149	3479665	39.491	ng/ul	100
90) Benzo(b)fluoranthene	22.659	252	3234887	34.367	ng/ul	99
91) Benzo(k)fluoranthene	22.700	252	2813228	32.952	ng/ul	99
93) Benzo(a)pyrene	23.206	252	2995822	33.641	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	25.424	276	3110551	31.600	ng/ul	99
95) Dibenzo(a,h)anthracene	25.430	278	2704336	32.018	ng/ul	100
96) Benzo(g,h,i)perylene	26.077	276	2630973	30.543	ng/ul	100





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 10.631 ng/uL
RT: 2.990 min Scan# 26
Delta R.T. -0.000 min
Lab File: BM032445.D
Acq: 12 Oct 2021 06:51




Instrument :
BNA_M
ClientSampleId :
SLCS789


Tgt Ion	Ratio	Lower	Upper
88	100		
43	26.1	22.2	33.2
58	72.5	59.5	89.3



#5
Pyridine
Concen: 29.140 ng/uL
RT: 3.402 min Scan# 96
Delta R.T. -0.006 min
Lab File: BM032445.D
Acq: 12 Oct 2021 06:51

Tgt Ion	Ratio	Lower	Upper
79	100		
52	70.7	59.4	89.0
51	32.8	28.1	42.1









#6
Benzaldehyde
Concen: 43.365 ng/ul
RT: 6.737 min Scan# 663
Delta R.T. -0.006 min
Lab File: BM032445.D
Acq: 12 Oct 2021 06:51




Instrument :
BNA_M
ClientSampleId :
SLCS789


Tgt Ion: 77 Resp: 436622
Ion Ratio Lower Upper
77 100
105 100.9 79.5 119.3
106 98.8 75.7 113.5



#8
Phenol
Concen: 33.372 ng/ul
RT: 6.831 min Scan# 679
Delta R.T. -0.000 min
Lab File: BM032445.D
Acq: 12 Oct 2021 06:51

Tgt Ion: 94 Resp: 723103
Ion Ratio Lower Upper
94 100
65 26.6 22.3 33.5
66 35.9 29.5 44.3









#10
Bis(2-Chloroethyl)ether
Concen: 32.929 ng/ul
RT: 7.031 min Scan# 713
Delta R.T. -0.006 min
Lab File: BM032445.D
Acq: 12 Oct 2021 06:51




Instrument :
BNA_M
ClientSampleId :
SLCS789


Tgt Ion: 93 Resp: 561419
Ion Ratio Lower Upper
93 100
63 73.6 59.1 88.7
95 32.5 26.0 39.0



#12
2-Chlorophenol
Concen: 33.902 ng/ul
RT: 7.184 min Scan# 739
Delta R.T. -0.006 min
Lab File: BM032445.D
Acq: 12 Oct 2021 06:51

Tgt Ion: 128 Resp: 571971
Ion Ratio Lower Upper
128 100
64 45.3 36.2 54.4
130 32.9 26.6 39.8









#13
2-Methylphenol
Concen: 34.112 ng/ul
RT: 8.078 min Scan# 891
Delta R.T. -0.000 min
Lab File: BM032445.D
Acq: 12 Oct 2021 06:51




Instrument :
BNA_M
ClientSampleId :
SLCS789


Tgt Ion:108 Resp: 566812
Ion Ratio Lower Upper
108 100
107 88.9 71.0 106.6



#14
2,2'-oxybis(1-Chloropropane)
Concen: 33.034 ng/ul
RT: 8.148 min Scan# 903
Delta R.T. -0.000 min
Lab File: BM032445.D
Acq: 12 Oct 2021 06:51

Tgt Ion: 45 Resp: 697477
Ion Ratio Lower Upper
45 100
77 15.7 12.8 19.2
79 12.2 9.6 14.4









#16
Acetophenone
Concen: 35.483 ng/ul
RT: 8.431 min Scan# 951
Delta R.T. -0.006 min
Lab File: BM032445.D
Acq: 12 Oct 2021 06:51




Instrument :
BNA_M
ClientSampleId :
SLCS789


Tgt Ion:105 Resp: 898056
Ion Ratio Lower Upper
105 100
77 75.9 60.9 91.3
51 24.7 20.5 30.7



#17
N-Nitroso-di-n-propylamine
Concen: 36.247 ng/ul
RT: 8.425 min Scan# 950
Delta R.T. -0.000 min
Lab File: BM032445.D
Acq: 12 Oct 2021 06:51

Tgt Ion: 70 Resp: 452950
Ion Ratio Lower Upper
70 100
42 53.0 44.2 66.2
101 10.4 8.2 12.2
130 22.3 17.4 26.2









#18
4-Methylphenol
Concen: 35.670 ng/ul
RT: 8.407 min Scan# 947
Delta R.T. -0.000 min
Lab File: BM032445.D
Acq: 12 Oct 2021 06:51




Instrument :
BNA_M
ClientSampleId :
SLCS789


Tgt Ion:108 Resp: 621233
Ion Ratio Lower Upper
108 100
107 110.6 87.9 131.9



#19
Hexachloroethane
Concen: 33.975 ng/ul
RT: 8.695 min Scan# 996
Delta R.T. -0.006 min
Lab File: BM032445.D
Acq: 12 Oct 2021 06:51

Tgt Ion:117 Resp: 228025
Ion Ratio Lower Upper
117 100
201 99.7 78.9 118.3
199 62.2 49.0 73.6









#22
Nitrobenzene
Concen: 33.030 ng/ul
RT: 8.807 min Scan# 1015
Delta R.T. -0.006 min
Lab File: BM032445.D
Acq: 12 Oct 2021 06:51




Instrument :
BNA_M
ClientSampleId :
SLCS789


Tgt Ion: 77 Resp: 644441
Ion Ratio Lower Upper
77 100
123 46.8 35.9 53.9
65 13.3 10.4 15.6



#23
Isophorone
Concen: 35.281 ng/ul
RT: 9.331 min Scan# 1104
Delta R.T. -0.000 min
Lab File: BM032445.D
Acq: 12 Oct 2021 06:51

Tgt Ion: 82 Resp: 1328004
Ion Ratio Lower Upper
82 100
95 7.3 5.8 8.6
138 18.0 13.9 20.9









#25
2-Nitrophenol
Concen: 35.971 ng/ul
RT: 9.519 min Scan# 1136
Delta R.T. -0.006 min
Lab File: BM032445.D
Acq: 12 Oct 2021 06:51




Instrument :
BNA_M
ClientSampleId :
SLCS789


Tgt Ion:139 Resp: 348172
Ion Ratio Lower Upper
139 100
65 48.4 40.2 60.4
109 30.5 24.7 37.1



#26
2,4-Dimethylphenol
Concen: 33.740 ng/ul
RT: 9.595 min Scan# 1149
Delta R.T. -0.006 min
Lab File: BM032445.D
Acq: 12 Oct 2021 06:51

Tgt Ion:107 Resp: 698243
Ion Ratio Lower Upper
107 100
121 50.5 40.9 61.3
122 84.2 69.8 104.6









#27
 Bis(2-Chloroethoxy)methane
 Concen: 32.868 ng/ul
 RT: 9.813 min Scan# 1186
 Delta R.T. -0.000 min
 Lab File: BM032445.D
 Acq: 12 Oct 2021 06:51




Instrument :
 BNA_M
 ClientSampleId :
 SLCS789


Tgt Ion: 93	Resp: 820497		
Ion Ratio	Lower	Upper	
93	100		
95	33.1	26.2	39.2
123	11.3	8.6	12.8



#29
 2,4-Dichlorophenol
 Concen: 33.942 ng/ul
 RT: 10.060 min Scan# 1228
 Delta R.T. -0.006 min
 Lab File: BM032445.D
 Acq: 12 Oct 2021 06:51

Tgt Ion:162	Resp: 589183
Ion Ratio	Lower Upper
162 100	
164 65.2	51.7 77.5
98 36.8	29.8 44.6









#30
Naphthalene
Concen: 32.245 ng/ul
RT: 10.448 min Scan# 1294
Delta R.T. -0.006 min
Lab File: BM032445.D
Acq: 12 Oct 2021 06:51




Instrument :
BNA_M
ClientSampleId :
SLCS789


Tgt Ion:128 Resp: 1924817
Ion Ratio Lower Upper
128 100
129 11.2 8.9 13.3
127 13.3 10.5 15.7



#32
4-Chloroaniline
Concen: 29.326 ng/ul
RT: 10.560 min Scan# 1313
Delta R.T. -0.006 min
Lab File: BM032445.D
Acq: 12 Oct 2021 06:51

Tgt Ion:127 Resp: 765815
Ion Ratio Lower Upper
127 100
129 31.9 25.1 37.7









#33
Hexachlorobutadiene
Concen: 32.326 ng/ul
RT: 10.754 min Scan# 1346
Delta R.T. -0.006 min
Lab File: BM032445.D
Acq: 12 Oct 2021 06:51




Instrument :
BNA_M
ClientSampleId :
SLCS789


Tgt Ion:225 Resp: 362019
Ion Ratio Lower Upper
225 100
223 62.8 51.1 76.7
227 62.6 52.0 78.0



#34
Caprolactam
Concen: 32.908 ng/ul
RT: 11.313 min Scan# 1441
Delta R.T. 0.012 min
Lab File: BM032445.D
Acq: 12 Oct 2021 06:51

Tgt Ion:113 Resp: 205924
Ion Ratio Lower Upper
113 100
55 151.4 123.8 185.8
56 118.4 98.6 147.8









#35
4-Chloro-3-methylphenol
Concen: 35.148 ng/ul
RT: 11.707 min Scan# 1508
Delta R.T. -0.000 min
Lab File: BM032445.D
Acq: 12 Oct 2021 06:51




Instrument :
BNA_M
ClientSampleId :
SLCS789


Tgt Ion:107 Resp: 671363
Ion Ratio Lower Upper
107 100
144 25.6 20.4 30.6
142 78.3 62.4 93.6



#36
2-Methylnaphthalene
Concen: 33.088 ng/ul
RT: 12.066 min Scan# 1569
Delta R.T. -0.006 min
Lab File: BM032445.D
Acq: 12 Oct 2021 06:51

Tgt Ion:142 Resp: 1343372
Ion Ratio Lower Upper
142 100
141 89.3 71.2 106.8









#37
1-Methylnaphthalene
Concen: 33.038 ng/ul
RT: 12.289 min Scan# 1607
Delta R.T. -0.006 min
Lab File: BM032445.D
Acq: 12 Oct 2021 06:51




Instrument :
BNA_M
ClientSampleId :
SLCS789


Tgt Ion:142 Resp: 1365884
Ion Ratio Lower Upper
142 100
141 91.6 73.0 109.6
116 3.9 3.0 4.6



#39
1,2,4,5-Tetrachlorobenzene
Concen: 32.530 ng/ul
RT: 12.448 min Scan# 1634
Delta R.T. -0.006 min
Lab File: BM032445.D
Acq: 12 Oct 2021 06:51

Tgt Ion:216 Resp: 699956
Ion Ratio Lower Upper
216 100
214 78.6 61.8 92.6
179 21.3 17.2 25.8
108 18.9 15.7 23.5









#40
Hexachlorocyclopentadiene
Concen: 32.407 ng/ul
RT: 12.442 min Scan# 1633
Delta R.T. -0.006 min
Lab File: BM032445.D
Acq: 12 Oct 2021 06:51




Instrument :
BNA_M
ClientSampleId :
SLCS789


Tgt Ion:	237	Resp:	406331
Ion	Ratio	Lower	Upper
237	100		
235	62.1	49.4	74.0
272	12.7	9.8	14.8



#41
2,4,6-Trichlorophenol
Concen: 35.612 ng/ul
RT: 12.689 min Scan# 1675
Delta R.T. -0.006 min
Lab File: BM032445.D
Acq: 12 Oct 2021 06:51

Tgt Ion:	196	Resp:	490732
Ion	Ratio	Lower	Upper
196	100		
198	94.3	75.9	113.9
200	30.1	24.7	37.1









#42
2,4,5-Trichlorophenol
Concen: 35.240 ng/ul
RT: 12.766 min Scan# 1688
Delta R.T. -0.006 min
Lab File: BM032445.D
Acq: 12 Oct 2021 06:51




Instrument :
BNA_M
ClientSampled :
SLCS789


Tgt Ion:196 Resp: 532883
Ion Ratio Lower Upper
196 100
198 96.4 76.4 114.6
200 31.1 25.0 37.4



#43
1,1'-Biphenyl
Concen: 32.411 ng/ul
RT: 13.089 min Scan# 1743
Delta R.T. -0.006 min
Lab File: BM032445.D
Acq: 12 Oct 2021 06:51

Tgt Ion:154 Resp: 1813350
Ion Ratio Lower Upper
154 100
153 41.0 32.1 48.1
76 12.6 10.2 15.2









#44
2-Chloronaphthalene
Concen: 32.487 ng/ul
RT: 13.130 min Scan# 1750
Delta R.T. -0.000 min
Lab File: BM032445.D
Acq: 12 Oct 2021 06:51




Instrument :
BNA_M
ClientSampleId :
SLCS789


Tgt Ion:162 Resp: 1385264
Ion Ratio Lower Upper
162 100
164 33.2 26.6 39.8
127 39.9 31.7 47.5



#45
2-Nitroaniline
Concen: 36.898 ng/ul
RT: 13.330 min Scan# 1784
Delta R.T. -0.006 min
Lab File: BM032445.D
Acq: 12 Oct 2021 06:51

Tgt Ion: 65 Resp: 420171
Ion Ratio Lower Upper
65 100
92 72.6 54.6 81.8
138 110.4 84.3 126.5









#47
Dimethylphthalate
Concen: 33.001 ng/ul
RT: 13.713 min Scan# 1849
Delta R.T. -0.006 min
Lab File: BM032445.D
Acq: 12 Oct 2021 06:51




Instrument :
BNA_M
ClientSampleId :
SLCS789


Tgt Ion:163 Resp: 1784906
Ion Ratio Lower Upper
163 100
194 4.0 3.1 4.7
164 10.3 8.1 12.1



#48
2,6-Dinitrotoluene
Concen: 37.151 ng/ul
RT: 13.830 min Scan# 1869
Delta R.T. -0.000 min
Lab File: BM032445.D
Acq: 12 Oct 2021 06:51

Tgt Ion:165 Resp: 374463
Ion Ratio Lower Upper
165 100
89 53.4 48.6 73.0
121 20.1 18.3 27.5









#50
Acenaphthylene
Concen: 33.197 ng/ul
RT: 13.977 min Scan# 1894
Delta R.T. -0.006 min
Lab File: BM032445.D
Acq: 12 Oct 2021 06:51




Instrument :
BNA_M
ClientSampleId :
SLCS789


Tgt Ion:152 Resp: 2304403
Ion Ratio Lower Upper
152 100
151 20.2 15.7 23.5
153 13.2 10.6 16.0



#51
3-Nitroaniline
Concen: 35.442 ng/ul
RT: 14.160 min Scan# 1925
Delta R.T. -0.000 min
Lab File: BM032445.D
Acq: 12 Oct 2021 06:51

Tgt Ion:138 Resp: 373408
Ion Ratio Lower Upper
138 100
108 11.0 8.9 13.3
92 115.1 93.3 139.9









#52
Acenaphthene
Concen: 32.559 ng/ul
RT: 14.324 min Scan# 1953
Delta R.T. -0.006 min
Lab File: BM032445.D
Acq: 12 Oct 2021 06:51




Instrument :
BNA_M
ClientSampleId :
SLCS789


Tgt Ion:153 Resp: 1492418
Ion Ratio Lower Upper
153 100
152 47.9 38.2 57.2
154 89.5 70.8 106.2



#53
2,4-Dinitrophenol
Concen: 33.986 ng/ul
RT: 14.360 min Scan# 1959
Delta R.T. -0.006 min
Lab File: BM032445.D
Acq: 12 Oct 2021 06:51

Tgt Ion:184 Resp: 220333
Ion Ratio Lower Upper
184 100
63 66.7 50.6 75.8
154 65.1 50.7 76.1









#55
4-Nitrophenol
Concen: 32.337 ng/ul
RT: 14.489 min Scan# 1981
Delta R.T. -0.000 min
Lab File: BM032445.D
Acq: 12 Oct 2021 06:51




Instrument :
BNA_M
ClientSampleId :
SLCS789


Tgt Ion:109 Resp: 276772
Ion Ratio Lower Upper
109 100
139 136.0 101.0 151.6
65 125.2 102.6 154.0



#56
Dibenzofuran
Concen: 32.026 ng/ul
RT: 14.660 min Scan# 2010
Delta R.T. -0.000 min
Lab File: BM032445.D
Acq: 12 Oct 2021 06:51

Tgt Ion:168 Resp: 2127403
Ion Ratio Lower Upper
168 100
139 37.5 30.0 45.0









#57
2,4-Dinitrotoluene
Concen: 36.126 ng/ul
RT: 14.618 min Scan# 2003
Delta R.T. -0.000 min
Lab File: BM032445.D
Acq: 12 Oct 2021 06:51




Instrument :
BNA_M
ClientSampleId :
SLCS789


Tgt Ion:165 Resp: 538208
Ion Ratio Lower Upper
165 100
63 40.6 34.4 51.6
182 2.7 2.2 3.4



#58
2,3,4,6-Tetrachlorophenol
Concen: 36.161 ng/ul
RT: 14.895 min Scan# 2050
Delta R.T. -0.006 min
Lab File: BM032445.D
Acq: 12 Oct 2021 06:51

Tgt Ion:232 Resp: 444069
Ion Ratio Lower Upper
232 100
131 45.0 37.4 56.2
130 2.5 1.8 2.8
166 30.3 25.5 38.3









#59
Diethylphthalate
Concen: 33.351 ng/ul
RT: 15.083 min Scan# 2082
Delta R.T. -0.000 min
Lab File: BM032445.D
Acq: 12 Oct 2021 06:51




Instrument :
BNA_M
ClientSampleId :
SLCS789


Tgt Ion:149 Resp: 1799542
Ion Ratio Lower Upper
149 100
177 22.4 17.9 26.9
150 12.3 10.0 15.0



#61
Fluorene
Concen: 32.145 ng/ul
RT: 15.307 min Scan# 2120
Delta R.T. -0.000 min
Lab File: BM032445.D
Acq: 12 Oct 2021 06:51

Tgt Ion:166 Resp: 1640728
Ion Ratio Lower Upper
166 100
165 97.5 77.0 115.6
167 13.3 10.9 16.3









#62
4-Chlorophenyl-phenylether
Concen: 32.008 ng/ul
RT: 15.295 min Scan# 2118
Delta R.T. -0.006 min
Lab File: BM032445.D
Acq: 12 Oct 2021 06:51




Instrument :
BNA_M
ClientSampleId :
SLCS789


Tgt Ion:204 Resp: 825197
Ion Ratio Lower Upper
204 100
206 33.0 26.2 39.4
141 61.9 48.6 73.0



#63
4-Nitroaniline
Concen: 42.552 ng/ul
RT: 15.324 min Scan# 2123
Delta R.T. -0.000 min
Lab File: BM032445.D
Acq: 12 Oct 2021 06:51

Tgt Ion:138 Resp: 414445
Ion Ratio Lower Upper
138 100
92 50.5 41.3 61.9
108 81.1 64.6 97.0









#66
4,6-Dinitro-2-methylphenol
Concen: 34.663 ng/ul
RT: 15.383 min Scan# 2133
Delta R.T. -0.006 min
Lab File: BM032445.D
Acq: 12 Oct 2021 06:51




Instrument :
BNA_M
ClientSampleId :
SLCS789


Tgt Ion:198 Resp: 309481
Ion Ratio Lower Upper
198 100
182 4.3 4.1 6.1
77 26.4 21.4 32.0



#67
N-Nitrosodiphenylamine
Concen: 33.560 ng/ul
RT: 15.513 min Scan# 2155
Delta R.T. -0.000 min
Lab File: BM032445.D
Acq: 12 Oct 2021 06:51

Tgt Ion:169 Resp: 1492461
Ion Ratio Lower Upper
169 100
168 66.2 53.3 79.9
167 36.1 29.3 43.9









#68
4-Bromophenyl-phenylether
Concen: 33.466 ng/ul
RT: 16.183 min Scan# 2269
Delta R.T. -0.006 min
Lab File: BM032445.D
Acq: 12 Oct 2021 06:51




Instrument :
BNA_M
ClientSampleId :
SLCS789


Tgt Ion:248	Resp: 514115
Ion Ratio	Lower Upper
248	100
250	97.9 77.0 115.4
141	78.8 63.7 95.5



#69
Hexachlorobenzene
Concen: 32.817 ng/ul
RT: 16.318 min Scan# 2292
Delta R.T. -0.006 min
Lab File: BM032445.D
Acq: 12 Oct 2021 06:51

Tgt Ion:284	Resp: 580687
Ion Ratio	Lower Upper
284	100
142	38.3 29.4 44.0
249	32.5 25.5 38.3









#70
Atrazine
Concen: 33.006 ng/ul
RT: 16.454 min Scan# 2315
Delta R.T. -0.006 min
Lab File: BM032445.D
Acq: 12 Oct 2021 06:51




Instrument :
BNA_M
ClientSampleId :
SLCS789


Tgt Ion:200 Resp: 554767
Ion Ratio Lower Upper
200 100
173 27.7 22.2 33.4
215 47.8 37.5 56.3



#71
Pentachlorophenol
Concen: 35.550 ng/ul
RT: 16.654 min Scan# 2349
Delta R.T. -0.006 min
Lab File: BM032445.D
Acq: 12 Oct 2021 06:51

Tgt Ion:266 Resp: 379198
Ion Ratio Lower Upper
266 100
264 62.8 49.9 74.9
268 64.2 50.5 75.7









#72
Phenanthrene
Concen: 32.600 ng/u1
RT: 17.030 min Scan# 2413
Delta R.T. -0.006 min
Lab File: BM032445.D
Acq: 12 Oct 2021 06:51




Instrument :
BNA_M
ClientSampleId :
SLCS789


Tgt Ion:178 Resp: 2681942
Ion Ratio Lower Upper
178 100
179 15.5 12.2 18.2
176 19.3 15.4 23.0



#74
Anthracene
Concen: 32.859 ng/u1
RT: 17.118 min Scan# 2428
Delta R.T. -0.006 min
Lab File: BM032445.D
Acq: 12 Oct 2021 06:51

Tgt Ion:178 Resp: 2715543
Ion Ratio Lower Upper
178 100
179 15.7 12.3 18.5
176 18.7 14.9 22.3









#75
1,2,3,4-Tetrachlorobenzene
Concen: 31.986 ng/uL
RT: 13.060 min Scan# 1738
Delta R.T. -0.006 min
Lab File: BM032445.D
Acq: 12 Oct 2021 06:51




Instrument :
BNA_M
ClientSampleId :
SLCS789


Tgt Ion:	216	Resp:	715829
Ion Ratio	Lower	Upper	
216	100		
214	78.4	62.8	94.2
218	47.8	38.4	57.6



#76
Pentachlorobenzene
Concen: 31.554 ng/uL
RT: 14.589 min Scan# 1998
Delta R.T. -0.006 min
Lab File: BM032445.D
Acq: 12 Oct 2021 06:51

Tgt Ion:	250	Resp:	701319
Ion Ratio	Lower	Upper	
250	100		
248	63.5	50.7	76.1
252	64.1	51.0	76.6









#77
 Carbazole
 Concen: 33.752 ng/ul
 RT: 17.383 min Scan# 2473
 Delta R.T. -0.006 min
 Lab File: BM032445.D
 Acq: 12 Oct 2021 06:51




Instrument :
 BNA_M
 ClientSampleId :
 SLCS789


Tgt Ion:167 Resp: 2487762
 Ion Ratio Lower Upper
 167 100
 166 21.4 17.4 26.2
 139 13.3 10.6 15.8



#78
 Di-n-butylphthalate
 Concen: 35.815 ng/ul
 RT: 17.942 min Scan# 2568
 Delta R.T. -0.006 min
 Lab File: BM032445.D
 Acq: 12 Oct 2021 06:51

Tgt Ion:149 Resp: 3049715
 Ion Ratio Lower Upper
 149 100
 150 9.3 7.4 11.0
 104 5.1 4.2 6.2









#80
Fluoranthene
Concen: 33.512 ng/ul
RT: 19.036 min Scan# 2754
Delta R.T. -0.006 min
Lab File: BM032445.D
Acq: 12 Oct 2021 06:51




Instrument :
BNA_M
ClientSampleId :
SLCS789


Tgt Ion:202 Resp: 3117004
Ion Ratio Lower Upper
202 100
101 11.8 9.1 13.7
100 9.1 7.0 10.4



#82
Pyrene
Concen: 32.992 ng/ul
RT: 19.401 min Scan# 2816
Delta R.T. -0.006 min
Lab File: BM032445.D
Acq: 12 Oct 2021 06:51

Tgt Ion:202 Resp: 3152465
Ion Ratio Lower Upper
202 100
101 13.9 10.3 15.5
100 11.3 8.1 12.1









#83
Butylbenzylphthalate
Concen: 37.292 ng/ul
RT: 20.289 min Scan# 2967
Delta R.T. -0.006 min
Lab File: BM032445.D
Acq: 12 Oct 2021 06:51




Instrument :
BNA_M
ClientSampleId :
SLCS789


Tgt Ion:149 Resp: 1352144
Ion Ratio Lower Upper
149 100
91 70.4 56.3 84.5
206 21.0 16.8 25.2



#84
3,3'-Dichlorobenzidine
Concen: 34.942 ng/ul
RT: 21.071 min Scan# 3100
Delta R.T. -0.006 min
Lab File: BM032445.D
Acq: 12 Oct 2021 06:51

Tgt Ion:252 Resp: 949417
Ion Ratio Lower Upper
252 100
254 64.2 50.9 76.3
126 11.6 9.0 13.4









#85
Benzo(a)anthracene
Concen: 32.861 ng/ul
RT: 21.136 min Scan# 3111
Delta R.T. -0.006 min
Lab File: BM032445.D
Acq: 12 Oct 2021 06:51




Instrument :
BNA_M
ClientSampleId :
SLCS789


Tgt Ion:228 Resp: 2969493
Ion Ratio Lower Upper
228 100
229 20.1 15.6 23.4
226 27.9 21.8 32.6



#86
Bis(2-ethylhexyl)phthalate
Concen: 37.822 ng/ul
RT: 21.059 min Scan# 3098
Delta R.T. -0.006 min
Lab File: BM032445.D
Acq: 12 Oct 2021 06:51

Tgt Ion:149 Resp: 1906429
Ion Ratio Lower Upper
149 100
167 27.7 22.0 33.0
279 4.1 3.3 4.9









#87
Chrysene
Concen: 32.535 ng/ul
RT: 21.189 min Scan# 3120
Delta R.T. -0.006 min
Lab File: BM032445.D
Acq: 12 Oct 2021 06:51




Instrument :
BNA_M
ClientSampleId :
SLCS789


Tgt Ion:228 Resp: 2894828
Ion Ratio Lower Upper
228 100
226 30.1 24.1 36.1
229 19.7 15.5 23.3



#89
Di-n-octyl phthalate
Concen: 39.491 ng/ul
RT: 21.900 min Scan# 3241
Delta R.T. -0.006 min
Lab File: BM032445.D
Acq: 12 Oct 2021 06:51

Tgt Ion:149 Resp: 3479665









#90
Benzo(b)fluoranthene
Concen: 34.367 ng/ul
RT: 22.659 min Scan# 3370
Delta R.T. -0.006 min
Lab File: BM032445.D
Acq: 12 Oct 2021 06:51




Instrument :
BNA_M
ClientSampleId :
SLCS789


Tgt Ion:252 Resp: 3234887
Ion Ratio Lower Upper
252 100
253 22.2 17.6 26.4
125 10.9 8.2 12.4



#91
Benzo(k)fluoranthene
Concen: 32.952 ng/ul
RT: 22.700 min Scan# 3377
Delta R.T. -0.006 min
Lab File: BM032445.D
Acq: 12 Oct 2021 06:51

Tgt Ion:252 Resp: 2813228
Ion Ratio Lower Upper
252 100
253 22.4 17.5 26.3
125 10.6 8.4 12.6









#93
Benzo(a)pyrene
Concen: 33.641 ng/ul
RT: 23.206 min Scan# 3463
Delta R.T. -0.006 min
Lab File: BM032445.D
Acq: 12 Oct 2021 06:51




Instrument :
BNA_M
ClientSampleId :
SLCS789


Tgt Ion:252 Resp: 2995822
Ion Ratio Lower Upper
252 100
253 21.9 17.8 26.6
125 11.8 9.2 13.8



#94
Indeno(1,2,3-cd)pyrene
Concen: 31.600 ng/ul
RT: 25.424 min Scan# 3840
Delta R.T. -0.006 min
Lab File: BM032445.D
Acq: 12 Oct 2021 06:51

Tgt Ion:276 Resp: 3110551
Ion Ratio Lower Upper
276 100
138 27.9 21.8 32.6
277 25.7 20.4 30.6









#95
Dibenzo(a,h)anthracene
Concen: 32.018 ng/ul
RT: 25.430 min Scan# 3841
Delta R.T. -0.006 min
Lab File: BM032445.D
Acq: 12 Oct 2021 06:51

Instrument :
BNA_M
ClientSampleId :
SLCS789

Tgt Ion:278 Resp: 2704336
Ion Ratio Lower Upper
278 100
139 17.0 13.4 20.2
279 23.9 19.0 28.4



#96
Benzo(g,h,i)perylene
Concen: 30.543 ng/ul
RT: 26.077 min Scan# 3951
Delta R.T. -0.006 min
Lab File: BM032445.D
Acq: 12 Oct 2021 06:51

Tgt Ion:276 Resp: 2630973
Ion Ratio Lower Upper
276 100
138 22.8 18.6 27.8
277 23.7 19.0 28.4

