

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM101121\
 Data File : BM032462.D
 Acq On : 12 Oct 2021 17:44
 Operator : CG/JU
 Sample : M4110-01
 Misc :
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 C00T3

Quant Time: Oct 12 18:23:38 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-BM100521.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Oct 12 09:54:27 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.607	152	238780	20.00	ng/ul	0.00
20) Naphthalene-d8	10.395	136	1061412	20.00	ng/ul	0.00
38) Acenaphthene-d10	14.259	164	711375	20.00	ng/ul	0.00
64) Phenanthrene-d10	16.983	188	1501463	20.00	ng/ul	0.00
79) Chrysene-d12	21.147	240	1448075	20.00	ng/ul	0.00
88) Perylene-d12	23.294	264	1349247	20.00	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	2.954	96	32522	5.46	ng/uL	0.00
4) Pyridine-d5	3.390	84	150995	8.97	ng/ul	0.00
7) Phenol-d5	6.801	99	150930	7.43	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	6.936	67	353073	30.28	ng/ul	0.00
11) 2-Chlorophenol-d4	7.148	132	377383	24.34	ng/ul	0.00
15) 4-Methylphenol-d8	8.336	113	270676	16.50	ng/ul	0.00
21) Nitrobenzene-d5	8.760	128	239869	32.03	ng/ul	0.00
24) 2-Nitrophenol-d4	9.483	143	252209	31.24	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.030	165	442870	27.56	ng/ul	0.00
31) 4-Chloroaniline-d4	10.530	131	488681	20.62	ng/ul	0.00
46) Dimethylphthalate-d6	13.660	166	1691045	34.35	ng/ul	0.00
49) Acenaphthylene-d8	13.948	160	1966553	32.88	ng/ul	0.00
54) 4-Nitrophenol-d4	14.465	143	52381	5.42	ng/ul	0.00
60) Fluorene-d10	15.248	176	1464174	35.07	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.365	200	228490	26.70	ng/ul	0.00
73) Anthracene-d10	17.083	188	2328384	35.63	ng/ul	0.00
81) Pyrene-d10	19.371	212	2674449	36.71	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.159	264	2296511	33.99	ng/ul	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM101121\
Data File : BM032462.D
Acq On : 12 Oct 2021 17:44
Operator : CG/JU
Sample : M4110-01
Misc :
ALS Vial : 48 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
C00T3

Quant Time: Oct 12 18:23:38 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-BM100521.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Oct 12 09:54:27 2021
Response via : Initial Calibration

