

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM101121\
 Data File : BM032466.D
 Acq On : 12 Oct 2021 20:10
 Operator : CG/JU
 Sample : M4110-04
 Misc :
 ALS Vial : 52 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 C00U1

Quant Time: Oct 13 01:09:45 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-BM100521.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Oct 12 09:54:27 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.607	152	207914	20.00	ng/ul	0.00
20) Naphthalene-d8	10.395	136	944595	20.00	ng/ul	0.00
38) Acenaphthene-d10	14.260	164	650604	20.00	ng/ul	0.00
64) Phenanthrene-d10	16.983	188	1374436	20.00	ng/ul	0.00
79) Chrysene-d12	21.147	240	1351335	20.00	ng/ul	0.00
88) Perylene-d12	23.288	264	1341542	20.00	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	2.954	96	24213	4.67	ng/uL	0.00
4) Pyridine-d5	3.390	84	113335	7.74	ng/ul	0.00
7) Phenol-d5	6.801	99	133669	7.56	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	6.936	67	307955	30.34	ng/ul	0.00
11) 2-Chlorophenol-d4	7.148	132	353604	26.19	ng/ul	0.00
15) 4-Methylphenol-d8	8.336	113	257697	18.05	ng/ul	0.00
21) Nitrobenzene-d5	8.760	128	222889	33.44	ng/ul	0.00
24) 2-Nitrophenol-d4	9.483	143	244086	33.98	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.030	165	433850	30.34	ng/ul	0.00
31) 4-Chloroaniline-d4	10.530	131	453381	21.49	ng/ul	0.00
46) Dimethylphthalate-d6	13.660	166	1727146	38.36	ng/ul	0.00
49) Acenaphthylene-d8	13.948	160	1880506	34.38	ng/ul	0.00
54) 4-Nitrophenol-d4	14.465	143	59150	6.69	ng/ul	0.00
60) Fluorene-d10	15.248	176	1421132	37.22	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.365	200	268779	34.31	ng/ul	0.00
73) Anthracene-d10	17.083	188	2397784	40.08	ng/ul	0.00
81) Pyrene-d10	19.371	212	2793598	41.09	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.159	264	2577603	38.37	ng/ul	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM101121\
Data File : BM032466.D
Acq On : 12 Oct 2021 20:10
Operator : CG/JU
Sample : M4110-04
Misc :
ALS Vial : 52 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
C00U1

Quant Time: Oct 13 01:09:45 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-BM100521.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Oct 12 09:54:27 2021
Response via : Initial Calibration

