

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM101121\
 Data File : BM032475.D
 Acq On : 13 Oct 2021 02:11
 Operator : CG/JU
 Sample : PB139820BS
 Misc :
 ALS Vial : 61 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SLCS820

Quant Time: Oct 13 04:30:58 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-BM100521.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Oct 12 09:54:27 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.607	152	237823	20.00	ng/ul	0.00
20) Naphthalene-d8	10.395	136	1055330	20.00	ng/ul	0.00
38) Acenaphthene-d10	14.254	164	689824	20.00	ng/ul	0.00
64) Phenanthrene-d10	16.983	188	1421578	20.00	ng/ul	0.00
79) Chrysene-d12	21.148	240	1345211	20.00	ng/ul	0.00
88) Perylene-d12	23.289	264	1353369	20.00	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	2.955	96	35245	5.94	ng/uL	0.00
4) Pyridine-d5	3.384	84	500969	29.90	ng/ul	0.00
7) Phenol-d5	6.801	99	678894	33.57	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	6.937	67	383632	33.04	ng/ul	0.00
11) 2-Chlorophenol-d4	7.148	132	533184	34.53	ng/ul	0.00
15) 4-Methylphenol-d8	8.342	113	570736	34.94	ng/ul	0.00
21) Nitrobenzene-d5	8.760	128	277040	37.21	ng/ul	0.00
24) 2-Nitrophenol-d4	9.484	143	308555	38.44	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.031	165	568553	35.58	ng/ul	0.00
31) 4-Chloroaniline-d4	10.531	131	717983	30.47	ng/ul	0.00
46) Dimethylphthalate-d6	13.660	166	1715023	35.92	ng/ul	0.00
49) Acenaphthylene-d8	13.948	160	2114220	36.45	ng/ul	0.00
54) 4-Nitrophenol-d4	14.472	143	331697	35.38	ng/ul	0.00
60) Fluorene-d10	15.242	176	1455061	35.94	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.366	200	298713	36.87	ng/ul	0.00
73) Anthracene-d10	17.083	188	2229279	36.03	ng/ul	0.00
81) Pyrene-d10	19.365	212	2509373	37.08	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.153	264	2323612	34.28	ng/ul	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	2.990	88	73057	11.32	ng/uL	96
5) Pyridine	3.402	79	509104	30.72	ng/ul	97
6) Benzaldehyde	6.737	77	417294	43.83	ng/ul	98
8) Phenol	6.825	94	707278	34.52	ng/ul	100
10) Bis(2-Chloroethyl)ether	7.031	93	541091	33.57	ng/ul	99
12) 2-Chlorophenol	7.178	128	561245	35.18	ng/ul	99
13) 2-Methylphenol	8.072	108	549704	34.99	ng/ul	99
14) 2,2'-oxybis(1-Chloropr...	8.143	45	667844	33.45	ng/ul	99
16) Acetophenone	8.425	105	870209	36.37	ng/ul	98
17) N-Nitroso-di-n-propyla...	8.419	70	438402	37.11	ng/ul	97
18) 4-Methylphenol	8.401	108	604296	36.70	ng/ul	98
19) Hexachloroethane	8.690	117	223766	35.26	ng/ul	100
22) Nitrobenzene	8.801	77	626782	35.45	ng/ul	98
23) Isophorone	9.325	82	1263052	37.03	ng/ul	99
25) 2-Nitrophenol	9.519	139	331652	37.82	ng/ul	97
26) 2,4-Dimethylphenol	9.589	107	672508	35.86	ng/ul	99
27) Bis(2-Chloroethoxy)met...	9.807	93	779088	34.44	ng/ul	99
29) 2,4-Dichlorophenol	10.054	162	571624	36.34	ng/ul	99
30) Naphthalene	10.448	128	1856991	34.33	ng/ul	99
32) 4-Chloroaniline	10.554	127	713422	30.15	ng/ul	98
33) Hexachlorobutadiene	10.748	225	348666	34.36	ng/ul	98
34) Caprolactam	11.307	113	195456	34.47	ng/ul	97
35) 4-Chloro-3-methylphenol	11.701	107	643184	37.16	ng/ul	99
36) 2-Methylnaphthalene	12.060	142	1304743	35.47	ng/ul	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM101121\
 Data File : BM032475.D
 Acq On : 13 Oct 2021 02:11
 Operator : CG/JU
 Sample : PB139820BS
 Misc :
 ALS Vial : 61 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SLCS820

Quant Time: Oct 13 04:30:58 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-BM100521.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Oct 12 09:54:27 2021
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.283	142	1308413	34.93	ng/ul	98
39) 1,2,4,5-Tetrachloroben...	12.448	216	674047	35.44	ng/ul	98
40) Hexachlorocyclopentadiene	12.436	237	381624	34.43	ng/ul	100
41) 2,4,6-Trichlorophenol	12.689	196	467340	38.36	ng/ul	98
42) 2,4,5-Trichlorophenol	12.760	196	506608	37.90	ng/ul	99
43) 1,1'-Biphenyl	13.083	154	1753128	35.45	ng/ul	100
44) 2-Chloronaphthalene	13.125	162	1334995	35.42	ng/ul	99
45) 2-Nitroaniline	13.330	65	396196	39.36	ng/ul	93
47) Dimethylphthalate	13.707	163	1713271	35.83	ng/ul	100
48) 2,6-Dinitrotoluene	13.825	165	356700	40.03	ng/ul	95
50) Acenaphthylene	13.977	152	2221285	36.20	ng/ul	99
51) 3-Nitroaniline	14.154	138	351395	37.73	ng/ul	99
52) Acenaphthene	14.319	153	1451436	35.82	ng/ul	100
53) 2,4-Dinitrophenol	14.360	184	200647	35.01	ng/ul	98
55) 4-Nitrophenol	14.483	109	267725	35.38	ng/ul	96
56) Dibenzofuran	14.654	168	2063265	35.14	ng/ul	99
57) 2,4-Dinitrotoluene	14.613	165	517494	39.29	ng/ul	96
58) 2,3,4,6-Tetrachlorophenol	14.889	232	426055	39.25	ng/ul	97
59) Diethylphthalate	15.077	149	1713252	35.92	ng/ul	100
61) Fluorene	15.301	166	1591513	35.27	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.289	204	795622	34.91	ng/ul	99
63) 4-Nitroaniline	15.319	138	396814	46.09	ng/ul	99
66) 4,6-Dinitro-2-methylph...	15.383	198	296968	37.29	ng/ul#	95
67) N-Nitrosodiphenylamine	15.507	169	1441157	36.34	ng/ul	99
68) 4-Bromophenyl-phenylether	16.177	248	495277	36.15	ng/ul	99
69) Hexachlorobenzene	16.313	284	561392	35.57	ng/ul	98
70) Atrazine	16.454	200	536584	35.80	ng/ul	98
71) Pentachlorophenol	16.648	266	359115	37.75	ng/ul	99
72) Phenanthrene	17.024	178	2616531	35.66	ng/ul	99
74) Anthracene	17.118	178	2634965	35.75	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.054	216	684579	34.30	ng/uL	100
76) Pentachlorobenzene	14.589	250	677218	34.16	ng/uL	100
77) Carbazole	17.383	167	2459130	37.41	ng/ul	100
78) Di-n-butylphthalate	17.936	149	2873936	37.84	ng/ul	100
80) Fluoranthene	19.036	202	3046694	36.49	ng/ul	100
82) Pyrene	19.401	202	3102581	36.17	ng/ul	99
83) Butylbenzylphthalate	20.283	149	1292725	39.72	ng/ul	97
84) 3,3'-Dichlorobenzidine	21.065	252	880848	36.11	ng/ul	99
85) Benzo(a)anthracene	21.136	228	2885892	35.58	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.053	149	1792051	39.60	ng/ul	100
87) Chrysene	21.183	228	2845389	35.62	ng/ul	99
89) Di-n-octyl phthalate	21.900	149	3229830	41.90	ng/ul	100
90) Benzo(b)fluoranthene	22.653	252	3092202	37.55	ng/ul	99
91) Benzo(k)fluoranthene	22.695	252	2688750	36.00	ng/ul	100
93) Benzo(a)pyrene	23.200	252	2844034	36.51	ng/ul	100
94) Indeno(1,2,3-cd)pyrene	25.412	276	2871066	33.34	ng/ul	99
95) Dibenzo(a,h)anthracene	25.418	278	2491467	33.72	ng/ul	100
96) Benzo(g,h,i)perylene	26.065	276	2450991	32.53	ng/ul	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM101121\
 Data File : BM032475.D
 Acq On : 13 Oct 2021 02:11
 Operator : CG/JU
 Sample : PB139820BS
 Misc :
 ALS Vial : 61 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SLCS820

Quant Time: Oct 13 04:30:58 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-BM100521.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Oct 12 09:54:27 2021
 Response via : Initial Calibration




Abundance Scan 26 (2.990 min): BM032471.D\data.ms (-20)




m/z-->

Abundance Scan 26 (2.990 min): BM032475.D\data.ms



m/z-->


Abundance Scan 26 (2.990 min): BM032475.D\data.ms (-1)




m/z-->

#2
1,4-Dioxane
Concen: 11.32 ng/uL
RT: 2.990 min Scan# 26
Delta R.T. 0.000 min
Lab File: BM032475.D
Acq: 13 Oct 2021 02:11

Tgt Ion:	88	Resp:	73057
Ion	Ratio	Lower	Upper
88	100		
43	25.8	22.2	33.2
58	71.2	59.5	89.3




Abundance Scan 96 (3.401 min): BM032471.D\data.ms (-91)




m/z-->

Abundance Scan 96 (3.402 min): BM032475.D\data.ms



m/z-->


Abundance Scan 96 (3.402 min): BM032475.D\data.ms (-79)




m/z-->

#5
Pyridine
Concen: 30.72 ng/uL
RT: 3.402 min Scan# 96
Delta R.T. 0.000 min
Lab File: BM032475.D
Acq: 13 Oct 2021 02:11

Tgt Ion:	79	Resp:	509104
Ion	Ratio	Lower	Upper
79	100		
52	71.1	59.4	89.0
51	33.6	28.1	42.1









#6
Benzaldehyde
Concen: 43.83 ng/ul
RT: 6.737 min Scan# 663
Delta R.T. 0.000 min
Lab File: BM032475.D
Acq: 13 Oct 2021 02:11




Instrument :
BNA_M
ClientSampleId :
SLCS820


Tgt Ion: 77 Resp: 417294
Ion Ratio Lower Upper
77 100
105 101.4 79.5 119.3
106 95.8 75.7 113.5



#8
Phenol
Concen: 34.52 ng/ul
RT: 6.825 min Scan# 678
Delta R.T. -0.006 min
Lab File: BM032475.D
Acq: 13 Oct 2021 02:11

Tgt Ion: 94 Resp: 707278
Ion Ratio Lower Upper
94 100
65 27.5 22.3 33.5
66 37.0 29.5 44.3









#10
Bis(2-Chloroethyl)ether
Concen: 33.57 ng/ul
RT: 7.031 min Scan# 713
Delta R.T. 0.000 min
Lab File: BM032475.D
Acq: 13 Oct 2021 02:11




Instrument :
BNA_M
ClientSampleId :
SLCS820


Tgt Ion: 93 Resp: 541091
Ion Ratio Lower Upper
93 100
63 75.3 59.1 88.7
95 32.5 26.0 39.0



#12
2-Chlorophenol
Concen: 35.18 ng/ul
RT: 7.178 min Scan# 738
Delta R.T. -0.006 min
Lab File: BM032475.D
Acq: 13 Oct 2021 02:11

Tgt Ion: 128 Resp: 561245
Ion Ratio Lower Upper
128 100
64 44.9 36.2 54.4
130 32.3 26.6 39.8









#13
2-Methylphenol
Concen: 34.99 ng/ul
RT: 8.072 min Scan# 890
Delta R.T. -0.006 min
Lab File: BM032475.D
Acq: 13 Oct 2021 02:11




Instrument :
BNA_M
ClientSampleId :
SLCS820


Tgt Ion:108 Resp: 549704
Ion Ratio Lower Upper
108 100
107 89.6 71.0 106.6



#14
2,2'-oxybis(1-Chloropropane)
Concen: 33.45 ng/ul
RT: 8.143 min Scan# 902
Delta R.T. 0.000 min
Lab File: BM032475.D
Acq: 13 Oct 2021 02:11

Tgt Ion: 45 Resp: 667844
Ion Ratio Lower Upper
45 100
77 15.7 12.8 19.2
79 12.5 9.6 14.4









#16
Acetophenone
Concen: 36.37 ng/ul
RT: 8.425 min Scan# 950
Delta R.T. -0.006 min
Lab File: BM032475.D
Acq: 13 Oct 2021 02:11




Instrument :
BNA_M
ClientSampleId :
SLCS820


Tgt Ion:105 Resp: 870209
Ion Ratio Lower Upper
105 100
77 77.7 60.9 91.3
51 25.2 20.5 30.7



#17
N-Nitroso-di-n-propylamine
Concen: 37.11 ng/ul
RT: 8.419 min Scan# 949
Delta R.T. 0.000 min
Lab File: BM032475.D
Acq: 13 Oct 2021 02:11

Tgt Ion: 70 Resp: 438402
Ion Ratio Lower Upper
70 100
42 52.3 44.2 66.2
101 10.0 8.2 12.2
130 21.5 17.4 26.2









#18
4-Methylphenol
Concen: 36.70 ng/ul
RT: 8.401 min Scan# 946
Delta R.T. 0.000 min
Lab File: BM032475.D
Acq: 13 Oct 2021 02:11




Instrument :
BNA_M
ClientSampleId :
SLCS820


Tgt Ion:108 Resp: 604296
Ion Ratio Lower Upper
108 100
107 111.8 87.9 131.9



#19
Hexachloroethane
Concen: 35.26 ng/ul
RT: 8.690 min Scan# 995
Delta R.T. -0.006 min
Lab File: BM032475.D
Acq: 13 Oct 2021 02:11

Tgt Ion:117 Resp: 223766
Ion Ratio Lower Upper
117 100
201 98.4 78.9 118.3
199 61.3 49.0 73.6









#22
Nitrobenzene
Concen: 35.45 ng/ul
RT: 8.801 min Scan# 1014
Delta R.T. -0.006 min
Lab File: BM032475.D
Acq: 13 Oct 2021 02:11




Instrument :
BNA_M
ClientSampleId :
SLCS820


Tgt Ion: 77 Resp: 626782
Ion Ratio Lower Upper
77 100
123 46.6 35.9 53.9
65 13.5 10.4 15.6



#23
Isophorone
Concen: 37.03 ng/ul
RT: 9.325 min Scan# 1103
Delta R.T. -0.006 min
Lab File: BM032475.D
Acq: 13 Oct 2021 02:11

Tgt Ion: 82 Resp: 1263052
Ion Ratio Lower Upper
82 100
95 7.4 5.8 8.6
138 17.8 13.9 20.9









#25
2-Nitrophenol
Concen: 37.82 ng/ul
RT: 9.519 min Scan# 1136
Delta R.T. 0.000 min
Lab File: BM032475.D
Acq: 13 Oct 2021 02:11




Instrument :
BNA_M
ClientSampleId :
SLCS820


Tgt Ion:139 Resp: 331652
Ion Ratio Lower Upper
139 100
65 47.9 40.2 60.4
109 30.4 24.7 37.1



#26
2,4-Dimethylphenol
Concen: 35.86 ng/ul
RT: 9.589 min Scan# 1148
Delta R.T. -0.006 min
Lab File: BM032475.D
Acq: 13 Oct 2021 02:11

Tgt Ion:107 Resp: 672508
Ion Ratio Lower Upper
107 100
121 51.1 40.9 61.3
122 85.8 69.8 104.6









#27
 Bis(2-Chloroethoxy)methane
 Concen: 34.44 ng/ul
 RT: 9.807 min Scan# 1185
 Delta R.T. -0.006 min
 Lab File: BM032475.D
 Acq: 13 Oct 2021 02:11

Instrument :
 BNA_M
 ClientSampleId :
 SLCS820

Tgt Ion: 93 Resp: 779088




Ion	Ratio	Lower	Upper
93	100		
95	32.4	26.2	39.2
123	10.9	8.6	12.8




#29
 2,4-Dichlorophenol
 Concen: 36.34 ng/ul
 RT: 10.054 min Scan# 1227
 Delta R.T. -0.006 min
 Lab File: BM032475.D
 Acq: 13 Oct 2021 02:11

Tgt Ion: 162 Resp: 571624

Ion	Ratio	Lower	Upper
162	100		
164	65.1	51.7	77.5
98	36.2	29.8	44.6









#30
Naphthalene
Concen: 34.33 ng/ul
RT: 10.448 min Scan# 1294
Delta R.T. 0.000 min
Lab File: BM032475.D
Acq: 13 Oct 2021 02:11




Instrument :
BNA_M
ClientSampleId :
SLCS820


Tgt Ion:128 Resp: 1856991
Ion Ratio Lower Upper
128 100
129 10.8 8.9 13.3
127 13.2 10.5 15.7



#32
4-Chloroaniline
Concen: 30.15 ng/ul
RT: 10.554 min Scan# 1312
Delta R.T. -0.006 min
Lab File: BM032475.D
Acq: 13 Oct 2021 02:11

Tgt Ion:127 Resp: 713422
Ion Ratio Lower Upper
127 100
129 32.3 25.1 37.7









#33
Hexachlorobutadiene
Concen: 34.36 ng/ul
RT: 10.748 min Scan# 1345
Delta R.T. -0.006 min
Lab File: BM032475.D
Acq: 13 Oct 2021 02:11




Instrument :
BNA_M
ClientSampleId :
SLCS820


Tgt Ion:225 Resp: 348666
Ion Ratio Lower Upper
225 100
223 63.6 51.1 76.7
227 62.6 52.0 78.0



#34
Caprolactam
Concen: 34.47 ng/ul
RT: 11.307 min Scan# 1440
Delta R.T. 0.006 min
Lab File: BM032475.D
Acq: 13 Oct 2021 02:11

Tgt Ion:113 Resp: 195456
Ion Ratio Lower Upper
113 100
55 159.1 123.8 185.8
56 120.2 98.6 147.8









#35
4-Chloro-3-methylphenol
Concen: 37.16 ng/ul
RT: 11.701 min Scan# 1507
Delta R.T. 0.000 min
Lab File: BM032475.D
Acq: 13 Oct 2021 02:11




Instrument :
BNA_M
ClientSampleId :
SLCS820


Tgt Ion:107 Resp: 643184
Ion Ratio Lower Upper
107 100
144 26.3 20.4 30.6
142 78.8 62.4 93.6



#36
2-Methylnaphthalene
Concen: 35.47 ng/ul
RT: 12.060 min Scan# 1568
Delta R.T. -0.006 min
Lab File: BM032475.D
Acq: 13 Oct 2021 02:11

Tgt Ion:142 Resp: 1304743
Ion Ratio Lower Upper
142 100
141 89.2 71.2 106.8









#37
1-Methylnaphthalene
Concen: 34.93 ng/ul
RT: 12.283 min Scan# 1606
Delta R.T. -0.006 min
Lab File: BM032475.D
Acq: 13 Oct 2021 02:11




Instrument :
BNA_M
ClientSampleId :
SLCS820


Tgt Ion:142 Resp: 1308413
Ion Ratio Lower Upper
142 100
141 93.0 73.0 109.6
116 3.8 3.0 4.6



#39
1,2,4,5-Tetrachlorobenzene
Concen: 35.44 ng/ul
RT: 12.448 min Scan# 1634
Delta R.T. 0.000 min
Lab File: BM032475.D
Acq: 13 Oct 2021 02:11

Tgt Ion:216 Resp: 674047
Ion Ratio Lower Upper
216 100
214 78.7 61.8 92.6
179 20.8 17.2 25.8
108 17.4 15.7 23.5









#40
Hexachlorocyclopentadiene
Concen: 34.43 ng/ul
RT: 12.436 min Scan# 1632
Delta R.T. -0.006 min
Lab File: BM032475.D
Acq: 13 Oct 2021 02:11

Instrument :
BNA_M
ClientSampleId :
SLCS820

Tgt Ion:237 Resp: 381624




Ion	Ratio	Lower	Upper
237	100		
235	61.3	49.4	74.0
272	12.3	9.8	14.8




#41
2,4,6-Trichlorophenol
Concen: 38.36 ng/ul
RT: 12.689 min Scan# 1675
Delta R.T. 0.000 min
Lab File: BM032475.D
Acq: 13 Oct 2021 02:11

Tgt Ion:196 Resp: 467340

Ion	Ratio	Lower	Upper
196	100		
198	96.5	75.9	113.9
200	31.6	24.7	37.1









#42
2,4,5-Trichlorophenol
Concen: 37.90 ng/ul
RT: 12.760 min Scan# 1687
Delta R.T. -0.006 min
Lab File: BM032475.D
Acq: 13 Oct 2021 02:11




Instrument :
BNA_M
ClientSampleId :
SLCS820


Tgt Ion:196 Resp: 506608
Ion Ratio Lower Upper
196 100
198 95.0 76.4 114.6
200 30.8 25.0 37.4



#43
1,1'-Biphenyl
Concen: 35.45 ng/ul
RT: 13.083 min Scan# 1742
Delta R.T. -0.006 min
Lab File: BM032475.D
Acq: 13 Oct 2021 02:11

Tgt Ion:154 Resp: 1753128
Ion Ratio Lower Upper
154 100
153 39.8 32.1 48.1
76 12.6 10.2 15.2









#44
2-Chloronaphthalene
Concen: 35.42 ng/ul
RT: 13.125 min Scan# 1749
Delta R.T. 0.000 min
Lab File: BM032475.D
Acq: 13 Oct 2021 02:11




Instrument :
BNA_M
ClientSampleId :
SLCS820


Tgt Ion:162 Resp: 1334995
Ion Ratio Lower Upper
162 100
164 32.6 26.6 39.8
127 40.0 31.7 47.5



#45
2-Nitroaniline
Concen: 39.36 ng/ul
RT: 13.330 min Scan# 1784
Delta R.T. 0.000 min
Lab File: BM032475.D
Acq: 13 Oct 2021 02:11

Tgt Ion: 65 Resp: 396196
Ion Ratio Lower Upper
65 100
92 71.4 54.6 81.8
138 114.6 84.3 126.5









#47
Dimethylphthalate
Concen: 35.83 ng/ul
RT: 13.707 min Scan# 1848
Delta R.T. -0.006 min
Lab File: BM032475.D
Acq: 13 Oct 2021 02:11




Instrument :
BNA_M
ClientSampleId :
SLCS820


Tgt Ion:163 Resp: 1713271
Ion Ratio Lower Upper
163 100
194 3.9 3.1 4.7
164 10.3 8.1 12.1



#48
2,6-Dinitrotoluene
Concen: 40.03 ng/ul
RT: 13.825 min Scan# 1868
Delta R.T. 0.000 min
Lab File: BM032475.D
Acq: 13 Oct 2021 02:11

Tgt Ion:165 Resp: 356700
Ion Ratio Lower Upper
165 100
89 56.5 48.6 73.0
121 20.6 18.3 27.5









#50
Acenaphthylene
Concen: 36.20 ng/ul
RT: 13.977 min Scan# 1894
Delta R.T. 0.000 min
Lab File: BM032475.D
Acq: 13 Oct 2021 02:11




Instrument :
BNA_M
ClientSampleId :
SLCS820


Tgt Ion:152 Resp: 2221285
Ion Ratio Lower Upper
152 100
151 20.1 15.7 23.5
153 13.1 10.6 16.0



#51
3-Nitroaniline
Concen: 37.73 ng/ul
RT: 14.154 min Scan# 1924
Delta R.T. -0.006 min
Lab File: BM032475.D
Acq: 13 Oct 2021 02:11

Tgt Ion:138 Resp: 351395
Ion Ratio Lower Upper
138 100
108 10.9 8.9 13.3
92 115.5 93.3 139.9









#52
Acenaphthene
Concen: 35.82 ng/ul
RT: 14.319 min Scan# 1952
Delta R.T. -0.006 min
Lab File: BM032475.D
Acq: 13 Oct 2021 02:11




Instrument :
BNA_M
ClientSampleId :
SLCS820


Tgt Ion:153 Resp: 1451436
Ion Ratio Lower Upper
153 100
152 47.6 38.2 57.2
154 88.6 70.8 106.2



#53
2,4-Dinitrophenol
Concen: 35.01 ng/ul
RT: 14.360 min Scan# 1959
Delta R.T. 0.000 min
Lab File: BM032475.D
Acq: 13 Oct 2021 02:11

Tgt Ion:184 Resp: 200647
Ion Ratio Lower Upper
184 100
63 61.8 50.6 75.8
154 64.9 50.7 76.1









#55
4-Nitrophenol
Concen: 35.38 ng/ul
RT: 14.483 min Scan# 1980
Delta R.T. 0.000 min
Lab File: BM032475.D
Acq: 13 Oct 2021 02:11




Instrument :
BNA_M
ClientSampleId :
SLCS820


Tgt Ion:109 Resp: 267725
Ion Ratio Lower Upper
109 100
139 130.7 101.0 151.6
65 124.1 102.6 154.0



#56
Dibenzofuran
Concen: 35.14 ng/ul
RT: 14.654 min Scan# 2009
Delta R.T. -0.006 min
Lab File: BM032475.D
Acq: 13 Oct 2021 02:11

Tgt Ion:168 Resp: 2063265
Ion Ratio Lower Upper
168 100
139 38.2 30.0 45.0









#57
2,4-Dinitrotoluene
Concen: 39.29 ng/ul
RT: 14.613 min Scan# 2002
Delta R.T. -0.006 min
Lab File: BM032475.D
Acq: 13 Oct 2021 02:11




Instrument :
BNA_M
ClientSampleId :
SLCS820


Tgt Ion:165 Resp: 517494
Ion Ratio Lower Upper
165 100
63 40.3 34.4 51.6
182 2.8 2.2 3.4



#58
2,3,4,6-Tetrachlorophenol
Concen: 39.25 ng/ul
RT: 14.889 min Scan# 2049
Delta R.T. -0.006 min
Lab File: BM032475.D
Acq: 13 Oct 2021 02:11

Tgt Ion:232 Resp: 426055
Ion Ratio Lower Upper
232 100
131 43.8 37.4 56.2
130 2.4 1.8 2.8
166 30.9 25.5 38.3









#59
Diethylphthalate
Concen: 35.92 ng/ul
RT: 15.077 min Scan# 2081
Delta R.T. 0.000 min
Lab File: BM032475.D
Acq: 13 Oct 2021 02:11




Instrument :
BNA_M
ClientSampleId :
SLCS820


Tgt Ion:149 Resp: 1713252
Ion Ratio Lower Upper
149 100
177 22.4 17.9 26.9
150 12.8 10.0 15.0



#61
Fluorene
Concen: 35.27 ng/ul
RT: 15.301 min Scan# 2119
Delta R.T. 0.000 min
Lab File: BM032475.D
Acq: 13 Oct 2021 02:11

Tgt Ion:166 Resp: 1591513
Ion Ratio Lower Upper
166 100
165 97.8 77.0 115.6
167 13.5 10.9 16.3









#62
4-Chlorophenyl-phenylether
Concen: 34.91 ng/ul
RT: 15.289 min Scan# 2117
Delta R.T. -0.006 min
Lab File: BM032475.D
Acq: 13 Oct 2021 02:11




Instrument :
BNA_M
ClientSampleId :
SLCS820


Tgt Ion:204 Resp: 795622
Ion Ratio Lower Upper
204 100
206 33.0 26.2 39.4
141 62.2 48.6 73.0



#63
4-Nitroaniline
Concen: 46.09 ng/ul
RT: 15.319 min Scan# 2122
Delta R.T. 0.000 min
Lab File: BM032475.D
Acq: 13 Oct 2021 02:11

Tgt Ion:138 Resp: 396814
Ion Ratio Lower Upper
138 100
92 52.5 41.3 61.9
108 81.9 64.6 97.0









#66
4,6-Dinitro-2-methylphenol
Concen: 37.29 ng/ul
RT: 15.383 min Scan# 2133
Delta R.T. 0.000 min
Lab File: BM032475.D
Acq: 13 Oct 2021 02:11




Instrument :
BNA_M
ClientSampleId :
SLCS820


Tgt Ion:198 Resp: 296968
Ion Ratio Lower Upper
198 100
182 3.9 4.1 6.1#
77 23.8 21.4 32.0



#67
N-Nitrosodiphenylamine
Concen: 36.34 ng/ul
RT: 15.507 min Scan# 2154
Delta R.T. 0.000 min
Lab File: BM032475.D
Acq: 13 Oct 2021 02:11

Tgt Ion:169 Resp: 1441157
Ion Ratio Lower Upper
169 100
168 66.2 53.3 79.9
167 36.1 29.3 43.9









#68
4-Bromophenyl-phenylether
Concen: 36.15 ng/ul
RT: 16.177 min Scan# 2268
Delta R.T. -0.006 min
Lab File: BM032475.D
Acq: 13 Oct 2021 02:11




Instrument :
BNA_M
ClientSampleId :
SLCS820


Tgt Ion:248 Resp: 495277
Ion Ratio Lower Upper
248 100
250 97.4 77.0 115.4
141 80.5 63.7 95.5



#69
Hexachlorobenzene
Concen: 35.57 ng/ul
RT: 16.313 min Scan# 2291
Delta R.T. -0.006 min
Lab File: BM032475.D
Acq: 13 Oct 2021 02:11

Tgt Ion:284 Resp: 561392
Ion Ratio Lower Upper
284 100
142 38.4 29.4 44.0
249 32.9 25.5 38.3









#70
Atrazine
Concen: 35.80 ng/ul
RT: 16.454 min Scan# 2315
Delta R.T. 0.000 min
Lab File: BM032475.D
Acq: 13 Oct 2021 02:11




Instrument :
BNA_M
ClientSampleId :
SLCS820


Tgt Ion:200 Resp: 536584
Ion Ratio Lower Upper
200 100
173 26.5 22.2 33.4
215 47.8 37.5 56.3



#71
Pentachlorophenol
Concen: 37.75 ng/ul
RT: 16.648 min Scan# 2348
Delta R.T. -0.006 min
Lab File: BM032475.D
Acq: 13 Oct 2021 02:11

Tgt Ion:266 Resp: 359115
Ion Ratio Lower Upper
266 100
264 63.4 49.9 74.9
268 63.3 50.5 75.7









#72
Phenanthrene
Concen: 35.66 ng/ul
RT: 17.024 min Scan# 2412
Delta R.T. -0.006 min
Lab File: BM032475.D
Acq: 13 Oct 2021 02:11




Instrument :
BNA_M
ClientSampleId :
SLCS820


Tgt Ion:178 Resp: 2616531
Ion Ratio Lower Upper
178 100
179 15.7 12.2 18.2
176 19.4 15.4 23.0



#74
Anthracene
Concen: 35.75 ng/ul
RT: 17.118 min Scan# 2428
Delta R.T. 0.000 min
Lab File: BM032475.D
Acq: 13 Oct 2021 02:11

Tgt Ion:178 Resp: 2634965
Ion Ratio Lower Upper
178 100
179 15.4 12.3 18.5
176 18.8 14.9 22.3









#75
1,2,3,4-Tetrachlorobenzene
Concen: 34.30 ng/uL
RT: 13.054 min Scan# 1737
Delta R.T. -0.006 min
Lab File: BM032475.D
Acq: 13 Oct 2021 02:11

Instrument :
BNA_M
ClientSampleId :
SLCS820

Tgt Ion:216 Resp: 684579




Ion	Ratio	Lower	Upper
216	100		
214	78.7	62.8	94.2
218	47.8	38.4	57.6




#76
Pentachlorobenzene
Concen: 34.16 ng/uL
RT: 14.589 min Scan# 1998
Delta R.T. 0.000 min
Lab File: BM032475.D
Acq: 13 Oct 2021 02:11

Tgt Ion:250 Resp: 677218

Ion	Ratio	Lower	Upper
250	100		
248	63.0	50.7	76.1
252	63.8	51.0	76.6









#77
 Carbazole
 Concen: 37.41 ng/ul
 RT: 17.383 min Scan# 2473
 Delta R.T. 0.000 min
 Lab File: BM032475.D
 Acq: 13 Oct 2021 02:11

Instrument :
 BNA_M
 ClientSampleId :
 SLCS820

Tgt Ion:167 Resp: 2459130




Ion	Ratio	Lower	Upper
167	100		
166	21.7	17.4	26.2
139	13.1	10.6	15.8




#78
 Di-n-butylphthalate
 Concen: 37.84 ng/ul
 RT: 17.936 min Scan# 2567
 Delta R.T. -0.006 min
 Lab File: BM032475.D
 Acq: 13 Oct 2021 02:11

Tgt Ion:149 Resp: 2873936

Ion	Ratio	Lower	Upper
149	100		
150	9.4	7.4	11.0
104	5.2	4.2	6.2









#80
Fluoranthene
Concen: 36.49 ng/ul
RT: 19.036 min Scan# 2754
Delta R.T. 0.000 min
Lab File: BM032475.D
Acq: 13 Oct 2021 02:11




Instrument :
BNA_M
ClientSampleId :
SLCS820


Tgt Ion:202 Resp: 3046694
Ion Ratio Lower Upper
202 100
101 11.4 9.1 13.7
100 8.7 7.0 10.4



#82
Pyrene
Concen: 36.17 ng/ul
RT: 19.401 min Scan# 2816
Delta R.T. 0.000 min
Lab File: BM032475.D
Acq: 13 Oct 2021 02:11

Tgt Ion:202 Resp: 3102581
Ion Ratio Lower Upper
202 100
101 12.5 10.3 15.5
100 10.4 8.1 12.1









#83
Butylbenzylphthalate
Concen: 39.72 ng/ul
RT: 20.283 min Scan# 2966
Delta R.T. -0.006 min
Lab File: BM032475.D
Acq: 13 Oct 2021 02:11




Instrument :
BNA_M
ClientSampleId :
SLCS820


Tgt Ion:149 Resp: 1292725
Ion Ratio Lower Upper
149 100
91 73.7 56.3 84.5
206 20.7 16.8 25.2



#84
3,3'-Dichlorobenzidine
Concen: 36.11 ng/ul
RT: 21.065 min Scan# 3099
Delta R.T. -0.006 min
Lab File: BM032475.D
Acq: 13 Oct 2021 02:11

Tgt Ion:252 Resp: 880848
Ion Ratio Lower Upper
252 100
254 64.5 50.9 76.3
126 11.8 9.0 13.4









#85
Benzo(a)anthracene
Concen: 35.58 ng/u1
RT: 21.136 min Scan# 3111
Delta R.T. 0.000 min
Lab File: BM032475.D
Acq: 13 Oct 2021 02:11




Instrument :
BNA_M
ClientSampleId :
SLCS820


Tgt Ion:228 Resp: 2885892
Ion Ratio Lower Upper
228 100
229 19.7 15.6 23.4
226 27.6 21.8 32.6



#86
Bis(2-ethylhexyl)phthalate
Concen: 39.60 ng/u1
RT: 21.053 min Scan# 3097
Delta R.T. -0.006 min
Lab File: BM032475.D
Acq: 13 Oct 2021 02:11

Tgt Ion:149 Resp: 1792051
Ion Ratio Lower Upper
149 100
167 27.4 22.0 33.0
279 3.8 3.3 4.9









#87
Chrysene
Concen: 35.62 ng/ul
RT: 21.183 min Scan# 3119
Delta R.T. -0.006 min
Lab File: BM032475.D
Acq: 13 Oct 2021 02:11




Instrument :
BNA_M
ClientSampleId :
SLCS820


Tgt Ion:228 Resp: 2845389
Ion Ratio Lower Upper
228 100
226 30.5 24.1 36.1
229 19.8 15.5 23.3



#89
Di-n-octyl phthalate
Concen: 41.90 ng/ul
RT: 21.900 min Scan# 3241
Delta R.T. 0.000 min
Lab File: BM032475.D
Acq: 13 Oct 2021 02:11

Tgt Ion:149 Resp: 3229830









#90
Benzo(b)fluoranthene
Concen: 37.55 ng/u1
RT: 22.653 min Scan# 3369
Delta R.T. -0.006 min
Lab File: BM032475.D
Acq: 13 Oct 2021 02:11




Instrument :
BNA_M
ClientSampleId :
SLCS820


Tgt Ion:252 Resp: 3092202
Ion Ratio Lower Upper
252 100
253 22.3 17.6 26.4
125 10.8 8.2 12.4



#91
Benzo(k)fluoranthene
Concen: 36.00 ng/u1
RT: 22.695 min Scan# 3376
Delta R.T. -0.006 min
Lab File: BM032475.D
Acq: 13 Oct 2021 02:11

Tgt Ion:252 Resp: 2688750
Ion Ratio Lower Upper
252 100
253 21.7 17.5 26.3
125 10.5 8.4 12.6









#93
Benzo(a)pyrene
Concen: 36.51 ng/ul
RT: 23.200 min Scan# 3462
Delta R.T. 0.000 min
Lab File: BM032475.D
Acq: 13 Oct 2021 02:11




Instrument :
BNA_M
ClientSampleId :
SLCS820


Tgt Ion:252 Resp: 2844034
Ion Ratio Lower Upper
252 100
253 22.0 17.8 26.6
125 11.4 9.2 13.8



#94
Indeno(1,2,3-cd)pyrene
Concen: 33.34 ng/ul
RT: 25.412 min Scan# 3838
Delta R.T. -0.006 min
Lab File: BM032475.D
Acq: 13 Oct 2021 02:11

Tgt Ion:276 Resp: 2871066
Ion Ratio Lower Upper
276 100
138 28.2 21.8 32.6
277 25.5 20.4 30.6









#95
Dibenzo(a,h)anthracene
Concen: 33.72 ng/ul
RT: 25.418 min Scan# 3839
Delta R.T. -0.006 min
Lab File: BM032475.D
Acq: 13 Oct 2021 02:11

Instrument :
BNA_M
ClientSampleId :
SLCS820

Tgt Ion:278 Resp: 2491467
Ion Ratio Lower Upper
278 100
139 16.6 13.4 20.2
279 23.5 19.0 28.4



#96
Benzo(g,h,i)perylene
Concen: 32.53 ng/ul
RT: 26.065 min Scan# 3949
Delta R.T. -0.006 min
Lab File: BM032475.D
Acq: 13 Oct 2021 02:11

Tgt Ion:276 Resp: 2450991
Ion Ratio Lower Upper
276 100
138 22.7 18.6 27.8
277 23.7 19.0 28.4

