

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM101121\
 Data File : BM032476.D
 Acq On : 13 Oct 2021 02:47
 Operator : CG/JU
 Sample : M4128-01
 Misc :
 ALS Vial : 62 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 C0AB6

Quant Time: Oct 13 04:31:19 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-BM100521.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Oct 12 09:54:27 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.607	152	467033	20.00	ng/ul	0.00
20) Naphthalene-d8	10.395	136	2042264	20.00	ng/ul	0.00
38) Acenaphthene-d10	14.254	164	1393384	20.00	ng/ul	0.00
64) Phenanthrene-d10	16.983	188	2891648	20.00	ng/ul	0.00
79) Chrysene-d12	21.147	240	2794991	20.00	ng/ul	0.00
88) Perylene-d12	23.288	264	2784038	20.00	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	2.955	96	31188	2.68	ng/uL	0.00
4) Pyridine-d5	3.384	84	434731	13.21	ng/ul	0.00
7) Phenol-d5	6.795	99	551152	13.88	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	6.931	67	316458	13.88	ng/ul	0.00
11) 2-Chlorophenol-d4	7.142	132	430812	14.21	ng/ul	0.00
15) 4-Methylphenol-d8	8.337	113	462500	14.42	ng/ul	0.00
21) Nitrobenzene-d5	8.760	128	222821	15.46	ng/ul	0.00
24) 2-Nitrophenol-d4	9.483	143	246588	15.88	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.025	165	440285	14.24	ng/ul	0.00
31) 4-Chloroaniline-d4	10.531	131	636778	13.96	ng/ul	0.00
46) Dimethylphthalate-d6	13.660	166	1540364	15.97	ng/ul	0.00
49) Acenaphthylene-d8	13.942	160	1847531	15.77	ng/ul	0.00
54) 4-Nitrophenol-d4	14.460	143	243624	12.86	ng/ul	-0.01
60) Fluorene-d10	15.242	176	1329641	16.26	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.360	200	199393	12.10	ng/ul	0.00
73) Anthracene-d10	17.077	188	2067397	16.42	ng/ul	0.00
81) Pyrene-d10	19.365	212	2337220	16.62	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.153	264	2119730	15.20	ng/ul	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM101121\
Data File : BM032476.D
Acq On : 13 Oct 2021 02:47
Operator : CG/JU
Sample : M4128-01
Misc :
ALS Vial : 62 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
C0AB6

Quant Time: Oct 13 04:31:19 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-BM100521.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Oct 12 09:54:27 2021
Response via : Initial Calibration

