

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM101121\
 Data File : BM032479.D
 Acq On : 13 Oct 2021 04:35
 Operator : CG/JU
 Sample : M4128-03
 Misc :
 ALS Vial : 65 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 C0AB8

Quant Time: Oct 13 05:44:14 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-BM100521.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Oct 12 09:54:27 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.607	152	206391	20.00	ng/ul	0.00
20) Naphthalene-d8	10.395	136	903257	20.00	ng/ul	0.00
38) Acenaphthene-d10	14.254	164	607487	20.00	ng/ul	0.00
64) Phenanthrene-d10	16.983	188	1275681	20.00	ng/ul	0.00
79) Chrysene-d12	21.142	240	1251551	20.00	ng/ul	-0.01
88) Perylene-d12	23.289	264	1210271	20.00	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	2.955	96	45035	8.75	ng/uL	0.00
4) Pyridine-d5	3.384	84	621356	42.73	ng/ul	0.00
7) Phenol-d5	6.796	99	804019	45.81	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	6.931	67	458911	45.54	ng/ul	0.00
11) 2-Chlorophenol-d4	7.143	132	637979	47.61	ng/ul	0.00
15) 4-Methylphenol-d8	8.337	113	670637	47.31	ng/ul	0.00
21) Nitrobenzene-d5	8.760	128	326466	51.23	ng/ul	0.00
24) 2-Nitrophenol-d4	9.484	143	364387	53.04	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.025	165	636743	46.56	ng/ul	0.00
31) 4-Chloroaniline-d4	10.531	131	925591	45.89	ng/ul	0.00
46) Dimethylphthalate-d6	13.660	166	2095308	49.84	ng/ul	0.00
49) Acenaphthylene-d8	13.942	160	2544268	49.82	ng/ul	0.00
54) 4-Nitrophenol-d4	14.466	143	360994	43.72	ng/ul	0.00
60) Fluorene-d10	15.242	176	1782413	49.99	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.360	200	292800	40.27	ng/ul	0.00
73) Anthracene-d10	17.077	188	2769417	49.87	ng/ul	0.00
81) Pyrene-d10	19.365	212	3112411	49.43	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.153	264	2826971	46.64	ng/ul	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM101121\
Data File : BM032479.D
Acq On : 13 Oct 2021 04:35
Operator : CG/JU
Sample : M4128-03
Misc :
ALS Vial : 65 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
C0AB8

Quant Time: Oct 13 05:44:14 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-BM100521.M
Quant Title : SVOA CALIBRATION
QLast Update : Tue Oct 12 09:54:27 2021
Response via : Initial Calibration

