



Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM101121\
 Data File : BM032413.D
 Acq On : 11 Oct 2021 09:42
 Operator : CG/JU
 Sample : DFTPP014
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP014

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-BM100521.M
 Title : SVOA CALIBRATION
 Last Update : Mon Oct 11 11:06:33 2021



AutoFind: Scans 2483, 2484, 2485; Background Corrected with Scan 2477

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	31.5	162162	PASS
68	69	0.00	2	1.7	2849	PASS
69	198	0.00	100	32.5	167749	PASS
70	69	0.00	2	0.5	828	PASS
127	198	10	80	44.2	227733	PASS
197	198	0.00	2	0.7	3673	PASS
198	198	100	100	100.0	515584	PASS
199	198	5	9	6.8	35115	PASS
275	198	10	60	26.6	137061	PASS
365	198	1	100	3.6	18314	PASS
441	198	0.01	100	13.4	69272	PASS
442	198	50	100	87.1	449003	PASS
443	442	15	24	19.3	86843	PASS

m/z	Abundance
38.10	165.0
39.00	445.0
40.00	774.0
41.00	859.0
42.00	484.0
43.10	1498.0
44.00	3716.0
45.00	553.0
46.80	161.0
49.00	481.0
51.10	475.0
52.10	182.0
53.00	331.0
54.00	181.0
55.00	1058.0
56.00	229.0
57.00	811.0
58.10	206.0
60.10	444.0
63.00	151.0
65.00	250.0
66.90	358.0
68.00	245.0
69.10	406.0
70.00	199.0
70.90	241.0
73.00	1154.0
74.90	175.0
77.00	572.0
78.00	311.0
79.10	333.0
81.10	302.0
82.90	219.0
84.00	454.0
84.90	329.0
85.90	380.0
91.00	400.0
95.10	290.0
96.00	340.0
97.20	191.0
97.90	179.0
105.00	178.0
108.90	212.0
113.90	223.0
115.00	166.0
131.10	186.0
132.90	233.0
135.10	155.0
147.00	283.0
165.00	225.0
167.00	267.0
191.10	231.0
193.00	390.0
207.00	1811.0
208.00	550.0
209.10	228.0
267.00	175.0
281.00	474.0
283.00	173.0

Instrument :
BNA_M
ClientSampleId :
DFTPP014

m/z	Abundance
37.00	793.0
38.10	2287.0
39.10	13497.0
40.10	1381.0
41.10	878.0
42.00	280.0
43.10	1260.0
44.00	2779.0
45.00	818.0
47.00	173.0
47.80	217.0
49.10	1754.0
50.10	49336.0
51.10	186560.0
52.10	9737.0
53.00	526.0
55.00	1420.0
56.00	5894.0
57.00	12863.0
58.00	830.0
59.10	290.0
60.00	498.0
61.00	2357.0
62.00	2636.0
63.00	7731.0
64.00	1236.0
65.00	3932.0
66.00	401.0
67.00	460.0
68.10	3227.0
69.00	191744.0
70.00	947.0
71.00	309.0
72.00	174.0
73.10	2187.0
74.10	20200.0
75.00	31440.0
76.10	10729.0
77.10	227200.0
78.10	15938.0
79.00	15594.0
80.00	11517.0
81.00	16071.0
82.00	3997.0
83.10	3545.0
84.00	826.0
85.00	3344.0
86.00	4718.0
87.10	2071.0
88.00	926.0
89.00	423.0
91.00	3619.0
92.00	4270.0
93.00	25288.0
94.00	1880.0
95.00	434.0
96.10	1525.0
96.90	624.0
98.00	20400.0
99.00	16920.0
100.00	1607.0
101.00	9510.0
102.00	531.0
103.00	2947.0
104.00	5495.0
105.00	5928.0
106.10	1937.0
107.00	72992.0
108.00	10705.0
109.10	2233.0
110.00	130288.0
111.00	19312.0
112.00	2373.0
113.00	730.0
114.10	301.0
115.00	509.0
116.00	3665.0
117.00	63344.0
118.00	4888.0
119.00	516.0
120.00	912.0
121.00	604.0
122.10	4318.0
123.00	7452.0
124.00	3196.0
125.00	3456.0
126.20	1771.0
127.00	242112.0
128.00	18056.0
129.00	95632.0
130.00	8243.0
131.00	1689.0

Instrument :
BNA_M
ClientSampleId :
DFTPP014

132.00	883.0
133.00	500.0
134.00	2934.0
135.00	7947.0
136.00	3321.0
137.10	4286.0
137.90	1141.0
138.90	716.0
140.00	1058.0
141.00	12843.0
142.10	3639.0
143.00	2925.0
143.90	662.0
145.10	644.0
146.00	2344.0
147.10	6657.0
148.00	15769.0
149.00	3216.0
150.00	1020.0
151.00	1569.0
152.00	1029.0
153.10	3424.0
154.10	3138.0
155.10	6885.0
156.10	11037.0
157.10	2152.0
157.90	2458.0
159.00	1791.0
160.00	4027.0
161.00	5933.0
162.00	1824.0
163.00	522.0
164.00	515.0
165.00	4627.0
166.10	3594.0
167.10	26464.0
168.00	16138.0
169.00	2327.0
170.00	1054.0
171.00	1572.0
172.00	2394.0
173.00	3136.0
174.10	5117.0
175.10	9872.0
176.10	2709.0
177.00	5130.0
177.90	1677.0
179.00	19280.0
180.00	13871.0
181.00	6642.0
182.00	1056.0
183.00	585.0
184.00	1821.0
185.10	9599.0
186.10	69680.0
187.10	21136.0
188.10	2393.0
189.00	4764.0
190.00	841.0
191.10	2344.0
192.00	6072.0
193.10	7061.0
194.10	1559.0
195.10	894.0
196.10	15529.0
197.10	3567.0
198.00	528448.0
199.00	36784.0
200.00	2497.0
201.30	1892.0
203.00	3464.0
204.10	18576.0
205.10	32024.0
206.10	138048.0
207.10	19544.0
208.00	4976.0
209.00	1981.0
210.00	2082.0
211.00	5401.0
212.90	455.0
214.00	254.0
214.90	1583.0
216.00	3154.0
217.00	35904.0
218.00	4639.0
219.20	391.0
221.10	19912.0
221.90	4646.0
223.00	8381.0
224.10	74792.0
225.10	19224.0
226.00	2321.0
227.00	32504.0
228.10	4668.0
229.10	6584.0
230.00	851.0

Instrument :
BNA_M
ClientSampleId :
DFTPP014

231.00	2646.0
232.10	493.0
232.90	608.0
234.10	2012.0
235.00	2353.0
236.10	1542.0
237.00	2922.0
238.10	404.0
239.00	1020.0
239.90	1118.0
241.00	1557.0
242.10	3422.0
243.00	4122.0
244.10	57600.0
245.10	8059.0
246.00	11621.0
247.00	2685.0
248.10	383.0
249.00	1749.0
249.90	473.0
251.00	614.0
252.10	629.0
253.10	1209.0
254.10	2266.0
255.00	291456.0
256.10	43616.0
257.00	3625.0
258.00	16464.0
259.00	2908.0
260.00	537.0
261.00	553.0
263.00	259.0
264.00	375.0
265.00	7444.0
266.00	899.0
267.10	205.0
269.10	166.0
269.90	250.0
271.00	721.0
272.10	698.0
273.10	8256.0
274.00	23416.0
275.10	129744.0
276.00	18024.0
277.00	11439.0
278.00	2227.0
279.10	412.0
281.10	489.0
282.00	346.0
283.10	1418.0
284.00	955.0
285.00	1860.0
286.00	353.0
288.00	157.0
288.90	409.0
290.00	312.0
291.00	190.0
291.90	459.0
293.00	2538.0
294.00	788.0
295.00	743.0
296.00	34784.0
297.00	5002.0
297.90	325.0
301.00	546.0
302.00	514.0
303.00	4263.0
304.00	1186.0
304.90	200.0
308.00	305.0
308.90	251.0
310.10	513.0
313.10	310.0
314.10	1857.0
315.00	4096.0
316.10	1874.0
317.00	359.0
321.00	1143.0
322.10	594.0
323.10	10679.0
324.00	2578.0
325.00	306.0
325.90	268.0
327.00	2207.0
328.00	960.0
329.10	165.0
332.10	893.0
333.00	1071.0
334.10	7462.0
335.00	1962.0
335.90	251.0
340.00	183.0
341.10	1473.0
342.00	367.0
346.00	2215.0
347.00	465.0

Instrument :
BNA_M
ClientSampleId :
DFTPP014

350.90	179.0
352.00	3615.0
353.10	2574.0
354.10	3405.0
355.00	844.0
359.00	197.0
365.00	15854.0
366.00	2373.0
366.90	292.0
369.90	309.0
371.00	821.0
372.10	5640.0
373.10	1374.0
374.00	198.0
377.10	242.0
383.10	1586.0
383.90	365.0
385.00	174.0
390.10	775.0
390.90	392.0
392.00	506.0
400.70	229.0
401.10	274.0
402.00	2501.0
403.00	3513.0
404.10	1260.0
405.00	261.0
421.00	3391.0
422.10	3259.0
423.10	19624.0
424.10	5685.0
425.00	641.0
440.10	319.0
441.10	55400.0
442.10	360960.0
443.10	71712.0
444.10	6742.0
445.00	487.0

Instrument :
BNA_M
ClientSampleId :
DFTPP014

m/z	Abundance
37.10	814.0
38.10	2165.0
39.10	12485.0
40.00	1378.0
41.10	743.0
42.10	351.0
43.10	1319.0
44.00	3377.0
45.00	797.0
46.90	153.0
49.10	1864.0
50.10	49344.0
51.10	187456.0
52.10	9960.0
53.00	637.0
55.00	1640.0
56.00	5755.0
57.00	13106.0
58.00	770.0
59.00	296.0
60.10	562.0
61.00	2445.0
62.10	2904.0
63.00	7273.0
64.10	989.0
65.10	3774.0
66.00	401.0
66.90	486.0
68.10	3581.0
69.00	194944.0
70.00	1325.0
71.10	250.0
72.00	178.0
73.00	2722.0
74.10	20680.0
75.00	32984.0
76.10	12132.0
77.10	239744.0
78.10	16170.0
79.00	14952.0
80.00	11461.0
81.00	16528.0
82.00	3849.0
83.00	3688.0
84.00	785.0
85.00	3322.0
86.00	5221.0
87.10	2361.0
88.00	1009.0
89.00	454.0
91.10	4118.0
92.00	4881.0
93.00	27608.0
94.00	1909.0
95.00	651.0
95.90	1192.0
97.10	890.0
98.00	22920.0
99.00	17464.0
100.00	1563.0
101.00	9664.0
102.10	492.0
103.00	3183.0
104.00	6278.0
105.10	5555.0
106.00	2163.0
107.00	77424.0
108.00	11301.0
109.10	2090.0
110.00	140800.0
111.00	21184.0
112.00	2829.0
113.00	853.0
114.00	312.0
114.90	509.0
116.10	4299.0
117.00	69272.0
118.00	5025.0
118.90	608.0
120.00	889.0
121.00	429.0
122.00	5360.0
123.10	8280.0
124.00	3413.0
125.10	3345.0
126.20	1690.0
127.00	268928.0
128.00	19848.0
129.00	102072.0
130.00	8529.0
131.00	1966.0
132.10	1144.0

Instrument :
BNA_M
ClientSampleId :
DFTPP014

132.90	668.0
134.00	3377.0
135.00	8827.0
136.00	3630.0
137.00	5092.0
137.90	802.0
138.80	603.0
140.00	1415.0
141.00	13840.0
142.10	4724.0
143.00	2956.0
144.10	1026.0
145.00	732.0
146.10	2565.0
147.00	7909.0
148.00	16504.0
149.00	3786.0
150.10	1125.0
151.10	1990.0
152.10	916.0
153.00	4099.0
154.00	3461.0
155.10	8103.0
156.10	12005.0
157.10	2111.0
158.00	2781.0
159.00	2072.0
160.00	4749.0
161.00	6855.0
162.10	2057.0
163.10	736.0
164.00	619.0
165.00	5811.0
166.00	4666.0
167.00	31984.0
168.10	16848.0
169.00	2380.0
170.10	946.0
171.00	1068.0
172.00	2847.0
173.00	3451.0
174.00	6441.0
175.10	12149.0
176.00	3373.0
177.00	5184.0
178.00	1921.0
179.00	22992.0
180.00	15207.0
181.00	6607.0
182.10	1326.0
183.00	719.0
184.10	1771.0
185.00	10213.0
186.10	81728.0
187.00	23720.0
188.00	2706.0
189.00	5451.0
190.00	967.0
191.00	2895.0
192.00	7439.0
193.00	8285.0
194.00	1859.0
195.00	1086.0
196.00	16504.0
197.10	4798.0
198.00	604288.0
199.00	41944.0
200.10	3213.0
201.60	2634.0
203.00	4404.0
204.10	22168.0
205.10	36936.0
206.10	161536.0
207.10	22656.0
208.00	5872.0
209.00	1919.0
210.10	2683.0
211.00	6788.0
213.00	392.0
213.80	243.0
215.00	1960.0
216.10	3776.0
217.00	41864.0
218.00	5740.0
219.00	659.0
221.10	24664.0
221.80	5988.0
223.00	9696.0
224.10	87720.0
225.10	23424.0
226.10	2605.0
227.00	35760.0
228.00	5225.0
229.00	8017.0
230.00	943.0
231.10	3333.0

Instrument :
BNA_M
ClientSampleId :
DFTPP014

232.10	540.0
233.10	783.0
234.00	2784.0
235.00	2845.0
236.00	2191.0
237.00	3122.0
237.90	451.0
239.00	1499.0
240.00	1209.0
241.00	1956.0
242.10	4266.0
243.10	5257.0
244.10	69776.0
245.10	9361.0
246.00	13658.0
247.00	2979.0
248.00	708.0
249.10	2482.0
250.00	500.0
251.00	651.0
251.90	765.0
253.00	1601.0
254.20	3306.0
255.00	360000.0
256.00	52568.0
257.00	4417.0
258.00	20584.0
259.00	3240.0
260.00	685.0
261.10	657.0
262.90	256.0
264.10	580.0
265.00	8693.0
266.00	1200.0
266.90	176.0
270.00	419.0
271.00	1016.0
272.00	926.0
273.00	10943.0
274.10	29616.0
275.10	164096.0
276.00	21312.0
277.00	14386.0
278.00	2320.0
279.10	612.0
281.20	475.0
282.00	466.0
283.00	1521.0
284.00	1081.0
285.00	2263.0
286.10	653.0
287.90	174.0
288.90	436.0
290.00	485.0
291.00	316.0
292.00	842.0
293.00	3134.0
294.00	882.0
295.00	985.0
296.00	45904.0
297.00	6225.0
298.10	408.0
301.10	605.0
302.10	824.0
303.10	5085.0
304.00	1615.0
305.00	214.0
308.00	622.0
309.00	518.0
310.00	609.0
310.80	157.0
313.00	483.0
314.10	2338.0
315.00	4826.0
316.10	2981.0
317.00	373.0
320.00	300.0
321.10	1397.0
322.00	963.0
323.10	15547.0
324.10	3065.0
325.00	254.0
325.90	352.0
327.00	3038.0
328.00	1569.0
329.00	338.0
332.10	912.0
333.00	1408.0
334.10	9629.0
335.10	2567.0
336.10	476.0
338.90	260.0
339.80	271.0
341.10	1713.0
341.90	480.0
346.10	3858.0

Instrument :
BNA_M
ClientSampleId :
DFTPP014

347.00	615.0
350.80	225.0
351.00	224.0
352.10	4262.0
353.10	3255.0
354.10	5233.0
355.00	982.0
359.10	287.0
365.00	22208.0
366.00	3584.0
366.90	196.0
370.00	462.0
371.00	1201.0
372.10	8143.0
373.10	2132.0
374.10	243.0
377.10	181.0
383.00	2183.0
384.00	773.0
385.10	224.0
390.00	1097.0
391.00	737.0
392.10	471.0
401.00	504.0
402.00	3362.0
403.00	4741.0
404.10	1705.0
405.10	258.0
410.00	178.0
414.80	185.0
421.00	4317.0
422.10	3983.0
423.10	29568.0
424.10	7867.0
425.00	799.0
437.90	197.0
438.90	202.0
441.10	82360.0
442.10	532608.0
443.10	104408.0
444.00	9674.0
445.00	570.0

Instrument :
BNA_M
ClientSampleId :
DFTPP014

m/z	Abundance
35.90	219.0
37.10	511.0
38.10	1304.0
39.10	7684.0
40.00	1017.0
41.10	925.0
42.00	310.0
43.00	1214.0
44.00	3176.0
45.00	654.0
49.00	1351.0
50.10	29752.0
51.10	113896.0
52.10	5800.0
52.90	464.0
53.90	161.0
55.10	1462.0
56.00	2975.0
57.10	8272.0
58.00	480.0
60.00	443.0
61.00	1525.0
62.10	1869.0
63.10	4671.0
64.10	707.0
65.10	2627.0
65.90	260.0
67.00	580.0
68.00	2473.0
69.00	117776.0
70.00	809.0
71.10	257.0
72.00	182.0
73.10	2245.0
74.10	12361.0
75.10	19800.0
76.10	6733.0
77.10	143424.0
78.10	10752.0
79.00	9485.0
80.00	7122.0
81.00	10115.0
82.00	2791.0
83.00	2582.0
83.90	852.0
85.00	2275.0
86.00	3279.0
87.00	1511.0
87.90	588.0
89.10	206.0
91.10	2516.0
92.10	2721.0
93.00	17080.0
94.00	1242.0
94.90	456.0
96.00	862.0
97.10	426.0
98.00	14157.0
99.00	11811.0
100.10	810.0
101.00	5760.0
102.00	302.0
103.00	2014.0
104.00	3849.0
105.00	3298.0
106.10	1221.0
107.00	48848.0
108.00	7278.0
109.10	1408.0
110.00	87768.0
111.00	12696.0
112.00	1552.0
112.90	546.0
114.10	234.0
114.90	408.0
116.00	2571.0
117.00	43112.0
118.00	3090.0
119.00	576.0
119.90	712.0
121.10	274.0
122.00	3401.0
123.10	5135.0
124.00	2295.0
125.00	2264.0
126.20	1270.0
127.00	172160.0
128.00	12724.0
129.00	66376.0
130.00	4959.0
131.00	1251.0
132.10	607.0

Instrument :
BNA_M
ClientSampleId :
DFTPP014

132.90	513.0
134.00	2231.0
135.00	5632.0
136.00	2122.0
137.00	2899.0
138.00	590.0
138.90	331.0
140.00	1023.0
141.00	9508.0
142.00	2956.0
143.00	1896.0
144.10	415.0
144.90	619.0
146.10	1706.0
147.00	4895.0
148.00	10682.0
149.00	2296.0
150.00	826.0
151.20	1166.0
152.10	996.0
153.00	2504.0
154.00	2295.0
155.00	5156.0
156.10	7635.0
157.10	1498.0
158.00	1505.0
158.90	1221.0
160.00	2873.0
161.00	4241.0
161.90	1003.0
163.00	404.0
164.10	430.0
165.00	3796.0
166.10	3151.0
167.00	21040.0
168.00	11688.0
169.00	1748.0
170.00	650.0
171.10	871.0
172.00	1548.0
173.00	2292.0
174.10	3935.0
175.10	7411.0
176.10	2143.0
177.00	3485.0
178.00	1213.0
179.00	14484.0
180.10	10602.0
181.00	4814.0
182.10	856.0
183.00	414.0
184.10	1246.0
185.00	7241.0
186.10	57064.0
187.00	16648.0
188.10	1535.0
189.00	3135.0
190.00	672.0
191.10	1713.0
192.00	4596.0
193.00	5804.0
194.10	1074.0
194.90	597.0
195.10	605.0
196.00	11557.0
197.10	2655.0
198.00	414016.0
199.00	26616.0
200.00	1944.0
201.60	1373.0
203.00	3142.0
204.10	15550.0
205.10	27576.0
206.10	107704.0
207.10	15513.0
208.00	4566.0
209.00	1504.0
211.00	4069.0
212.90	427.0
215.00	1194.0
216.00	2264.0
217.00	29424.0
218.00	3528.0
218.90	297.0
221.10	17216.0
221.90	3602.0
223.00	5959.0
224.10	62376.0
225.10	16005.0
226.10	1800.0
227.00	26168.0
228.00	3722.0
229.00	5196.0
230.00	843.0
231.10	2152.0
232.00	394.0

Instrument :
BNA_M
ClientSampleId :
DFTPP014

233.00	507.0
234.00	1768.0
235.00	1909.0
235.90	1176.0
237.00	2465.0
238.10	337.0
239.00	1216.0
240.00	932.0
241.00	1464.0
242.10	3245.0
243.10	3687.0
244.10	50072.0
245.10	6245.0
246.00	9995.0
247.00	2321.0
248.00	493.0
249.00	1713.0
250.00	337.0
251.00	565.0
252.00	648.0
253.00	1385.0
254.10	2124.0
255.00	252992.0
256.10	37144.0
257.00	2662.0
258.00	14775.0
259.00	2426.0
260.10	489.0
261.10	424.0
263.00	238.0
264.00	382.0
265.00	5740.0
266.00	772.0
267.00	301.0
270.00	279.0
271.00	563.0
272.10	733.0
273.00	7937.0
274.00	20568.0
275.10	117344.0
276.10	15408.0
277.00	10150.0
278.00	2039.0
279.00	352.0
281.10	515.0
282.10	310.0
283.00	1125.0
284.10	629.0
285.00	1909.0
286.00	313.0
288.10	166.0
289.00	393.0
289.80	293.0
291.00	226.0
292.00	502.0
293.00	2558.0
294.00	824.0
295.10	651.0
296.00	34160.0
297.00	4781.0
298.00	413.0
301.00	446.0
302.00	546.0
303.10	3783.0
304.10	1024.0
308.00	450.0
308.90	368.0
309.90	430.0
312.10	150.0
312.90	287.0
314.10	1497.0
315.00	4049.0
316.00	2087.0
317.10	363.0
320.00	202.0
321.00	1071.0
322.10	598.0
323.10	10607.0
324.00	1768.0
325.00	266.0
326.10	205.0
327.00	2038.0
328.00	1154.0
332.00	643.0
333.00	1100.0
334.00	7758.0
335.10	1953.0
336.10	259.0
339.90	221.0
341.00	1409.0
342.00	444.0
346.00	2918.0
347.00	578.0
351.10	225.0
352.00	3399.0
353.10	2491.0

Instrument :
BNA_M
ClientSampleId :
DFTPP014

354.10	3868.0
355.10	794.0
356.10	159.0
359.00	318.0
365.00	16880.0
366.00	2682.0
367.00	184.0
370.00	298.0
371.00	1012.0
372.00	6430.0
373.00	1727.0
374.20	171.0
383.10	1653.0
384.00	508.0
385.00	213.0
390.00	724.0
391.10	633.0
392.00	468.0
401.00	413.0
402.00	2604.0
403.10	3780.0
404.00	1258.0
404.90	196.0
415.10	182.0
421.00	3398.0
422.00	3461.0
423.10	23344.0
424.10	5644.0
425.00	599.0
439.20	175.0
441.10	70056.0
442.10	453440.0
443.10	84408.0
444.10	7425.0
445.00	681.0

Instrument :
BNA_M
ClientSampleId :
DFTPP014