


Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM101124\
Data File : BM048010.D
Acq On : 11 Oct 2024 13:26
Operator : RC/JU
Sample : DFTPP048
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP048

Integration File: LSCINT.P

Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM092724.MA.M
Title : SVOA CALIBRATION
Last Update : Thu Oct 10 05:10:14 2024



AutoFind: Scans 2484, 2485, 2486; Background Corrected with Scan 2477

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	34.4	233062	PASS
68	69	0.00	2	1.5	3632	PASS
69	198	0.00	100	35.9	243434	PASS
70	69	0.00	2	0.8	2068	PASS
127	198	10	80	44.1	298576	PASS
197	198	0.00	2	0.5	3167	PASS
198	198	100	100	100.0	677205	PASS
199	198	5	9	6.6	44597	PASS
275	198	10	60	24.4	164949	PASS
365	198	1	100	3.5	23389	PASS
441	198	0.01	100	13.7	92659	PASS
442	198	50	100	84.0	568704	PASS
443	442	15	24	19.4	110571	PASS

m/z	Abundance
36.00	163.0
39.00	483.0
40.00	2531.0
41.10	1884.0
42.00	607.0
43.10	4747.0
44.00	7340.0
45.00	370.0
49.00	441.0
51.00	325.0
53.00	167.0
53.90	209.0
55.10	2367.0
56.00	1075.0
57.10	6053.0
58.10	467.0
59.00	169.0
59.90	183.0
60.10	178.0
60.90	180.0
67.10	343.0
68.20	502.0
69.10	1515.0
70.10	910.0
71.10	4257.0
72.10	357.0
73.00	919.0
77.00	403.0
78.00	311.0
79.10	236.0
81.00	298.0
82.00	353.0
83.10	1027.0
84.00	817.0
85.10	3327.0
86.00	418.0
88.80	158.0
91.00	401.0
93.10	233.0
95.20	171.0
96.10	475.0
97.10	861.0
98.00	347.0
99.10	1081.0
105.00	223.0
106.90	172.0
111.10	484.0
112.10	282.0
113.10	489.0
115.10	185.0
121.10	201.0
125.10	188.0
126.10	302.0
127.20	389.0
132.90	297.0
135.10	215.0
140.20	172.0
143.30	158.0
147.10	273.0
149.10	188.0
154.10	162.0
155.20	177.0
167.00	397.0
168.00	168.0
169.10	158.0
190.90	406.0
193.10	307.0
207.10	2993.0
208.00	640.0
209.00	316.0
210.10	169.0
221.30	191.0
281.10	618.0
282.10	177.0

Instrument :
BNA_M
ClientSampleId :
DFTPP048

m/z	Abundance
36.00	314.0
37.00	1156.0
38.00	3527.0
39.10	22128.0
40.00	3720.0
41.10	4258.0
42.00	1154.0
43.10	10220.0
44.00	8435.0
45.00	952.0
47.00	287.0
48.20	154.0
49.00	1973.0
50.10	73216.0
51.10	298816.0
52.10	16736.0
53.00	1080.0
54.00	454.0
55.10	5214.0
56.00	10723.0
57.10	33680.0
58.00	1595.0
59.10	371.0
60.00	391.0
61.10	3190.0
62.00	4246.0
63.10	11029.0
64.00	1703.0
65.10	5888.0
66.00	623.0
67.00	1196.0
68.10	5377.0
69.00	309312.0
70.10	3240.0
71.10	10290.0
72.10	991.0
73.10	2969.0
74.10	29024.0
75.10	46784.0
76.10	16488.0
77.10	353344.0
78.10	24784.0
79.00	20112.0
80.00	16158.0
81.10	23632.0
82.00	7115.0
83.10	8119.0
84.10	1999.0
85.10	10955.0
86.00	7391.0
87.10	3164.0
88.00	1131.0
89.00	708.0
90.20	166.0
91.00	4985.0
92.10	5576.0
93.00	35760.0
94.00	2633.0
95.00	792.0
96.10	2341.0
97.10	2356.0
98.00	28488.0
99.00	26328.0
100.10	1976.0
101.10	13470.0
102.00	805.0
103.00	4060.0
104.00	7262.0
105.00	7758.0
106.10	2830.0
107.10	103424.0
108.00	16346.0
109.10	2652.0
110.00	190976.0
111.00	28816.0
112.00	4313.0
113.10	2187.0
114.10	312.0
115.00	595.0
116.10	5586.0
117.00	82440.0
118.00	6838.0
119.00	748.0
120.00	1467.0
121.00	379.0
122.00	6083.0
123.00	10128.0
124.00	5137.0
125.00	4657.0
126.10	1529.0
127.10	364224.0
128.10	27016.0

Instrument :
BNA_M
ClientSampleId :
DFTPP048

129.00	137600.0
130.10	12106.0
131.10	2433.0
132.10	1307.0
133.00	647.0
134.10	3518.0
135.00	11511.0
136.00	4770.0
137.10	6779.0
138.00	1388.0
139.00	959.0
140.10	1747.0
141.00	17696.0
142.00	5284.0
143.00	4040.0
144.10	972.0
145.00	1142.0
146.00	3065.0
147.00	8998.0
148.10	21488.0
149.00	4801.0
150.00	1171.0
151.10	2727.0
151.80	1216.0
153.00	5148.0
154.00	4097.0
155.10	10496.0
156.10	14485.0
157.10	2731.0
158.00	2973.0
159.00	2497.0
160.00	5574.0
161.00	9113.0
162.10	2284.0
163.00	883.0
164.00	1447.0
165.00	6386.0
166.00	4905.0
167.10	36568.0
168.00	23272.0
169.10	3340.0
170.10	1176.0
171.00	1581.0
172.00	2801.0
173.10	3623.0
174.10	8016.0
175.10	13961.0
176.10	4167.0
177.10	6292.0
178.00	2197.0
179.00	26808.0
180.00	19768.0
181.00	9330.0
182.10	1621.0
183.10	868.0
184.10	2152.0
185.10	12990.0
186.10	98824.0
187.10	28104.0
188.00	3285.0
189.10	5550.0
190.00	1038.0
191.10	2824.0
192.00	9108.0
193.10	9764.0
194.00	2155.0
195.10	1024.0
196.00	20168.0
197.00	4088.0
198.00	772736.0
199.00	52160.0
200.10	3891.0
201.50	3150.0
203.00	4915.0
204.10	27104.0
205.10	45344.0
206.10	195264.0
207.10	28392.0
208.10	6664.0
209.10	2414.0
210.20	2730.0
211.00	7190.0
211.90	1493.0
213.00	499.0
214.10	296.0
215.10	2187.0
216.10	4283.0
217.00	52800.0
218.10	6216.0
219.10	676.0
221.10	27160.0
221.90	6868.0
223.00	11345.0
224.10	109656.0
225.10	28448.0
226.10	3231.0

Instrument :
BNA_M
ClientSampleId :
DFTPP048

227.00	43368.0
228.00	5644.0
229.00	9255.0
230.00	1129.0
231.10	3899.0
232.00	626.0
233.00	829.0
234.00	2654.0
235.10	2971.0
236.00	2380.0
237.10	4138.0
238.10	590.0
239.00	1797.0
240.00	1409.0
241.10	2328.0
242.10	5108.0
243.10	6077.0
244.10	87464.0
245.10	10724.0
246.00	15203.0
247.10	3376.0
248.00	745.0
249.10	2809.0
250.10	518.0
251.00	765.0
252.10	838.0
253.10	1787.0
254.10	2350.0
255.10	425536.0
256.10	65392.0
257.00	4744.0
258.00	24016.0
259.00	3789.0
260.10	585.0
261.00	742.0
263.00	196.0
264.00	458.0
265.00	9246.0
266.10	1350.0
267.10	347.0
268.00	175.0
269.90	324.0
271.10	1408.0
272.00	1066.0
273.10	12911.0
274.10	32088.0
275.10	181248.0
276.10	24056.0
277.10	16266.0
278.00	2844.0
279.00	424.0
281.20	851.0
282.10	614.0
283.00	1781.0
284.00	1341.0
285.10	3149.0
286.10	726.0
288.10	246.0
289.00	437.0
290.00	577.0
291.00	649.0
292.00	602.0
293.10	4033.0
294.10	1277.0
295.00	998.0
296.10	54656.0
297.10	7002.0
298.00	546.0
301.00	756.0
302.00	991.0
303.10	6392.0
304.10	1309.0
305.10	330.0
308.10	944.0
309.10	483.0
310.00	672.0
310.90	159.0
312.00	278.0
313.00	628.0
314.10	2434.0
315.10	6173.0
316.10	3647.0
317.00	471.0
319.00	184.0
321.10	1687.0
322.00	859.0
323.10	17776.0
324.10	3124.0
325.10	313.0
325.90	364.0
327.10	3270.0
328.00	1577.0
329.00	269.0
332.00	1375.0
333.00	1551.0
334.10	11348.0

Instrument :
BNA_M
ClientSampleId :
DFTPP048

335.10	2765.0
336.20	249.0
339.10	360.0
340.00	318.0
341.10	2396.0
342.10	474.0
343.10	185.0
346.00	4708.0
347.00	745.0
351.10	305.0
352.10	5392.0
353.10	3123.0
354.10	5013.0
355.20	864.0
358.90	337.0
364.20	166.0
365.10	24272.0
366.00	3914.0
366.90	194.0
370.10	396.0
371.00	1658.0
372.10	9294.0
373.10	2176.0
374.00	209.0
377.00	325.0
383.10	2434.0
384.10	788.0
385.00	225.0
390.10	814.0
391.10	829.0
392.10	697.0
393.30	211.0
397.00	235.0
401.00	677.0
402.10	4084.0
403.10	4982.0
404.10	2019.0
405.10	318.0
410.20	229.0
415.00	268.0
421.10	4264.0
422.10	4788.0
423.10	29176.0
424.10	7389.0
425.10	1068.0
438.00	208.0
439.10	249.0
440.00	185.0
441.10	89720.0
442.10	540608.0
443.10	110472.0
444.10	9768.0
445.10	433.0

Instrument :
BNA_M
ClientSampleId :
DFTPP048

m/z	Abundance
36.00	399.0
37.00	1108.0
38.00	2542.0
39.10	18992.0
40.00	4649.0
41.10	4726.0
42.10	1124.0
43.10	9982.0
44.00	7791.0
45.00	737.0
47.00	238.0
48.10	261.0
49.00	1766.0
50.10	63400.0
51.10	261952.0
52.10	13022.0
53.10	883.0
54.00	377.0
55.00	5438.0
56.10	10007.0
57.10	31000.0
58.00	1380.0
59.00	288.0
60.10	598.0
61.00	3375.0
62.00	3549.0
63.10	9867.0
64.10	1177.0
65.00	5195.0
66.20	416.0
67.10	1013.0
68.10	4358.0
69.00	275008.0
70.10	3298.0
71.10	10060.0
72.10	764.0
73.10	2487.0
74.10	25688.0
75.10	42936.0
76.10	13763.0
77.10	309888.0
78.10	21456.0
79.10	18216.0
80.00	13870.0
81.10	19848.0
82.00	5635.0
83.10	7667.0
84.10	1425.0
85.10	10746.0
86.10	5765.0
87.10	3108.0
88.00	1144.0
89.10	619.0
90.00	241.0
91.10	4523.0
92.00	4887.0
93.00	31368.0
94.10	2434.0
95.00	1008.0
96.10	2013.0
97.10	2140.0
98.00	25832.0
99.10	23416.0
100.00	1925.0
101.00	12702.0
102.00	672.0
103.00	3836.0
104.00	6490.0
105.00	6787.0
106.00	2175.0
107.10	91104.0
108.00	14882.0
109.10	2189.0
110.00	174272.0
111.00	26512.0
112.10	3862.0
113.10	2008.0
114.00	447.0
115.00	697.0
116.00	4887.0
117.00	74840.0
118.00	5791.0
119.00	755.0
120.00	1326.0
121.00	683.0
122.00	5373.0
123.00	9012.0
124.00	4532.0
125.00	4170.0
126.10	1363.0
127.10	337728.0
128.00	26128.0

Instrument :
BNA_M
ClientSampleId :
DFTPP048

129.00	128152.0
130.10	11212.0
131.00	2080.0
132.10	1218.0
133.00	623.0
134.00	3268.0
135.00	10782.0
136.00	4232.0
137.00	6168.0
138.00	1270.0
139.00	672.0
140.00	1917.0
141.00	16360.0
142.10	5610.0
143.00	4108.0
144.00	1011.0
144.90	816.0
146.00	2698.0
147.00	9024.0
148.00	20264.0
149.00	4781.0
150.00	1416.0
151.10	2472.0
152.00	1272.0
153.00	4489.0
154.00	4311.0
155.00	9899.0
156.10	13170.0
157.10	2668.0
158.00	2870.0
159.00	2484.0
160.00	5077.0
161.10	7727.0
162.00	2399.0
163.00	930.0
164.00	1048.0
165.00	6142.0
166.00	4914.0
167.10	33920.0
168.00	21008.0
169.00	3396.0
170.10	982.0
171.00	1456.0
172.00	3072.0
173.00	4155.0
174.10	7598.0
175.10	14748.0
176.10	3775.0
177.00	5663.0
178.00	2196.0
179.00	25856.0
180.10	17752.0
181.00	9136.0
181.90	1726.0
183.00	876.0
184.00	1903.0
185.10	13081.0
186.10	95816.0
187.10	27192.0
188.10	3194.0
189.00	5653.0
190.00	848.0
191.00	3081.0
192.00	9279.0
193.10	9401.0
194.00	1986.0
195.00	1312.0
196.00	19256.0
196.90	3513.0
198.00	782976.0
199.00	50080.0
200.00	3810.0
201.50	3047.0
203.00	5138.0
204.10	26560.0
205.10	43776.0
206.10	195136.0
207.10	28184.0
208.10	7175.0
209.00	2199.0
210.30	2942.0
211.20	7952.0
212.00	1416.0
213.00	514.0
215.00	2327.0
216.00	4009.0
217.00	54240.0
218.00	6967.0
219.00	639.0
221.10	27840.0
221.90	6140.0
223.00	11210.0
224.10	109544.0
225.10	27416.0
226.10	3084.0
227.00	43824.0

Instrument :
BNA_M
ClientSampleId :
DFTPP048

228.00	6270.0
229.00	9365.0
230.00	1071.0
231.10	3973.0
232.00	755.0
233.00	971.0
234.10	3167.0
235.00	3180.0
236.00	2458.0
237.10	3966.0
238.00	607.0
239.00	1709.0
240.00	1643.0
241.10	2415.0
242.10	5407.0
243.10	5793.0
244.10	86488.0
245.10	12581.0
246.00	15817.0
247.00	3010.0
248.00	658.0
249.00	2637.0
250.00	662.0
251.00	779.0
252.00	855.0
253.10	1836.0
254.10	2441.0
255.10	436160.0
256.10	63328.0
257.10	4814.0
258.10	25288.0
259.00	3490.0
259.90	435.0
260.20	450.0
261.00	618.0
262.10	269.0
263.10	312.0
264.00	682.0
265.00	9805.0
266.00	1407.0
267.10	312.0
267.90	195.0
270.10	449.0
271.10	1570.0
272.10	1064.0
273.10	13303.0
274.10	34400.0
275.10	192000.0
276.10	25528.0
277.00	15894.0
278.10	2976.0
279.00	545.0
281.10	902.0
282.10	682.0
283.10	2458.0
284.10	1237.0
285.10	2735.0
286.10	700.0
288.10	202.0
289.10	490.0
289.90	547.0
291.10	432.0
292.10	986.0
293.00	3923.0
294.10	971.0
295.10	831.0
296.10	58384.0
297.10	7994.0
298.10	545.0
301.10	789.0
302.00	912.0
303.10	7131.0
304.00	1695.0
305.10	292.0
308.10	778.0
309.00	433.0
310.00	581.0
311.10	235.0
312.00	286.0
313.10	543.0
314.10	2598.0
315.00	6202.0
316.10	3760.0
317.00	509.0
319.10	167.0
320.00	202.0
321.10	1975.0
322.00	878.0
323.10	17992.0
324.10	3736.0
325.00	451.0
326.00	440.0
327.10	3254.0
328.10	1686.0
329.00	536.0
332.10	1499.0

Instrument :
BNA_M
ClientSampleId :
DFTPP048

333.10	1561.0
334.10	12159.0
335.10	2626.0
336.20	316.0
339.10	164.0
340.00	286.0
341.10	2480.0
342.10	707.0
346.10	4867.0
347.10	657.0
350.00	234.0
351.10	506.0
352.10	5675.0
353.10	3837.0
354.10	6153.0
355.10	1213.0
359.10	414.0
365.10	26704.0
366.10	4156.0
367.00	264.0
370.00	725.0
371.10	1424.0
372.10	10794.0
373.10	2594.0
374.20	168.0
377.00	257.0
383.00	2865.0
384.00	815.0
385.00	327.0
390.10	1412.0
391.00	937.0
392.10	808.0
393.20	161.0
397.10	201.0
401.10	802.0
402.10	4641.0
403.10	6987.0
404.00	2281.0
405.10	226.0
415.00	223.0
421.10	4839.0
422.10	5158.0
423.10	34912.0
424.10	8097.0
425.10	865.0
438.10	263.0
439.10	176.0
440.00	355.0
441.10	107440.0
442.10	667584.0
443.10	127208.0
444.10	11109.0
445.10	623.0

Instrument :
BNA_M
ClientSampleId :
DFTPP048

m/z	Abundance
36.00	267.0
37.00	554.0
38.00	1852.0
39.10	10848.0
40.00	4525.0
41.00	3845.0
42.10	1255.0
43.10	10498.0
44.00	9031.0
45.00	496.0
47.00	180.0
49.00	1391.0
50.10	35296.0
51.10	139392.0
52.10	6994.0
53.00	529.0
54.10	594.0
55.10	4830.0
56.00	6307.0
57.10	23912.0
58.10	1462.0
59.00	185.0
60.00	212.0
61.00	1924.0
62.10	2090.0
63.10	4982.0
64.00	985.0
65.10	3131.0
66.10	299.0
67.10	1179.0
68.00	2666.0
69.00	150528.0
70.10	2397.0
71.10	9548.0
72.00	732.0
73.10	1841.0
74.10	14275.0
75.00	22888.0
76.10	7608.0
77.10	173824.0
78.10	11833.0
79.10	8861.0
80.00	8319.0
81.00	11040.0
82.00	3498.0
83.10	5120.0
84.00	1639.0
85.10	9755.0
86.00	3942.0
87.10	1655.0
88.00	548.0
89.00	449.0
90.00	176.0
91.00	2894.0
92.00	2931.0
93.00	17504.0
94.00	1452.0
95.00	715.0
96.10	1079.0
97.10	2037.0
98.00	14403.0
99.10	13698.0
100.10	1292.0
101.00	5962.0
102.00	632.0
103.00	2258.0
104.00	4142.0
105.00	4008.0
106.00	1441.0
107.00	52048.0
108.10	8204.0
109.00	1362.0
110.00	98496.0
111.00	15541.0
112.00	1924.0
113.10	1856.0
114.20	459.0
115.00	363.0
116.00	2580.0
117.00	43648.0
118.00	3064.0
118.90	510.0
120.00	675.0
121.10	240.0
122.00	3419.0
123.00	5182.0
124.00	2411.0
125.10	2780.0
126.10	1022.0
127.00	194944.0
128.10	13351.0
129.00	72368.0

Instrument :
BNA_M
ClientSampleId :
DFTPP048

130.00	6121.0
131.00	1437.0
132.10	541.0
132.90	538.0
134.00	2184.0
135.10	6813.0
136.00	2576.0
137.00	3759.0
138.10	630.0
139.10	659.0
140.00	903.0
141.00	10113.0
142.00	2726.0
143.00	2161.0
144.00	612.0
145.00	588.0
145.90	1660.0
147.00	4490.0
148.00	11436.0
149.00	2636.0
150.00	911.0
151.10	1388.0
151.90	838.0
153.00	2965.0
154.00	2703.0
155.10	6157.0
156.10	8625.0
157.00	1404.0
158.10	1458.0
159.00	1662.0
160.00	3130.0
161.10	4873.0
162.00	1376.0
163.10	574.0
164.00	685.0
165.00	3765.0
166.10	3481.0
167.10	21680.0
168.00	13406.0
169.10	1802.0
170.00	709.0
171.00	898.0
172.00	1626.0
173.00	2170.0
174.00	4205.0
175.00	7870.0
176.10	2278.0
177.00	3535.0
178.00	1191.0
179.00	15792.0
180.10	11614.0
181.00	5150.0
182.10	955.0
183.10	678.0
184.00	1069.0
185.10	7397.0
186.10	59312.0
187.00	15536.0
188.00	1815.0
189.00	3338.0
189.90	630.0
191.00	2033.0
192.10	5313.0
193.00	6287.0
194.00	1405.0
195.10	683.0
196.10	11260.0
196.90	1899.0
198.00	475904.0
199.00	31552.0
200.00	2275.0
201.50	1882.0
203.00	3645.0
204.10	16984.0
205.10	27360.0
206.10	121624.0
207.10	17920.0
208.00	4626.0
209.00	1312.0
210.20	1779.0
211.10	4447.0
211.80	913.0
213.20	240.0
215.00	1426.0
216.10	2568.0
217.00	33264.0
218.10	4051.0
219.00	416.0
221.10	17400.0
222.00	4810.0
223.10	7198.0
224.10	69184.0
225.10	16968.0
226.00	1855.0
227.00	28216.0
228.00	3970.0

Instrument :
BNA_M
ClientSampleId :
DFTPP048

229.00	6198.0
230.00	767.0
231.00	2643.0
232.10	399.0
233.10	573.0
234.00	1900.0
235.00	1871.0
236.10	1474.0
237.00	2259.0
238.10	355.0
239.10	1151.0
240.00	867.0
241.00	1685.0
242.10	4143.0
243.10	4086.0
244.10	56744.0
245.10	7617.0
246.10	11158.0
247.00	2032.0
248.00	552.0
249.00	1791.0
250.10	444.0
251.00	482.0
252.00	680.0
253.00	1050.0
254.10	1650.0
255.10	281344.0
256.10	40200.0
257.10	2836.0
258.10	16078.0
259.00	2428.0
260.10	381.0
261.00	589.0
263.10	162.0
264.10	327.0
265.10	6950.0
265.90	871.0
267.20	172.0
268.10	157.0
270.20	301.0
271.00	1044.0
272.10	643.0
273.10	8424.0
274.00	22672.0
275.10	121600.0
276.10	15694.0
277.10	10008.0
278.00	1630.0
279.10	515.0
281.10	616.0
282.00	414.0
283.00	1411.0
284.10	791.0
285.10	2170.0
285.90	329.0
289.10	547.0
289.90	447.0
291.00	426.0
292.00	566.0
293.00	3058.0
294.00	707.0
295.10	856.0
296.10	39072.0
297.10	5246.0
297.90	316.0
301.00	434.0
302.00	704.0
303.10	4075.0
304.00	1254.0
305.00	182.0
308.10	564.0
309.00	456.0
310.10	582.0
310.90	195.0
312.10	153.0
313.10	195.0
314.10	1519.0
315.10	3910.0
316.10	2221.0
317.00	578.0
321.10	1341.0
322.10	748.0
323.10	13450.0
324.10	2318.0
325.10	200.0
326.10	204.0
327.00	2712.0
328.10	1543.0
329.10	333.0
331.90	957.0
333.00	987.0
334.10	8497.0
335.10	2010.0
336.20	257.0
340.00	240.0
341.10	1540.0

Instrument :
BNA_M
ClientSampleId :
DFTPP048

342.00	374.0
343.10	230.0
346.00	3222.0
347.00	498.0
350.00	259.0
351.20	330.0
352.10	3768.0
353.10	2662.0
354.10	4128.0
355.10	803.0
359.00	463.0
365.10	19192.0
366.10	3120.0
367.00	257.0
370.00	349.0
371.10	1140.0
372.10	7104.0
373.00	1664.0
374.00	199.0
383.10	2182.0
384.00	553.0
390.00	1063.0
391.10	647.0
392.20	416.0
401.10	363.0
402.10	3520.0
403.10	4454.0
404.10	1605.0
405.10	466.0
415.20	176.0
421.10	4215.0
422.10	3883.0
423.10	25648.0
424.10	5935.0
425.00	587.0
439.10	192.0
440.00	227.0
441.10	80816.0
442.10	497920.0
443.10	94032.0
444.10	8435.0
445.10	258.0

Instrument :
BNA_M
ClientSampleId :
DFTPP048