

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM101218\
 Data File : BM017132.D
 Acq On : 12 Oct 2018 16:19
 Operator : MJ/SJ
 Sample : SSTDC0020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02064

Quant Time: Oct 12 23:35:41 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM100418MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Oct 12 08:22:37 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	125	0.00
2	1,4-Dioxane	0.423	0.310	26.7#	94	0.00
3 S	1,4-Dioxane-d8	0.402	0.317	21.1	101	0.00
4	Benzaldehyde	1.038	1.025	1.3	111	0.00
5 S	Phenol-d5	1.672	1.560	6.7	118	0.00
6	Phenol	1.673	1.548	7.5	117	0.00
7 S	Bis-(2-Chloroethyl)ether-d8	1.014	0.925	8.8	113	0.00
8	Bis(2-Chloroethyl)ether	1.295	1.217	6.0	116	0.00
9 S	2-Chlorophenol-d4	1.390	1.354	2.6	122	0.00
10	2-Chlorophenol	1.384	1.335	3.5	120	0.00
11	2-Methylphenol	1.259	1.244	1.2	123	0.00
12	2,2'-oxybis(1-Chloropropane	1.863	1.749	6.1	115	0.00
13 S	4-Methylphenol-d8	1.310	1.308	0.2	126	0.00
14	Acetophenone	2.082	2.148	-3.2	128	0.00
15 P	N-Nitroso-di-n-propylamine	0.969	0.946	2.4	119	0.00
16	4-Methylphenol	1.349	1.362	-1.0	125	0.00
17	Hexachloroethane	0.565	0.569	-0.7	124	0.00
18 I	Naphthalene-d8	1.000	1.000	0.0	134	0.00
19 S	Nitrobenzene-d5	0.146	0.143	2.1	129	0.00
20	Nitrobenzene	0.346	0.329	4.9	127	0.00
21	Isophorone	0.627	0.609	2.9	128	0.00
22 S	2-Nitrophenol-d4	0.149	0.147	1.3	132	0.00
23 C	2-Nitrophenol	0.168	0.167	0.6	132	0.00
24	2,4-Dimethylphenol	0.341	0.336	1.5	132	0.00
25	Bis(2-Chloroethoxy)methane	0.402	0.387	3.7	129	0.00
26 S	2,4-Dichlorophenol-d3	0.306	0.316	-3.3	136	0.00
27 C	2,4-Dichlorophenol	0.297	0.305	-2.7	137	0.00
28	Naphthalene	1.028	1.020	0.8	133	0.00
29 S	4-Chloroaniline-d4	0.368	0.394	-7.1	130	0.00
30	4-Chloroaniline	0.373	0.399	-7.0	131	0.00
31 C	Hexachlorobutadiene	0.202	0.207	-2.5	139	0.00
32	Caprolactam	0.091	0.084	7.7	126	0.00
33 C	4-Chloro-3-methylphenol	0.305	0.313	-2.6	135	0.00
34	2-Methylnaphthalene	0.726	0.750	-3.3	138	0.00
35	1-Methylnaphthalene	0.707	0.734	-3.8	138	0.00
36 I	Acenaphthene-d10	1.000	1.000	0.0	142	0.00
37	1,2,4,5-Tetrachlorobenzene	0.687	0.687	0.0	141	0.00
38	Hexachlorocyclopentadiene	0.440	0.329	25.2#	111	0.00
39 C	2,4,6-Trichlorophenol	0.414	0.421	-1.7	143	0.00
40	2,4,5-Trichlorophenol	0.445	0.450	-1.1	143	0.00
41	1,1'-Biphenyl	1.649	1.624	1.5	140	0.00
42	2-Chloronaphthalene	1.299	1.302	-0.2	142	0.00
43	2-Nitroaniline	0.283	0.251	11.3	123	0.00
44 S	Dimethylphthalate-d6	1.587	1.585	0.1	140	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM101218\
 Data File : BM017132.D
 Acq On : 12 Oct 2018 16:19
 Operator : MJ/SJ
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTD02064

Quant Time: Oct 12 23:35:41 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM100418MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Oct 12 08:22:37 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
45	Dimethylphthalate	1.571	1.573	-0.1	141	0.00
46	2,6-Dinitrotoluene	0.265	0.270	-1.9	139	0.00
47 S	Acenaphthylene-d8	2.028	2.061	-1.6	141	0.00
48	Acenaphthylene	1.943	1.962	-1.0	141	0.00
49	3-Nitroaniline	0.294	0.279	5.1	127	0.00
50 C	Acenaphthene	1.380	1.381	-0.1	141	0.00
51	2,4-Dinitrophenol	0.171	0.131	23.4	127	0.00
52 S	4-Nitrophenol-d4	0.300	0.256	14.7	124	0.00
53	4-Nitrophenol	0.218	0.195	10.6	124	0.00
54	Dibenzofuran	1.949	1.974	-1.3	143	0.00
55	2,4-Dinitrotoluene	0.428	0.450	-5.1	142	0.00
56	2,3,4,6-Tetrachlorophenol	0.399	0.411	-3.0	144	0.00
57	Diethylphthalate	1.537	1.539	-0.1	139	0.00
58 S	Fluorene-d10	1.413	1.433	-1.4	141	0.00
59	Fluorene	1.589	1.616	-1.7	142	0.00
60	4-Chlorophenyl-phenylether	0.828	0.848	-2.4	143	0.00
61	4-Nitroaniline	0.347	0.301	13.3	122	0.00
62 I	Phenanthrene-d10	1.000	1.000	0.0	139	0.00
63 S	4,6-Dinitro-2-methylphenol-	0.118	0.103	12.7	135	0.00
64	4,6-Dinitro-2-methylphenol	0.125	0.110	12.0	133	0.00
65	N-Nitrosodiphenylamine	0.593	0.611	-3.0	142	0.00
66	4-Bromophenyl-phenylether	0.225	0.235	-4.4	145	0.00
67	Hexachlorobenzene	0.254	0.263	-3.5	145	0.00
68	Atrazine	0.206	0.194	5.8	133	0.00
69 C	Pentachlorophenol	0.153	0.142	7.2	135	0.00
70	Phenanthrene	1.121	1.142	-1.9	141	0.00
71 S	Anthracene-d10	0.955	0.982	-2.8	141	0.00
72	Anthracene	1.139	1.156	-1.5	139	0.00
73	1,2,3,4-Tetrachlorobenzene	0.319	0.324	-1.6	143	0.00
74	Pentachlorobenzene	0.300	0.309	-3.0	144	0.00
75	Carbazole	0.986	0.948	3.9	132	0.00
76	Di-n-butylphthalate	1.103	1.098	0.5	136	0.00
77 C	Fluoranthene	1.299	1.201	7.5	127	0.00
78 I	Chrysene-d12	1.000	1.000	0.0	96	0.00
79 S	Pyrene-d10	0.902	1.163	-28.9#	123	0.00
80	Pyrene	1.159	1.493	-28.8#	123	0.00
81	Butylbenzylphthalate	0.400	0.420	-5.0	100	0.00
82	3,3'-Dichlorobenzidine	0.355	0.285	19.7	75	0.00
83	Benzo(a)anthracene	1.200	1.226	-2.2	98	0.00
84	Bis(2-ethylhexyl)phthalate	0.620	0.659	-6.3	100	0.00
85	Chrysene	1.160	1.166	-0.5	96	0.00
86 I	Perylene-d12	1.000	1.000	0.0	84	0.00
87	Di-n-octyl phthalate	1.069	1.049	1.9	85	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM101218\
Data File : BM017132.D
Acq On : 12 Oct 2018 16:19
Operator : MJ/SJ
Sample : SSTDCCC020
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTD02064

Quant Time: Oct 12 23:35:41 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM100418MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Oct 12 08:22:37 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
88	Benzo(b)fluoranthene	1.191	1.240	-4.1	89	0.00
89	Benzo(k)fluoranthene	1.144	1.171	-2.4	84	0.00
90 S	Benzo(a)pyrene-d12	1.024	1.048	-2.3	85	0.00
91 C	Benzo(a)pyrene	1.148	1.160	-1.0	84	0.00
92	Indeno(1,2,3-cd)pyrene	1.355	1.383	-2.1	86	0.00
93	Dibenzo(a,h)anthracene	1.128	1.146	-1.6	85	0.00
94	Benzo(g,h,i)perylene	1.128	1.159	-2.7	87	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0