


Data Path : Z:\SVOASRV\HPCHEM1\BNA_M\DATA\BM101218\
Data File : BM017131.D
Acq On : 12 Oct 2018 15:43
Operator : MJ/SJ
Sample : DFTPP17
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP17

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SOM-EPA-BM100418MA.M
Title : SVOA CALIBRATION
Last Update : Mon Oct 15 10:51:41 2018



AutoFind: Scans 2471, 2472, 2473; Background Corrected with Scan 2466

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	33.3	123773	PASS
68	69	0.00	2	1.7	2237	PASS
69	198	0.00	100	35.4	131522	PASS
70	69	0.00	2	0.5	678	PASS
127	198	10	80	45.6	169514	PASS
197	198	0.00	2	0.5	1978	PASS
198	198	100	100	100.0	371840	PASS
199	198	5	9	6.9	25656	PASS
275	198	10	60	29.5	109848	PASS
365	198	1	100	3.7	13701	PASS
441	198	0.01	100	12.4	45944	PASS
442	198	50	100	83.3	309674	PASS
443	442	15	24	19.5	60485	PASS

m/z	Abundance
37.10	459.0
38.00	1355.0
39.10	8669.0
40.00	1172.0
44.00	8716.0
44.90	327.0
49.00	1001.0
50.00	26200.0
51.00	105624.0
52.10	5434.0
54.90	659.0
56.00	3347.0
57.00	7230.0
58.00	389.0
61.00	1276.0
62.00	1718.0
63.00	4708.0
64.00	804.0
65.10	1997.0
68.10	1700.0
69.00	109720.0
70.00	663.0
73.00	1110.0
74.00	10958.0
75.00	17616.0
76.10	5356.0
77.10	116720.0
78.00	8006.0
79.00	9136.0
80.00	6692.0
81.00	8875.0
82.00	2037.0
83.00	2084.0
83.90	328.0
85.00	1358.0
86.00	2936.0
87.00	1200.0
87.90	641.0
90.90	1887.0
92.00	2342.0
93.00	13648.0
93.90	881.0
96.00	843.0
97.20	649.0
98.00	10447.0
99.00	9338.0
99.90	744.0
101.00	4466.0
103.00	1895.0
104.00	2976.0
105.00	3023.0
106.10	953.0
107.00	37176.0
108.00	5710.0
109.10	917.0
110.00	67184.0
111.00	10652.0
112.00	1093.0
113.00	397.0
116.00	2087.0
117.00	29912.0
118.00	2265.0
120.00	383.0
122.00	2250.0
123.00	4124.0
124.00	1605.0
125.00	1568.0
126.10	847.0
127.00	132544.0
128.00	10536.0
129.00	51856.0
130.00	4679.0
131.10	488.0
132.00	585.0
133.00	413.0
134.00	1605.0
135.00	4003.0
136.00	1650.0
137.00	2397.0
137.70	443.0
138.10	347.0
140.10	586.0
141.00	6319.0
142.00	1994.0
142.90	1636.0
144.00	344.0
144.90	429.0
146.00	1210.0
146.90	2913.0
148.00	6974.0
149.00	1915.0

Instrument :
BNA_M
ClientSampleId :
DFTPP17

149.90	318.0
150.90	788.0
151.20	774.0
151.70	468.0
153.00	1682.0
154.00	1741.0
155.00	3343.0
156.00	5212.0
157.00	818.0
158.00	1355.0
159.00	967.0
159.90	2252.0
161.00	3065.0
162.00	787.0
163.90	464.0
165.00	2130.0
166.00	2212.0
167.00	11541.0
168.00	9162.0
169.00	1305.0
169.90	445.0
170.90	631.0
171.90	1417.0
173.00	1577.0
174.00	2873.0
175.00	5307.0
176.00	2129.0
176.90	2239.0
178.00	705.0
179.00	10038.0
180.00	7101.0
181.00	3244.0
182.00	458.0
184.00	708.0
185.00	5343.0
186.00	37800.0
187.00	11056.0
188.00	1373.0
189.00	2273.0
189.90	596.0
191.00	1041.0
192.00	3610.0
193.00	3336.0
194.00	815.0
195.00	589.0
196.00	8410.0
197.10	1430.0
198.00	265280.0
199.00	18600.0
200.00	1499.0
201.60	904.0
202.90	1530.0
204.00	9576.0
205.00	15916.0
206.00	71248.0
207.00	11649.0
208.00	2806.0
209.10	956.0
210.00	1111.0
211.00	2836.0
211.80	678.0
214.90	890.0
216.00	1401.0
217.00	19760.0
218.00	2373.0
219.10	319.0
219.90	324.0
221.00	11099.0
221.80	3172.0
223.00	4167.0
224.00	39104.0
225.00	9898.0
226.00	1241.0
227.00	16392.0
228.00	2411.0
229.00	3207.0
229.90	521.0
231.00	1331.0
234.10	1068.0
235.00	1135.0
235.90	648.0
237.00	1617.0
238.90	650.0
239.90	554.0
241.00	982.0
242.00	2115.0
243.10	1644.0
244.00	29208.0
245.00	4239.0
246.00	5579.0
247.00	1258.0
247.90	379.0
249.00	1115.0
249.90	432.0
252.90	664.0

Instrument :
BNA_M
ClientSampleId :
DFTPP17

254.10	1074.0
255.00	156224.0
256.00	23680.0
257.10	1755.0
258.00	8856.0
259.00	1563.0
265.00	3554.0
265.90	462.0
271.00	980.0
272.00	435.0
273.00	4741.0
274.00	11076.0
275.00	70696.0
276.00	8664.0
277.00	6013.0
278.00	727.0
281.10	642.0
282.90	818.0
284.00	541.0
285.00	1112.0
288.90	305.0
293.00	1324.0
293.90	422.0
295.10	422.0
296.00	17384.0
297.00	2815.0
302.10	468.0
303.00	1931.0
304.00	501.0
313.90	766.0
315.00	1846.0
316.00	773.0
316.90	336.0
321.00	598.0
323.00	5876.0
324.10	1230.0
327.00	928.0
327.90	471.0
331.90	336.0
333.10	428.0
334.00	4026.0
335.00	949.0
341.00	482.0
345.90	1136.0
347.00	505.0
352.00	1694.0
353.00	1361.0
354.00	1653.0
355.10	360.0
365.00	7664.0
366.00	1073.0
371.00	438.0
372.10	3082.0
373.00	619.0
383.10	815.0
390.00	320.0
391.10	311.0
402.00	1116.0
403.00	1351.0
404.10	634.0
421.10	1264.0
422.10	1419.0
423.00	8280.0
424.10	2520.0
425.00	314.0
440.20	309.0
441.10	23576.0
442.10	152448.0
443.10	30528.0
444.00	3096.0

Instrument :
BNA_M
ClientSampleId :
DFTPP17

m/z	Abundance
37.00	511.0
38.10	2089.0
39.00	10961.0
40.00	1337.0
41.00	354.0
44.00	9412.0
45.00	444.0
49.10	1021.0
50.00	37992.0
51.00	151040.0
52.00	8176.0
54.90	655.0
56.00	4631.0
57.00	10054.0
58.00	371.0
61.00	2131.0
62.00	2434.0
63.00	6685.0
64.00	821.0
65.00	3166.0
67.00	316.0
68.00	2752.0
69.00	162944.0
70.00	657.0
73.00	1587.0
74.00	16323.0
75.00	25616.0
76.00	8842.0
77.10	173248.0
78.10	11479.0
79.00	11949.0
80.00	9367.0
81.00	13614.0
82.00	2976.0
83.00	2904.0
84.00	396.0
85.00	2216.0
86.00	3971.0
87.00	1565.0
88.00	652.0
88.90	375.0
91.00	2766.0
92.00	3308.0
93.00	21096.0
93.90	1497.0
95.00	352.0
96.00	1084.0
97.00	334.0
98.00	16944.0
99.00	12673.0
100.00	948.0
101.00	7663.0
103.00	2577.0
104.00	5031.0
105.00	4141.0
106.10	1616.0
107.00	55416.0
108.00	8286.0
109.10	1468.0
110.00	104560.0
111.00	14926.0
112.00	1772.0
113.00	518.0
115.00	358.0
116.00	3413.0
117.00	44504.0
118.00	3259.0
119.00	444.0
120.00	766.0
121.00	331.0
122.00	3128.0
123.00	6877.0
124.10	2631.0
125.00	2470.0
126.20	1029.0
127.00	202496.0
128.00	15427.0
129.00	79344.0
130.00	6879.0
131.00	1149.0
132.00	759.0
133.00	375.0
134.00	2465.0
135.00	6205.0
136.00	2409.0
137.00	3798.0
137.90	573.0
138.90	506.0
139.90	866.0
141.00	10145.0
142.00	3532.0

Instrument :
BNA_M
ClientSampleId :
DFTPP17

143.00	2663.0
144.10	589.0
146.00	2087.0
147.00	5323.0
148.00	12674.0
149.00	2765.0
150.00	568.0
151.00	1325.0
152.00	814.0
153.00	3219.0
153.90	2642.0
155.00	5668.0
156.10	8365.0
156.90	2171.0
158.10	1741.0
159.00	1699.0
160.00	2953.0
161.00	5491.0
161.90	1359.0
163.10	431.0
163.90	484.0
165.00	3741.0
166.00	2954.0
167.00	20384.0
168.00	11532.0
169.00	1763.0
169.90	788.0
171.00	825.0
172.00	1857.0
173.00	2519.0
174.00	4041.0
175.00	7905.0
175.90	2483.0
176.90	4454.0
177.90	1429.0
179.00	15783.0
180.00	11550.0
181.00	5255.0
182.10	838.0
182.90	565.0
184.00	1359.0
185.10	7404.0
186.00	61752.0
187.00	18800.0
188.10	1839.0
189.00	3680.0
190.00	670.0
191.00	2227.0
192.00	5567.0
193.00	6156.0
194.00	1118.0
195.00	532.0
196.00	13166.0
197.10	2497.0
198.00	444928.0
199.00	31584.0
200.00	2359.0
201.50	2249.0
203.00	2975.0
204.00	16350.0
205.00	27536.0
206.00	119528.0
207.00	16528.0
208.00	4510.0
209.00	1443.0
211.00	4945.0
213.10	407.0
214.90	1535.0
216.00	2270.0
217.00	30856.0
218.00	4380.0
221.00	19032.0
221.80	4327.0
223.00	7301.0
224.00	65608.0
225.00	17680.0
226.00	2040.0
227.00	27616.0
228.00	3662.0
229.00	5815.0
230.00	772.0
231.00	2272.0
232.00	398.0
233.00	451.0
234.00	2137.0
235.00	2086.0
236.00	1387.0
237.00	2223.0
239.00	1010.0
240.00	851.0
241.00	1759.0
242.00	3438.0
243.00	3131.0
244.00	51040.0
245.00	6951.0

Instrument :
BNA_M
ClientSampleId :
DFTPP17

246.00	10180.0
247.00	2099.0
248.00	514.0
249.00	1924.0
250.00	385.0
250.90	386.0
252.10	332.0
253.00	1375.0
254.20	2143.0
255.00	279104.0
256.00	39664.0
257.10	3035.0
258.00	15603.0
259.00	2345.0
259.90	459.0
260.90	445.0
264.00	414.0
265.00	6496.0
266.00	938.0
270.00	344.0
271.00	1460.0
272.00	762.0
273.00	8089.0
274.00	22120.0
275.00	127712.0
276.00	16744.0
277.00	10199.0
278.00	1456.0
278.90	443.0
281.00	535.0
281.90	433.0
282.90	1455.0
284.10	690.0
284.90	1758.0
285.90	343.0
289.00	671.0
290.00	454.0
291.00	334.0
291.90	499.0
293.00	2234.0
293.90	795.0
295.10	790.0
296.00	32320.0
297.00	4749.0
298.10	421.0
301.00	428.0
302.00	643.0
303.10	3866.0
304.00	1093.0
305.00	383.0
308.00	553.0
310.10	590.0
314.00	1689.0
315.00	3962.0
316.00	2036.0
321.00	1064.0
322.00	624.0
323.00	10821.0
324.00	2221.0
327.00	2225.0
327.90	1030.0
332.00	926.0
333.00	995.0
334.00	7482.0
335.00	1570.0
341.00	1206.0
341.90	555.0
346.00	2281.0
346.90	456.0
352.10	3285.0
353.10	2396.0
354.10	3716.0
355.10	768.0
365.00	16640.0
366.00	2074.0
370.00	478.0
371.00	870.0
372.00	5776.0
373.00	1250.0
383.00	1366.0
384.00	559.0
390.00	551.0
390.90	548.0
400.90	341.0
402.00	2328.0
403.00	3231.0
403.90	1235.0
421.00	2406.0
422.10	2811.0
423.10	18672.0
424.10	4458.0
425.00	596.0
440.10	379.0
441.10	53336.0
442.10	351872.0

Instrument :
BNA_M
ClientSampleId :
DFTPP17

443.10	70944.0
444.10	6333.0
444.90	398.0

Instrument :
BNA_M
ClientSampleId :
DFTPP17

m/z	Abundance
37.00	433.0
38.00	1509.0
39.00	8602.0
40.00	1368.0
41.00	349.0
44.00	9346.0
45.10	411.0
49.00	853.0
50.00	29888.0
51.00	114656.0
52.00	5787.0
55.00	683.0
56.00	3667.0
57.00	7981.0
61.00	2164.0
62.00	1848.0
63.00	4643.0
64.00	649.0
65.00	2399.0
68.00	2260.0
69.00	121904.0
70.00	714.0
73.00	1427.0
74.00	12319.0
75.00	18512.0
76.10	6183.0
77.00	138432.0
78.00	8847.0
79.00	9955.0
80.00	7044.0
81.00	10153.0
82.00	2591.0
83.00	2328.0
85.00	1642.0
86.00	3053.0
87.00	1207.0
88.00	581.0
89.10	319.0
91.00	2961.0
92.00	2725.0
93.00	16512.0
94.00	1283.0
96.00	907.0
98.00	12466.0
99.00	10726.0
100.00	942.0
101.00	6281.0
101.90	404.0
102.90	2027.0
104.00	3918.0
105.00	4000.0
106.00	1182.0
107.00	45616.0
108.00	6755.0
109.10	1333.0
110.00	83544.0
111.00	12438.0
112.00	1684.0
112.90	368.0
116.00	2425.0
117.00	36984.0
118.00	2576.0
118.90	338.0
120.00	414.0
122.00	2949.0
123.00	5283.0
124.00	2124.0
125.00	1999.0
126.20	1073.0
127.00	173504.0
128.00	13407.0
129.00	66600.0
130.00	5134.0
130.90	1256.0
132.00	621.0
133.00	610.0
134.00	1744.0
135.00	5336.0
136.00	2477.0
137.00	2696.0
138.00	759.0
139.00	399.0
139.90	1003.0
141.00	9043.0
142.00	3121.0
142.90	1835.0
143.90	518.0
145.00	344.0
146.00	1546.0
147.00	4665.0
148.00	10031.0

Instrument :
BNA_M
ClientSampleId :
DFTPP17

149.00	2380.0
150.00	425.0
150.90	1200.0
151.90	630.0
152.90	2836.0
154.00	2226.0
155.00	5250.0
156.00	6993.0
157.00	1265.0
157.90	1766.0
159.00	1657.0
160.00	2595.0
161.00	4155.0
161.90	1334.0
163.10	504.0
163.90	577.0
165.00	3440.0
166.00	2802.0
167.00	16616.0
168.00	10762.0
169.00	1442.0
170.10	489.0
170.90	939.0
172.00	1624.0
173.00	2319.0
174.00	4249.0
175.00	7706.0
176.00	2104.0
176.90	3312.0
178.00	1393.0
179.00	14222.0
180.00	10544.0
181.00	4582.0
181.90	985.0
182.80	333.0
184.00	1097.0
185.00	6885.0
186.00	55920.0
187.00	16368.0
188.00	1351.0
189.00	3741.0
190.00	577.0
191.00	1702.0
192.00	4598.0
193.00	5133.0
193.90	1036.0
195.00	464.0
196.00	11905.0
197.00	2008.0
198.00	405312.0
198.90	26784.0
200.00	2174.0
201.50	2013.0
202.90	2882.0
204.00	14760.0
205.00	24752.0
206.00	109176.0
207.00	15394.0
208.00	3580.0
209.00	1533.0
210.10	1777.0
211.00	4591.0
213.00	313.0
215.00	1298.0
216.00	2752.0
217.00	28320.0
218.00	4104.0
221.00	18800.0
221.80	3981.0
223.00	6154.0
224.00	63176.0
225.00	16880.0
226.00	1443.0
227.00	26584.0
228.00	3778.0
229.00	5762.0
230.00	671.0
231.00	2548.0
233.10	432.0
233.90	1581.0
235.00	1753.0
236.00	994.0
237.00	2395.0
237.90	356.0
239.00	985.0
240.00	670.0
240.90	1460.0
242.00	3392.0
243.10	2920.0
244.00	49720.0
245.10	6128.0
246.00	10172.0
247.00	2119.0
248.00	497.0
249.00	2189.0

Instrument :
BNA_M
ClientSampleId :
DFTPP17

249.90	310.0
250.90	362.0
252.10	467.0
253.10	975.0
254.10	1874.0
255.00	270336.0
256.00	37552.0
257.00	2893.0
258.00	14801.0
259.00	2349.0
260.10	505.0
260.90	489.0
264.00	359.0
265.00	6029.0
265.90	891.0
271.00	1471.0
272.00	915.0
273.00	7355.0
274.00	21784.0
275.00	131136.0
276.00	16488.0
277.00	11036.0
278.00	1852.0
278.90	376.0
281.00	675.0
282.10	375.0
283.00	974.0
284.00	980.0
285.00	1892.0
286.00	443.0
289.00	628.0
290.00	452.0
291.00	592.0
292.00	498.0
293.00	2275.0
294.10	770.0
295.10	1027.0
296.00	35440.0
297.00	5025.0
301.10	462.0
302.00	554.0
303.10	3791.0
304.00	1046.0
307.90	518.0
309.90	648.0
313.00	398.0
314.00	1779.0
315.00	3763.0
316.00	2247.0
317.10	300.0
320.90	1114.0
322.00	593.0
323.00	11718.0
324.00	2452.0
327.00	1934.0
328.00	1265.0
332.00	731.0
333.00	1289.0
334.00	7923.0
335.10	1750.0
336.10	306.0
341.00	1702.0
342.00	401.0
346.00	2661.0
347.00	418.0
352.00	3460.0
353.00	2731.0
354.00	3643.0
355.00	717.0
359.00	375.0
365.00	16800.0
366.00	2437.0
371.10	816.0
372.10	5634.0
373.00	1755.0
383.00	1785.0
383.90	578.0
390.10	764.0
391.00	762.0
392.00	488.0
400.80	471.0
402.00	2178.0
403.10	3447.0
404.00	1445.0
421.00	2827.0
422.00	2887.0
423.10	21504.0
424.00	5644.0
425.00	757.0
440.10	314.0
441.10	60920.0
442.10	424704.0
443.10	79984.0
444.10	7154.0
445.00	560.0

Instrument :
BNA_M
ClientSampleId :
DFTPP17

Scan 2466 (17.492 min): BM017131.D
DFTPP17

m/z	Abundance
39.90	964.0
44.00	7838.0
72.90	353.0
207.00	1804.0
208.00	518.0
280.90	544.0

Instrument :
BNA_M
ClientSampleId :
DFTPP17