

Data Path : Z:\HPCHEM1\BNA M\DATA\BM101315\
 Data File : BM002897.D
 Acq On : 13 Oct 2015 21:35
 Operator : UM/IZ
 Sample : SSTDICC005
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTDICC005

Quant Time: Oct 14 18:07:40 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\SIM-ACID-BM0101315.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 14 17:20:35 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.65	152	88582	5.00	ng	-0.01
8) Naphthalene-d8	11.50	136	380584	5.00	ng	0.00
14) Acenaphthene-d10	15.25	164	200693	5.00	ng	0.00
21) Phenanthrene-d10	17.95	188	442781	5.00	ng	0.00
24) Chrysene-d12	22.04	240	423849	5.00	ng	0.00
26) Perylene-d12	24.61	264	390241	5.00	ng	0.00
System Monitoring Compounds						
2) 2-Fluorophenol	6.12	112	86587	5.44	ng	-0.02
3) Phenol-d6	7.77	99	116408	4.78	ng	-0.02
9) Nitrobenzene-d5	9.85	82	102048	4.29	ng	-0.01
15) 2,4,6-Tribromophenol	16.71	330	41585	5.67	ng	0.00
18) 2-Fluorobiphenyl	13.87	172	309319	4.71	ng	0.00
25) Terphenyl-d14	20.48	244	306638	4.76	ng	0.00
Target Compounds						
4) 2-Chlorophenol	8.20	128	112297	4.98	ng	# 72
5) Phenol	7.80	94	122972	4.76	ng	90
6) 2-Methylphenol	9.09	107	86638	4.73	ng	# 79
7) 3+4-Methylphenols	9.43	107	119374	5.27	ng	# 34
10) 2-Nitrophenol	10.62	139	51257	5.35	ng	# 77
11) 2,4-Dimethylphenol	10.64	122	109197	5.20	ng	97
12) 2,4-Dichlorophenol	11.15	162	93390	5.71	ng	# 77
13) 4-Chloro-3-methylphenol	12.74	107	109859	5.31	ng	91
16) 2,4,6-Trichlorophenol	13.70	196	72170	5.52	ng	# 81
17) 2,4,5-Trichlorophenol	13.80	196	79123	5.67	ng	# 90
19) 2,4-Dinitrophenol	15.40	184	8664	8.41	ng	# 90
20) 4-Nitrophenol	15.47	139	37035	1.09	ng	# 58
23) Pentachlorophenol	17.62	266	21399	6.37	ng	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA M\DATA\BM0101315\
Data File : BM002897.D
Acq On : 13 Oct 2015 21:35
Operator : UM/IZ
Sample : SSTDICC005
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTDICC005

Quant Time: Oct 14 18:07:40 2015
Quant Method : Z:\HPCHEM1\BNA M\METHODS\SIM-ACID-BM0101315.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 14 17:20:35 2015
Response via : Initial Calibration

