

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM101321\
 Data File : BM032496.D
 Acq On : 13 Oct 2021 15:14
 Operator : CG/JU
 Sample : M4156-02
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 20-20211008-WC

Quant Time: Oct 13 16:42:07 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM100221.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Oct 13 12:02:28 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.601	152	247680	20.00	ng	0.00
21) Naphthalene-d8	10.389	136	936655	20.00	ng	# 0.00
39) Acenaphthene-d10	14.248	164	463902	20.00	ng	0.00
64) Phenanthrene-d10	16.977	188	943374	20.00	ng	# 0.00
76) Chrysene-d12	21.142	240	771378	20.00	ng	0.00
86) Perylene-d12	23.283	264	110623	20.00	ng	# 0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.178	112	1711963	116.46	ng	0.00
7) Phenol-d6	6.795	99	1817214	90.40	ng	0.00
23) Nitrobenzene-d5	8.754	82	1484008	88.12	ng	0.00
42) 2,4,6-Tribromophenol	15.736	330	686910	132.27	ng	0.00
45) 2-Fluorobiphenyl	12.866	172	3014247	96.61	ng	0.00
79) Terphenyl-d14	19.595	244	2847559	77.74	ng	0.00
Target Compounds						
2) 1,4-Dioxane	2.996	88	140227	22.42	ng	# 90
50) Dimethylphthalate	13.701	163	91016	2.78	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM101321\
Data File : BM032496.D
Acq On : 13 Oct 2021 15:14
Operator : CG/JU
Sample : M4156-02
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
20-20211008-WC

Quant Time: Oct 13 16:42:07 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\8270-BM100221.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Oct 13 12:02:28 2021
Response via : Initial Calibration

