

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM101322\
 Data File : BM037022.D
 Acq On : 13 Oct 2022 18:09
 Operator : CG/JU
 Sample : N5026-07
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 C0BX5

Quant Time: Oct 13 23:36:00 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM100622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Oct 11 09:03:52 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.943	152	6800	0.400	ng/ul	0.00
4) Naphthalene-d8	10.749	136	25112	0.400	ng/ul #	0.00
9) Acenaphthene-d10	14.571	164	13794	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.309	188	27115	0.400	ng/ul	0.00
17) Chrysene-d12	21.479	240	16745	0.400	ng/ul	0.00
23) Perylene-d12	23.873	264	13855	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.344	96	33802	3.525	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.332	152	13496	0.356	ng/ul	0.00
18) Fluoranthene-d10	19.331	212	25924	0.518	ng/ul	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM101322\
Data File : BM037022.D
Acq On : 13 Oct 2022 18:09
Operator : CG/JU
Sample : N5026-07
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
C0BX5

Quant Time: Oct 13 23:36:00 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM100622.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Oct 11 09:03:52 2022
Response via : Initial Calibration

