

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM101322\
 Data File : BM037047.D
 Acq On : 14 Oct 2022 10:08
 Operator : CG/JU
 Sample : N4984-11
 Misc :
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 BGNM3

Quant Time: Oct 14 23:01:15 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM100622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 14 07:11:07 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.934	152	6212	0.400	ng/ul	0.00
4) Naphthalene-d8	10.737	136	22432	0.400	ng/ul #	0.00
9) Acenaphthene-d10	14.562	164	12457	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.301	188	23194	0.400	ng/ul	0.00
17) Chrysene-d12	21.475	240	13127	0.400	ng/ul	0.00
23) Perylene-d12	23.864	264	11383	0.400	ng/ul #	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.339	96	34234	3.908	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.321	152	8733	0.258	ng/ul	0.00
18) Fluoranthene-d10	19.322	212	15290	0.389	ng/ul	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM101322\
Data File : BM037047.D
Acq On : 14 Oct 2022 10:08
Operator : CG/JU
Sample : N4984-11
Misc :
ALS Vial : 42 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
BGNM3

Quant Time: Oct 14 23:01:15 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM100622.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 14 07:11:07 2022
Response via : Initial Calibration

