

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM101322\
 Data File : BM037054.D
 Acq On : 14 Oct 2022 14:19
 Operator : CG/JU
 Sample : N4980-18
 Misc :
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 C1BC8

Quant Time: Oct 14 23:02:53 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM100622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 14 07:11:07 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.934	152	6365	0.400	ng/ul	0.00
4) Naphthalene-d8	10.738	136	22939	0.400	ng/ul #	0.00
9) Acenaphthene-d10	14.561	164	12557	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.301	188	24407	0.400	ng/ul	0.00
17) Chrysene-d12	21.474	240	13649	0.400	ng/ul	0.00
23) Perylene-d12	23.859	264	12329	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.340	96	39768	4.430	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.321	152	12241	0.353	ng/ul	0.00
18) Fluoranthene-d10	19.322	212	24646	0.604	ng/ul	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.373	88	11419	1.301	ng/ul#	80
5) Naphthalene	10.787	128	1329	0.020	ng/ul#	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

BNA M

ClientSampleId :

C1BC8

Abundance

TIC: BM037054.D\data.ms

Time-->

1,4-Dichlorobenzene-d4, I

1,4-Dichlorobenzene-d4, II

Naphthalene-d8, I

2-Methylnaphthalene-d10, SURR

Acenaphthene-d10, I

Phenanthrene-d10, I

Fluoranthene-d10, SURR

Chrysene-d12

Perylene-d12, I