

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM101322\
 Data File : BM037031.D
 Acq On : 14 Oct 2022 00:07
 Operator : CG/JU
 Sample : N5049-21
 Misc :
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 C0BE1

Quant Time: Oct 14 00:59:31 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM100622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Oct 11 09:03:52 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.938	152	6380	0.400	ng/ul	-0.01
4) Naphthalene-d8	10.743	136	23122	0.400	ng/ul	#-0.01
9) Acenaphthene-d10	14.566	164	12558	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.305	188	24531	0.400	ng/ul	0.00
17) Chrysene-d12	21.476	240	14606	0.400	ng/ul	0.00
23) Perylene-d12	23.867	264	12379	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.344	96	35599	3.957	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.327	152	11428	0.327	ng/ul	-0.01
18) Fluoranthene-d10	19.326	212	21360	0.489	ng/ul	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM101322\
Data File : BM037031.D
Acq On : 14 Oct 2022 00:07
Operator : CG/JU
Sample : N5049-21
Misc :
ALS Vial : 25 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
C0BE1

Quant Time: Oct 14 00:59:31 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM100622.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Oct 11 09:03:52 2022
Response via : Initial Calibration

