

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM101322\
 Data File : BM037090.D
 Acq On : 15 Oct 2022 13:24
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 87 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 SSTD0.4150

Quant Time: Oct 17 00:55:46 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM100622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Oct 15 04:13:53 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.930	152	6348	0.400	ng/ul	0.00
4) Naphthalene-d8	10.732	136	23029	0.400	ng/ul #	0.00
9) Acenaphthene-d10	14.557	164	12358	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.297	188	23412	0.400	ng/ul	0.00
17) Chrysene-d12	21.467	240	14902	0.400	ng/ul	0.00
23) Perylene-d12	23.855	264	12396	0.400	ng/ul #	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.340	96	3309	0.370	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.316	152	14152	0.407	ng/ul	0.00
18) Fluoranthene-d10	19.318	212	22619	0.507	ng/ul	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.373	88	3265	0.373	ng/ul	90
5) Naphthalene	10.781	128	26898	0.409	ng/ul	96
7) 2-Methylnaphthalene	12.387	142	16653	0.414	ng/ul	92
8) 1-Methylnaphthalene	12.607	142	17014	0.416	ng/ul	100
10) Acenaphthylene	14.280	152	23354	0.455	ng/ul	99
11) Acenaphthene	14.617	153	18168	0.408	ng/ul	100
12) Fluorene	15.603	166	19686	0.390	ng/ul	99
14) Pentachlorophenol	16.951	266	2765	0.488	ng/ul	98
15) Phenanthrene	17.339	178	30209	0.404	ng/ul	99
16) Anthracene	17.428	178	26505	0.424	ng/ul	99
19) Fluoranthene	19.350	202	30161	0.509	ng/ul	98
20) Pyrene	19.708	202	30365	0.491	ng/ul	99
21) Benzo(a)anthracene	21.452	228	23059	0.428	ng/ul	99
22) Chrysene	21.505	228	22975	0.397	ng/ul	99
24) Benzo(b)fluoranthene	23.124	252	21204	0.383	ng/ul	95
25) Benzo(k)fluoranthene	23.171	252	20803	0.386	ng/ul	93
26) Benzo(a)pyrene	23.750	252	18278	0.407	ng/ul#	88
27) Indeno(1,2,3-cd)pyrene	26.329	276	24150	0.379	ng/ul#	97
28) Dibenzo(a,h)anthracene	26.345	278	18897	0.367	ng/ul#	90
29) Benzo(g,h,i)perylene	27.090	276	20443	0.365	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM101322\
Data File : BM037090.D
Acq On : 15 Oct 2022 13:24
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 87 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
SSTD0.4150

Quant Time: Oct 17 00:55:46 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_M\METHODS\SFAM-EPA-SIM-BM100622.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Oct 15 04:13:53 2022
Response via : Initial Calibration

