



Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM101424\
Data File : BM048023.D
Acq On : 14 Oct 2024 10:55
Operator : RC/JU
Sample : DFTPP049
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
DFTPP049

Integration File: LSCINT.P

Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM092724.MA.M
Title : SVOA CALIBRATION
Last Update : Thu Oct 10 05:10:14 2024



AutoFind: Scans 2483, 2484, 2485; Background Corrected with Scan 2477

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	33.4	120425	PASS
68	69	0.00	2	1.7	2135	PASS
69	198	0.00	100	35.7	128659	PASS
70	69	0.00	2	0.4	558	PASS
127	198	10	80	44.8	161387	PASS
197	198	0.00	2	0.4	1562	PASS
198	198	100	100	100.0	360085	PASS
199	198	5	9	7.0	25120	PASS
275	198	10	60	24.5	88261	PASS
365	198	1	100	3.3	11830	PASS
441	198	0.01	100	12.4	44531	PASS
442	198	50	100	78.9	284203	PASS
443	442	15	24	19.3	54824	PASS

m/z	Abundance
39.00	367.0
40.00	3619.0
41.10	525.0
42.10	152.0
43.00	332.0
44.00	6936.0
44.90	237.0
49.00	488.0
50.00	226.0
50.90	298.0
55.10	369.0
56.00	171.0
57.20	175.0
67.00	211.0
73.10	695.0
75.00	262.0
76.90	334.0
78.10	271.0
81.00	174.0
83.90	441.0
85.90	303.0
91.10	211.0
96.20	301.0
96.90	275.0
133.00	437.0
135.00	162.0
147.00	262.0
190.90	426.0
192.00	264.0
193.00	255.0
207.10	3379.0
208.00	724.0
209.00	440.0
267.00	168.0
281.10	778.0
281.90	178.0
355.10	243.0

Instrument :
BNA_M
ClientSampleId :
DFTPP049

m/z	Abundance
36.00	153.0
37.00	534.0
38.10	1065.0
39.10	7988.0
40.00	3159.0
41.00	731.0
42.00	232.0
43.00	339.0
44.00	6933.0
45.00	490.0
49.10	1059.0
50.10	26736.0
51.10	108304.0
52.10	5730.0
53.00	340.0
55.00	783.0
56.00	2977.0
57.00	7692.0
58.00	614.0
58.90	258.0
60.00	278.0
61.00	1420.0
62.00	1446.0
63.10	4216.0
64.00	635.0
65.00	2512.0
66.00	274.0
67.10	249.0
68.10	2128.0
69.00	114264.0
70.10	470.0
71.00	241.0
73.00	1451.0
74.00	10045.0
75.10	18248.0
76.00	5746.0
77.10	128064.0
78.10	9562.0
79.00	8200.0
80.10	5600.0
81.00	8361.0
82.00	2015.0
83.00	2037.0
84.00	501.0
85.10	1432.0
86.00	2368.0
87.00	1208.0
88.00	498.0
89.00	427.0
91.00	1862.0
92.10	2651.0
93.00	13760.0
94.00	967.0
94.90	325.0
96.00	985.0
96.90	468.0
98.00	10101.0
99.00	9762.0
100.00	852.0
101.00	5545.0
101.90	282.0
102.90	1377.0
104.00	2864.0
105.00	2810.0
106.00	834.0
107.00	36992.0
108.00	6284.0
109.00	944.0
110.00	73128.0
111.00	11270.0
112.00	1375.0
113.00	554.0
115.10	323.0
116.00	2143.0
117.00	29144.0
118.00	2608.0
119.00	414.0
120.00	459.0
121.00	167.0
122.10	2135.0
123.10	3321.0
124.00	1769.0
125.00	1697.0
126.20	765.0
127.10	137152.0
128.00	10364.0
129.00	50920.0
130.00	4560.0
131.00	876.0
132.00	511.0
133.10	735.0
134.00	1299.0

Instrument :
BNA_M
ClientSampleId :
DFTPP049

135.00	4379.0
136.00	1626.0
137.00	2561.0
138.10	390.0
139.00	388.0
140.00	420.0
141.00	6357.0
142.00	2070.0
143.10	1496.0
144.00	336.0
144.90	438.0
146.00	1156.0
147.00	3600.0
148.10	7540.0
149.00	1923.0
150.00	414.0
151.20	882.0
151.80	527.0
153.00	1995.0
154.00	1366.0
155.00	3481.0
156.10	5158.0
157.10	882.0
158.10	936.0
159.10	1035.0
160.00	1989.0
161.00	3177.0
162.10	878.0
162.90	329.0
164.10	479.0
165.00	2531.0
166.00	1892.0
167.00	13579.0
168.00	9416.0
169.00	1344.0
170.00	471.0
171.00	707.0
172.00	1406.0
173.00	1498.0
174.00	2463.0
175.10	5372.0
176.00	1432.0
177.00	2109.0
177.90	778.0
179.00	10000.0
180.00	7181.0
181.00	3123.0
182.10	699.0
183.00	281.0
184.00	665.0
185.00	4385.0
186.10	36912.0
187.10	10513.0
188.10	1184.0
189.00	2298.0
190.10	401.0
191.00	1416.0
192.00	3147.0
193.10	3387.0
194.00	1057.0
195.10	299.0
196.10	7533.0
196.80	1226.0
198.00	280768.0
199.00	20056.0
200.00	1300.0
201.60	1174.0
203.00	1899.0
204.00	9369.0
205.10	16704.0
206.10	68360.0
207.10	12414.0
208.10	3198.0
209.00	1361.0
210.20	785.0
211.00	2671.0
211.90	511.0
212.90	249.0
215.00	972.0
216.00	1721.0
217.00	18320.0
218.00	2219.0
219.10	287.0
220.10	285.0
221.10	9528.0
222.00	2507.0
223.00	3977.0
224.10	37896.0
225.10	10413.0
226.00	1068.0
227.10	14907.0
228.00	2235.0
229.10	3452.0
230.10	495.0
231.00	1367.0
232.00	288.0

Instrument :
BNA_M
ClientSampleId :
DFTPP049

233.00	328.0
234.00	750.0
235.00	857.0
236.10	864.0
237.00	1711.0
237.90	207.0
239.00	510.0
240.00	481.0
241.00	735.0
242.10	2016.0
243.00	2041.0
244.10	29672.0
245.10	3773.0
246.10	4883.0
247.10	1217.0
248.10	241.0
249.00	1073.0
249.90	211.0
251.00	277.0
252.00	324.0
253.10	580.0
254.10	821.0
255.00	143616.0
256.10	22080.0
257.10	1887.0
258.00	7856.0
259.00	1345.0
260.00	205.0
260.90	276.0
263.10	189.0
263.90	217.0
265.00	3372.0
266.00	555.0
266.90	235.0
270.00	169.0
271.10	734.0
271.90	325.0
273.00	4438.0
274.10	11502.0
275.10	62824.0
276.00	8433.0
277.00	4706.0
278.00	1047.0
279.10	172.0
281.00	915.0
281.90	388.0
283.00	716.0
284.10	296.0
285.10	998.0
286.10	187.0
289.10	220.0
291.00	187.0
292.00	345.0
293.00	947.0
294.10	481.0
295.00	483.0
296.00	17040.0
297.10	2249.0
298.00	329.0
301.10	153.0
302.00	436.0
303.10	2292.0
304.00	435.0
308.10	260.0
309.10	206.0
310.00	263.0
313.10	167.0
314.00	774.0
315.00	1837.0
316.00	1038.0
317.00	223.0
320.90	519.0
322.00	304.0
323.10	5310.0
324.10	1078.0
326.00	271.0
327.00	931.0
328.00	342.0
332.00	508.0
333.10	626.0
334.10	3596.0
335.10	1010.0
341.10	781.0
342.20	176.0
346.00	1096.0
347.10	186.0
352.10	1862.0
353.00	1073.0
354.20	1309.0
355.10	515.0
359.00	179.0
365.00	6977.0
366.00	1100.0
371.10	663.0
372.10	2783.0
373.00	710.0

Instrument :
BNA_M
ClientSampleId :
DFTPP049

377.00	151.0
383.00	754.0
383.80	165.0
390.00	388.0
391.00	292.0
392.00	243.0
400.90	237.0
402.00	1083.0
403.10	1687.0
404.10	627.0
405.00	201.0
421.10	1418.0
422.10	1265.0
423.10	7996.0
424.10	2414.0
425.10	272.0
441.10	25256.0
442.10	159424.0
443.10	31792.0
444.10	3009.0

Instrument :
BNA_M
ClientSampleId :
DFTPP049

m/z	Abundance
36.10	308.0
37.10	562.0
38.10	1556.0
39.10	10740.0
40.00	2945.0
41.00	754.0
42.00	316.0
43.00	332.0
44.00	6514.0
45.00	297.0
47.10	187.0
49.00	1113.0
50.10	35688.0
51.10	145344.0
52.10	7098.0
53.00	445.0
55.00	581.0
56.00	4671.0
57.00	10045.0
58.00	594.0
59.00	285.0
60.00	263.0
61.00	1701.0
62.00	2268.0
63.00	5832.0
64.00	1025.0
65.10	2686.0
66.10	217.0
66.90	382.0
68.10	2505.0
69.00	156096.0
70.10	660.0
73.10	1556.0
74.00	14717.0
75.10	23056.0
76.10	7294.0
77.10	174400.0
78.10	12853.0
79.00	10355.0
80.00	8435.0
81.00	11871.0
82.00	3177.0
83.00	2957.0
84.00	513.0
85.00	2179.0
86.00	3262.0
87.10	1289.0
88.00	712.0
89.00	366.0
91.00	2657.0
92.00	3081.0
93.00	18256.0
94.10	1584.0
95.00	516.0
96.10	980.0
97.00	432.0
98.00	15032.0
99.00	12696.0
100.00	1179.0
101.00	6998.0
102.00	420.0
103.00	2389.0
104.00	4105.0
105.00	3848.0
106.00	1374.0
107.10	53640.0
108.10	7762.0
109.10	1519.0
110.00	99080.0
111.00	14965.0
112.00	1563.0
112.90	586.0
113.90	174.0
115.00	285.0
116.00	2938.0
117.00	40192.0
118.00	3115.0
118.90	390.0
120.00	608.0
121.00	222.0
122.00	3544.0
123.10	5206.0
124.00	2395.0
125.10	2381.0
126.00	606.0
127.10	189504.0
128.10	13613.0
129.00	71816.0
130.00	6240.0
131.00	1476.0
132.00	535.0
133.00	834.0

Instrument :
BNA_M
ClientSampleId :
DFTPP049

134.00	1818.0
135.00	5419.0
136.00	2627.0
137.00	3234.0
137.90	610.0
138.30	529.0
139.00	291.0
140.00	841.0
141.10	9032.0
142.00	3269.0
143.00	2017.0
144.00	610.0
145.10	600.0
146.10	1752.0
147.00	4576.0
148.10	10401.0
149.00	2385.0
150.00	568.0
151.10	1412.0
151.80	735.0
153.00	2688.0
154.00	2226.0
155.10	5688.0
156.10	8564.0
157.00	1397.0
157.90	1433.0
159.00	1071.0
160.00	2759.0
161.00	4206.0
162.00	1421.0
163.00	469.0
163.90	463.0
165.00	4015.0
166.00	3076.0
167.10	18080.0
168.10	11215.0
169.10	1674.0
170.00	560.0
171.00	877.0
172.00	1576.0
173.10	2298.0
174.10	4146.0
175.10	7233.0
176.00	2256.0
177.00	3498.0
178.00	1309.0
179.00	14397.0
180.10	10032.0
181.00	4702.0
182.00	930.0
182.80	434.0
183.90	968.0
185.00	6242.0
186.10	55496.0
187.10	15474.0
188.10	1690.0
189.00	3158.0
190.00	505.0
191.00	1795.0
192.00	4863.0
193.10	5329.0
194.00	1116.0
195.10	619.0
196.00	11074.0
196.80	1829.0
198.00	428736.0
199.00	30416.0
200.10	2305.0
201.60	1411.0
203.10	2760.0
204.10	14317.0
205.10	24624.0
206.10	108952.0
207.10	17800.0
208.10	3990.0
209.00	1606.0
210.20	1677.0
211.10	3918.0
213.10	315.0
214.10	282.0
215.00	1350.0
216.10	2144.0
217.00	28800.0
218.00	4028.0
219.00	272.0
221.10	16381.0
221.80	3864.0
223.10	6537.0
224.10	60072.0
225.10	15190.0
226.10	1549.0
227.00	24592.0
228.00	3210.0
229.00	5324.0
229.90	723.0
231.10	2060.0

Instrument :
BNA_M
ClientSampleId :
DFTPP049

232.00	348.0
233.10	320.0
234.10	1270.0
235.10	1663.0
236.10	1315.0
237.10	2365.0
238.10	255.0
239.00	894.0
240.10	796.0
241.00	1256.0
242.10	3366.0
243.10	3489.0
244.10	46376.0
245.10	6325.0
246.00	9222.0
247.00	1924.0
248.00	404.0
249.00	1630.0
250.00	323.0
250.90	365.0
251.90	419.0
253.00	800.0
254.00	1347.0
255.10	238336.0
256.10	35584.0
257.10	2791.0
258.00	13557.0
259.00	2138.0
260.00	468.0
261.10	290.0
263.90	336.0
265.00	4808.0
266.00	717.0
267.00	253.0
269.00	151.0
270.00	242.0
271.10	942.0
272.10	716.0
273.10	7242.0
274.10	18312.0
275.10	103400.0
276.10	13517.0
277.10	8736.0
278.10	1616.0
278.80	229.0
279.20	218.0
281.00	975.0
282.00	749.0
283.10	1185.0
284.00	758.0
285.10	1318.0
286.00	294.0
289.00	288.0
290.00	271.0
291.00	304.0
292.10	499.0
293.00	1808.0
294.00	784.0
295.00	557.0
296.10	29800.0
297.00	4837.0
297.90	263.0
301.10	437.0
302.00	474.0
303.10	3665.0
304.10	888.0
305.00	202.0
308.00	578.0
309.00	269.0
310.00	394.0
313.00	353.0
314.00	1439.0
315.00	3020.0
316.10	1779.0
317.00	340.0
321.10	935.0
322.00	555.0
323.10	9060.0
324.10	2037.0
325.00	342.0
326.10	232.0
327.00	1991.0
328.10	1265.0
329.10	179.0
332.10	651.0
333.00	997.0
334.10	6384.0
335.10	1417.0
336.10	155.0
339.10	227.0
340.00	165.0
341.10	970.0
342.10	286.0
346.00	2368.0
347.00	502.0
351.00	215.0

Instrument :
BNA_M
ClientSampleId :
DFTPP049

352.00	3174.0
353.00	1992.0
354.10	2701.0
355.10	866.0
365.10	14608.0
366.00	1943.0
367.00	154.0
370.10	339.0
371.00	740.0
372.10	5281.0
373.10	1065.0
383.00	1766.0
384.00	418.0
390.10	780.0
391.10	495.0
391.90	247.0
400.80	326.0
401.30	257.0
402.10	2071.0
403.10	3308.0
404.00	927.0
405.10	262.0
421.10	2953.0
422.10	2348.0
423.10	17336.0
424.10	4413.0
425.20	438.0
440.00	202.0
441.10	51432.0
442.10	325248.0
443.10	63688.0
444.10	6109.0
445.20	325.0

Instrument :
BNA_M
ClientSampleId :
DFTPP049

m/z	Abundance
36.20	151.0
37.00	346.0
38.00	1176.0
39.10	7831.0
40.00	3239.0
41.00	803.0
42.10	195.0
43.10	455.0
44.10	6835.0
45.10	308.0
48.00	150.0
49.00	1308.0
50.10	26456.0
51.10	108520.0
52.00	5556.0
53.00	391.0
54.10	188.0
55.00	639.0
56.00	3513.0
57.00	7661.0
58.10	451.0
61.00	1157.0
62.00	1805.0
63.10	4534.0
64.00	750.0
65.00	2346.0
67.00	241.0
68.10	1773.0
69.00	115616.0
70.00	544.0
71.00	221.0
73.10	1296.0
74.10	10614.0
75.00	17576.0
76.10	5381.0
77.10	133952.0
78.10	9180.0
79.00	7214.0
80.00	6019.0
81.00	9228.0
82.00	2400.0
83.10	2321.0
84.00	570.0
85.00	1814.0
86.00	2619.0
87.00	1100.0
88.00	614.0
89.10	207.0
91.00	2199.0
92.00	2528.0
93.00	14009.0
94.00	1203.0
95.10	259.0
96.00	918.0
97.10	307.0
98.00	11259.0
99.00	8952.0
100.10	754.0
101.00	5134.0
101.90	371.0
103.00	1473.0
104.00	3063.0
105.00	3153.0
106.00	1222.0
107.00	40360.0
108.10	6488.0
109.00	753.0
110.00	81344.0
111.00	11851.0
112.00	1180.0
112.90	419.0
115.00	309.0
116.00	2209.0
117.00	31864.0
118.00	2583.0
118.90	338.0
120.10	576.0
120.90	196.0
122.10	2233.0
123.00	3600.0
124.00	2092.0
125.00	1755.0
126.10	342.0
127.00	157504.0
128.10	11874.0
129.00	58912.0
130.00	4797.0
131.10	1111.0
131.90	473.0
132.90	456.0
134.00	1480.0
135.00	4941.0

Instrument :

BNA_M

ClientSampleId :

DFTPP049

136.00	1956.0
137.10	2419.0
137.90	479.0
139.00	419.0
140.00	762.0
141.00	7520.0
142.00	2504.0
143.10	1615.0
144.10	567.0
145.00	402.0
146.00	1649.0
147.00	3732.0
148.00	9049.0
149.10	1886.0
150.10	547.0
151.10	937.0
151.90	540.0
153.00	2256.0
154.00	1851.0
155.00	4328.0
156.10	6334.0
157.00	1126.0
158.00	1303.0
159.00	701.0
160.00	2382.0
161.10	3937.0
162.10	1140.0
162.80	405.0
164.00	377.0
165.00	3029.0
166.10	2000.0
167.00	16496.0
168.00	9602.0
169.00	1292.0
170.00	502.0
170.80	548.0
172.00	1289.0
173.00	1730.0
174.00	3147.0
175.10	6685.0
176.10	1554.0
176.90	2810.0
178.00	1084.0
179.00	12204.0
180.00	8291.0
181.00	3557.0
182.20	620.0
183.00	414.0
183.90	850.0
185.10	6147.0
186.10	46056.0
187.00	13005.0
188.10	1356.0
189.10	2853.0
189.90	469.0
190.90	1460.0
192.00	4379.0
193.10	4882.0
194.00	943.0
195.20	704.0
196.00	9655.0
196.90	1631.0
198.00	370752.0
199.00	24888.0
200.00	2014.0
201.50	1339.0
203.00	2550.0
204.00	12490.0
205.10	22360.0
206.10	90904.0
207.10	14729.0
208.00	3375.0
209.00	1312.0
210.00	1373.0
211.10	3443.0
213.10	320.0
215.10	1019.0
216.10	2019.0
217.00	25104.0
218.10	3408.0
219.10	449.0
221.10	13648.0
221.90	3074.0
223.00	5819.0
224.10	53992.0
225.10	13154.0
226.00	1485.0
227.10	20768.0
228.10	2714.0
229.00	4107.0
230.00	593.0
231.10	1875.0
232.00	340.0
233.00	326.0
234.00	1382.0
235.00	1569.0

Instrument :
BNA_M
ClientSampleId :
DFTPP049

236.00	1024.0
237.10	2053.0
238.10	343.0
239.00	1094.0
240.10	501.0
241.00	1141.0
242.10	2681.0
243.10	3151.0
244.10	42472.0
245.10	5157.0
246.00	7993.0
247.00	1767.0
248.00	482.0
249.00	1701.0
250.00	310.0
251.00	455.0
252.00	379.0
253.10	889.0
254.00	1326.0
255.10	209728.0
256.10	31136.0
257.10	2427.0
258.00	11948.0
259.00	2102.0
260.00	386.0
261.00	345.0
264.00	315.0
265.10	4935.0
266.00	723.0
267.10	196.0
270.00	332.0
271.10	712.0
272.00	544.0
273.10	6504.0
274.10	16448.0
275.10	98560.0
276.10	12355.0
277.00	7730.0
278.00	1234.0
279.00	279.0
281.10	1029.0
282.10	541.0
283.10	1084.0
284.10	764.0
285.00	1201.0
286.10	270.0
289.10	392.0
290.10	263.0
291.10	366.0
291.90	312.0
293.00	1986.0
294.00	525.0
295.10	714.0
296.10	28360.0
297.10	3718.0
301.10	216.0
302.10	497.0
303.10	3416.0
304.10	884.0
305.10	188.0
308.00	484.0
309.10	248.0
310.00	367.0
313.10	261.0
314.00	1281.0
315.10	3418.0
316.10	1961.0
317.10	239.0
319.90	321.0
321.10	995.0
322.00	481.0
323.10	9681.0
324.10	2004.0
325.00	166.0
325.90	224.0
327.00	1785.0
328.00	1510.0
329.10	247.0
332.10	511.0
333.10	904.0
334.10	6468.0
335.10	1622.0
336.00	206.0
339.10	196.0
341.00	1099.0
342.10	502.0
346.00	2538.0
347.00	487.0
351.10	151.0
352.10	3092.0
353.10	2249.0
354.10	3080.0
355.10	766.0
359.00	162.0
365.10	13905.0
366.10	2051.0

Instrument :
BNA_M
ClientSampleId :
DFTPP049

367.10	168.0
370.00	357.0
371.00	797.0
372.10	5099.0
373.10	1549.0
377.00	172.0
383.10	1699.0
384.00	612.0
385.10	243.0
390.20	813.0
391.00	436.0
392.00	249.0
401.10	430.0
402.00	2481.0
403.10	2700.0
404.10	1270.0
405.00	258.0
421.10	2829.0
422.10	2734.0
423.10	18400.0
424.10	4442.0
425.10	433.0
441.10	56904.0
442.10	367936.0
443.10	68992.0
444.10	6628.0
445.20	299.0

Instrument :
BNA_M
ClientSampleId :
DFTPP049