

Data Path : Z:\HPCHEM1\BNA M\DATA\BM101515\
 Data File : BM002915.D
 Acq On : 15 Oct 2015 15:26
 Operator : UM/IZ
 Sample : SSTDICV001
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 ICVBM101515

Quant Time: Oct 16 06:51:15 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\8270-SIM-BM101515.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 15 16:21:18 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	78	0.00
2	1,4-Dioxane	0.456	0.437	4.2	77	0.00
3	n-Nitrosodimethylamine	0.355	0.363	-2.3	78	-0.02
4 S	2-Fluorophenol	0.935	0.886	5.2	76	-0.02
5 S	Phenol-d6	1.193	1.191	0.2	77	-0.02
6	bis(2-Chloroethyl)ether	1.140	1.142	-0.2	77	-0.02
7 I	Naphthalene-d8	1.000	1.000	0.0	80	0.00
8 S	Nitrobenzene-d5	0.240	0.231	3.7	77	-0.02
9	Nitrobenzene	0.254	0.243	4.3	76	-0.02
10	Naphthalene	1.110	1.075	3.2	79	0.00
11	Hexachlorobutadiene	0.179	0.172	3.9	78	-0.01
12	2-Methylnaphthalene	0.748	0.741	0.9	80	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	80	0.00
14 S	2,4,6-Tribromophenol	0.158	0.154	2.5	80	0.00
15 S	2-Fluorobiphenyl	1.601	1.593	0.5	81	0.00
16	Acenaphthylene	2.185	2.152	1.5	80	0.00
17	Acenaphthene	1.499	1.450	3.3	81	0.00
18	Fluorene	1.787	1.774	0.7	81	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	90	0.00
20	4-Bromophenyl-phenylether	0.207	0.224	-8.2	100	0.00
21	Hexachlorobenzene	0.246	0.212	13.8	79	0.02
22	Pentachlorophenol	0.014	0.018	-28.6#	89	0.00
23	Phenanthrene	1.165	1.089	6.5	85	0.02
24	Anthracene	1.110	0.943	15.0	76	0.00
25	Fluoranthene	1.289	1.256	2.6	90	0.00
26 I	Chrysene-d12	1.000	1.000	0.0	87	0.01
27	Benzydine	0.590	0.597	-1.2	96	0.00
28	Pvrene	1.474	1.467	0.5	88	0.00
29 S	Terphenyl-d14	0.749	0.743	0.8	88	0.00
30	Benzo(a)anthracene	1.445	1.322	8.5	77	0.01
31	3,3'-Dichlorobenzidine	0.489	0.458	6.3	86	0.01
32	Chrysene	1.351	1.343	0.6	95	0.00
33	Bis(2-ethylhexyl)phthalate	0.841	0.700	16.8	70	0.00
34	Indeno(1,2,3-cd)pyrene	1.390	1.405	-1.1	96	0.00
35 I	Perylene-d12	1.000	1.000	0.0	97	0.00
36	Benzo(b)fluoranthene	1.435	1.451	-1.1	97	0.00
37	Benzo(k)fluoranthene	1.404	1.313	6.5	93	0.00
38 C	Benzo(a)pyrene	1.269	1.207	4.9	94	0.00
39	Dibenzo(a,h)anthracene	1.215	1.167	4.0	97	0.00
40	Benzo(a,h,i)perylene	1.270	1.242	2.2	99	0.00

Data Path : Z:\HPCHEM1\BNA M\DATA\BM101515\
Data File : BM002915.D
Acq On : 15 Oct 2015 15:26
Operator : UM/IZ
Sample : SSTDICV001
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_M
ClientSampleId :
ICVBM101515

Quant Time: Oct 16 06:51:15 2015
Quant Method : Z:\HPCHEM1\BNA M\METHODS\8270-SIM-BM101515.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 15 16:21:18 2015
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0