

Data Path : Z:\HPCHEM1\BNA M\DATA\BM101515\
 Data File : BM002917.D
 Acq On : 15 Oct 2015 16:54
 Operator : UM/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC001

Quant Time: Oct 16 06:55:12 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\8270-SIM-BM101515.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 15 16:21:18 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	97	0.00
2	1,4-Dioxane	0.456	0.420	7.9	92	0.00
3	n-Nitrosodimethylamine	0.355	0.356	-0.3	95	-0.02
4 S	2-Fluorophenol	0.935	0.905	3.2	96	-0.02
5 S	Phenol-d6	1.193	1.173	1.7	94	-0.02
6	bis(2-Chloroethyl)ether	1.140	1.146	-0.5	96	-0.02
7 I	Naphthalene-d8	1.000	1.000	0.0	97	0.00
8 S	Nitrobenzene-d5	0.240	0.222	7.5	90	-0.02
9	Nitrobenzene	0.254	0.234	7.9	90	-0.02
10	Naphthalene	1.110	1.074	3.2	97	0.00
11	Hexachlorobutadiene	0.179	0.175	2.2	97	-0.01
12	2-Methylnaphthalene	0.748	0.724	3.2	96	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	92	0.00
14 S	2,4,6-Tribromophenol	0.158	0.143	9.5	85	0.00
15 S	2-Fluorobiphenyl	1.601	1.626	-1.6	96	0.00
16	Acenaphthylene	2.185	2.085	4.6	90	0.00
17	Acenaphthene	1.499	1.464	2.3	95	0.00
18	Fluorene	1.787	1.783	0.2	94	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	103	0.00
20	4-Bromophenyl-phenylether	0.207	0.215	-3.9	109	0.00
21	Hexachlorobenzene	0.246	0.207	15.9	87	0.02
22	Pentachlorophenol	0.014	0.016	-14.3	89	0.00
23	Phenanthrene	1.165	1.033	11.3	92	0.02
24	Anthracene	1.110	0.931	16.1	86	0.00
25	Fluoranthene	1.289	1.161	9.9	95	0.00
26 I	Chrysene-d12	1.000	1.000	0.0	90	0.00
27	Benzydine	0.590	0.522	11.5	86	0.00
28	Pvrene	1.474	1.519	-3.1	93	0.00
29 S	Terphenyl-d14	0.749	0.787	-5.1	96	0.00
30	Benzo(a)anthracene	1.445	1.350	6.6	81	0.00
31	3,3'-Dichlorobenzidine	0.489	0.422	13.7	81	0.00
32	Chrysene	1.351	1.370	-1.4	99	0.00
33	Bis(2-ethylhexyl)phthalate	0.841	0.661	21.4	68	0.00
34	Indeno(1,2,3-cd)pyrene	1.390	1.357	2.4	95	0.00
35 I	Perylene-d12	1.000	1.000	0.0	96	0.00
36	Benzo(b)fluoranthene	1.435	1.493	-4.0	99	0.00
37	Benzo(k)fluoranthene	1.404	1.277	9.0	90	0.00
38 C	Benzo(a)pyrene	1.269	1.196	5.8	92	0.00
39	Dibenzo(a,h)anthracene	1.215	1.169	3.8	96	0.00
40	Benzo(a,h,i)perylene	1.270	1.235	2.8	97	0.00

Data Path : Z:\HPCHEM1\BNA M\DATA\BM101515\
Data File : BM002917.D
Acq On : 15 Oct 2015 16:54
Operator : UM/IZ
Sample : SSTDCCC001
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTDCCC001

Quant Time: Oct 16 06:55:12 2015
Quant Method : Z:\HPCHEM1\BNA M\METHODS\8270-SIM-BM101515.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 15 16:21:18 2015
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range				
SPCC's out = 0 CCC's out = 0				