

Data Path : Z:\HPCHEM1\BNA M\DATA\BM101515\
 Data File : BM002919.D
 Acq On : 15 Oct 2015 18:08
 Operator : UM/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC001

Quant Time: Oct 16 06:59:44 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\8270-SIM-BM101515.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Oct 15 16:21:18 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	82	0.00
2	1,4-Dioxane	0.456	0.439	3.7	81	0.00
3	n-Nitrosodimethylamine	0.355	0.349	1.7	79	-0.02
4 S	2-Fluorophenol	0.935	0.892	4.6	80	-0.02
5 S	Phenol-d6	1.193	1.168	2.1	79	0.00
6	bis(2-Chloroethyl)ether	1.140	1.090	4.4	77	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	83	0.00
8 S	Nitrobenzene-d5	0.240	0.223	7.1	77	-0.02
9	Nitrobenzene	0.254	0.235	7.5	77	-0.01
10	Naphthalene	1.110	1.078	2.9	82	0.00
11	Hexachlorobutadiene	0.179	0.172	3.9	82	0.00
12	2-Methylnaphthalene	0.748	0.731	2.3	83	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	87	0.00
14 S	2,4,6-Tribromophenol	0.158	0.137	13.3	77	0.00
15 S	2-Fluorobiphenyl	1.601	1.500	6.3	84	0.00
16	Acenaphthylene	2.185	2.033	7.0	83	0.00
17	Acenaphthene	1.499	1.343	10.4	82	0.00
18	Fluorene	1.787	1.674	6.3	83	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	78	0.00
20	4-Bromophenyl-phenylether	0.207	0.209	-1.0	80	0.00
21	Hexachlorobenzene	0.246	0.289	-17.5	92	0.00
22	Pentachlorophenol	0.014	0.021	-50.0#	90	-0.02
23	Phenanthrene	1.165	1.347	-15.6	91	0.00
24	Anthracene	1.110	1.294	-16.6	90	0.00
25	Fluoranthene	1.289	1.415	-9.8	88	-0.01
26 I	Chrysene-d12	1.000	1.000	0.0	92	0.00
27	Benzydine	0.590	0.492	16.6	83	0.00
28	Pvrene	1.474	1.381	6.3	87	-0.01
29 S	Terphenyl-d14	0.749	0.700	6.5	87	0.00
30	Benzo(a)anthracene	1.445	1.291	10.7	80	0.00
31	3,3'-Dichlorobenzidine	0.489	0.424	13.3	84	0.00
32	Chrysene	1.351	1.289	4.6	96	0.00
33	Bis(2-ethylhexyl)phthalate	0.841	0.570	32.2#	60	0.00
34	Indeno(1,2,3-cd)pyrene	1.390	1.320	5.0	95	0.00
35 I	Perylene-d12	1.000	1.000	0.0	95	0.00
36	Benzo(b)fluoranthene	1.435	1.362	5.1	90	0.00
37	Benzo(k)fluoranthene	1.404	1.387	1.2	97	0.00
38 C	Benzo(a)pyrene	1.269	1.190	6.2	91	0.00
39	Dibenzo(a,h)anthracene	1.215	1.173	3.5	96	0.00
40	Benzo(a,h,i)perylene	1.270	1.240	2.4	97	-0.01

Data Path : Z:\HPCHEM1\BNA M\DATA\BM101515\
Data File : BM002919.D
Acq On : 15 Oct 2015 18:08
Operator : UM/IZ
Sample : SSTDCCC001
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTDCCC001

Quant Time: Oct 16 06:59:44 2015
Quant Method : Z:\HPCHEM1\BNA M\METHODS\8270-SIM-BM101515.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Oct 15 16:21:18 2015
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range				
SPCC's out = 0 CCC's out = 0				