

Data Path : Z:\HPCHEM1\BNA M\DATA\BM101515\
 Data File : BM002928.D
 Acq On : 15 Oct 2015 23:34
 Operator : UM/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_M
 LabSampleId :
 SSTDCCC001

Quant Time: Oct 16 07:32:31 2015
 Quant Method : Z:\HPCHEM1\BNA M\METHODS\8270-SIM-BM101515.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Oct 16 06:59:27 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	101	0.00
2	1,4-Dioxane	0.456	0.430	5.7	98	0.00
3	n-Nitrosodimethylamine	0.355	0.372	-4.8	103	-0.03
4 S	2-Fluorophenol	0.935	0.953	-1.9	105	-0.03
5 S	Phenol-d6	1.193	1.274	-6.8	106	-0.02
6	bis(2-Chloroethyl)ether	1.140	1.163	-2.0	101	-0.02
7 I	Naphthalene-d8	1.000	1.000	0.0	102	-0.01
8 S	Nitrobenzene-d5	0.240	0.247	-2.9	105	-0.03
9	Nitrobenzene	0.254	0.263	-3.5	105	-0.03
10	Naphthalene	1.110	1.072	3.4	100	-0.01
11	Hexachlorobutadiene	0.179	0.176	1.7	102	-0.01
12	2-Methylnaphthalene	0.748	0.728	2.7	101	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	99	0.00
14 S	2,4,6-Tribromophenol	0.158	0.181	-14.6	116	0.00
15 S	2-Fluorobiphenyl	1.601	1.579	1.4	100	-0.01
16	Acenaphthylene	2.185	2.248	-2.9	104	0.00
17	Acenaphthene	1.499	1.430	4.6	99	0.00
18	Fluorene	1.787	1.782	0.3	101	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	106	0.00
20	4-Bromophenyl-phenylether	0.207	0.211	-1.9	111	0.00
21	Hexachlorobenzene	0.246	0.206	16.3	90	0.00
22	Pentachlorophenol	0.014	0.029	-107.1#	175#	-0.03
23	Phenanthrene	1.165	1.020	12.4	94	0.02
24	Anthracene	1.110	1.005	9.5	96	0.00
25	Fluoranthene	1.289	1.177	8.7	100	0.00
26 I	Chrysene-d12	1.000	1.000	0.0	95	0.00
27	Benzydine	0.590	0.646	-9.5	113	0.00
28	Pvrene	1.474	1.505	-2.1	98	0.00
29 S	Terphenyl-d14	0.749	0.772	-3.1	99	0.00
30	Benzo(a)anthracene	1.445	1.560	-8.0	100	0.00
31	3,3'-Dichlorobenzidine	0.489	0.545	-11.5	112	0.00
32	Chrysene	1.351	1.215	10.1	93	0.00
33	Bis(2-ethylhexyl)phthalate	0.841	1.125	-33.8#	123	0.00
34	Indeno(1,2,3-cd)pyrene	1.390	1.208	13.1	90	-0.01
35 I	Perylene-d12	1.000	1.000	0.0	93	0.00
36	Benzo(b)fluoranthene	1.435	1.521	-6.0	98	-0.01
37	Benzo(k)fluoranthene	1.404	1.348	4.0	92	-0.01
38 C	Benzo(a)pyrene	1.269	1.275	-0.5	96	-0.01
39	Dibenzo(a,h)anthracene	1.215	1.135	6.6	91	-0.01
40	Benzo(a,h,i)perylene	1.270	1.140	10.2	88	-0.01

Data Path : Z:\HPCHEM1\BNA M\DATA\BM101515\
Data File : BM002928.D
Acq On : 15 Oct 2015 23:34
Operator : UM/IZ
Sample : SSTDCCC001
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_M
LabSampleId :
SSTDCCC001

Quant Time: Oct 16 07:32:31 2015
Quant Method : Z:\HPCHEM1\BNA M\METHODS\8270-SIM-BM101515.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Oct 16 06:59:27 2015
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range				
SPCC's out = 0 CCC's out = 0				