


Data Path : Z:\HPCHEM1\BNA M\DATA\BM101515\
 Data File : BM002918.D
 Acq On : 15 Oct 2015 17:32
 Operator : UM/IZ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_M
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\HPCHEM1\BNA_M\METHODS\8270-SIM-BM101515.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Thu Oct 15 16:21:18 2015



AutoFind: Scans 2586, 2587, 2588; Background Corrected with Scan 2580

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	30.6	16608	PASS
68	69	0.00	2	1.3	226	PASS
69	198	0.00	100	32.1	17432	PASS
70	69	0.00	2	0.0	0	PASS
127	198	10	80	48.8	26469	PASS
197	198	0.00	2	0.6	336	PASS
198	198	100	100	100.0	54229	PASS
199	198	5	9	6.7	3646	PASS
275	198	10	60	25.9	14058	PASS
365	198	1	100	3.0	1637	PASS
441	442	0.01	24	15.1	7696	PASS
442	198	50	100	94.2	51074	PASS
443	442	15	24	19.3	9866	PASS